

Data File : H:\VOL1\HPLC\HPLC L\120316\HL002170.D Vial: 45
Acq On : 17-Mar-2012, 13:17 Operator: HLM
Sample : D1801-21 Inst : G1314A
Misc : Multiplr: 1.00
IntFile : events.e
Quant Time: Mar 19 11:18 2012 Quant Results File: HL022712.RES

Quant Method : H:\VOL1\HPLC\HPLC L\METHODS\HL022712.M (Chemstation Integrator)
Title : SW8330A/B Primary (Extend C18 3.5u 4.6x10mm)
Last Update : Mon Mar 19 11:16:51 2012
Response via : Initial Calibration
DataAcq Meth : 8330AGIP.M

Volume Inj. :
Signal Phase :
Signal Info :

Compound	R.T.	Response	Conc Units
----------	------	----------	------------

System Monitoring Compounds

1) S 4-Nitroaniline	1.39	630133	166.653 ug/L
Spiked Amount 300.000	Range 37 - 149	Recovery =	55.55%

Target Compounds

Quantitation Report

Data File : H:\VOL1\HPLC\HPLC L\120316\HL002170.D Vial: 45
Acq On : 17-Mar-2012, 13:17 Operator: HLM
Sample : D1801-21 Inst : G1314A
Misc : Multiplr: 1.00
IntFile : events.e
Quant Time: Mar 19 11:18 2012 Quant Results File: HL022712.RES

Quant Method : H:\VOL1\HPLC\HPLC L\METHODS\HL022712.M (Chemstation Integrator)
Title : SW8330A/B Primary (Extend C18 3.5u 4.6x10mm)
Last Update : Mon Mar 19 11:16:51 2012
Response via : Multiple Level Calibration
DataAcq Meth : 8330AGIP.M

Volume Inj. :
Signal Phase :
Signal Info :

