

Data File : H:\VOL1\HPLC\HPLC L\120427\HL002786.D Vial: 45  
Acq On : 27-Apr-2012, 22:11 Operator: HLM  
Sample : D2332-05 Inst : G1314A  
Misc : Multiplr: 1.00  
IntFile : events.e  
Quant Time: May 4 11:12 2012 Quant Results File: HL032912.RES

Quant Method : H:\VOL1\HPLC\HPLC L\METHODS\HL032912.M (Chemstation Integrator)  
Title : SW8330A/B Primary (Extend C18 3.5u 4.6x10mm)  
Last Update : Mon Apr 30 14:06:15 2012  
Response via : Initial Calibration  
DataAcq Meth : 8330AGIP.M

Volume Inj. :  
Signal Phase :  
Signal Info :

| Compound | R.T. | Response | Conc Units |
|----------|------|----------|------------|
|----------|------|----------|------------|

-----  
System Monitoring Compounds

|                       |                |          |              |
|-----------------------|----------------|----------|--------------|
| 1) S 4-Nitroaniline   | 1.40           | 1817754  | 477.269 ug/L |
| Spiked Amount 300.000 | Range 37 - 149 | Recovery | = 159.09%#   |

Target Compounds

# Quantitation Report

Data File : H:\VOL1\HPLC\HPLC L\120427\HL002786.D Vial: 45  
 Acq On : 27-Apr-2012, 22:11 Operator: HLM  
 Sample : D2332-05 Inst : G1314A  
 Misc : Multiplr: 1.00  
 IntFile : events.e  
 Quant Time: May 4 11:12 2012 Quant Results File: HL032912.RES

Quant Method : H:\VOL1\HPLC\HPLC L\METHODS\HL032912.M (Chemstation Integrator)  
 Title : SW8330A/B Primary (Extend C18 3.5u 4.6x10mm)  
 Last Update : Mon Apr 30 14:06:15 2012  
 Response via : Multiple Level Calibration  
 DataAcq Meth : 8330AGIP.M

Volume Inj. :  
 Signal Phase :  
 Signal Info :

