

Data File : H:\VOL1\HPLC\HPLC L\120427\HL002785.D Vial: 44
Acq On : 27-Apr-2012, 21:44 Operator: HLM
Sample : D2332-04 Inst : G1314A
Misc : Multiplr: 1.00
IntFile : events.e
Quant Time: May 4 11:11 2012 Quant Results File: HL032912.RES

Instrument :
HPLC_L
ClientSampleId :
TW-26

Quant Method : H:\VOL1\HPLC\HPLC L\METHODS\HL032912.M (Chemstation Integrator)
Title : SW8330A/B Primary (Extend C18 3.5u 4.6x10mm)
Last Update : Mon Apr 30 14:06:15 2012
Response via : Initial Calibration
DataAcq Meth : 8330AGIP.M

Volume Inj. :
Signal Phase :
Signal Info :

Compound	R.T.	Response	Conc Units
----------	------	----------	------------

System Monitoring Compounds

1) S	4-Nitroaniline	1.42	1883989	494.660 ug/L
Spiked Amount	300.000	Range	37 - 149	Recovery = 164.89%#

Target Compounds

Data File : H:\VOL1\HPLC\HPLC L\120427\HL002785.D

Vial: 44

Acq On : 27-Apr-2012, 21:44

Operator: HLM

Sample : D2332-04

Inst : G1314A

Misc :

Multiplr: 1.00

IntFile : events.e

Quant Time: May 4 11:11 2012 Quant Results File: HL032912.RES

Instrument :

HPLC_L

ClientSampleId :

TW-26

Quant Method : H:\VOL1\HPLC\HPLC L\METHODS\HL032912.M (Chemstation Integrator)

Title : SW8330A/B Primary (Extend C18 3.5u 4.6x10mm)

Last Update : Mon Apr 30 14:06:15 2012

Response via : Multiple Level Calibration

DataAcq Meth : 8330AGIP.M

Volume Inj. :

Signal Phase :

Signal Info :

