

Evaluate Continuing Calibration Report

Data File : Z:\HPCHEM1\HPLC_L\DATA\HL052213\HL005294.D Vial: 1
 Acq On : 23-May-2013, 02:16 Operator: QM
 Sample : HSTDCCC0500 Inst : G1314A
 Misc : Multiplr: 1.00
 IntFile : autoint1.e

Instrument :
 HPLC_L
 LabSampleId :
 HSTDCCC0500

Method : Z:\HPCHEM1\HPLC_L\METHODS\HL052213.M
 Title :
 Last Update : Thu May 23 10:04:01 2013
 Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 S	4-Nitroaniline	3.970	3.721 E3	6.3	100	0.02
2 T	HMX	4.536	4.439 E3	2.1	97	0.02
3 T	RDX	5.298	5.104 E3	3.7	100	0.03
4 T	1,3,5-Trinitrobenzene	10.818	11.067 E3	-2.3	101	0.05
5 T	1,3-Dinitrobenzene	15.438	15.798 E3	-2.3	99	0.06
6 T	3,5-Dinitroaniline	13.530	13.416 E3	0.8	100	0.04
7 T	Nitrobenzene	10.244	9.876 E3	3.6	101	0.06
8 T	Tetryl	6.730	6.974 E3	-3.6	102	0.08
9 T	2,4,6-Trinitrotoluene	11.258	11.046 E3	1.9	101	0.08
10 T	2-Amino-4,6-dinitrotoluene	10.714	10.751 E3	-0.3	99	0.04
11 T	4-Amino-2,6-dinitrotoluene	5.576	6.391 E3	-14.6	101	0.08
12 T	2,4-Dinitrotoluene	15.136	14.537 E3	4.0	99	0.09
13 T	2,6-Dinitrotoluene	8.132	7.772 E3	4.4	97	0.08
14 T	2-Nitrotoluene	7.093	6.574 E3	7.3	98	0.02
15 T	4-Nitrotoluene	5.594	5.510 E3	1.5	102	0.00
16 T	3-Nitrotoluene	6.630	7.095 E3	-7.0	102	0.05
17 T	PETN	11.369	8.910 E3	21.6	115	-0.02

Evaluate Continuing Calibration Report - Not Found

Data File : Z:\HPCHEM1\HPLC_L\DATA\HL052213\HL005294.D Vial: 1
 Acq On : 23-May-2013, 02:16 Operator: QM
 Sample : HSTDCCC0500 Inst : G1314A
 Misc : Multiplr: 1.00
 IntFile : autoint1.e

Instrument :
 HPLC_L
 LabSampleId :
 HSTDCCC0500

Method : Z:\HPCHEM1\HPLC_L\METHODS\HL052213.M
 Title :
 Last Update : Thu May 23 10:04:01 2013
 Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------