

Data File : H:\VOL1\HPLC\HPLC L\120601\HL003153.D Vial: 54
Acq On : 02-Jun-2012, 01:03 Operator: HLM
Sample : D2816-02 Inst : G1314A
Misc : Multiplr: 1.00
IntFile : events.e
Quant Time: Jun 4 10:17 2012 Quant Results File: HL050312.RES

Quant Method : H:\VOL1\HPLC\HPLC L\METHODS\HL050312.M (Chemstation Integrator)
Title : SW8330A/B Primary (Extend C18 3.5u 4.6x10mm)
Last Update : Mon Jun 04 10:17:03 2012
Response via : Initial Calibration
DataAcq Meth : 8330AGIP.M

Volume Inj. :
Signal Phase :
Signal Info :

Compound	R.T.	Response	Conc Units
----------	------	----------	------------

System Monitoring Compounds

1) S 4-Nitroaniline	1.37	554233	151.477 ug/L
Spiked Amount 300.000	Range 37 - 149	Recovery =	50.49%

Target Compounds

Data File : H:\VOL1\HPLC\HPLC L\120601\HL003153.D Vial: 54
Acq On : 02-Jun-2012, 01:03 Operator: HLM
Sample : D2816-02 Inst : G1314A
Misc : Multiplr: 1.00
IntFile : events.e
Quant Time: Jun 4 10:17 2012 Quant Results File: HL050312.RES

Quant Method : H:\VOL1\HPLC\HPLC L\METHODS\HL050312.M (Chemstation Integrator)
Title : SW8330A/B Primary (Extend C18 3.5u 4.6x10mm)
Last Update : Mon Jun 04 10:17:03 2012
Response via : Multiple Level Calibration
DataAcq Meth : 8330AGIP.M

Volume Inj. :
Signal Phase :
Signal Info :

