

Data File : H:\VOL1\HPLC\HPLC L\120615\HL003292.D Vial: 15
Acq On : 15-Jun-2012, 15:40 Operator: HLM
Sample : D3002-02 Inst : G1314A
Misc : Multiplr: 1.00
IntFile : events.e
Quant Time: Jun 22 17:39 2012 Quant Results File: HL060712.RES

Quant Method : H:\VOL1\HPLC\HPLC L\METHODS\HL060712.M (Chemstation Integrator)
Title : SW8330A/B Primary (Extend C18 3.5u 4.6x10mm)
Last Update : Fri Jun 22 17:39:50 2012
Response via : Initial Calibration
DataAcq Meth : 8330AGIP.M

Volume Inj. :
Signal Phase :
Signal Info :

Compound	R.T.	Response	Conc Units
----------	------	----------	------------

System Monitoring Compounds

1) S	4-Nitroaniline	1.39	1391973	382.739 ug/L
Spiked Amount	300.000	Range	37 - 149	Recovery = 127.58%

Target Compounds

Quantitation Report

Data File : H:\VOL1\HPLC\HPLC L\120615\HL003292.D Vial: 15
 Acq On : 15-Jun-2012, 15:40 Operator: HLM
 Sample : D3002-02 Inst : G1314A
 Misc : Multiplr: 1.00
 IntFile : events.e
 Quant Time: Jun 22 17:39 2012 Quant Results File: HL060712.RES

Quant Method : H:\VOL1\HPLC\HPLC L\METHODS\HL060712.M (Chemstation Integrator)
 Title : SW8330A/B Primary (Extend C18 3.5u 4.6x10mm)
 Last Update : Fri Jun 22 17:39:50 2012
 Response via : Multiple Level Calibration
 DataAcq Meth : 8330AGIP.M

Volume Inj. :
 Signal Phase :
 Signal Info :

