

# Evaluate Continuing Calibration Report

Data File : S:\HPCHEM1\HPLC\_L\DATA\HL070913\HL005529.D Vial: 1  
 Acq On : 09-Jul-2013, 17:54 Operator: QM  
 Sample : HSTDCCC0500 Inst : G1314A  
 Misc : Multiplr: 1.00  
 IntFile : autoint1.e

Instrument :  
 HPLC\_L  
 LabSampleId :  
 HSTDCCC0500

Method : Z:\HPCHEM1\HPLC\_L\METHODS\HL070213.M  
 Title :  
 Last Update : Tue Jul 02 16:59:51 2013  
 Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|      | Compound                   | AvgRF  | CCRF      | %Dev | Area% | Dev(min) |
|------|----------------------------|--------|-----------|------|-------|----------|
| 1 S  | 4-Nitroaniline             | 4.050  | 4.309 E3  | -6.4 | 105   | -0.02    |
| 2 T  | HMX                        | 4.606  | 4.128 E3  | 10.4 | 93    | -0.01    |
| 3 T  | RDX                        | 5.046  | 5.302 E3  | -5.1 | 103   | -0.03    |
| 4 T  | 1,3,5-Trinitrobenzene      | 11.148 | 10.842 E3 | 2.7  | 98    | -0.06    |
| 5 T  | 1,3-Dinitrobenzene         | 15.624 | 15.741 E3 | -0.7 | 98    | -0.12    |
| 6 T  | 3,5-Dinitroaniline         | 13.968 | 13.442 E3 | 3.8  | 99    | -0.15    |
| 7 T  | Nitrobenzene               | 9.954  | 9.764 E3  | 1.9  | 97    | -0.16    |
| 8 T  | Tetryl                     | 6.356  | 5.393 E3  | 15.2 | 88    | -0.41    |
| 9 T  | 2,4,6-Trinitrotoluene      | 11.742 | 12.526 E3 | -6.7 | 99    | -0.35    |
| 10 T | 2-Amino-4,6-dinitrotoluene | 11.757 | 11.076 E3 | 5.8  | 99    | -0.44    |
| 11 T | 4-Amino-2,6-dinitrotoluene | 5.764  | 6.249 E3  | -8.4 | 99    | -0.48    |
| 12 T | 2,4-Dinitrotoluene         | 14.968 | 14.910 E3 | 0.4  | 98    | -0.51#   |
| 13 T | 2,6-Dinitrotoluene         | 8.365  | 8.152 E3  | 2.5  | 99    | -0.55#   |
| 14 T | 2-Nitrotoluene             | 6.383  | 6.627 E3  | -3.8 | 100   | -0.74#   |
| 15 T | 4-Nitrotoluene             | 5.395  | 5.404 E3  | -0.2 | 100   | -0.79#   |
| 16 T | 3-Nitrotoluene             | 6.949  | 6.894 E3  | 0.8  | 101   | -0.81#   |
| 17 T | PETN                       | 10.159 | 10.062 E3 | 1.0  | 86    | -1.14#   |

Evaluate Continuing Calibration Report - Not Found

Data File : S:\HPCHEM1\HPLC\_L\DATA\HL070913\HL005529.D Vial: 1  
 Acq On : 09-Jul-2013, 17:54 Operator: QM  
 Sample : HSTDCCC0500 Inst : G1314A  
 Misc : Multiplr: 1.00  
 IntFile : autoint1.e

Instrument :  
 HPLC\_L  
 LabSampleId :  
 HSTDCCC0500

Method : Z:\HPCHEM1\HPLC\_L\METHODS\HL070213.M  
 Title :  
 Last Update : Tue Jul 02 16:59:51 2013  
 Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

| Compound | AvgRF | CCRF | %Dev | Area% | Dev(min) |
|----------|-------|------|------|-------|----------|
|----------|-------|------|------|-------|----------|