

# Evaluate Continuing Calibration Report

Data File : S:\HPCHEM1\HPLC\_L\DATA\HL070913\HL005534.D Vial: 1  
 Acq On : 09-Jul-2013, 20:09 Operator: QM  
 Sample : HSTDCCC0500 Inst : G1314A  
 Misc : Multiplr: 1.00  
 IntFile : autoint1.e

Instrument :  
 HPLC\_L  
 LabSampleId :  
 HSTDCCC0500

Method : Z:\HPCHEM1\HPLC\_L\METHODS\HL070213.M  
 Title :  
 Last Update : Tue Jul 02 16:59:51 2013  
 Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|      | Compound                   | AvgRF  | CCRF      | %Dev | Area% | Dev(min) |
|------|----------------------------|--------|-----------|------|-------|----------|
| 1 S  | 4-Nitroaniline             | 4.050  | 4.127 E3  | -1.9 | 101   | -0.06    |
| 2 T  | HMX                        | 4.606  | 3.793 E3  | 17.7 | 85    | -0.03    |
| 3 T  | RDX                        | 5.046  | 5.261 E3  | -4.3 | 103   | -0.07    |
| 4 T  | 1,3,5-Trinitrobenzene      | 11.148 | 10.931 E3 | 1.9  | 99    | -0.10    |
| 5 T  | 1,3-Dinitrobenzene         | 15.624 | 15.969 E3 | -2.2 | 99    | -0.15    |
| 6 T  | 3,5-Dinitroaniline         | 13.968 | 13.543 E3 | 3.0  | 100   | -0.17    |
| 7 T  | Nitrobenzene               | 9.954  | 9.752 E3  | 2.0  | 97    | -0.18    |
| 8 T  | Tetryl                     | 6.356  | 5.176 E3  | 18.6 | 85    | -0.39    |
| 9 T  | 2,4,6-Trinitrotoluene      | 11.742 | 12.584 E3 | -7.2 | 100   | -0.36    |
| 10 T | 2-Amino-4,6-dinitrotoluene | 11.757 | 10.972 E3 | 6.7  | 98    | -0.44    |
| 11 T | 4-Amino-2,6-dinitrotoluene | 5.764  | 6.286 E3  | -9.1 | 100   | -0.48    |
| 12 T | 2,4-Dinitrotoluene         | 14.968 | 15.093 E3 | -0.8 | 99    | -0.48    |
| 13 T | 2,6-Dinitrotoluene         | 8.365  | 7.971 E3  | 4.7  | 97    | -0.50#   |
| 14 T | 2-Nitrotoluene             | 6.383  | 6.535 E3  | -2.4 | 98    | -0.70#   |
| 15 T | 4-Nitrotoluene             | 5.395  | 5.357 E3  | 0.7  | 99    | -0.76#   |
| 16 T | 3-Nitrotoluene             | 6.949  | 6.696 E3  | 3.6  | 98    | -0.80#   |
| 17 T | PETN                       | 10.159 | 10.443 E3 | -2.8 | 89    | -1.11#   |

Evaluate Continuing Calibration Report - Not Found

Data File : S:\HPCHEM1\HPLC\_L\DATA\HL070913\HL005534.D Vial: 1  
 Acq On : 09-Jul-2013, 20:09 Operator: QM  
 Sample : HSTDCCC0500 Inst : G1314A  
 Misc : Multiplr: 1.00  
 IntFile : autoint1.e

Instrument :  
 HPLC\_L  
 LabSampleId :  
 HSTDCCC0500

Method : Z:\HPCHEM1\HPLC\_L\METHODS\HL070213.M  
 Title :  
 Last Update : Tue Jul 02 16:59:51 2013  
 Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

| Compound | AvgRF | CCRF | %Dev | Area% | Dev(min) |
|----------|-------|------|------|-------|----------|
|----------|-------|------|------|-------|----------|