

Data File : H:\VOL1\HPLC\HPLC L\120925\HL003535.D Vial: 43
Acq On : 25-Sep-2012, 18:13 Operator: HLM
Sample : D4282-03 Inst : G1314A
Misc : Multiplr: 1.00
IntFile : events.e
Quant Time: Sep 25 18:42 2012 Quant Results File: HL091912.RES

Instrument :
HPLC_L
ClientSampleId :
AI-S-3

Quant Method : H:\VOL1\HPLC\HPLC L\METHODS\HL091912.M (Chemstation Integrator)
Title : SW8330A/B Primary (Extend C18 3.5u 4.6x10mm)
Last Update : Tue Sep 25 17:44:26 2012
Response via : Initial Calibration
DataAcq Meth : 8330AGIP.M

Volume Inj. :
Signal Phase :
Signal Info :

Compound	R.T.	Response	Conc Units
----------	------	----------	------------

System Monitoring Compounds

1) S 4-Nitroaniline	1.34	546980	147.215 ug/L
Spiked Amount 300.000	Range 37 - 149	Recovery =	49.07%

Target Compounds

Quantitation Report

Data File : H:\VOL1\HPLC\HPLC L\120925\HL003535.D Vial: 43
 Acq On : 25-Sep-2012, 18:13 Operator: HLM
 Sample : D4282-03 Inst : G1314A
 Misc : Multiplr: 1.00
 IntFile : events.e
 Quant Time: Sep 25 18:42 2012 Quant Results File: HL091912.RES

Instrument :
 HPLC_L
 ClientSampleId :
 AI-S-3

Quant Method : H:\VOL1\HPLC\HPLC L\METHODS\HL091912.M (Chemstation Integrator)
 Title : SW8330A/B Primary (Extend C18 3.5u 4.6x10mm)
 Last Update : Tue Sep 25 17:44:26 2012
 Response via : Multiple Level Calibration
 DataAcq Meth : 8330AGIP.M

Volume Inj. :
 Signal Phase :
 Signal Info :

