

Data File : Z:\HPCHEM1\HPLC_L\DATA\HL110414\HL005824.D Vial: 2
Acq On : 04-Nov-2014, 15:39 Operator: TP
Sample : F4368-04 Inst : G1314A
Misc : Multiplr: 1.00
IntFile : AUTOINT1.E
Quant Time: Nov 5 7:08 2014 Quant Results File: HL102214.RES

Quant Method : Z:\HPCHEM1\HPLC_L\METHODS\HL102214.M
Title :
Last Update : Wed Nov 05 07:04:01 2014
Response via : Initial Calibration
DataAcq Meth : TEST.M

Volume Inj. : 50 ul
Signal Phase : ZORBAX Extend-C18
Signal Info : 4.6 X 100mm 3.5-Micron 80A

Compound	R.T.	Response	Conc Units
----------	------	----------	------------

System Monitoring Compounds

Target Compounds

Analyst Signature:_____ Analyst Name:_____ Date:_____

-----REASONS FOR MANUAL INTEGRATIONS-----

___Poor resolution of peaks exhibited on chromatogram.Compound #:_____

___Peak Integrated by software incorrectly.Compound #:_____

___OTHER:_____Compound #:_____

(f)=RT Delta > 1/2 Window (m)=manual int.

HL005824.D HL102214.M Fri Nov 07 23:01:57 2014

Data File : Z:\HPCHEM1\HPLC_L\DATA\HL110414\HL005824.D Vial: 2
Acq On : 04-Nov-2014, 15:39 Operator: TP
Sample : F4368-04 Inst : G1314A
Misc : Multiplr: 1.00
IntFile : AUTOINT1.E
Quant Time: Nov 5 7:08 2014 Quant Results File: HL102214.RES

Quant Method : Z:\HPCHEM1\HPLC_L\METHODS\HL102214.M
Title :
Last Update : Wed Nov 05 07:04:01 2014
Response via : Multiple Level Calibration
DataAcq Meth : TEST.M

Volume Inj. : 50 ul
Signal Phase : ZORBAX Extend-C18
Signal Info : 4.6 X 100mm 3.5-Micron 80A

