

Evaluate Continuing Calibration Report

Data File : H:\VOL1\HPLC\HPLC_N\120926\HN000507.D Vial: 64
Acq On : 26-Sep-2012, 21:06 Operator: HLM
Sample : NO-CCV Inst : G1314A
Misc : Multiplr: 1.00
IntFile : events.e

Method : H:\VOL1\HPLC\HPLC_N\METHODS\HN092512.M (Chemstation Integrator)
Title : SW8330 / 8330A confirmation
Last Update : Wed Sep 26 15:07:54 2012
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 T Nitroguanidine	139.020	142.318 E3	-2.4	0#	-0.03

Data File : H:\VOL1\HPLC\HPLC_N\120926\HN000507.D Vial: 64
Acq On : 26-Sep-2012, 21:06 Operator: HLM
Sample : NO-CCV Inst : G1314A
Misc : Multiplr: 1.00
IntFile : events.e

Method : H:\VOL1\HPLC\HPLC_N\METHODS\HN092512.M (Chemstation Integrator)
Title : SW8330 / 8330A confirmation
Last Update : Wed Sep 26 15:07:54 2012
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------