


4-Amino-2,6-dinitrotoluene




$$\text{Response} = 6.298\text{e}+003 * \text{Amt} + 9.189\text{e}+005$$

Coef of Det (r^2) = 0.992213 Curve Fit: Linear

Method Name: Z:\HPCHEM1\HPLC_N\Methods\HN052015.M

Calibration Table Last Updated: Thu May 21 13:50:13 2015

2,4-Dinitrotoluene




$$\text{Response} = 1.400\text{e}+004 * \text{Amt} + 2.304\text{e}+006$$

Coef of Det (r^2) = 0.999889 Curve Fit: Linear

Method Name: Z:\HPCHEM1\HPLC_N\Methods\HN052015.M

Calibration Table Last Updated: Thu May 21 13:50:13 2015

2,6-Dinitrotoluene



$$\text{Response} = 7.790\text{e}+003 * \text{Amt} + 1.544\text{e}+006$$

Coef of Det (r^2) = 0.999626 Curve Fit: Linear

Method Name: Z:\HPCHEM1\HPLC_N\Methods\HN052015.M

Calibration Table Last Updated: Thu May 21 13:50:13 2015