

Data File : Z:\HPCHEM1\HPLC_N\DATA\HN052914\HN006060.D Vial: 9
Acq On : 29-May-2014, 15:33 Operator: HNP
Sample : PB76451BL Inst : G1314A
Misc : Multiplr: 1.00
IntFile : AUTOINT1.E
Quant Time: May 30 8:48 2014 Quant Results File: HN0529NG.RES

Instrument :
HPLC_N
ClientSampleId :
PB76451BL

Manual Integrations
APPROVED

TEJASKUMAR
10/13/2014 4:09:58 PM

Quant Method : Z:\HPCHEM1\HPLC_N\METHODS\HN0529NG.M
Title :
Last Update : Fri May 30 03:32:30 2014
Response via : Initial Calibration
DataAcq Meth : 8330ANG.M

Volume Inj. : 50 ul
Signal Phase : ZORBAX Extend-C18
Signal Info : 4.6 X 100mm 3.5-Micron 80A

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
1) S 4-Nitroaniline	1.42	668750	0.181 ng/ulm

Target Compounds

Analyst Signature:_____ Analyst Name:_____ Date:_____

-----REASONS FOR MANUAL INTEGRATIONS-----

___Poor resolution of peaks exhibited on chromatogram.Compound #:_____

___Peak Integrated by software incorrectly.Compound #:_____

___OTHER:_____Compound #:_____

(f)=RT Delta > 1/2 Window (m)=manual int.

HN006060.D HN0529NG.M

Wed Jul 16 18:00:50 2014

RPT1

Quantitation Report

Data File : Z:\HPCHEM1\HPLC_N\DATA\HN052914\HN006060.D Vial: 9
Acq On : 29-May-2014, 15:33 Operator: HNP
Sample : PB76451BL Inst : G1314A
Misc : Multiplr: 1.00
IntFile : AUTOINT1.E
Quant Time: May 30 8:48 2014 Quant Results File: HN0529NG.RES


Instrument :
HPLC_N
ClientSampled :
PB76451BL


Manual Integrations
APPROVED

TEJASKUMAR
10/13/2014 4:09:58 PM

Quant Method : Z:\HPCHEM1\HPLC_N\METHODS\HN0529NG.M
Title :
Last Update : Fri May 30 03:32:30 2014
Response via : Multiple Level Calibration
DataAcq Meth : 8330ANG.M

Volume Inj. : 50 ul
Signal Phase : ZORBAX Extend-C18
Signal Info : 4.6 X 100mm 3.5-Micron 80A





#1 4-Nitroaniline

R.T.: 1.423 min
Delta R.T.: 0.003 min
Response: 668750
Conc: 0.18 ng/ul

Instrument :
HPLC_N
ClientSampleId :
PB76451BL

Manual Integrations
APPROVED

TEJASKUMAR
10/13/2014 4:09:58 PM