

# Evaluate Continuing Calibration Report

Data File : Z:\HPCHEM1\HPLC\_N\DATA\HN053013\HN005339.D Vial: 1  
 Acq On : 30-May-2013, 15:41 Operator: QM  
 Sample : HSTDCCC0500 Inst : G1314A  
 Misc : Multiplr: 1.00  
 IntFile : autoint1.e

Instrument :  
 HPLC\_N  
 LabSampleId :  
 HSTDCCC0500

Method : Z:\HPCHEM1\HPLC\_N\METHOD\HN052213.M  
 Title :  
 Last Update : Fri May 24 05:23:33 2013  
 Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|      | Compound                   | AvgRF  | CCRF      | %Dev   | Area% | Dev(min) |
|------|----------------------------|--------|-----------|--------|-------|----------|
| 1 S  | 4-Nitroaniline             | 4.250  | 4.349 E3  | -2.3   | 101   | 0.04     |
| 2 T  | HMX                        | 4.254  | 4.317 E3  | -1.5   | 104   | 0.03     |
| 3 T  | RDX                        | 5.143  | 5.050 E3  | 1.8    | 96    | 0.03     |
| 4 T  | 1,3,5-Trinitrobenzene      | 10.966 | 10.474 E3 | 4.5    | 96    | -0.03    |
| 5 T  | 1,3-Dinitrobenzene         | 13.023 | 15.250 E3 | -17.1  | 118   | -0.09    |
| 6 T  | 3,5-Dinitroaniline         | 15.568 | 12.585 E3 | 19.2   | 81    | -0.14    |
| 7 T  | Nitrobenzene               | 9.308  | 9.009 E3  | 3.2    | 98    | -0.15    |
| 8 T  | Tetryl                     | 6.006  | 4.888 E3  | 18.6   | 79    | -0.28    |
| 9 T  | 2,4,6-Trinitrotoluene      | 11.690 | 11.596 E3 | 0.8    | 99    | -0.21    |
| 10 T | 2-Amino-4,6-dinitrotoluene | 10.677 | 9.973 E3  | 6.6    | 96    | -0.27    |
| 11 T | 4-Amino-2,6-dinitrotoluene | 6.107  | 5.887 E3  | 3.6    | 90    | -0.27    |
| 12 T | 2,4-Dinitrotoluene         | 14.546 | 14.011 E3 | 3.7    | 98    | -0.25    |
| 13 T | 2,6-Dinitrotoluene         | 8.250  | 7.996 E3  | 3.1    | 97    | -0.26    |
| 14 T | 2-Nitrotoluene             | 6.237  | 6.229 E3  | 0.1    | 98    | -0.32    |
| 15 T | 4-Nitrotoluene             | 5.245  | 5.260 E3  | -0.3   | 98    | -0.33    |
| 16 T | 3-Nitrotoluene             | 6.642  | 6.856 E3  | -3.2   | 103   | -0.30    |
| 17 T | PETN                       | 8.590  | 16.434 E3 | -91.3# | 189#  | -0.26    |

Evaluate Continuing Calibration Report - Not Found

Data File : Z:\HPCHEM1\HPLC\_N\DATA\HN053013\HN005339.D Vial: 1  
 Acq On : 30-May-2013, 15:41 Operator: QM  
 Sample : HSTDCCC0500 Inst : G1314A  
 Misc : Multiplr: 1.00  
 IntFile : autoint1.e

Instrument :  
 HPLC\_N  
 LabSampleId :  
 HSTDCCC0500

Method : Z:\HPCHEM1\HPLC\_N\METHOD\HN052213.M  
 Title :  
 Last Update : Fri May 24 05:23:33 2013  
 Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

| Compound | AvgRF | CCRF | %Dev | Area% | Dev(min) |
|----------|-------|------|------|-------|----------|
|----------|-------|------|------|-------|----------|