

Evaluate Continuing Calibration Report

Signal #1 : Z:\HPCHEM1\HPLC_N\DATA\HN103114\HN006961.D\VWD1A.CH Vial: 91
 Signal #2 : Z:\HPCHEM1\HPLC_N\DATA\HN103114\HN006961.D\VWD1A.CH
 Acq On : 02-Nov-2014, 03:53 Operator: TP
 Sample : HSTDCCC005 Inst : G1314A
 Misc : Multiplr: 1.00
 IntFile Signal #1: AUTOINT1.E IntFile Signal #2: autoint2.e

Method : Z:\HPCHEM1\HPLC_N\METHODS\HN1029NG.M
 Title :
 Last Update : Fri Oct 31 04:50:16 2014
 Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 S	4-Nitroaniline	3.683	0.123 E6	96.7#	3#	0.00
2 T	Nitroglycerin	7.712	0.265 E6	96.6#	3#	0.02

Signal #1 : Z:\HPCHEM1\HPLC_N\DATA\HN103114\HN006961.D\VWD1A.CH Vial: 91
Signal #2 : Z:\HPCHEM1\HPLC_N\DATA\HN103114\HN006961.D\VWD1A.CH
Acq On : 02-Nov-2014, 03:53 Operator: TP
Sample : HSTDCCC005 Inst : G1314A
Misc : Multiplr: 1.00
IntFile Signal #1: AUTOINT1.E IntFile Signal #2: autoint2.e

Method : Z:\HPCHEM1\HPLC_N\METHODS\HN1029NG.M
Title :
Last Update : Fri Oct 31 04:50:16 2014
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------