

Data Path : Z:\HPCHEM1\HPLC_N\Data\HN110414\
Data File : HN007023.D
Signal(s) : VWD1A.CH
Acq On : 05-Nov-2014, 00:20
Operator : TP
Sample : HSTDCCC005
Misc :
ALS Vial : 21 Sample Multiplier: 1

Integration File: AUTOINT1.E
Quant Time: Mar 07 11:49:59 2015
Quant Method : Z:\HPCHEM1\HPLC_N\METHODS\HN1029NG.M
Quant Title :
QLast Update : Wed Nov 05 02:20:18 2014
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 50 ul
Signal Phase : ZORBAX Extend-C18
Signal Info : 4.6 X 100mm 3.5-Micron 80A

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
1) S 4-Nitroaniline	1.126	1899801	0.516 ng/ulm
Target Compounds			

(f)=RT Delta > 1/2 Window

(m)=manual int.

Instrument :
HPLC_N
ClientSampleId :
HSTDCCC005

Manual Integrations
APPROVED

mohammad
3/16/2015 12:49:11 PM

Data Path : Z:\HPCHEM1\HPLC_N\Data\HN110414\
Data File : HN007023.D
Signal(s) : VWD1A.CH
Acq On : 05-Nov-2014, 00:20
Operator : TP
Sample : HSTDCCC005
Misc :
ALS Vial : 21 Sample Multiplier: 1


Instrument :
HPLC_N
ClientSampleId :
HSTDCCC005


Manual Integrations
APPROVED

mohammad
3/16/2015 12:49:11 PM

Integration File: AUTOINT1.E
Quant Time: Mar 07 11:49:59 2015
Quant Method : Z:\HPCHEM1\HPLC_N\METHODS\HN1029NG.M
Quant Title :
QLast Update : Wed Nov 05 02:20:18 2014
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 50 ul
Signal Phase : ZORBAX Extend-C18
Signal Info : 4.6 X 100mm 3.5-Micron 80A





#1 4-Nitroaniline

R.T.: 1.126 min
Delta R.T.: -0.170 min
Response: 1899801
Conc: 0.52 ng/u1 m

Instrument :
HPLC_N
ClientSampleId :
HSTDCCC005

Manual Integrations
APPROVED

mohammad
3/16/2015 12:49:11 PM