

Evaluate Continuing Calibration Report

Data File : Z:\HPCHEM1\HPLC_N\DATA\HN110813\HN005679.D Vial: 7
 Acq On : 08-Nov-2013, 20:41 Operator: QM
 Sample : HSTDICV0500 Inst : G1314A
 Misc : Multiplr: 1.00
 IntFile : autoint1.e

Instrument :
 HPLC_N
 ClientSampleId :
 ICVHN110813

Method : Z:\HPCHEM1\HPLC_N\METHOD\HN110813.M
 Title :
 Last Update : Mon Jul 22 14:51:19 2013
 Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 S	4-Nitroaniline	3.691	3.780 E3	-2.4	101	0.00
2 T	HMX	4.015	4.045 E3	-0.7	101	0.00
3 T	RDX	5.470	5.351 E3	2.2	99	0.00
4 T	1,3,5-Trinitrobenzene	10.397	10.466 E3	-0.7	101	0.00
5 T	3,5-Dinitroaniline	16.128	16.088 E3	0.2	101	0.00
6 T	1,3-Dinitrobenzene	12.692	12.605 E3	0.7	101	0.00
7 T	Nitrobenzene	10.192	10.014 E3	1.7	100	0.00
8 T	Tetryl	6.865	7.432 E3	-8.3	102	0.01
9 T	2,4,6-Trinitrotoluene	11.359	11.105 E3	2.2	101	0.01
10 T	2-Amino-4,6-dinitrotoluene	10.912	10.673 E3	2.2	101	0.01
11 T	4-Amino-2,6-dinitrotoluene	5.864	6.687 E3	-14.0	102	0.02
12 T	2,4-Dinitrotoluene	14.768	14.772 E3	-0.0	103	0.01
13 T	2,6-Dinitrotoluene	7.883	8.315 E3	-5.5	105	0.02
14 T	2-Nitrotoluene	6.398	6.626 E3	-3.6	101	0.01
15 T	4-Nitrotoluene	5.392	5.296 E3	1.8	100	0.00
16 T	3-Nitrotoluene	6.858	6.676 E3	2.7	101	0.02
17 T	PETN	8.659	8.759 E3	-1.2	99	0.02

Evaluate Continuing Calibration Report - Not Found

Data File : Z:\HPCHEM1\HPLC_N\DATA\HN110813\HN005679.D Vial: 7
 Acq On : 08-Nov-2013, 20:41 Operator: QM
 Sample : HSTDICV0500 Inst : G1314A
 Misc : Multiplr: 1.00
 IntFile : autoint1.e

Instrument :
 HPLC_N
 ClientSampleId :
 ICVHN110813

Method : Z:\HPCHEM1\HPLC_N\METHOD\HN110813.M
 Title :
 Last Update : Mon Jul 22 14:51:19 2013
 Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------