

Evaluate Continuing Calibration Report

Data File : Z:\HPCHEM1\HPLC_N\DATA\HN110813\HN005680.D Vial: 4
 Acq On : 08-Nov-2013, 21:09 Operator: QM
 Sample : HSTDCCC500 Inst : G1314A
 Misc : Multiplr: 1.00
 IntFile : autoint1.e

Instrument :
 HPLC_N
 LabSampleId :
 HSTDCCC500

Method : Z:\HPCHEM1\HPLC_N\METHOD\HN110813.M
 Title :
 Last Update : Mon Jul 22 14:51:19 2013
 Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 S	4-Nitroaniline	3.691	3.782 E3	-2.5	101	0.00
2 T	HMX	4.015	4.023 E3	-0.2	100	0.00
3 T	RDX	5.470	5.350 E3	2.2	99	0.00
4 T	1,3,5-Trinitrobenzene	10.397	10.447 E3	-0.5	100	0.01
5 T	3,5-Dinitroaniline	16.128	16.027 E3	0.6	100	0.01
6 T	1,3-Dinitrobenzene	12.692	12.566 E3	1.0	100	0.02
7 T	Nitrobenzene	10.192	10.002 E3	1.9	100	0.02
8 T	Tetryl	6.865	7.346 E3	-7.0	101	0.03
9 T	2,4,6-Trinitrotoluene	11.359	11.011 E3	3.1	101	0.03
10 T	2-Amino-4,6-dinitrotoluene	10.912	10.588 E3	3.0	100	0.03
11 T	4-Amino-2,6-dinitrotoluene	5.864	6.610 E3	-12.7	101	0.03
12 T	2,4-Dinitrotoluene	14.768	14.491 E3	1.9	101	0.03
13 T	2,6-Dinitrotoluene	7.883	7.834 E3	0.6	99	0.03
14 T	2-Nitrotoluene	6.398	6.636 E3	-3.7	101	0.02
15 T	4-Nitrotoluene	5.392	5.298 E3	1.7	100	0.02
16 T	3-Nitrotoluene	6.858	6.629 E3	3.3	101	0.02
17 T	PETN	8.659	8.436 E3	2.6	96	0.05

Evaluate Continuing Calibration Report - Not Found

Data File : Z:\HPCHEM1\HPLC_N\DATA\HN110813\HN005680.D Vial: 4
 Acq On : 08-Nov-2013, 21:09 Operator: QM
 Sample : HSTDCCC500 Inst : G1314A
 Misc : Multiplr: 1.00
 IntFile : autoint1.e

Instrument :
 HPLC_N
 LabSampleId :
 HSTDCCC500

Method : Z:\HPCHEM1\HPLC_N\METHOD\HN110813.M
 Title :
 Last Update : Mon Jul 22 14:51:19 2013
 Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------