

Evaluate Continuing Calibration Report

Data File : Z:\HPCHEM1\HPLC_N\DATA\HN111113\HN005702.D Vial: 4
 Acq On : 11-Nov-2013, 21:29 Operator: QM
 Sample : HSTDCCC500 Inst : G1314A
 Misc : Multiplr: 1.00
 IntFile : autoint1.e

Instrument :
 HPLC_N
 LabSampleId :
 HSTDCCC500

Method : Z:\HPCHEM1\HPLC_N\METHOD\HN111113.M
 Title :
 Last Update : Mon Nov 11 16:27:41 2013
 Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 S	4-Nitroaniline	3.921	3.830 E3	2.3	102	0.01
2 T	HMX	4.036	3.984 E3	1.3	99	0.00
3 T	RDX	5.574	5.301 E3	4.9	103	0.02
4 T	1,3,5-Trinitrobenzene	10.401	10.341 E3	0.6	101	0.03
5 T	3,5-Dinitroaniline	16.901	16.235 E3	3.9	102	0.05
6 T	1,3-Dinitrobenzene	13.811	12.857 E3	6.9	101	0.06
7 T	Nitrobenzene	10.368	10.230 E3	1.3	103	0.07
8 T	Tetryl	7.009	7.002 E3	0.1	100	0.12
9 T	2,4,6-Trinitrotoluene	11.488	10.689 E3	7.0	99	0.10
10 T	2-Amino-4,6-dinitrotoluene	10.757	9.873 E3	8.2	97	0.12
11 T	4-Amino-2,6-dinitrotoluene	6.315	6.337 E3	-0.3	98	0.13
12 T	2,4-Dinitrotoluene	14.734	15.094 E3	-2.4	104	0.13
13 T	2,6-Dinitrotoluene	8.211	8.373 E3	-2.0	105	0.14
14 T	2-Nitrotoluene	6.409	6.318 E3	1.4	100	0.19
15 T	4-Nitrotoluene	5.189	5.264 E3	-1.4	102	0.22
16 T	3-Nitrotoluene	19.675	19.729 E3	-0.3	100	0.23
17 T	PETN	9.274	8.967 E3	3.3	95	0.35

Evaluate Continuing Calibration Report - Not Found

Data File : Z:\HPCHEM1\HPLC_N\DATA\HN111113\HN005702.D Vial: 4
 Acq On : 11-Nov-2013, 21:29 Operator: QM
 Sample : HSTDCCC500 Inst : G1314A
 Misc : Multiplr: 1.00
 IntFile : autoint1.e

Instrument :
 HPLC_N
 LabSampleId :
 HSTDCCC500

Method : Z:\HPCHEM1\HPLC_N\METHOD\HN111113.M
 Title :
 Last Update : Mon Nov 11 16:27:41 2013
 Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------