


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010422\
 Data File : BN018327.D
 Acq On : 04 Jan 2022 09:28
 Operator : CG/JU
 Sample : DFTPP222
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP222

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN010422.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Jan 04 13:43:36 2022



AutoFind: Scans 2458, 2459, 2460; Background Corrected with Scan 2451

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
51	198	10	80	34.2	641126	PASS
68	69	0.00	2	1.8	11255	PASS
69	198	0.00	100	33.7	631529	PASS
70	69	0.00	2	0.6	3673	PASS
127	198	10	80	38.8	727019	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	1874261	PASS
199	198	5	9	7.0	131616	PASS
275	198	10	60	29.9	560853	PASS
365	198	1	100	3.6	68267	PASS
441	198	0.01	100	13.3	249280	PASS
442	198	50	100	91.9	1722539	PASS
443	442	15	24	18.4	317717	PASS

m/z	Abundance
35.90	666.0
39.00	615.0
40.00	3611.0
41.00	829.0
42.00	989.0
43.00	1798.0
44.00	10777.0
44.90	371.0
48.90	2151.0
50.00	444.0
51.00	815.0
55.00	1205.0
57.10	921.0
69.00	578.0
71.00	742.0
72.00	654.0
73.00	1213.0
74.90	538.0
76.90	460.0
78.00	479.0
78.90	386.0
80.20	316.0
80.90	442.0
82.90	329.0
84.00	1670.0
85.20	327.0
86.00	1190.0
91.00	390.0
96.10	734.0
97.00	404.0
115.00	350.0
129.00	606.0
133.10	628.0
135.10	521.0
141.00	402.0
153.80	330.0
163.10	443.0
191.00	758.0
193.00	611.0
207.00	3738.0
208.00	1248.0
208.90	590.0
267.00	334.0
281.00	1628.0
282.10	382.0
340.90	314.0

Instrument :
BNA_N
ClientSampleId :
DFTPP222

m/z	Abundance
36.00	1039.0
37.10	3782.0
38.10	9321.0
39.10	47536.0
40.00	4231.0
41.00	2085.0
42.00	527.0
42.90	752.0
44.00	8825.0
45.00	1070.0
46.70	374.0
47.90	1163.0
48.90	6734.0
50.00	174528.0
51.00	682496.0
52.00	36032.0
53.10	1673.0
54.90	3601.0
56.00	17816.0
57.00	42568.0
57.90	2181.0
59.00	894.0
59.90	1243.0
61.00	8221.0
62.00	9452.0
63.00	24344.0
64.00	4637.0
65.00	13585.0
66.00	1011.0
67.00	1822.0
68.10	10840.0
69.00	660544.0
70.00	3699.0
70.90	626.0
72.10	310.0
73.00	6322.0
74.00	74104.0
75.00	104824.0
76.00	42400.0
77.00	796800.0
78.00	56712.0
79.00	54792.0
80.00	39168.0
81.00	54072.0
82.00	13028.0
83.00	10912.0
84.00	2397.0
85.00	11469.0
86.00	15167.0
87.00	6998.0
87.90	3026.0
89.00	1380.0
91.00	12766.0
92.00	14265.0
93.00	84768.0
93.90	7149.0
94.90	1528.0
96.00	3795.0
97.00	1501.0
98.00	67856.0
99.00	59336.0
100.00	4282.0
101.00	29656.0
101.90	1267.0
103.00	9440.0
104.00	21264.0
105.00	16416.0
106.00	6683.0
107.00	218688.0
108.00	33904.0
109.00	7352.0
110.00	397312.0
111.00	62576.0
112.00	7766.0
113.00	3280.0
114.20	851.0
115.00	1108.0
116.00	11706.0
117.00	204416.0
118.00	17056.0
119.00	2502.0
120.10	2638.0
120.80	1422.0
122.00	13956.0
123.00	22704.0
124.00	10176.0
125.00	8603.0
127.00	738304.0
128.00	54408.0
129.00	259200.0
130.00	27064.0

Instrument :
BNA_N
ClientSampleId :
DFTPP222

131.00	4496.0
132.00	2359.0
133.00	1381.0
134.00	7920.0
135.00	24792.0
136.10	9549.0
137.00	14472.0
138.00	3134.0
139.00	1945.0
140.00	3449.0
141.00	41096.0
142.00	11803.0
143.00	9590.0
144.00	2611.0
145.00	1651.0
146.00	5474.0
147.00	21288.0
148.00	46104.0
149.00	11602.0
150.00	3737.0
151.10	4914.0
151.90	2842.0
153.00	12424.0
154.00	9038.0
155.00	24128.0
156.00	35752.0
157.10	6863.0
158.00	7010.0
159.00	6424.0
160.00	13571.0
161.00	19928.0
162.00	5824.0
162.90	1829.0
164.00	2399.0
165.00	15213.0
166.00	12461.0
167.00	89208.0
168.00	66696.0
169.00	8531.0
170.00	3109.0
171.00	4481.0
172.00	7443.0
172.90	9447.0
174.00	17336.0
175.00	34424.0
176.00	9954.0
177.00	15595.0
178.00	4259.0
179.00	62136.0
180.00	44976.0
181.00	18184.0
182.00	4020.0
182.80	2362.0
184.00	5214.0
185.00	32080.0
186.00	241280.0
187.00	69648.0
188.00	8226.0
189.00	13486.0
189.90	3459.0
191.00	7085.0
192.00	20144.0
193.00	22992.0
194.00	4824.0
195.00	1852.0
196.00	49352.0
198.00	1671168.0
199.00	130904.0
200.00	11251.0
201.30	7827.0
203.00	11705.0
204.00	69368.0
205.00	109976.0
206.00	490368.0
207.10	68856.0
208.00	17736.0
209.00	5405.0
210.20	7360.0
211.00	19112.0
213.20	1075.0
215.00	5691.0
216.10	9156.0
217.00	124344.0
218.00	17616.0
219.00	1877.0
221.00	79728.0
221.90	14587.0
223.00	30056.0
224.00	260160.0
225.00	71000.0
226.00	9169.0
227.00	117024.0
228.00	16368.0
229.00	24448.0
230.00	3011.0

Instrument :
BNA_N
ClientSampleId :
DFTPP222

231.00	10324.0
232.00	1720.0
233.00	2047.0
234.10	6328.0
235.00	6961.0
235.90	4979.0
237.00	12099.0
238.00	2409.0
238.90	3550.0
240.00	3902.0
240.90	7144.0
242.00	17264.0
243.00	18024.0
244.00	214272.0
245.00	31520.0
246.00	39200.0
247.00	7697.0
247.90	1831.0
249.10	7112.0
250.00	1086.0
251.10	1486.0
252.00	2199.0
253.00	5020.0
255.00	1106432.0
256.00	163584.0
257.00	13526.0
258.00	58432.0
258.90	8512.0
259.80	1902.0
260.90	1968.0
261.90	468.0
263.10	898.0
264.10	2041.0
265.00	23352.0
266.00	4067.0
266.90	678.0
267.90	371.0
270.10	1476.0
271.00	2493.0
272.00	3882.0
273.00	30552.0
274.00	87864.0
275.00	486400.0
276.00	65768.0
277.00	36640.0
278.00	6490.0
278.90	2204.0
281.10	1127.0
281.90	1135.0
283.00	5793.0
284.00	3338.0
285.00	7966.0
286.10	1300.0
288.00	523.0
288.90	1918.0
289.80	876.0
291.10	1132.0
292.00	2368.0
293.00	9574.0
294.00	3444.0
295.10	2885.0
296.00	126592.0
297.00	20720.0
298.00	1776.0
301.00	1651.0
302.00	3193.0
303.00	13699.0
304.10	4648.0
304.90	833.0
307.10	373.0
307.90	2100.0
308.90	1352.0
310.00	1654.0
311.10	516.0
312.00	540.0
312.90	935.0
314.00	6424.0
315.00	13729.0
316.00	7529.0
317.00	1528.0
319.20	455.0
319.90	596.0
321.10	4317.0
322.10	2052.0
323.00	39592.0
324.10	7619.0
325.00	869.0
326.00	1027.0
327.00	8100.0
328.00	5586.0
329.10	988.0
331.90	3404.0
333.10	4960.0
334.00	24160.0
335.00	6526.0

Instrument :
BNA_N
ClientSampleId :
DFTPP222

336.00	1286.0
338.90	876.0
340.10	937.0
341.00	5163.0
342.00	1532.0
343.00	401.0
346.00	8701.0
346.90	2109.0
348.10	427.0
349.90	471.0
350.90	1374.0
352.00	12518.0
353.00	9650.0
354.00	15607.0
355.10	1899.0
356.90	452.0
357.90	402.0
359.10	912.0
361.10	308.0
365.00	55872.0
365.90	7418.0
366.90	966.0
370.00	1462.0
371.00	3513.0
372.00	19704.0
372.90	6609.0
373.90	788.0
376.90	386.0
377.80	360.0
382.30	331.0
383.00	5241.0
383.90	2260.0
385.10	1033.0
389.90	3298.0
390.90	2066.0
392.00	1060.0
392.90	305.0
396.90	604.0
400.90	1038.0
402.00	7392.0
403.00	13069.0
404.00	4724.0
404.90	1258.0
405.80	419.0
414.90	545.0
421.00	8445.0
422.10	11005.0
423.00	77224.0
424.10	28616.0
424.90	3098.0
438.70	455.0
439.70	384.0
441.00	188864.0
442.00	1248768.0
443.00	238144.0
444.00	25600.0
445.00	932.0

Instrument :
BNA_N
ClientSampleId :
DFTPP222

m/z	Abundance
36.00	1333.0
37.00	2826.0
38.00	10040.0
39.10	54936.0
40.00	4808.0
41.00	1815.0
43.00	1350.0
44.00	9397.0
44.90	1206.0
47.00	556.0
47.90	580.0
49.00	6839.0
50.00	206656.0
51.00	763456.0
52.00	37976.0
53.00	1791.0
55.00	4860.0
56.00	20360.0
57.00	46024.0
58.00	2100.0
59.00	802.0
59.90	1209.0
61.00	8965.0
62.00	11358.0
63.00	29808.0
64.00	4171.0
65.00	14500.0
66.20	834.0
67.10	948.0
68.00	14939.0
69.00	748288.0
70.00	3943.0
71.10	919.0
73.00	7972.0
74.00	87352.0
75.00	127704.0
76.00	45352.0
77.00	914496.0
78.00	60480.0
79.00	59112.0
80.00	46064.0
81.00	65472.0
82.00	15755.0
83.00	13346.0
83.90	2771.0
85.00	14220.0
86.00	17448.0
87.00	7752.0
87.90	3777.0
88.90	1658.0
89.90	385.0
91.00	12282.0
92.00	17472.0
93.00	108304.0
94.00	7425.0
94.90	1517.0
96.00	5109.0
97.00	1426.0
98.00	83336.0
99.00	66040.0
100.00	4865.0
101.00	34248.0
101.90	1547.0
103.00	9691.0
104.00	18912.0
105.00	18800.0
106.00	8443.0
107.00	265408.0
108.00	41784.0
109.00	8422.0
110.00	467648.0
111.00	69576.0
112.00	8658.0
112.80	2907.0
113.70	430.0
114.30	576.0
115.20	807.0
116.00	15027.0
117.00	241984.0
117.90	20552.0
119.00	2163.0
120.10	3588.0
121.00	1190.0
122.00	14910.0
123.00	23920.0
124.00	10855.0
125.00	9828.0
127.00	872064.0
128.00	62696.0
129.00	317568.0
130.00	28832.0

Instrument :
BNA_N
ClientSampleId :
DFTPP222

130.90	5212.0
132.10	2439.0
132.80	1503.0
134.00	8025.0
135.00	27000.0
136.00	12207.0
137.00	16576.0
138.00	1929.0
139.00	1607.0
140.00	4472.0
141.00	44416.0
142.00	12846.0
143.00	9925.0
143.90	3658.0
145.10	2398.0
146.00	9728.0
147.00	25072.0
148.00	66320.0
149.00	13339.0
150.00	2986.0
151.10	6785.0
151.90	2916.0
152.00	2909.0
153.10	14320.0
154.00	11828.0
155.10	26896.0
156.00	43960.0
157.10	8909.0
157.90	8504.0
159.10	6229.0
160.00	15294.0
161.00	22344.0
162.00	6596.0
163.10	1935.0
163.80	2312.0
164.90	17824.0
166.00	15279.0
167.00	111592.0
168.00	76288.0
168.90	11610.0
170.00	2558.0
171.00	4166.0
172.00	8521.0
173.00	11334.0
174.00	23904.0
175.00	40904.0
176.00	13401.0
177.00	18624.0
178.10	6117.0
179.00	76048.0
180.00	52448.0
181.00	29208.0
182.00	4526.0
182.80	2437.0
184.00	7518.0
185.00	43416.0
186.00	306176.0
187.00	86224.0
188.00	9715.0
189.00	18640.0
189.90	2313.0
191.00	7644.0
192.00	23080.0
193.00	30736.0
194.00	7800.0
194.90	3343.0
196.00	64712.0
198.00	2260480.0
199.00	155392.0
199.90	11288.0
201.40	12220.0
203.00	15166.0
204.00	80096.0
205.00	132800.0
206.00	614464.0
207.00	86496.0
208.00	19952.0
209.00	6663.0
210.20	9949.0
211.00	24416.0
213.00	1623.0
214.20	786.0
215.00	7052.0
216.00	11442.0
217.00	166592.0
218.00	20728.0
219.00	1853.0
220.00	3618.0
221.00	104944.0
221.80	23392.0
223.00	38280.0
224.00	359488.0
225.00	98976.0
226.00	11024.0
227.00	145792.0

Instrument :
BNA_N
ClientSampleId :
DFTPP222

228.00	20400.0
229.00	30344.0
230.00	5187.0
231.00	11600.0
231.90	1454.0
232.90	2692.0
234.00	8861.0
235.10	11483.0
235.90	8728.0
237.00	14609.0
238.00	2212.0
239.10	5317.0
240.10	4906.0
241.00	6483.0
242.00	18512.0
243.10	24784.0
244.00	291776.0
245.00	35688.0
246.00	54256.0
247.00	13406.0
248.00	2238.0
249.00	9875.0
249.90	1834.0
251.00	2787.0
252.00	2725.0
253.10	4371.0
255.00	1435136.0
256.00	212352.0
257.00	17608.0
258.00	82384.0
259.00	11750.0
260.00	2567.0
261.00	2298.0
263.00	718.0
264.10	2364.0
265.00	30992.0
266.00	4796.0
267.00	833.0
268.10	526.0
270.10	1127.0
271.00	4105.0
272.20	4915.0
273.00	39912.0
274.00	109536.0
275.00	643328.0
276.00	86216.0
277.00	50448.0
278.00	9108.0
279.10	2489.0
280.80	928.0
281.20	493.0
282.00	2101.0
283.00	6753.0
284.00	3888.0
285.00	10671.0
286.10	1720.0
287.90	960.0
289.00	1998.0
290.00	2466.0
290.90	1925.0
292.00	3038.0
293.00	12351.0
294.00	4344.0
296.00	172480.0
297.00	24800.0
297.90	1807.0
299.00	415.0
301.00	2397.0
301.90	2193.0
303.00	19240.0
304.00	6482.0
304.90	797.0
306.90	332.0
308.00	2067.0
309.00	1301.0
310.00	3150.0
310.90	476.0
311.90	899.0
313.10	1280.0
313.90	7876.0
315.00	20624.0
316.00	10906.0
317.00	2291.0
319.00	653.0
320.30	780.0
321.00	5054.0
322.10	2970.0
323.00	54912.0
324.10	12366.0
324.90	1215.0
326.00	1107.0
327.00	10848.0
328.00	5070.0
329.00	822.0
330.10	346.0

Instrument :
BNA_N
ClientSampleId :
DFTPP222

332.10	3721.0
333.00	5640.0
334.10	40312.0
335.10	10514.0
336.10	1487.0
339.00	1044.0
339.90	694.0
341.00	6954.0
342.10	2307.0
346.00	13286.0
346.90	2653.0
348.10	815.0
351.00	1361.0
352.00	20096.0
353.00	11266.0
354.10	19448.0
355.00	4681.0
355.90	423.0
358.10	391.0
359.00	1464.0
361.10	636.0
363.70	648.0
364.90	77888.0
366.00	12601.0
366.90	828.0
368.90	324.0
370.00	1363.0
371.00	4792.0
372.00	32792.0
373.00	6963.0
373.90	985.0
377.00	1032.0
381.80	373.0
383.00	11191.0
384.00	3103.0
385.00	847.0
388.90	357.0
390.00	3245.0
390.80	3084.0
392.00	1639.0
392.90	308.0
394.80	455.0
396.90	598.0
398.90	318.0
401.00	2748.0
401.90	13886.0
403.00	21624.0
404.00	7310.0
405.10	2347.0
409.90	627.0
415.00	1268.0
415.90	351.0
421.00	15703.0
422.00	15484.0
423.00	109240.0
424.00	34696.0
425.00	3321.0
426.00	445.0
438.60	366.0
439.30	720.0
441.00	282112.0
442.00	2078720.0
443.00	408576.0
444.00	37504.0
444.90	2070.0

Instrument :
BNA_N
ClientSampleId :
DFTPP222

m/z	Abundance
36.00	859.0
37.00	2219.0
38.10	6745.0
39.00	36400.0
40.00	4445.0
41.00	2004.0
42.00	782.0
43.00	1546.0
44.00	9303.0
45.00	1722.0
47.00	727.0
48.00	862.0
49.10	5860.0
50.00	125648.0
51.00	479872.0
52.00	25248.0
53.00	1006.0
54.90	2773.0
56.00	14808.0
57.00	29640.0
57.90	1215.0
59.00	814.0
59.90	711.0
60.90	6174.0
62.00	8221.0
63.00	18456.0
64.00	1872.0
65.00	9376.0
66.00	794.0
67.20	979.0
68.00	7985.0
69.00	487488.0
69.90	3376.0
71.00	1153.0
72.00	565.0
73.00	5576.0
74.00	55312.0
75.00	82136.0
76.00	28792.0
77.00	605312.0
78.00	38976.0
79.00	37528.0
80.00	28576.0
81.00	37848.0
82.00	10273.0
83.00	8658.0
83.90	2024.0
85.00	9634.0
86.00	11088.0
87.00	7366.0
88.00	1661.0
89.00	796.0
90.10	347.0
91.00	8180.0
92.00	11148.0
93.00	69152.0
94.00	4755.0
95.00	1479.0
96.00	2608.0
97.10	2189.0
98.00	55112.0
99.00	40176.0
100.00	3861.0
101.00	22296.0
101.90	1175.0
102.90	7100.0
104.00	13330.0
104.90	13422.0
106.00	4511.0
107.00	180096.0
108.00	29096.0
109.10	6692.0
110.00	303872.0
111.00	44424.0
112.00	5236.0
112.90	2008.0
113.80	319.0
114.80	796.0
116.00	9958.0
117.00	165952.0
118.00	11792.0
118.90	1793.0
120.00	1945.0
121.20	1393.0
122.00	10721.0
123.00	18496.0
124.00	9128.0
125.00	7020.0
127.00	570688.0
128.00	42936.0
129.00	223808.0

Instrument :
BNA_N
ClientSampleId :
DFTPP222

130.00	19744.0
131.00	3083.0
132.00	1855.0
133.20	1496.0
134.00	6846.0
135.00	16760.0
136.00	8424.0
137.10	11070.0
138.00	2673.0
139.20	1055.0
140.10	2838.0
141.00	26256.0
142.00	12332.0
143.00	5998.0
144.00	1957.0
145.00	1598.0
146.00	5917.0
147.00	16020.0
148.00	43432.0
149.00	8394.0
149.90	2584.0
151.00	4555.0
151.50	3219.0
153.00	9248.0
154.00	6656.0
155.00	19704.0
156.00	28496.0
156.90	6723.0
158.00	6971.0
159.00	5245.0
160.00	11095.0
161.00	16424.0
162.10	3899.0
162.90	1475.0
164.00	1725.0
165.00	11848.0
166.00	10176.0
167.00	84528.0
168.00	47648.0
169.10	7689.0
170.10	2306.0
171.00	1916.0
171.90	6450.0
173.00	8895.0
174.00	16014.0
175.00	28232.0
176.00	9006.0
177.00	13307.0
178.00	4818.0
178.90	51736.0
180.00	41416.0
181.00	19616.0
181.90	3122.0
183.10	2145.0
184.00	4925.0
185.00	29448.0
186.00	219328.0
187.00	58176.0
188.00	6127.0
189.00	12240.0
190.00	1737.0
191.00	6784.0
192.00	18616.0
193.00	22960.0
194.10	4949.0
194.90	3628.0
196.00	46056.0
198.00	1691136.0
198.90	108552.0
200.00	8424.0
201.50	7259.0
203.00	12327.0
204.00	64680.0
205.00	106080.0
206.00	463616.0
207.00	59704.0
208.00	17752.0
209.10	5752.0
210.30	7827.0
211.00	17944.0
212.20	2577.0
213.00	2104.0
213.80	726.0
215.00	4516.0
216.00	11136.0
217.00	128176.0
218.00	16030.0
219.00	2540.0
221.00	77352.0
221.90	16187.0
223.00	27480.0
224.00	287168.0
225.00	67392.0
226.10	8710.0
227.00	114904.0

Instrument :
BNA_N
ClientSampleId :
DFTPP222

228.00	14582.0
229.00	23928.0
230.00	3225.0
231.00	9482.0
232.00	1888.0
232.90	2666.0
233.90	7815.0
235.00	7417.0
236.00	5036.0
237.00	9217.0
238.00	1854.0
239.00	3457.0
240.00	3656.0
241.00	5794.0
242.00	15768.0
243.10	15822.0
244.00	209408.0
245.00	30240.0
246.00	41992.0
246.90	6958.0
248.10	1988.0
248.90	8026.0
250.10	1677.0
251.10	1559.0
251.80	1788.0
253.10	4247.0
254.10	10464.0
255.00	1099264.0
256.00	174208.0
257.00	12694.0
258.00	65392.0
259.00	11455.0
259.90	1759.0
261.00	3013.0
262.00	333.0
262.80	385.0
263.80	1948.0
264.20	1557.0
265.00	28144.0
266.00	3494.0
267.10	743.0
268.00	710.0
269.90	1066.0
271.00	3510.0
272.10	3046.0
273.00	30080.0
274.00	95632.0
275.00	552832.0
276.00	65720.0
277.00	43040.0
278.00	8262.0
279.00	2020.0
280.00	718.0
281.00	1560.0
282.00	948.0
283.00	5354.0
284.00	3616.0
285.00	6225.0
286.10	1414.0
288.00	626.0
288.90	1724.0
289.90	1159.0
290.90	936.0
292.00	1507.0
293.00	10491.0
294.00	3502.0
295.00	3530.0
296.00	151296.0
297.00	20560.0
297.90	1095.0
299.20	350.0
300.10	373.0
301.10	2437.0
301.90	3243.0
303.00	15050.0
304.00	5561.0
304.80	767.0
307.20	550.0
307.90	1251.0
308.90	1295.0
309.90	2217.0
311.00	451.0
311.80	631.0
312.90	858.0
314.10	7701.0
315.00	15974.0
316.00	8978.0
317.00	1650.0
319.10	361.0
319.90	560.0
321.00	5499.0
322.00	3286.0
323.10	53352.0
324.10	8391.0
325.10	701.0

Instrument :
BNA_N
ClientSampleId :
DFTPP222

326.00	857.0
326.90	9006.0
328.00	4832.0
328.80	814.0
329.90	387.0
331.20	551.0
332.00	2701.0
333.10	3187.0
334.00	30232.0
335.00	7202.0
336.10	1036.0
337.00	454.0
338.90	843.0
339.80	554.0
341.00	7655.0
342.00	1635.0
345.90	11264.0
346.90	2295.0
351.00	876.0
352.00	17928.0
353.00	9153.0
354.00	17528.0
354.90	3490.0
355.80	396.0
359.00	1461.0
361.10	452.0
363.90	1382.0
365.00	71040.0
366.00	9756.0
367.20	804.0
370.10	1247.0
370.90	4694.0
372.00	26240.0
373.00	6602.0
374.00	1406.0
377.00	1122.0
383.00	6624.0
383.90	2450.0
385.00	1124.0
388.90	333.0
390.00	4666.0
391.00	3022.0
392.00	2010.0
393.00	525.0
395.10	498.0
400.90	1626.0
402.00	12287.0
403.00	18752.0
404.00	5251.0
404.90	1200.0
409.00	456.0
410.00	617.0
415.00	491.0
415.90	580.0
420.00	685.0
421.00	16456.0
422.00	15293.0
423.00	111296.0
424.00	34224.0
424.90	4388.0
425.90	514.0
439.00	1285.0
441.10	276864.0
442.10	1840128.0
443.00	306432.0
444.00	31232.0
445.20	2163.0

Instrument :
BNA_N
ClientSampleId :
DFTPP222