

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010523\
 Data File : BN023553.D
 Acq On : 04 Jan 2023 15:19
 Operator : CG/JU
 Sample : SST00553
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SST005217

Quant Time: Jan 05 02:35:56 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN010523.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 05 02:34:34 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.952	152	161615	20.000	ng/u1	0.00
20) Naphthalene-d8	10.769	136	708381	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.604	164	456322	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.351	188	1038452	20.000	ng/u1	0.00
79) Chrysene-d12	21.539	240	958813	20.000	ng/u1	0.00
88) Perylene-d12	23.968	264	998132	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.299	96	7017	1.796	ng/uL	0.00
4) Pyridine-d5	3.740	84	47027	4.078	ng/u1	0.01
7) Phenol-d5	7.110	99	59584	3.964	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.275	67	41629	4.759	ng/u1	0.00
11) 2-Chlorophenol-d4	7.475	132	48554	4.287	ng/u1	0.00
15) 4-Methylphenol-d8	8.663	113	47896	3.969	ng/u1	0.00
21) Nitrobenzene-d5	9.122	128	22515	4.114	ng/u1	0.00
24) 2-Nitrophenol-d4	9.846	143	21668	3.759	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.387	165	47088	4.327	ng/u1	0.00
31) 4-Chloroaniline-d4	10.910	131	77942	4.451	ng/u1	0.00
46) Dimethylphthalate-d6	14.010	166	156575	4.579	ng/u1	0.00
49) Acenaphthylene-d8	14.298	160	162307	4.232	ng/u1	0.00
54) 4-Nitrophenol-d4	14.798	143	23763	3.111	ng/u1	0.00
60) Fluorene-d10	15.592	176	139163	4.777	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.710	200	16724	2.736	ng/u1	0.00
73) Anthracene-d10	17.451	188	214053	4.599	ng/u1	0.00
81) Pyrene-d10	19.745	212	256815	4.924	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.804	264	217748	4.451	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.340	88	7742	1.912	ng/uL	89
5) Pyridine	3.758	79	49836	4.230	ng/u1	97
6) Benzaldehyde	7.087	77	25500	4.736	ng/u1	98
8) Phenol	7.134	94	64245	4.231	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.369	93	54439	4.616	ng/u1	98
12) 2-Chlorophenol	7.510	128	52233	4.490	ng/u1	99
13) 2-Methylphenol	8.399	108	45445	3.934	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.487	45	74796	5.008	ng/u1	97
16) Acetophenone	8.781	105	81535	4.465	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.763	70	39432	4.312	ng/u1	99
18) 4-Methylphenol	8.728	108	52442	4.091	ng/u1	99
19) Hexachloroethane	9.034	117	21134	4.717	ng/u1	97
22) Nitrobenzene	9.163	77	60629	4.620	ng/u1	97
23) Isophorone	9.687	82	102725	4.150	ng/u1	99
25) 2-Nitrophenol	9.875	139	25658	4.081	ng/u1	98
26) 2,4-Dimethylphenol	9.940	107	57672	4.532	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.175	93	74462	4.702	ng/u1	100
29) 2,4-Dichlorophenol	10.410	162	48894	4.540	ng/u1	98
30) Naphthalene	10.816	128	185800	4.910	ng/u1	99
32) 4-Chloroaniline	10.934	127	80238	4.593	ng/u1	99
33) Hexachlorobutadiene	11.104	225	32538	5.300	ng/u1	98
34) Caprolactam	11.698	113	12070	2.883	ng/u1	90
35) 4-Chloro-3-methylphenol	12.057	107	51436	4.217	ng/u1	96
36) 2-Methylnaphthalene	12.428	142	121234	4.753	ng/u1	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010523\
 Data File : BN023553.D
 Acq On : 04 Jan 2023 15:19
 Operator : CG/JU
 Sample : SST00553
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SST005217

Quant Time: Jan 05 02:35:56 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN010523.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 05 02:34:34 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.645	142	123733	4.707	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.798	216	64019	4.960	ng/ul	98
40) Hexachlorocyclopentadiene	12.775	237	29449	3.793	ng/ul	96
41) 2,4,6-Trichlorophenol	13.040	196	37633	4.239	ng/ul	89
42) 2,4,5-Trichlorophenol	13.104	196	43484	4.426	ng/ul	91
43) 1,1'-Biphenyl	13.439	154	171659	4.978	ng/ul	96
44) 2-Chloronaphthalene	13.481	162	131568	4.900	ng/ul	97
45) 2-Nitroaniline	13.687	65	28271	3.633	ng/ul	94
47) Dimethylphthalate	14.057	163	158120	4.582	ng/ul	99
48) 2,6-Dinitrotoluene	14.181	165	23610	3.534	ng/ul#	89
50) Acenaphthylene	14.322	152	189841	4.502	ng/ul	99
51) 3-Nitroaniline	14.510	138	24686	3.629	ng/ul	94
52) Acenaphthene	14.669	153	142232	4.851	ng/ul	97
53) 2,4-Dinitrophenol	14.716	184	10034	2.357	ng/ul	98
55) 4-Nitrophenol	14.810	109	21533	3.812	ng/ul	96
56) Dibenzofuran	14.998	168	202235	4.954	ng/ul	96
57) 2,4-Dinitrotoluene	14.969	165	34805	3.529	ng/ul	97
58) 2,3,4,6-Tetrachlorophenol	15.228	232	36396	4.278	ng/ul	95
59) Diethylphthalate	15.422	149	157218	4.580	ng/ul	99
61) Fluorene	15.651	166	165724	5.013	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.645	204	79964	5.047	ng/ul	99
63) 4-Nitroaniline	15.669	138	16882	2.548	ng/ul	93
66) 4,6-Dinitro-2-methylph...	15.728	198	18168	3.044	ng/ul#	88
67) N-Nitrosodiphenylamine	15.857	169	134389	4.676	ng/ul	99
68) 4-Bromophenyl-phenylether	16.539	248	46830	4.674	ng/ul	99
69) Hexachlorobenzene	16.651	284	56688	4.839	ng/ul	97
70) Atrazine	16.810	200	44254	4.214	ng/ul	98
71) Pentachlorophenol	16.998	266	26052	3.368	ng/ul	94
72) Phenanthrene	17.392	178	269914	4.912	ng/ul	100
74) Anthracene	17.486	178	260329	4.748	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.404	216	64280	4.977	ng/uL	100
76) Pentachlorobenzene	14.922	250	66355	5.071	ng/uL	95
77) Carbazole	17.757	167	231618	4.588	ng/ul	98
78) Di-n-butylphthalate	18.322	149	244499	4.079	ng/ul	99
80) Fluoranthene	19.410	202	297362	4.902	ng/ul	99
82) Pyrene	19.774	202	333953	5.241	ng/ul	99
83) Butylbenzylphthalate	20.674	149	102803	3.899	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.457	252	88908	4.220	ng/ul	99
85) Benzo(a)anthracene	21.521	228	316384	4.925	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.451	149	159507	4.278	ng/ul	99
87) Chrysene	21.580	228	307674	5.075	ng/ul	98
89) Di-n-octyl phthalate	22.380	149	240600	3.980	ng/ul	100
90) Benzo(b)fluoranthene	23.221	252	290408	4.613	ng/ul	98
91) Benzo(k)fluoranthene	23.274	252	300432	5.015	ng/ul	98
93) Benzo(a)pyrene	23.857	252	251276	4.644	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.492	276	309773	4.520	ng/ul	98
95) Dibenzo(a,h)anthracene	26.509	278	265375	4.749	ng/ul	95
96) Benzo(g,h,i)perylene	27.274	276	246545	4.546	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

BNA N

BNA N

ClientSampleId :

SSTD005217


Quant Time: Jan 05 02:35:56 2023


Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN010523.M

Quant Title : SVOA CALIBRATION

QLast Update : Thu Jan 05 02:34:34 2023

Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 1.912 ng/uL
RT: 3.340 min Scan# 41
Delta R.T. 0.006 min
Lab File: BN023553.D
Acq: 04 Jan 2023 15:19




Instrument :
BNA_N
ClientSampleId :
SSTD005217


Tgt Ion: 88 Resp: 7742
Ion Ratio Lower Upper
88 100
43 33.4 24.6 36.8
58 59.6 56.2 84.4



#5
Pyridine
Concen: 4.230 ng/uL
RT: 3.758 min Scan# 114
Delta R.T. 0.012 min
Lab File: BN023553.D
Acq: 04 Jan 2023 15:19

Tgt Ion: 79 Resp: 49836
Ion Ratio Lower Upper
79 100
52 71.0 59.9 89.9
51 35.7 28.6 42.8









#6
Benzaldehyde
Concen: 4.736 ng/ul
RT: 7.087 min Scan# 679
Delta R.T. 0.006 min
Lab File: BN023553.D
Acq: 04 Jan 2023 15:19




Instrument :
BNA_N
ClientSampleId :
SSTD005217


Tgt Ion: 77 Resp: 25500
Ion Ratio Lower Upper
77 100
105 91.4 75.1 112.7
106 93.0 72.8 109.2



#8
Phenol
Concen: 4.231 ng/ul
RT: 7.134 min Scan# 688
Delta R.T. -0.006 min
Lab File: BN023553.D
Acq: 04 Jan 2023 15:19

Tgt Ion: 94 Resp: 64245
Ion Ratio Lower Upper
94 100
65 29.8 23.7 35.5
66 37.2 29.9 44.9









#10
 Bis(2-Chloroethyl)ether
 Concen: 4.616 ng/ul
 RT: 7.369 min Scan# 728
 Delta R.T. -0.000 min
 Lab File: BN023553.D
 Acq: 04 Jan 2023 15:19




Instrument :
 BNA_N
 ClientSampleId :
 SSTD005217


Tgt Ion: 93 Resp: 54439
 Ion Ratio Lower Upper
 93 100
 63 75.6 61.2 91.8
 95 30.4 26.6 40.0



#12
 2-Chlorophenol
 Concen: 4.490 ng/ul
 RT: 7.510 min Scan# 752
 Delta R.T. -0.000 min
 Lab File: BN023553.D
 Acq: 04 Jan 2023 15:19

Tgt Ion: 128 Resp: 52233
 Ion Ratio Lower Upper
 128 100
 64 47.3 37.4 56.2
 130 32.8 25.6 38.4









#13
2-Methylphenol
Concen: 3.934 ng/ul
RT: 8.399 min Scan# 903
Delta R.T. -0.000 min
Lab File: BN023553.D
Acq: 04 Jan 2023 15:19




Instrument :
BNA_N
ClientSampleId :
SSTD005217


Tgt Ion:108 Resp: 45445
Ion Ratio Lower Upper
108 100
107 88.4 72.3 108.5



#14
2,2'-oxybis(1-Chloropropane)
Concen: 5.008 ng/ul
RT: 8.487 min Scan# 918
Delta R.T. -0.000 min
Lab File: BN023553.D
Acq: 04 Jan 2023 15:19

Tgt Ion: 45 Resp: 74796
Ion Ratio Lower Upper
45 100
77 15.5 11.4 17.2
79 12.6 8.8 13.2









#16
Acetophenone
Concen: 4.465 ng/ul
RT: 8.781 min Scan# 968
Delta R.T. -0.000 min
Lab File: BN023553.D
Acq: 04 Jan 2023 15:19




Instrument :
BNA_N
ClientSampleId :
SSTD005217


Tgt Ion:105 Resp: 81535
Ion Ratio Lower Upper
105 100
77 81.1 64.2 96.2
51 27.5 22.0 33.0



#17
N-Nitroso-di-n-propylamine
Concen: 4.312 ng/ul
RT: 8.763 min Scan# 965
Delta R.T. -0.006 min
Lab File: BN023553.D
Acq: 04 Jan 2023 15:19

Tgt Ion: 70 Resp: 39432
Ion Ratio Lower Upper
70 100
42 60.1 47.8 71.6
101 10.3 7.8 11.6
130 19.2 16.3 24.5









#18
4-Methylphenol
Concen: 4.091 ng/ul
RT: 8.728 min Scan# 91
Delta R.T. -0.000 min
Lab File: BN023553.D
Acq: 04 Jan 2023 15:19




Instrument :
BNA_N
ClientSampleId :
SSTD005217


Tgt Ion:108 Resp: 52442
Ion Ratio Lower Upper
108 100
107 112.0 90.7 136.1



#19
Hexachloroethane
Concen: 4.717 ng/ul
RT: 9.034 min Scan# 1011
Delta R.T. -0.000 min
Lab File: BN023553.D
Acq: 04 Jan 2023 15:19

Tgt Ion:117 Resp: 21134
Ion Ratio Lower Upper
117 100
201 98.4 76.9 115.3
199 59.2 44.6 67.0





#22

Nitrobenzene

Concen: 4.620 ng/ul

RT: 9.163 min Scan# 1033

Delta R.T. -0.000 min

Lab File: BN023553.D

Acq: 04 Jan 2023 15:19

Instrument :

BNA_N

ClientSampleId :

SSTD005217





Tgt Ion: 77 Resp: 60629

Ion Ratio Lower Upper

77 100

123 42.2 35.6 53.4

65 13.2 11.0 16.4



#23

Isophorone

Concen: 4.150 ng/ul

RT: 9.687 min Scan# 1122

Delta R.T. -0.006 min

Lab File: BN023553.D

Acq: 04 Jan 2023 15:19




Tgt Ion: 82 Resp: 102725


Ion Ratio Lower Upper

82 100

95 7.2 5.5 8.3

138 17.7 13.8 20.8









#25
2-Nitrophenol
Concen: 4.081 ng/ul
RT: 9.875 min Scan# 1154
Delta R.T. -0.006 min
Lab File: BN023553.D
Acq: 04 Jan 2023 15:19




Instrument :
BNA_N
ClientSampleId :
SSTD005217


Tgt Ion:139 Resp: 25658
Ion Ratio Lower Upper
139 100
65 52.0 40.3 60.5
109 31.0 24.6 37.0



#26
2,4-Dimethylphenol
Concen: 4.532 ng/ul
RT: 9.940 min Scan# 1165
Delta R.T. -0.000 min
Lab File: BN023553.D
Acq: 04 Jan 2023 15:19

Tgt Ion:107 Resp: 57672
Ion Ratio Lower Upper
107 100
121 50.7 40.4 60.6
122 88.5 67.0 100.4









#27
 Bis(2-Chloroethoxy)methane
 Concen: 4.702 ng/ul
 RT: 10.175 min Scan# 1205
 Delta R.T. -0.006 min
 Lab File: BN023553.D
 Acq: 04 Jan 2023 15:19




Instrument :
 BNA_N
 ClientSampleId :
 SSTD005217


Tgt Ion: 93 Resp: 74462
 Ion Ratio Lower Upper
 93 100
 95 31.9 25.5 38.3
 123 11.2 9.4 14.2



#29
 2,4-Dichlorophenol
 Concen: 4.540 ng/ul
 RT: 10.410 min Scan# 1245
 Delta R.T. -0.006 min
 Lab File: BN023553.D
 Acq: 04 Jan 2023 15:19

Tgt Ion: 162 Resp: 48894
 Ion Ratio Lower Upper
 162 100
 164 64.9 53.8 80.6
 98 37.1 28.8 43.2









#30
Naphthalene
Concen: 4.910 ng/ul
RT: 10.816 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN023553.D
Acq: 04 Jan 2023 15:19




Instrument :
BNA_N
ClientSampleId :
SSTD005217


Tgt Ion:128 Resp: 185800
Ion Ratio Lower Upper
128 100
129 11.3 9.1 13.7
127 13.1 10.8 16.2



#32
4-Chloroaniline
Concen: 4.593 ng/ul
RT: 10.934 min Scan# 1334
Delta R.T. -0.000 min
Lab File: BN023553.D
Acq: 04 Jan 2023 15:19

Tgt Ion:127 Resp: 80238
Ion Ratio Lower Upper
127 100
129 32.4 26.4 39.6





#33

Hexachlorobutadiene

Concen: 5.300 ng/ul

RT: 11.104 min Scan# 1363

Delta R.T. -0.000 min

Lab File: BN023553.D

Acq: 04 Jan 2023 15:19

Instrument :

BNA_N

ClientSampleId :

SSTD005217





Tgt Ion:225 Resp: 32538

Ion Ratio Lower Upper

225 100

223 64.5 51.7 77.5

227 66.0 50.8 76.2



#34

Caprolactam

Concen: 2.883 ng/ul

RT: 11.698 min Scan# 1464

Delta R.T. -0.012 min

Lab File: BN023553.D

Acq: 04 Jan 2023 15:19




Tgt Ion:113 Resp: 12070


Ion Ratio Lower Upper

113 100

55 174.2 129.1 193.7

56 128.2 94.2 141.4









#35
4-Chloro-3-methylphenol
Concen: 4.217 ng/ul
RT: 12.057 min Scan# 1525
Delta R.T. -0.000 min
Lab File: BN023553.D
Acq: 04 Jan 2023 15:19




Instrument :
BNA_N
ClientSampleId :
SSTD005217


Tgt Ion:107 Resp: 51436
Ion Ratio Lower Upper
107 100
144 28.1 21.9 32.9
142 84.1 64.0 96.0



#36
2-Methylnaphthalene
Concen: 4.753 ng/ul
RT: 12.428 min Scan# 1588
Delta R.T. -0.000 min
Lab File: BN023553.D
Acq: 04 Jan 2023 15:19

Tgt Ion:142 Resp: 121234
Ion Ratio Lower Upper
142 100
141 88.8 73.6 110.4









#37
1-Methylnaphthalene
Concen: 4.707 ng/ul
RT: 12.645 min Scan# 1625
Delta R.T. -0.006 min
Lab File: BN023553.D
Acq: 04 Jan 2023 15:19




Instrument :
BNA_N
ClientSampleId :
SSTD005217


Tgt Ion:142 Resp: 123733
Ion Ratio Lower Upper
142 100
141 90.2 73.8 110.6
116 3.8 3.0 4.4



#39
1,2,4,5-Tetrachlorobenzene
Concen: 4.960 ng/ul
RT: 12.798 min Scan# 1651
Delta R.T. -0.000 min
Lab File: BN023553.D
Acq: 04 Jan 2023 15:19

Tgt Ion:216 Resp: 64019
Ion Ratio Lower Upper
216 100
214 80.2 62.4 93.6
179 20.5 16.2 24.2
108 19.7 14.6 21.8





#40

Hexachlorocyclopentadiene

Concen: 3.793 ng/ul

RT: 12.775 min Scan# 1647

Delta R.T. -0.000 min

Lab File: BN023553.D

Acq: 04 Jan 2023 15:19

Instrument :

BNA_N

ClientSampleId :

SSTD005217





Tgt Ion:237 Resp: 29449

Ion Ratio Lower Upper

237 100

235 67.6 51.0 76.4

272 12.5 9.8 14.8



#41

2,4,6-Trichlorophenol

Concen: 4.239 ng/ul

RT: 13.040 min Scan# 1692

Delta R.T. -0.000 min

Lab File: BN023553.D

Acq: 04 Jan 2023 15:19




Tgt Ion:196 Resp: 37633


Ion Ratio Lower Upper

196 100

198 100.3 71.4 107.0

200 31.4 22.5 33.7





#42

2,4,5-Trichlorophenol

Concen: 4.426 ng/ul

RT: 13.104 min Scan# 1703

Delta R.T. -0.000 min

Lab File: BN023553.D

Acq: 04 Jan 2023 15:19

Instrument :

BNA_N

ClientSampleId :

SSTD005217





Tgt Ion:196 Resp: 43484

Ion Ratio Lower Upper

196 100

198 96.0 68.1 102.1

200 29.4 24.1 36.1



#43

1,1'-Biphenyl

Concen: 4.978 ng/ul

RT: 13.439 min Scan# 1760

Delta R.T. -0.000 min

Lab File: BN023553.D

Acq: 04 Jan 2023 15:19




Tgt Ion:154 Resp: 171659


Ion Ratio Lower Upper

154 100

153 37.8 32.5 48.7

76 13.2 11.0 16.6





#44

2-Chloronaphthalene

Concen: 4.900 ng/ul

RT: 13.481 min Scan# 1767

Delta R.T. -0.000 min

Lab File: BN023553.D

Acq: 04 Jan 2023 15:19

Instrument :

BNA_N

ClientSampleId :

SSTD005217





Tgt Ion:162 Resp: 131568

Ion Ratio Lower Upper

162 100

164 33.5 26.2 39.2

127 42.0 31.8 47.8



#45

2-Nitroaniline

Concen: 3.633 ng/ul

RT: 13.687 min Scan# 1802

Delta R.T. -0.000 min

Lab File: BN023553.D

Acq: 04 Jan 2023 15:19




Tgt Ion: 65 Resp: 28271


Ion Ratio Lower Upper

65 100

92 72.6 53.5 80.3

138 109.6 84.0 126.0









#47
Dimethylphthalate
Concen: 4.582 ng/ul
RT: 14.057 min Scan# 1865
Delta R.T. -0.006 min
Lab File: BN023553.D
Acq: 04 Jan 2023 15:19




Instrument :
BNA_N
ClientSampleId :
SSTD005217


Tgt Ion:163 Resp: 158120
Ion Ratio Lower Upper
163 100
194 3.8 3.2 4.8
164 10.3 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 3.534 ng/ul
RT: 14.181 min Scan# 1886
Delta R.T. -0.006 min
Lab File: BN023553.D
Acq: 04 Jan 2023 15:19

Tgt Ion:165 Resp: 23610
Ion Ratio Lower Upper
165 100
89 70.2 49.9 74.9
121 29.6 19.1 28.7#









#50
 Acenaphthylene
 Concen: 4.502 ng/ul
 RT: 14.322 min Scan# 1910
 Delta R.T. -0.006 min
 Lab File: BN023553.D
 Acq: 04 Jan 2023 15:19

Instrument :
 BNA_N
 ClientSampleId :
 SSTD005217

Tgt Ion:152 Resp: 189841




Ion	Ratio	Lower	Upper
152	100		
151	19.1	15.8	23.8
153	12.6	10.2	15.4




#51
 3-Nitroaniline
 Concen: 3.629 ng/ul
 RT: 14.510 min Scan# 1942
 Delta R.T. -0.000 min
 Lab File: BN023553.D
 Acq: 04 Jan 2023 15:19

Tgt Ion:138 Resp: 24686

Ion	Ratio	Lower	Upper
138	100		
108	12.1	8.1	12.1
92	123.0	93.2	139.8









#52
 Acenaphthene
 Concen: 4.851 ng/ul
 RT: 14.669 min Scan# 1969
 Delta R.T. -0.000 min
 Lab File: BN023553.D
 Acq: 04 Jan 2023 15:19




Instrument :
 BNA_N
 ClientSampleId :
 SSTD005217


Tgt Ion:153	Resp:	142232
Ion Ratio	Lower	Upper
153	100	
152	47.0	38.5 57.7
154	85.5	71.5 107.3



#53
 2,4-Dinitrophenol
 Concen: 2.357 ng/ul
 RT: 14.716 min Scan# 1977
 Delta R.T. -0.000 min
 Lab File: BN023553.D
 Acq: 04 Jan 2023 15:19

Tgt Ion:184	Resp:	10034
Ion Ratio	Lower	Upper
184	100	
63	75.1	58.2 87.4
154	63.3	49.4 74.0









#55
4-Nitrophenol
Concen: 3.812 ng/ul
RT: 14.810 min Scan# 1994
Delta R.T. -0.006 min
Lab File: BN023553.D
Acq: 04 Jan 2023 15:19




Instrument :
BNA_N
ClientSampleId :
SSTD005217


Tgt Ion:109 Resp: 21533
Ion Ratio Lower Upper
109 100
139 133.8 104.8 157.2
65 139.6 107.3 160.9



#56
Dibenzofuran
Concen: 4.954 ng/ul
RT: 14.998 min Scan# 2025
Delta R.T. -0.006 min
Lab File: BN023553.D
Acq: 04 Jan 2023 15:19

Tgt Ion:168 Resp: 202235
Ion Ratio Lower Upper
168 100
139 39.1 29.6 44.4









#57
 2,4-Dinitrotoluene
 Concen: 3.529 ng/ul
 RT: 14.969 min Scan# 2020
 Delta R.T. -0.000 min
 Lab File: BN023553.D
 Acq: 04 Jan 2023 15:19




Instrument :
 BNA_N
 ClientSampleId :
 SST005217


Tgt Ion:165 Resp: 34805
 Ion Ratio Lower Upper
 165 100
 63 48.9 37.2 55.8
 182 2.8 2.2 3.2



#58
 2,3,4,6-Tetrachlorophenol
 Concen: 4.278 ng/ul
 RT: 15.228 min Scan# 2064
 Delta R.T. -0.000 min
 Lab File: BN023553.D
 Acq: 04 Jan 2023 15:19

Tgt Ion:232 Resp: 36396
 Ion Ratio Lower Upper
 232 100
 131 45.5 39.5 59.3
 130 2.5 2.2 3.4
 166 29.2 25.1 37.7





#59

Diethylphthalate

Concen: 4.580 ng/ul

RT: 15.422 min Scan# 2098

Delta R.T. -0.006 min

Lab File: BN023553.D

Acq: 04 Jan 2023 15:19

Instrument :

BNA_N

ClientSampleId :

SSTD005217





Tgt Ion:149 Resp: 157218

Ion Ratio Lower Upper

149 100

177 21.7 17.9 26.9

150 12.6 9.9 14.9



#61

Fluorene

Concen: 5.013 ng/ul

RT: 15.651 min Scan# 2136

Delta R.T. -0.000 min

Lab File: BN023553.D

Acq: 04 Jan 2023 15:19




Tgt Ion:166 Resp: 165724


Ion Ratio Lower Upper

166 100

165 98.3 77.8 116.6

167 14.2 10.4 15.6





#62

4-Chlorophenyl-phenylether

Concen: 5.047 ng/ul

RT: 15.645 min Scan# 2135

Delta R.T. -0.000 min

Lab File: BN023553.D

Acq: 04 Jan 2023 15:19

Instrument :

BNA_N

ClientSampleId :

SSTD005217





Tgt Ion:204 Resp: 79964

Ion Ratio Lower Upper

204 100

206 35.3 27.0 40.6

141 64.3 51.0 76.4



#63

4-Nitroaniline

Concen: 2.548 ng/ul

RT: 15.669 min Scan# 2139

Delta R.T. -0.006 min

Lab File: BN023553.D

Acq: 04 Jan 2023 15:19




Tgt Ion:138 Resp: 16882


Ion Ratio Lower Upper

138 100

92 58.2 41.3 61.9

108 84.4 64.2 96.2









#66
4,6-Dinitro-2-methylphenol
Concen: 3.044 ng/ul
RT: 15.728 min Scan# 2149
Delta R.T. -0.006 min
Lab File: BN023553.D
Acq: 04 Jan 2023 15:19




Instrument :
BNA_N
ClientSampleId :
SSTD005217


Tgt Ion:198 Resp: 18168
Ion Ratio Lower Upper
198 100
182 7.6 3.6 5.4#
77 32.9 21.2 31.8#



#67
N-Nitrosodiphenylamine
Concen: 4.676 ng/ul
RT: 15.857 min Scan# 2171
Delta R.T. -0.006 min
Lab File: BN023553.D
Acq: 04 Jan 2023 15:19

Tgt Ion:169 Resp: 134389
Ion Ratio Lower Upper
169 100
168 67.3 53.4 80.2
167 36.3 29.0 43.4









#68
4-Bromophenyl-phenylether
Concen: 4.674 ng/ul
RT: 16.539 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN023553.D
Acq: 04 Jan 2023 15:19




Instrument :
BNA_N
ClientSampleId :
SSTD005217


Tgt Ion:248 Resp: 46830
Ion Ratio Lower Upper
248 100
250 95.6 76.4 114.6
141 75.1 58.0 87.0



#69
Hexachlorobenzene
Concen: 4.839 ng/ul
RT: 16.651 min Scan# 2306
Delta R.T. -0.006 min
Lab File: BN023553.D
Acq: 04 Jan 2023 15:19

Tgt Ion:284 Resp: 56688
Ion Ratio Lower Upper
284 100
142 38.5 28.9 43.3
249 31.5 24.2 36.2





#70

Atrazine

Concen: 4.214 ng/ul

RT: 16.810 min Scan# 21

Delta R.T. -0.006 min

Lab File: BN023553.D

Acq: 04 Jan 2023 15:19

Instrument :

BNA_N

ClientSampleId :

SSTD005217





Tgt Ion:200 Resp: 44254

Ion Ratio Lower Upper

200 100

173 28.5 21.4 32.2

215 49.2 39.0 58.6



#71

Pentachlorophenol

Concen: 3.368 ng/ul

RT: 16.998 min Scan# 2365

Delta R.T. -0.000 min

Lab File: BN023553.D

Acq: 04 Jan 2023 15:19




Tgt Ion:266 Resp: 26052


Ion Ratio Lower Upper

266 100

264 66.8 51.4 77.0

268 60.8 54.5 81.7









#72
 Phenanthrene
 Concen: 4.912 ng/ul
 RT: 17.392 min Scan# 2432
 Delta R.T. -0.000 min
 Lab File: BN023553.D
 Acq: 04 Jan 2023 15:19

Instrument :
 BNA_N
 ClientSampleId :
 SSTD005217

Tgt Ion:178 Resp: 269914




Ion	Ratio	Lower	Upper
178	100		
179	15.2	12.3	18.5
176	18.9	15.2	22.8




#74
 Anthracene
 Concen: 4.748 ng/ul
 RT: 17.486 min Scan# 2448
 Delta R.T. -0.000 min
 Lab File: BN023553.D
 Acq: 04 Jan 2023 15:19

Tgt Ion:178 Resp: 260329

Ion	Ratio	Lower	Upper
178	100		
179	15.8	12.6	19.0
176	19.2	15.1	22.7





#75

1,2,3,4-Tetrachlorobenzene

Concen: 4.977 ng/uL

RT: 13.404 min Scan# 1754

Delta R.T. -0.000 min

Lab File: BN023553.D

Acq: 04 Jan 2023 15:19

Instrument :

BNA_N

ClientSampleId :

SSTD005217





Tgt Ion:216 Resp: 64280

Ion Ratio Lower Upper

216 100

214 76.4 61.0 91.4

218 49.7 39.8 59.8



#76

Pentachlorobenzene

Concen: 5.071 ng/uL

RT: 14.922 min Scan# 1212

Delta R.T. -0.000 min

Lab File: BN023553.D

Acq: 04 Jan 2023 15:19




Tgt Ion:250 Resp: 66355


Ion Ratio Lower Upper

250 100

248 65.0 50.3 75.5

252 66.8 48.9 73.3









#77
 Carbazole
 Concen: 4.588 ng/ul
 RT: 17.757 min Scan# 2494
 Delta R.T. -0.000 min
 Lab File: BN023553.D
 Acq: 04 Jan 2023 15:19




Instrument :
 BNA_N
 ClientSampleId :
 SSTD005217


Tgt Ion:167 Resp: 231618
 Ion Ratio Lower Upper
 167 100
 166 20.8 17.4 26.2
 139 13.0 11.0 16.6



#78
 Di-n-butylphthalate
 Concen: 4.079 ng/ul
 RT: 18.322 min Scan# 2590
 Delta R.T. -0.000 min
 Lab File: BN023553.D
 Acq: 04 Jan 2023 15:19

Tgt Ion:149 Resp: 244499
 Ion Ratio Lower Upper
 149 100
 150 9.0 7.4 11.2
 104 5.7 4.3 6.5









#80
Fluoranthene
Concen: 4.902 ng/ul
RT: 19.410 min Scan# 2775
Delta R.T. -0.000 min
Lab File: BN023553.D
Acq: 04 Jan 2023 15:19




Instrument :
BNA_N
ClientSampleId :
SSTD005217


Tgt Ion:202 Resp: 297362
Ion Ratio Lower Upper
202 100
101 11.6 9.5 14.3
100 8.7 7.3 10.9



#82
Pyrene
Concen: 5.241 ng/ul
RT: 19.774 min Scan# 2837
Delta R.T. -0.000 min
Lab File: BN023553.D
Acq: 04 Jan 2023 15:19

Tgt Ion:202 Resp: 333953
Ion Ratio Lower Upper
202 100
101 13.9 10.9 16.3
100 10.7 8.9 13.3





#83

Butylbenzylphthalate

Concen: 3.899 ng/ul

RT: 20.674 min Scan# 2990

Delta R.T. -0.000 min

Lab File: BN023553.D

Acq: 04 Jan 2023 15:19

Instrument :

BNA_N

ClientSampleId :

SSTD005217





Tgt Ion:149 Resp: 102803

Ion Ratio Lower Upper

149 100

91 72.9 57.9 86.9

206 20.1 16.8 25.2



#84

3,3'-Dichlorobenzidine

Concen: 4.220 ng/ul

RT: 21.457 min Scan# 3123

Delta R.T. -0.006 min

Lab File: BN023553.D

Acq: 04 Jan 2023 15:19




Tgt Ion:252 Resp: 88908


Ion Ratio Lower Upper

252 100

254 63.7 51.2 76.8

126 12.2 8.8 13.2









#85
Benzo(a)anthracene
Concen: 4.925 ng/ul
RT: 21.521 min Scan# 3135
Delta R.T. -0.006 min
Lab File: BN023553.D
Acq: 04 Jan 2023 15:19




Instrument :
BNA_N
ClientSampleId :
SSTD005217


Tgt Ion:228 Resp: 316384
Ion Ratio Lower Upper
228 100
229 19.8 15.0 22.4
226 27.2 22.1 33.1



#86
Bis(2-ethylhexyl)phthalate
Concen: 4.278 ng/ul
RT: 21.451 min Scan# 3122
Delta R.T. -0.000 min
Lab File: BN023553.D
Acq: 04 Jan 2023 15:19

Tgt Ion:149 Resp: 159507
Ion Ratio Lower Upper
149 100
167 26.6 22.0 33.0
279 4.4 3.4 5.2





#87

Chrysene

Concen: 5.075 ng/ul

RT: 21.580 min Scan# 3144

Delta R.T. -0.000 min

Lab File: BN023553.D

Acq: 04 Jan 2023 15:19

Instrument :

BNA_N

ClientSampleId :

SSTD005217





Tgt Ion:228 Resp: 307674

Ion Ratio Lower Upper

228 100

226 29.8 24.6 37.0

229 19.4 16.8 25.2



#89

Di-n-octyl phthalate

Concen: 3.980 ng/ul




RT: 22.380 min Scan# 3280


Delta R.T. -0.000 min

Lab File: BN023553.D

Acq: 04 Jan 2023 15:19

Tgt Ion:149 Resp: 240600





#90

Benzo(b)fluoranthene

Concen: 4.613 ng/ul

RT: 23.221 min Scan# 3423

Delta R.T. -0.006 min

Lab File: BN023553.D

Acq: 04 Jan 2023 15:19

Instrument :

BNA_N

ClientSampleId :

SSTD005217





Tgt Ion:252 Resp: 290408

Ion Ratio Lower Upper

252 100

253 22.1 17.2 25.8

125 12.2 8.6 13.0



#91

Benzo(k)fluoranthene

Concen: 5.015 ng/ul

RT: 23.274 min Scan# 3432

Delta R.T. -0.000 min

Lab File: BN023553.D

Acq: 04 Jan 2023 15:19




Tgt Ion:252 Resp: 300432


Ion Ratio Lower Upper

252 100

253 21.3 17.8 26.8

125 11.0 8.6 12.8





#93

Benzo(a)pyrene

Concen: 4.644 ng/ul

RT: 23.857 min Scan# 3

Delta R.T. -0.000 min

Lab File: BN023553.D

Acq: 04 Jan 2023 15:19

Instrument :

BNA_N

ClientSampleId :

SSTD005217





Tgt Ion:252 Resp: 251276

Ion Ratio Lower Upper

252 100

253 22.1 16.6 25.0

125 12.5 9.8 14.6



#94

Indeno(1,2,3-cd)pyrene

Concen: 4.520 ng/ul

RT: 26.492 min Scan# 3979

Delta R.T. -0.012 min

Lab File: BN023553.D

Acq: 04 Jan 2023 15:19




Tgt Ion:276 Resp: 309773


Ion Ratio Lower Upper

276 100

138 25.2 20.6 30.8

277 22.2 18.9 28.3





#95

Dibenzo(a,h)anthracene

Concen: 4.749 ng/ul

RT: 26.509 min Scan# 3982

Delta R.T. -0.012 min

Lab File: BN023553.D

Acq: 04 Jan 2023 15:19

Instrument :

BNA_N

ClientSampleId :

SSTD005217





Tgt Ion:278 Resp: 265375

Ion Ratio Lower Upper

278 100

139 15.3 14.1 21.1

279 20.5 18.0 27.0



#96

Benzo(g,h,i)perylene

Concen: 4.546 ng/ul

RT: 27.274 min Scan# 4112

Delta R.T. -0.012 min

Lab File: BN023553.D

Acq: 04 Jan 2023 15:19

Tgt Ion:276 Resp: 246545

Ion Ratio Lower Upper

276 100

138 22.3 17.0 25.6

277 21.6 17.6 26.4

