

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010523\
 Data File : BN023555.D
 Acq On : 04 Jan 2023 16:34
 Operator : CG/JU
 Sample : SST02055
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SST020219

Quant Time: Jan 05 02:36:56 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN010523.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 05 02:34:34 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.952	152	176770	20.000	ng/ul	0.00
20) Naphthalene-d8	10.769	136	782493	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.604	164	505061	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.351	188	1117757	20.000	ng/ul	0.00
79) Chrysene-d12	21.545	240	1082969	20.000	ng/ul	0.00
88) Perylene-d12	23.968	264	1123116	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.299	96	34047	7.969	ng/uL	0.00
4) Pyridine-d5	3.728	84	238141	18.882	ng/ul	0.00
7) Phenol-d5	7.110	99	303482	18.459	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.275	67	192696	20.141	ng/ul	0.00
11) 2-Chlorophenol-d4	7.475	132	241317	19.479	ng/ul	0.00
15) 4-Methylphenol-d8	8.669	113	242428	18.366	ng/ul	0.00
21) Nitrobenzene-d5	9.122	128	117891	19.500	ng/ul	0.00
24) 2-Nitrophenol-d4	9.846	143	124765	19.594	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.387	165	241763	20.112	ng/ul	0.00
31) 4-Chloroaniline-d4	10.910	131	377461	19.515	ng/ul	0.00
46) Dimethylphthalate-d6	14.016	166	757391	20.011	ng/ul	0.00
49) Acenaphthylene-d8	14.298	160	846031	19.932	ng/ul	0.00
54) 4-Nitrophenol-d4	14.804	143	149048	17.632	ng/ul	0.00
60) Fluorene-d10	15.598	176	646278	20.045	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.716	200	126098	19.164	ng/ul	0.00
73) Anthracene-d10	17.451	188	1026706	20.496	ng/ul	0.00
81) Pyrene-d10	19.745	212	1222528	20.752	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.810	264	1107610	20.120	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.334	88	35260	7.963	ng/uL	100
5) Pyridine	3.746	79	252475	19.591	ng/ul	100
6) Benzaldehyde	7.081	77	121013	20.549	ng/ul	100
8) Phenol	7.140	94	309424	18.632	ng/ul	100
10) Bis(2-Chloroethyl)ether	7.369	93	251438	19.494	ng/ul	100
12) 2-Chlorophenol	7.510	128	248166	19.504	ng/ul	100
13) 2-Methylphenol	8.399	108	234310	18.544	ng/ul	100
14) 2,2'-oxybis(1-Chloropr...	8.487	45	337104	20.635	ng/ul	100
16) Acetophenone	8.781	105	392874	19.671	ng/ul	100
17) N-Nitroso-di-n-propyla...	8.769	70	197818	19.780	ng/ul	100
18) 4-Methylphenol	8.728	108	259542	18.510	ng/ul	100
19) Hexachloroethane	9.034	117	98949	20.192	ng/ul	100
22) Nitrobenzene	9.163	77	297990	20.556	ng/ul	100
23) Isophorone	9.693	82	541882	19.817	ng/ul	100
25) 2-Nitrophenol	9.881	139	142357	20.498	ng/ul	100
26) 2,4-Dimethylphenol	9.940	107	285860	20.334	ng/ul	100
27) Bis(2-Chloroethoxy)met...	10.181	93	348242	19.909	ng/ul	100
29) 2,4-Dichlorophenol	10.416	162	239834	20.162	ng/ul	100
30) Naphthalene	10.816	128	837444	20.033	ng/ul	100
32) 4-Chloroaniline	10.934	127	380369	19.711	ng/ul	100
33) Hexachlorobutadiene	11.104	225	146927	21.666	ng/ul	100
34) Caprolactam	11.710	113	42255	9.136	ng/ul	100
35) 4-Chloro-3-methylphenol	12.057	107	266066	19.748	ng/ul	100
36) 2-Methylnaphthalene	12.428	142	561977	19.946	ng/ul	100


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010523\
 Data File : BN023555.D
 Acq On : 04 Jan 2023 16:34
 Operator : CG/JU
 Sample : SSTD02055
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020219

Quant Time: Jan 05 02:36:56 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN010523.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 05 02:34:34 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.651	142	577360	19.885	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.798	216	294786	20.636	ng/ul	100
40) Hexachlorocyclopentadiene	12.775	237	165988	19.317	ng/ul	100
41) 2,4,6-Trichlorophenol	13.040	196	199508	20.304	ng/ul	100
42) 2,4,5-Trichlorophenol	13.104	196	215080	19.777	ng/ul	100
43) 1,1'-Biphenyl	13.440	154	765068	20.047	ng/ul	100
44) 2-Chloronaphthalene	13.481	162	596509	20.071	ng/ul	100
45) 2-Nitroaniline	13.687	65	169802	19.716	ng/ul	100
47) Dimethylphthalate	14.063	163	760301	19.907	ng/ul	100
48) 2,6-Dinitrotoluene	14.187	165	143810	19.449	ng/ul	100
50) Acenaphthylene	14.328	152	940481	20.149	ng/ul	100
51) 3-Nitroaniline	14.510	138	134000	17.798	ng/ul	100
52) Acenaphthene	14.669	153	650069	20.033	ng/ul	100
53) 2,4-Dinitrophenol	14.716	184	80801	17.149	ng/ul	100
55) 4-Nitrophenol	14.816	109	122087	19.527	ng/ul	100
56) Dibenzofuran	15.004	168	910143	20.145	ng/ul	100
57) 2,4-Dinitrotoluene	14.969	165	218056	19.978	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.228	232	186569	19.815	ng/ul	100
59) Diethylphthalate	15.428	149	764049	20.110	ng/ul	100
61) Fluorene	15.651	166	749502	20.486	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.645	204	361739	20.629	ng/ul	100
63) 4-Nitroaniline	15.675	138	128226	17.488	ng/ul	100
66) 4,6-Dinitro-2-methylph...	15.734	198	127293	19.816	ng/ul	100
67) N-Nitrosodiphenylamine	15.863	169	626074	20.239	ng/ul	100
68) 4-Bromophenyl-phenylether	16.539	248	231496	21.464	ng/ul	100
69) Hexachlorobenzene	16.657	284	263064	20.862	ng/ul	100
70) Atrazine	16.816	200	231016	20.436	ng/ul	100
71) Pentachlorophenol	16.998	266	157161	18.875	ng/ul	100
72) Phenanthrene	17.392	178	1218939	20.609	ng/ul	100
74) Anthracene	17.486	178	1202680	20.379	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.404	216	295865	21.285	ng/uL	100
76) Pentachlorobenzene	14.922	250	297716	21.136	ng/uL	100
77) Carbazole	17.757	167	1116238	20.541	ng/ul	100
78) Di-n-butylphthalate	18.322	149	1328821	20.594	ng/ul	100
80) Fluoranthene	19.410	202	1426177	20.815	ng/ul	100
82) Pyrene	19.775	202	1515354	21.054	ng/ul	100
83) Butylbenzylphthalate	20.674	149	617683	20.739	ng/ul	100
84) 3,3'-Dichlorobenzidine	21.463	252	473720	19.905	ng/ul	100
85) Benzo(a)anthracene	21.527	228	1434724	19.774	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.451	149	903333	21.450	ng/ul	100
87) Chrysene	21.580	228	1352566	19.753	ng/ul	100
89) Di-n-octyl phthalate	22.380	149	1504752	22.124	ng/ul	100
90) Benzo(b)fluoranthene	23.227	252	1410218	19.908	ng/ul	100
91) Benzo(k)fluoranthene	23.274	252	1387657	20.584	ng/ul	100
93) Benzo(a)pyrene	23.857	252	1234007	20.270	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.504	276	1549413	20.090	ng/ul	100
95) Dibenzo(a,h)anthracene	26.521	278	1264468	20.108	ng/ul	100
96) Benzo(g,h,i)perylene	27.286	276	1204366	19.737	ng/ul	100





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 7.963 ng/uL
RT: 3.334 min Scan# 42
Delta R.T. 0.000 min
Lab File: BN023555.D
Acq: 04 Jan 2023 16:34




Instrument :
BNA_N
ClientSampleId :
SSTD020219


Tgt Ion: 88 Resp: 35260
Ion Ratio Lower Upper
88 100
43 30.7 24.6 36.8
58 70.3 56.2 84.4



#5
Pyridine
Concen: 19.591 ng/uL
RT: 3.746 min Scan# 112
Delta R.T. 0.000 min
Lab File: BN023555.D
Acq: 04 Jan 2023 16:34

Tgt Ion: 79 Resp: 252475
Ion Ratio Lower Upper
79 100
52 74.9 59.9 89.9
51 35.7 28.6 42.8









#6
Benzaldehyde
Concen: 20.549 ng/ul
RT: 7.081 min Scan# 679
Delta R.T. 0.000 min
Lab File: BN023555.D
Acq: 04 Jan 2023 16:34




Instrument :
BNA_N
ClientSampleId :
SSTD020219


Tgt Ion: 77 Resp: 121013
Ion Ratio Lower Upper
77 100
105 93.9 75.1 112.7
106 91.0 72.8 109.2



#8
Phenol
Concen: 18.632 ng/ul
RT: 7.140 min Scan# 689
Delta R.T. 0.000 min
Lab File: BN023555.D
Acq: 04 Jan 2023 16:34

Tgt Ion: 94 Resp: 309424
Ion Ratio Lower Upper
94 100
65 29.6 23.7 35.5
66 37.4 29.9 44.9









#10
Bis(2-Chloroethyl)ether
Concen: 19.494 ng/ul
RT: 7.369 min Scan# 728
Delta R.T. 0.000 min
Lab File: BN023555.D
Acq: 04 Jan 2023 16:34




Instrument :
BNA_N
ClientSampleId :
SSTD020219


Tgt Ion: 93 Resp: 251438
Ion Ratio Lower Upper
93 100
63 76.5 61.2 91.8
95 33.3 26.6 40.0



#12
2-Chlorophenol
Concen: 19.504 ng/ul
RT: 7.510 min Scan# 752
Delta R.T. 0.000 min
Lab File: BN023555.D
Acq: 04 Jan 2023 16:34

Tgt Ion: 128 Resp: 248166
Ion Ratio Lower Upper
128 100
64 46.8 37.4 56.2
130 32.0 25.6 38.4









#13
2-Methylphenol
Concen: 18.544 ng/ul
RT: 8.399 min Scan# 903
Delta R.T. 0.000 min
Lab File: BN023555.D
Acq: 04 Jan 2023 16:34




Instrument :
BNA_N
ClientSampleId :
SSTD020219


Tgt Ion:108 Resp: 234310
Ion Ratio Lower Upper
108 100
107 90.4 72.3 108.5



#14
2,2'-oxybis(1-Chloropropane)
Concen: 20.635 ng/ul
RT: 8.487 min Scan# 918
Delta R.T. 0.000 min
Lab File: BN023555.D
Acq: 04 Jan 2023 16:34

Tgt Ion: 45 Resp: 337104
Ion Ratio Lower Upper
45 100
77 14.3 11.4 17.2
79 11.0 8.8 13.2









#16
Acetophenone
Concen: 19.671 ng/ul
RT: 8.781 min Scan# 968
Delta R.T. 0.000 min
Lab File: BN023555.D
Acq: 04 Jan 2023 16:34




Instrument :
BNA_N
ClientSampleId :
SSTD020219


Tgt Ion:105 Resp: 392874
Ion Ratio Lower Upper
105 100
77 80.2 64.2 96.2
51 27.5 22.0 33.0



#17
N-Nitroso-di-n-propylamine
Concen: 19.780 ng/ul
RT: 8.769 min Scan# 966
Delta R.T. 0.000 min
Lab File: BN023555.D
Acq: 04 Jan 2023 16:34

Tgt Ion: 70 Resp: 197818
Ion Ratio Lower Upper
70 100
42 59.7 47.8 71.6
101 9.7 7.8 11.6
130 20.4 16.3 24.5









#18
4-Methylphenol
Concen: 18.510 ng/ul
RT: 8.728 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN023555.D
Acq: 04 Jan 2023 16:34




Instrument :
BNA_N
ClientSampleId :
SSTD020219


Tgt Ion:108 Resp: 259542
Ion Ratio Lower Upper
108 100
107 113.4 90.7 136.1



#19
Hexachloroethane
Concen: 20.192 ng/ul
RT: 9.034 min Scan# 1011
Delta R.T. 0.000 min
Lab File: BN023555.D
Acq: 04 Jan 2023 16:34

Tgt Ion:117 Resp: 98949
Ion Ratio Lower Upper
117 100
201 96.1 76.9 115.3
199 55.8 44.6 67.0





#22

Nitrobenzene

Concen: 20.556 ng/ul

RT: 9.163 min Scan# 1033

Delta R.T. 0.000 min

Lab File: BN023555.D

Acq: 04 Jan 2023 16:34

Instrument :





BNA_N

ClientSampleId :

SSTD020219

Tgt Ion: 77 Resp: 297990

Ion	Ratio	Lower	Upper
77	100		
123	44.5	35.6	53.4
65	13.7	11.0	16.4



#23

Isophorone

Concen: 19.817 ng/ul

RT: 9.693 min Scan# 1123




Delta R.T. 0.000 min


Lab File: BN023555.D

Acq: 04 Jan 2023 16:34

Tgt Ion: 82 Resp: 541882

Ion	Ratio	Lower	Upper
82	100		
95	6.9	5.5	8.3
138	17.3	13.8	20.8





#25

2-Nitrophenol

Concen: 20.498 ng/ul

RT: 9.881 min Scan# 1155

Delta R.T. 0.000 min

Lab File: BN023555.D

Acq: 04 Jan 2023 16:34

Instrument :

BNA_N

ClientSampleId :

SSTD020219





Tgt Ion:139 Resp: 142357

Ion Ratio Lower Upper

139 100

65 50.4 40.3 60.5

109 30.8 24.6 37.0



#26

2,4-Dimethylphenol

Concen: 20.334 ng/ul

RT: 9.940 min Scan# 1165

Delta R.T. 0.000 min

Lab File: BN023555.D

Acq: 04 Jan 2023 16:34




Tgt Ion:107 Resp: 285860


Ion Ratio Lower Upper

107 100

121 50.5 40.4 60.6

122 83.7 67.0 100.4









#27
 Bis(2-Chloroethoxy)methane
 Concen: 19.909 ng/ul
 RT: 10.181 min Scan# 1206
 Delta R.T. 0.000 min
 Lab File: BN023555.D
 Acq: 04 Jan 2023 16:34




Instrument :
 BNA_N
 ClientSampleId :
 SST020219


Tgt Ion: 93 Resp: 348242
 Ion Ratio Lower Upper
 93 100
 95 31.9 25.5 38.3
 123 11.8 9.4 14.2



#29
 2,4-Dichlorophenol
 Concen: 20.162 ng/ul
 RT: 10.416 min Scan# 1246
 Delta R.T. 0.000 min
 Lab File: BN023555.D
 Acq: 04 Jan 2023 16:34

Tgt Ion: 162 Resp: 239834
 Ion Ratio Lower Upper
 162 100
 164 67.2 53.8 80.6
 98 36.0 28.8 43.2









#30
Naphthalene
Concen: 20.033 ng/ul
RT: 10.816 min Scan# 1314
Delta R.T. 0.000 min
Lab File: BN023555.D
Acq: 04 Jan 2023 16:34




Instrument :
BNA_N
ClientSampleId :
SSTD020219


Tgt Ion:128 Resp: 837444
Ion Ratio Lower Upper
128 100
129 11.4 9.1 13.7
127 13.5 10.8 16.2



#32
4-Chloroaniline
Concen: 19.711 ng/ul
RT: 10.934 min Scan# 1334
Delta R.T. 0.000 min
Lab File: BN023555.D
Acq: 04 Jan 2023 16:34

Tgt Ion:127 Resp: 380369
Ion Ratio Lower Upper
127 100
129 33.0 26.4 39.6









#33
Hexachlorobutadiene
Concen: 21.666 ng/ul
RT: 11.104 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN023555.D
Acq: 04 Jan 2023 16:34




Instrument :
BNA_N
ClientSampleId :
SSTD020219


Tgt Ion:225 Resp: 146927
Ion Ratio Lower Upper
225 100
223 64.6 51.7 77.5
227 63.5 50.8 76.2



#34
Caprolactam
Concen: 9.136 ng/ul
RT: 11.710 min Scan# 1466
Delta R.T. 0.000 min
Lab File: BN023555.D
Acq: 04 Jan 2023 16:34

Tgt Ion:113 Resp: 42255
Ion Ratio Lower Upper
113 100
55 161.4 129.1 193.7
56 117.8 94.2 141.4









#35
 4-Chloro-3-methylphenol
 Concen: 19.748 ng/ul
 RT: 12.057 min Scan# 1525
 Delta R.T. 0.000 min
 Lab File: BN023555.D
 Acq: 04 Jan 2023 16:34




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020219


Tgt Ion:107 Resp: 266066
 Ion Ratio Lower Upper
 107 100
 144 27.4 21.9 32.9
 142 80.0 64.0 96.0



#36
 2-Methylnaphthalene
 Concen: 19.946 ng/ul
 RT: 12.428 min Scan# 1588
 Delta R.T. 0.000 min
 Lab File: BN023555.D
 Acq: 04 Jan 2023 16:34

Tgt Ion:142 Resp: 561977
 Ion Ratio Lower Upper
 142 100
 141 92.0 73.6 110.4





#37

1-Methylnaphthalene

Concen: 19.885 ng/ul

RT: 12.651 min Scan# 1626

Delta R.T. 0.000 min

Lab File: BN023555.D

Acq: 04 Jan 2023 16:34

Instrument :

BNA_N

ClientSampleId :

SSTD020219





Tgt Ion:142 Resp: 577360

Ion Ratio Lower Upper

142 100

141 92.2 73.8 110.6

116 3.7 3.0 4.4



#39

1,2,4,5-Tetrachlorobenzene

Concen: 20.636 ng/ul

RT: 12.798 min Scan# 1651

Delta R.T. 0.000 min

Lab File: BN023555.D

Acq: 04 Jan 2023 16:34

Tgt Ion:216 Resp: 294786




Ion Ratio Lower Upper


216 100

214 78.0 62.4 93.6

179 20.2 16.2 24.2

108 18.2 14.6 21.8









#40
Hexachlorocyclopentadiene
Concen: 19.317 ng/ul
RT: 12.775 min Scan# 1647
Delta R.T. 0.000 min
Lab File: BN023555.D
Acq: 04 Jan 2023 16:34




Instrument :
BNA_N
ClientSampleId :
SSTD020219


Tgt Ion:237 Resp: 165988
Ion Ratio Lower Upper
237 100
235 63.7 51.0 76.4
272 12.3 9.8 14.8



#41
2,4,6-Trichlorophenol
Concen: 20.304 ng/ul
RT: 13.040 min Scan# 1692
Delta R.T. 0.000 min
Lab File: BN023555.D
Acq: 04 Jan 2023 16:34

Tgt Ion:196 Resp: 199508
Ion Ratio Lower Upper
196 100
198 89.2 71.4 107.0
200 28.1 22.5 33.7





#42

2,4,5-Trichlorophenol

Concen: 19.777 ng/ul

RT: 13.104 min Scan# 1703

Delta R.T. 0.000 min

Lab File: BN023555.D

Acq: 04 Jan 2023 16:34

Instrument :

BNA_N

ClientSampleId :

SSTD020219





Tgt Ion:196 Resp: 215080

Ion Ratio Lower Upper

196 100

198 85.1 68.1 102.1

200 30.1 24.1 36.1



#43

1,1'-Biphenyl

Concen: 20.047 ng/ul

RT: 13.440 min Scan# 1760

Delta R.T. 0.000 min

Lab File: BN023555.D

Acq: 04 Jan 2023 16:34




Tgt Ion:154 Resp: 765068


Ion Ratio Lower Upper

154 100

153 40.6 32.5 48.7

76 13.8 11.0 16.6





#44

2-Chloronaphthalene

Concen: 20.071 ng/ul

RT: 13.481 min Scan# 11

Delta R.T. 0.000 min

Lab File: BN023555.D

Acq: 04 Jan 2023 16:34

Instrument :

BNA_N

ClientSampleId :

SSTD020219





Tgt Ion:162 Resp: 596509

Ion Ratio Lower Upper

162 100

164 32.7 26.2 39.2

127 39.8 31.8 47.8



#45

2-Nitroaniline

Concen: 19.716 ng/ul

RT: 13.687 min Scan# 1802

Delta R.T. 0.000 min

Lab File: BN023555.D

Acq: 04 Jan 2023 16:34




Tgt Ion: 65 Resp: 169802


Ion Ratio Lower Upper

65 100

92 66.9 53.5 80.3

138 105.0 84.0 126.0





#47

Dimethylphthalate

Concen: 19.907 ng/ul

RT: 14.063 min Scan# 1866

Delta R.T. 0.000 min

Lab File: BN023555.D

Acq: 04 Jan 2023 16:34

Instrument :

BNA_N

ClientSampleId :

SSTD020219





Tgt Ion:163 Resp: 760301

Ion Ratio Lower Upper

163 100

194 4.0 3.2 4.8

164 9.9 7.9 11.9



#48

2,6-Dinitrotoluene

Concen: 19.449 ng/ul

RT: 14.187 min Scan# 1887

Delta R.T. 0.000 min

Lab File: BN023555.D

Acq: 04 Jan 2023 16:34




Tgt Ion:165 Resp: 143810


Ion Ratio Lower Upper

165 100

89 62.4 49.9 74.9

121 23.9 19.1 28.7









#50
Acenaphthylene
Concen: 20.149 ng/ul
RT: 14.328 min Scan# 1911
Delta R.T. 0.000 min
Lab File: BN023555.D
Acq: 04 Jan 2023 16:34




Instrument :
BNA_N
ClientSampleId :
SSTD020219


Tgt Ion:152 Resp: 940481
Ion Ratio Lower Upper
152 100
151 19.8 15.8 23.8
153 12.8 10.2 15.4



#51
3-Nitroaniline
Concen: 17.798 ng/ul
RT: 14.510 min Scan# 1942
Delta R.T. 0.000 min
Lab File: BN023555.D
Acq: 04 Jan 2023 16:34

Tgt Ion:138 Resp: 134000
Ion Ratio Lower Upper
138 100
108 10.1 8.1 12.1
92 116.5 93.2 139.8









#52
Acenaphthene
Concen: 20.033 ng/ul
RT: 14.669 min Scan# 1969
Delta R.T. 0.000 min
Lab File: BN023555.D
Acq: 04 Jan 2023 16:34




Instrument :
BNA_N
ClientSampleId :
SSTD020219


Tgt Ion:153 Resp: 650069
Ion Ratio Lower Upper
153 100
152 48.1 38.5 57.7
154 89.4 71.5 107.3



#53
2,4-Dinitrophenol
Concen: 17.149 ng/ul
RT: 14.716 min Scan# 1977
Delta R.T. 0.000 min
Lab File: BN023555.D
Acq: 04 Jan 2023 16:34

Tgt Ion:184 Resp: 80801
Ion Ratio Lower Upper
184 100
63 72.8 58.2 87.4
154 61.7 49.4 74.0









#55
4-Nitrophenol
Concen: 19.527 ng/ul
RT: 14.816 min Scan# 1994
Delta R.T. 0.000 min
Lab File: BN023555.D
Acq: 04 Jan 2023 16:34




Instrument :
BNA_N
ClientSampleId :
SSTD020219


Tgt Ion:109 Resp: 122087
Ion Ratio Lower Upper
109 100
139 131.0 104.8 157.2
65 134.1 107.3 160.9



#56
Dibenzofuran
Concen: 20.145 ng/ul
RT: 15.004 min Scan# 2026
Delta R.T. 0.000 min
Lab File: BN023555.D
Acq: 04 Jan 2023 16:34

Tgt Ion:168 Resp: 910143
Ion Ratio Lower Upper
168 100
139 37.0 29.6 44.4









#57
2,4-Dinitrotoluene
Concen: 19.978 ng/ul
RT: 14.969 min Scan# 2020
Delta R.T. 0.000 min
Lab File: BN023555.D
Acq: 04 Jan 2023 16:34




Instrument :
BNA_N
ClientSampleId :
SSTD020219


Tgt Ion:165 Resp: 218056
Ion Ratio Lower Upper
165 100
63 46.5 37.2 55.8
182 2.7 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 19.815 ng/ul
RT: 15.228 min Scan# 2064
Delta R.T. 0.000 min
Lab File: BN023555.D
Acq: 04 Jan 2023 16:34

Tgt Ion:232 Resp: 186569
Ion Ratio Lower Upper
232 100
131 49.4 39.5 59.3
130 2.8 2.2 3.4
166 31.4 25.1 37.7





#59

Diethylphthalate

Concen: 20.110 ng/ul

RT: 15.428 min Scan# 2098

Delta R.T. 0.000 min

Lab File: BN023555.D

Acq: 04 Jan 2023 16:34

Instrument :

BNA_N

ClientSampleId :

SSTD020219





Tgt Ion:149 Resp: 764049

Ion Ratio Lower Upper

149 100

177 22.4 17.9 26.9

150 12.4 9.9 14.9



#61

Fluorene

Concen: 20.486 ng/ul

RT: 15.651 min Scan# 2136

Delta R.T. 0.000 min

Lab File: BN023555.D

Acq: 04 Jan 2023 16:34




Tgt Ion:166 Resp: 749502


Ion Ratio Lower Upper

166 100

165 97.2 77.8 116.6

167 13.0 10.4 15.6





#62

4-Chlorophenyl-phenylether

Concen: 20.629 ng/ul

RT: 15.645 min Scan# 2135

Delta R.T. 0.000 min

Lab File: BN023555.D

Acq: 04 Jan 2023 16:34

Instrument :

BNA_N

ClientSampleId :

SSTD020219





Tgt Ion:204 Resp: 361739

Ion Ratio Lower Upper

204 100

206 33.8 27.0 40.6

141 63.7 51.0 76.4



#63

4-Nitroaniline

Concen: 17.488 ng/ul

RT: 15.675 min Scan# 2140

Delta R.T. 0.000 min

Lab File: BN023555.D

Acq: 04 Jan 2023 16:34




Tgt Ion:138 Resp: 128226


Ion Ratio Lower Upper

138 100

92 51.6 41.3 61.9

108 80.2 64.2 96.2









#66
 4,6-Dinitro-2-methylphenol
 Concen: 19.816 ng/ul
 RT: 15.734 min Scan# 2150
 Delta R.T. 0.000 min
 Lab File: BN023555.D
 Acq: 04 Jan 2023 16:34




Instrument :
 BNA_N
 ClientSampleId :
 SST020219


Tgt Ion:198 Resp: 127293
 Ion Ratio Lower Upper
 198 100
 182 4.5 3.6 5.4
 77 26.5 21.2 31.8



#67
 N-Nitrosodiphenylamine
 Concen: 20.239 ng/ul
 RT: 15.863 min Scan# 2172
 Delta R.T. 0.000 min
 Lab File: BN023555.D
 Acq: 04 Jan 2023 16:34

Tgt Ion:169 Resp: 626074
 Ion Ratio Lower Upper
 169 100
 168 66.8 53.4 80.2
 167 36.2 29.0 43.4









#68
4-Bromophenyl-phenylether
Concen: 21.464 ng/ul
RT: 16.539 min Scan# 21496
Delta R.T. 0.000 min
Lab File: BN023555.D
Acq: 04 Jan 2023 16:34




Instrument :
BNA_N
ClientSampleId :
SSTD020219


Tgt Ion:248 Resp: 231496
Ion Ratio Lower Upper
248 100
250 95.5 76.4 114.6
141 72.5 58.0 87.0



#69
Hexachlorobenzene
Concen: 20.862 ng/ul
RT: 16.657 min Scan# 2307
Delta R.T. 0.000 min
Lab File: BN023555.D
Acq: 04 Jan 2023 16:34

Tgt Ion:284 Resp: 263064
Ion Ratio Lower Upper
284 100
142 36.1 28.9 43.3
249 30.2 24.2 36.2





#70

Atrazine

Concen: 20.436 ng/ul

RT: 16.816 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN023555.D

Acq: 04 Jan 2023 16:34

Instrument :

BNA_N

ClientSampleId :

SSTD020219





Tgt Ion:200 Resp: 231016

Ion Ratio Lower Upper

200 100

173 26.8 21.4 32.2

215 48.8 39.0 58.6



#71

Pentachlorophenol

Concen: 18.875 ng/ul

RT: 16.998 min Scan# 2365

Delta R.T. 0.000 min

Lab File: BN023555.D

Acq: 04 Jan 2023 16:34




Tgt Ion:266 Resp: 157161


Ion Ratio Lower Upper

266 100

264 64.2 51.4 77.0

268 68.1 54.5 81.7





#72

Phenanthrene

Concen: 20.609 ng/uI

RT: 17.392 min Scan# 2432

Delta R.T. 0.000 min

Lab File: BN023555.D

Acq: 04 Jan 2023 16:34

Instrument :

BNA_N

ClientSampleId :

SSTD020219





Tgt Ion:178 Resp: 1218939

Ion Ratio Lower Upper

178 100

179 15.4 12.3 18.5

176 19.0 15.2 22.8



#74

Anthracene

Concen: 20.379 ng/uI

RT: 17.486 min Scan# 2448

Delta R.T. 0.000 min

Lab File: BN023555.D

Acq: 04 Jan 2023 16:34




Tgt Ion:178 Resp: 1202680


Ion Ratio Lower Upper

178 100

179 15.8 12.6 19.0

176 18.9 15.1 22.7





#75

1,2,3,4-Tetrachlorobenzene

Concen: 21.285 ng/uL

RT: 13.404 min Scan# 11

Delta R.T. 0.000 min

Lab File: BN023555.D

Acq: 04 Jan 2023 16:34

Instrument :

BNA_N

ClientSampleId :

SSTD020219





Tgt Ion:216 Resp: 295865

Ion Ratio Lower Upper

216 100

214 76.2 61.0 91.4

218 49.8 39.8 59.8



#76

Pentachlorobenzene

Concen: 21.136 ng/uL

RT: 14.922 min Scan# 2012

Delta R.T. 0.000 min

Lab File: BN023555.D

Acq: 04 Jan 2023 16:34




Tgt Ion:250 Resp: 297716


Ion Ratio Lower Upper

250 100

248 62.9 50.3 75.5

252 61.1 48.9 73.3





#77

Carbazole

Concen: 20.541 ng/ul

RT: 17.757 min Scan# 2494

Delta R.T. 0.000 min

Lab File: BN023555.D

Acq: 04 Jan 2023 16:34

Instrument :

BNA_N

ClientSampleId :

SSTD020219





Tgt Ion:167 Resp: 1116238

Ion Ratio Lower Upper

167 100

166 21.8 17.4 26.2

139 13.8 11.0 16.6



#78

Di-n-butylphthalate

Concen: 20.594 ng/ul

RT: 18.322 min Scan# 2590

Delta R.T. 0.000 min

Lab File: BN023555.D

Acq: 04 Jan 2023 16:34




Tgt Ion:149 Resp: 1328821


Ion Ratio Lower Upper

149 100

150 9.3 7.4 11.2

104 5.4 4.3 6.5





#80

Fluoranthene

Concen: 20.815 ng/ul

RT: 19.410 min Scan# 2775

Delta R.T. 0.000 min

Lab File: BN023555.D

Acq: 04 Jan 2023 16:34

Instrument :

BNA_N

ClientSampleId :

SSTD020219





Tgt Ion:202 Resp: 1426177

Ion Ratio Lower Upper

202 100

101 11.9 9.5 14.3

100 9.1 7.3 10.9



#82

Pyrene

Concen: 21.054 ng/ul

RT: 19.775 min Scan# 2837

Delta R.T. 0.000 min

Lab File: BN023555.D

Acq: 04 Jan 2023 16:34




Tgt Ion:202 Resp: 1515354


Ion Ratio Lower Upper

202 100

101 13.6 10.9 16.3

100 11.1 8.9 13.3





#83

Butylbenzylphthalate

Concen: 20.739 ng/ul

RT: 20.674 min Scan# 2990

Delta R.T. 0.000 min

Lab File: BN023555.D

Acq: 04 Jan 2023 16:34

Instrument :

BNA_N

ClientSampleId :

SSTD020219





Tgt Ion:149 Resp: 617683

Ion Ratio Lower Upper

149 100

91 72.4 57.9 86.9

206 21.0 16.8 25.2



#84

3,3'-Dichlorobenzidine

Concen: 19.905 ng/ul

RT: 21.463 min Scan# 3124

Delta R.T. 0.000 min

Lab File: BN023555.D

Acq: 04 Jan 2023 16:34




Tgt Ion:252 Resp: 473720


Ion Ratio Lower Upper

252 100

254 64.0 51.2 76.8

126 11.0 8.8 13.2





#85

Benzo(a)anthracene

Concen: 19.774 ng/ul

RT: 21.527 min Scan# 3135

Delta R.T. 0.000 min

Lab File: BN023555.D

Acq: 04 Jan 2023 16:34

Instrument :

BNA_N

ClientSampleId :

SSTD020219





Tgt Ion:228 Resp: 1434724

Ion Ratio Lower Upper

228 100

229 18.7 15.0 22.4

226 27.6 22.1 33.1



#86

Bis(2-ethylhexyl)phthalate

Concen: 21.450 ng/ul

RT: 21.451 min Scan# 3122

Delta R.T. 0.000 min

Lab File: BN023555.D

Acq: 04 Jan 2023 16:34




Tgt Ion:149 Resp: 903333


Ion Ratio Lower Upper

149 100

167 27.5 22.0 33.0

279 4.3 3.4 5.2





#87

Chrysene

Concen: 19.753 ng/ul

RT: 21.580 min Scan# 3144

Delta R.T. 0.000 min

Lab File: BN023555.D

Acq: 04 Jan 2023 16:34

Instrument :

BNA_N

ClientSampleId :

SSTD020219





Tgt Ion:228 Resp: 1352566

Ion Ratio Lower Upper

228 100

226 30.8 24.6 37.0

229 21.0 16.8 25.2



#89

Di-n-octyl phthalate

Concen: 22.124 ng/ul




RT: 22.380 min Scan# 3280


Delta R.T. 0.000 min

Lab File: BN023555.D

Acq: 04 Jan 2023 16:34

Tgt Ion:149 Resp: 1504752









#90
Benzo(b)fluoranthene
Concen: 19.908 ng/ul
RT: 23.227 min Scan# 3424
Delta R.T. 0.000 min
Lab File: BN023555.D
Acq: 04 Jan 2023 16:34




Instrument :
BNA_N
ClientSampleId :
SSTD020219


Tgt Ion:252 Resp: 1410218
Ion Ratio Lower Upper
252 100
253 21.5 17.2 25.8
125 10.8 8.6 13.0



#91
Benzo(k)fluoranthene
Concen: 20.584 ng/ul
RT: 23.274 min Scan# 3432
Delta R.T. 0.000 min
Lab File: BN023555.D
Acq: 04 Jan 2023 16:34

Tgt Ion:252 Resp: 1387657
Ion Ratio Lower Upper
252 100
253 22.3 17.8 26.8
125 10.7 8.6 12.8





#93

Benzo(a)pyrene

Concen: 20.270 ng/ul

RT: 23.857 min Scan# 3531

Delta R.T. 0.000 min

Lab File: BN023555.D

Acq: 04 Jan 2023 16:34

Instrument :

BNA_N

ClientSampleId :

SSTD020219





Tgt Ion:252 Resp: 1234007

Ion Ratio Lower Upper

252 100

253 20.8 16.6 25.0

125 12.2 9.8 14.6



#94

Indeno(1,2,3-cd)pyrene

Concen: 20.090 ng/ul

RT: 26.504 min Scan# 3981

Delta R.T. 0.000 min

Lab File: BN023555.D

Acq: 04 Jan 2023 16:34




Tgt Ion:276 Resp: 1549413


Ion Ratio Lower Upper

276 100

138 25.7 20.6 30.8

277 23.6 18.9 28.3





#95

Dibenzo(a,h)anthracene

Concen: 20.108 ng/ul

RT: 26.521 min Scan# 3984

Delta R.T. 0.000 min

Lab File: BN023555.D

Acq: 04 Jan 2023 16:34

Instrument :

BNA_N

ClientSampleId :

SSTD020219





Tgt Ion:278 Resp: 1264468

Ion Ratio Lower Upper

278 100

139 17.6 14.1 21.1

279 22.5 18.0 27.0



#96

Benzo(g,h,i)perylene

Concen: 19.737 ng/ul

RT: 27.286 min Scan# 4114

Delta R.T. 0.000 min

Lab File: BN023555.D

Acq: 04 Jan 2023 16:34

Tgt Ion:276 Resp: 1204366

Ion Ratio Lower Upper

276 100

138 21.3 17.0 25.6

277 22.0 17.6 26.4

