

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010523\
 Data File : BN023556.D
 Acq On : 04 Jan 2023 17:13
 Operator : CG/JU
 Sample : SST04056
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SST040220

Quant Time: Jan 05 02:37:26 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN010523.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 05 02:34:34 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.952	152	170829	20.000	ng/u1	0.00
20) Naphthalene-d8	10.769	136	766686	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.604	164	481317	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.357	188	1042467	20.000	ng/u1	0.00
79) Chrysene-d12	21.545	240	962969	20.000	ng/u1	0.00
88) Perylene-d12	23.974	264	984584	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.299	96	62695	15.184	ng/uL	0.00
4) Pyridine-d5	3.728	84	491132	40.295	ng/u1	0.00
7) Phenol-d5	7.110	99	616306	38.790	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.281	67	377158	40.792	ng/u1	0.00
11) 2-Chlorophenol-d4	7.475	132	487193	40.694	ng/u1	0.00
15) 4-Methylphenol-d8	8.669	113	493862	38.715	ng/u1	0.00
21) Nitrobenzene-d5	9.122	128	248897	42.018	ng/u1	0.00
24) 2-Nitrophenol-d4	9.846	143	274786	44.043	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.387	165	497426	42.233	ng/u1	0.00
31) 4-Chloroaniline-d4	10.910	131	750751	39.614	ng/u1	0.00
46) Dimethylphthalate-d6	14.016	166	1484430	41.154	ng/u1	0.00
49) Acenaphthylene-d8	14.298	160	1695448	41.914	ng/u1	0.00
54) 4-Nitrophenol-d4	14.810	143	298611	37.067	ng/u1	0.00
60) Fluorene-d10	15.598	176	1242736	40.446	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.722	200	266124	43.367	ng/u1	0.00
73) Anthracene-d10	17.457	188	1951128	41.763	ng/u1	0.00
81) Pyrene-d10	19.751	212	2265868	43.256	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.815	264	2070822	42.910	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.334	88	66277	15.489	ng/uL	97
5) Pyridine	3.746	79	489029	39.267	ng/u1	97
6) Benzaldehyde	7.081	77	265923	46.726	ng/u1	98
8) Phenol	7.140	94	627750	39.115	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.375	93	489445	39.265	ng/u1	99
12) 2-Chlorophenol	7.510	128	502796	40.891	ng/u1	98
13) 2-Methylphenol	8.399	108	473577	38.783	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.493	45	671882	42.558	ng/u1	98
16) Acetophenone	8.781	105	774526	40.129	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.775	70	401746	41.567	ng/u1	98
18) 4-Methylphenol	8.734	108	530619	39.158	ng/u1	99
19) Hexachloroethane	9.034	117	199190	42.061	ng/u1	95
22) Nitrobenzene	9.163	77	606395	42.693	ng/u1	99
23) Isophorone	9.699	82	1126221	42.037	ng/u1	99
25) 2-Nitrophenol	9.881	139	296120	43.517	ng/u1	98
26) 2,4-Dimethylphenol	9.940	107	577942	41.958	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.181	93	690786	40.307	ng/u1	99
29) 2,4-Dichlorophenol	10.416	162	486598	41.750	ng/u1	97
30) Naphthalene	10.822	128	1675688	40.911	ng/u1	98
32) 4-Chloroaniline	10.934	127	749589	39.645	ng/u1	99
33) Hexachlorobutadiene	11.104	225	289671	43.597	ng/u1	98
34) Caprolactam	11.722	113	84783	18.709	ng/u1	98
35) 4-Chloro-3-methylphenol	12.063	107	544500	41.247	ng/u1	97
36) 2-Methylnaphthalene	12.428	142	1128133	40.866	ng/u1	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010523\
 Data File : BN023556.D
 Acq On : 04 Jan 2023 17:13
 Operator : CG/JU
 Sample : SST04056
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SST040220

Quant Time: Jan 05 02:37:26 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN010523.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 05 02:34:20 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.651	142	1153808	40.559	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.798	216	593260	43.578	ng/ul	99
40) Hexachlorocyclopentadiene	12.775	237	367587	44.889	ng/ul	97
41) 2,4,6-Trichlorophenol	13.040	196	399926	42.708	ng/ul	94
42) 2,4,5-Trichlorophenol	13.110	196	434147	41.890	ng/ul	88
43) 1,1'-Biphenyl	13.440	154	1514152	41.632	ng/ul	98
44) 2-Chloronaphthalene	13.481	162	1165893	41.165	ng/ul	97
45) 2-Nitroaniline	13.692	65	359717	43.828	ng/ul	96
47) Dimethylphthalate	14.069	163	1474969	40.525	ng/ul	100
48) 2,6-Dinitrotoluene	14.187	165	312505	44.348	ng/ul	97
50) Acenaphthylene	14.328	152	1833385	41.216	ng/ul	100
51) 3-Nitroaniline	14.516	138	290812	40.531	ng/ul	100
52) Acenaphthene	14.669	153	1259860	40.740	ng/ul	99
53) 2,4-Dinitrophenol	14.722	184	181981	40.528	ng/ul	97
55) 4-Nitrophenol	14.828	109	241980	40.613	ng/ul	97
56) Dibenzofuran	15.004	168	1725284	40.070	ng/ul	98
57) 2,4-Dinitrotoluene	14.975	165	449658	43.229	ng/ul	97
58) 2,3,4,6-Tetrachlorophenol	15.234	232	375635	41.864	ng/ul	95
59) Diethylphthalate	15.428	149	1492721	41.228	ng/ul	99
61) Fluorene	15.657	166	1400855	40.178	ng/ul	97
62) 4-Chlorophenyl-phenyle...	15.645	204	686025	41.053	ng/ul	97
63) 4-Nitroaniline	15.681	138	275135	39.375	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.739	198	258616	43.168	ng/ul	96
67) N-Nitrosodiphenylamine	15.863	169	1222593	42.376	ng/ul	98
68) 4-Bromophenyl-phenylether	16.539	248	430758	42.824	ng/ul	98
69) Hexachlorobenzene	16.657	284	506763	43.091	ng/ul	98
70) Atrazine	16.822	200	458346	43.474	ng/ul	99
71) Pentachlorophenol	17.004	266	329465	42.426	ng/ul	93
72) Phenanthrene	17.398	178	2242608	40.655	ng/ul	99
74) Anthracene	17.486	178	2329663	42.327	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.404	216	589512	45.473	ng/ul	94
76) Pentachlorobenzene	14.922	250	573851	43.682	ng/ul	95
77) Carbazole	17.763	167	2083718	41.114	ng/ul	98
78) Di-n-butylphthalate	18.322	149	2620637	43.548	ng/ul	100
80) Fluoranthene	19.416	202	2666052	43.761	ng/ul	100
82) Pyrene	19.780	202	2724121	42.565	ng/ul	100
83) Butylbenzylphthalate	20.674	149	1185291	44.755	ng/ul	95
84) 3,3'-Dichlorobenzidine	21.463	252	892019	42.153	ng/ul	96
85) Benzo(a)anthracene	21.527	228	2691655	41.720	ng/ul	97
86) Bis(2-ethylhexyl)phtha...	21.451	149	1724765	46.059	ng/ul	99
87) Chrysene	21.580	228	2482805	40.778	ng/ul	96
89) Di-n-octyl phthalate	22.386	149	2909506	48.797	ng/ul	100
90) Benzo(b)fluoranthene	23.233	252	2607116	41.982	ng/ul	99
91) Benzo(k)fluoranthene	23.286	252	2563520	43.377	ng/ul	95
93) Benzo(a)pyrene	23.874	252	2249202	42.144	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	26.515	276	2823627	41.763	ng/ul	97
95) Dibenzo(a,h)anthracene	26.527	278	2348794	42.607	ng/ul	98
96) Benzo(g,h,i)perylene	27.292	276	1460069	27.294	ng/ul	96





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
 1,4-Dioxane
 Concen: 15.489 ng/uL
 RT: 3.334 min Scan# 42
 Delta R.T. -0.000 min
 Lab File: BN023556.D
 Acq: 04 Jan 2023 17:13




Instrument :
 BNA_N
 ClientSampleId :
 SST040220


Tgt Ion: 88 Resp: 66277
 Ion Ratio Lower Upper
 88 100
 43 29.5 24.6 36.8
 58 72.7 56.2 84.4



#5
 Pyridine
 Concen: 39.267 ng/uL
 RT: 3.746 min Scan# 112
 Delta R.T. -0.000 min
 Lab File: BN023556.D
 Acq: 04 Jan 2023 17:13

Tgt Ion: 79 Resp: 489029
 Ion Ratio Lower Upper
 79 100
 52 71.8 59.9 89.9
 51 34.1 28.6 42.8









#6
Benzaldehyde
Concen: 46.726 ng/ul
RT: 7.081 min Scan# 679
Delta R.T. -0.000 min
Lab File: BN023556.D
Acq: 04 Jan 2023 17:13




Instrument :
BNA_N
ClientSampleId :
SSTD040220


Tgt Ion: 77 Resp: 265923
Ion Ratio Lower Upper
77 100
105 93.3 75.1 112.7
106 88.3 72.8 109.2



#8
Phenol
Concen: 39.115 ng/ul
RT: 7.140 min Scan# 689
Delta R.T. -0.000 min
Lab File: BN023556.D
Acq: 04 Jan 2023 17:13

Tgt Ion: 94 Resp: 627750
Ion Ratio Lower Upper
94 100
65 28.4 23.7 35.5
66 36.8 29.9 44.9









#10
Bis(2-Chloroethyl)ether
Concen: 39.265 ng/ul
RT: 7.375 min Scan# 729
Delta R.T. 0.006 min
Lab File: BN023556.D
Acq: 04 Jan 2023 17:13




Instrument :
BNA_N
ClientSampleId :
SSTD040220


Tgt Ion: 93 Resp: 489445
Ion Ratio Lower Upper
93 100
63 76.1 61.2 91.8
95 32.9 26.6 40.0



#12
2-Chlorophenol
Concen: 40.891 ng/ul
RT: 7.510 min Scan# 752
Delta R.T. -0.000 min
Lab File: BN023556.D
Acq: 04 Jan 2023 17:13

Tgt Ion: 128 Resp: 502796
Ion Ratio Lower Upper
128 100
64 44.6 37.4 56.2
130 31.9 25.6 38.4





#13

2-Methylphenol

Concen: 38.783 ng/ul

RT: 8.399 min Scan# 903

Delta R.T. -0.000 min

Lab File: BN023556.D

Acq: 04 Jan 2023 17:13

Instrument :

BNA_N

ClientSampleId :





SSTD040220

Tgt Ion:108 Resp: 473577

Ion Ratio Lower Upper

108 100

107 91.1 72.3 108.5



#14

2,2'-oxybis(1-Chloropropane)

Concen: 42.558 ng/ul

RT: 8.493 min Scan# 919

Delta R.T. 0.006 min

Lab File: BN023556.D

Acq: 04 Jan 2023 17:13




Tgt Ion: 45 Resp: 671882


Ion Ratio Lower Upper

45 100

77 15.3 11.4 17.2

79 11.5 8.8 13.2









#16
Acetophenone
Concen: 40.129 ng/ul
RT: 8.781 min Scan# 968
Delta R.T. -0.000 min
Lab File: BN023556.D
Acq: 04 Jan 2023 17:13




Instrument :
BNA_N
ClientSampleId :
SSTD040220


Tgt Ion:105 Resp: 774526
Ion Ratio Lower Upper
105 100
77 81.0 64.2 96.2
51 27.1 22.0 33.0



#17
N-Nitroso-di-n-propylamine
Concen: 41.567 ng/ul
RT: 8.775 min Scan# 967
Delta R.T. 0.006 min
Lab File: BN023556.D
Acq: 04 Jan 2023 17:13

Tgt Ion: 70 Resp: 401746
Ion Ratio Lower Upper
70 100
42 61.1 47.8 71.6
101 9.8 7.8 11.6
130 21.4 16.3 24.5









#18
4-Methylphenol
Concen: 39.158 ng/ul
RT: 8.734 min Scan# 90
Delta R.T. 0.006 min
Lab File: BN023556.D
Acq: 04 Jan 2023 17:13




Instrument :
BNA_N
ClientSampleId :
SSTD040220


Tgt Ion:108 Resp: 530619
Ion Ratio Lower Upper
108 100
107 114.0 90.7 136.1



#19
Hexachloroethane
Concen: 42.061 ng/ul
RT: 9.034 min Scan# 1011
Delta R.T. -0.000 min
Lab File: BN023556.D
Acq: 04 Jan 2023 17:13

Tgt Ion:117 Resp: 199190
Ion Ratio Lower Upper
117 100
201 92.7 76.9 115.3
199 61.6 44.6 67.0





#22

Nitrobenzene

Concen: 42.693 ng/ul

RT: 9.163 min Scan# 1033

Delta R.T. -0.000 min

Lab File: BN023556.D

Acq: 04 Jan 2023 17:13

Instrument :

BNA_N

ClientSampleId :

SSTD040220





Tgt Ion: 77 Resp: 606395

Ion Ratio Lower Upper

77 100

123 44.1 35.6 53.4

65 13.0 11.0 16.4



#23

Isophorone

Concen: 42.037 ng/ul

RT: 9.699 min Scan# 1124

Delta R.T. 0.006 min

Lab File: BN023556.D

Acq: 04 Jan 2023 17:13




Tgt Ion: 82 Resp: 1126221


Ion Ratio Lower Upper

82 100

95 6.9 5.5 8.3

138 18.1 13.8 20.8





#25

2-Nitrophenol

Concen: 43.517 ng/ul

RT: 9.881 min Scan# 1155

Delta R.T. -0.000 min

Lab File: BN023556.D

Acq: 04 Jan 2023 17:13

Instrument :

BNA_N

ClientSampleId :

SSTD040220





Tgt Ion:139 Resp: 296120

Ion Ratio Lower Upper

139 100

65 51.0 40.3 60.5

109 28.7 24.6 37.0



#26

2,4-Dimethylphenol

Concen: 41.958 ng/ul

RT: 9.940 min Scan# 1165

Delta R.T. -0.000 min

Lab File: BN023556.D

Acq: 04 Jan 2023 17:13




Tgt Ion:107 Resp: 577942


Ion Ratio Lower Upper

107 100

121 51.2 40.4 60.6

122 85.6 67.0 100.4









#27
 Bis(2-Chloroethoxy)methane
 Concen: 40.307 ng/ul
 RT: 10.181 min Scan# 1206
 Delta R.T. -0.000 min
 Lab File: BN023556.D
 Acq: 04 Jan 2023 17:13




Instrument :
 BNA_N
 ClientSampleId :
 SST040220


Tgt Ion: 93 Resp: 690786
 Ion Ratio Lower Upper
 93 100
 95 32.8 25.5 38.3
 123 11.8 9.4 14.2



#29
 2,4-Dichlorophenol
 Concen: 41.750 ng/ul
 RT: 10.416 min Scan# 1246
 Delta R.T. -0.000 min
 Lab File: BN023556.D
 Acq: 04 Jan 2023 17:13

Tgt Ion: 162 Resp: 486598
 Ion Ratio Lower Upper
 162 100
 164 64.1 53.8 80.6
 98 34.7 28.8 43.2









#30
Naphthalene
Concen: 40.911 ng/ul
RT: 10.822 min Scan# 1315
Delta R.T. 0.006 min
Lab File: BN023556.D
Acq: 04 Jan 2023 17:13




Instrument :
BNA_N
ClientSampleId :
SSTD040220


Tgt Ion:128 Resp: 1675688
Ion Ratio Lower Upper
128 100
129 10.8 9.1 13.7
127 12.7 10.8 16.2



#32
4-Chloroaniline
Concen: 39.645 ng/ul
RT: 10.934 min Scan# 1334
Delta R.T. -0.000 min
Lab File: BN023556.D
Acq: 04 Jan 2023 17:13

Tgt Ion:127 Resp: 749589
Ion Ratio Lower Upper
127 100
129 33.4 26.4 39.6









#33
Hexachlorobutadiene
Concen: 43.597 ng/ul
RT: 11.104 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN023556.D
Acq: 04 Jan 2023 17:13




Instrument :
BNA_N
ClientSampleId :
SSTD040220


Tgt Ion:225 Resp: 289671
Ion Ratio Lower Upper
225 100
223 62.1 51.7 77.5
227 63.8 50.8 76.2



#34
Caprolactam
Concen: 18.709 ng/ul
RT: 11.722 min Scan# 1468
Delta R.T. 0.012 min
Lab File: BN023556.D
Acq: 04 Jan 2023 17:13

Tgt Ion:113 Resp: 84783
Ion Ratio Lower Upper
113 100
55 157.8 129.1 193.7
56 117.7 94.2 141.4









#35
4-Chloro-3-methylphenol
Concen: 41.247 ng/ul
RT: 12.063 min Scan# 1525
Delta R.T. 0.006 min
Lab File: BN023556.D
Acq: 04 Jan 2023 17:13




Instrument :
BNA_N
ClientSampleId :
SSTD040220


Tgt Ion:107 Resp: 544500
Ion Ratio Lower Upper
107 100
144 26.5 21.9 32.9
142 82.6 64.0 96.0



#36
2-Methylnaphthalene
Concen: 40.866 ng/ul
RT: 12.428 min Scan# 1588
Delta R.T. -0.000 min
Lab File: BN023556.D
Acq: 04 Jan 2023 17:13

Tgt Ion:142 Resp: 1128133
Ion Ratio Lower Upper
142 100
141 91.2 73.6 110.4









#37
1-Methylnaphthalene
Concen: 40.559 ng/ul
RT: 12.651 min Scan# 1626
Delta R.T. -0.000 min
Lab File: BN023556.D
Acq: 04 Jan 2023 17:13




Instrument :
BNA_N
ClientSampleId :
SSTD040220


Tgt Ion:142 Resp: 1153808
Ion Ratio Lower Upper
142 100
141 91.4 73.8 110.6
116 3.7 3.0 4.4



#39
1,2,4,5-Tetrachlorobenzene
Concen: 43.578 ng/ul
RT: 12.798 min Scan# 1651
Delta R.T. -0.000 min
Lab File: BN023556.D
Acq: 04 Jan 2023 17:13

Tgt Ion:216 Resp: 593260
Ion Ratio Lower Upper
216 100
214 78.4 62.4 93.6
179 20.2 16.2 24.2
108 17.1 14.6 21.8





#40

Hexachlorocyclopentadiene

Concen: 44.889 ng/ul

RT: 12.775 min Scan# 1647

Delta R.T. -0.000 min

Lab File: BN023556.D

Acq: 04 Jan 2023 17:13

Instrument :

BNA_N

ClientSampleId :

SSTD040220





Tgt Ion:237 Resp: 367587

Ion Ratio Lower Upper

237 100

235 61.8 51.0 76.4

272 13.6 9.8 14.8



#41

2,4,6-Trichlorophenol

Concen: 42.708 ng/ul

RT: 13.040 min Scan# 1692

Delta R.T. -0.000 min

Lab File: BN023556.D

Acq: 04 Jan 2023 17:13




Tgt Ion:196 Resp: 399926


Ion Ratio Lower Upper

196 100

198 95.3 71.4 107.0

200 30.1 22.5 33.7









#42
2,4,5-Trichlorophenol
Concen: 41.890 ng/ul
RT: 13.110 min Scan# 1704
Delta R.T. 0.006 min
Lab File: BN023556.D
Acq: 04 Jan 2023 17:13




Instrument :
BNA_N
ClientSampleId :
SSTD040220


Tgt Ion:196 Resp: 434147
Ion Ratio Lower Upper
196 100
198 99.0 68.1 102.1
200 31.6 24.1 36.1



#43
1,1'-Biphenyl
Concen: 41.632 ng/ul
RT: 13.440 min Scan# 1760
Delta R.T. -0.000 min
Lab File: BN023556.D
Acq: 04 Jan 2023 17:13

Tgt Ion:154 Resp: 1514152
Ion Ratio Lower Upper
154 100
153 39.3 32.5 48.7
76 13.7 11.0 16.6









#44
2-Chloronaphthalene
Concen: 41.165 ng/ul
RT: 13.481 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN023556.D
Acq: 04 Jan 2023 17:13




Instrument :
BNA_N
ClientSampleId :
SSTD040220


Tgt Ion:162 Resp: 1165893
Ion Ratio Lower Upper
162 100
164 33.7 26.2 39.2
127 42.0 31.8 47.8



#45
2-Nitroaniline
Concen: 43.828 ng/ul
RT: 13.692 min Scan# 1803
Delta R.T. 0.006 min
Lab File: BN023556.D
Acq: 04 Jan 2023 17:13

Tgt Ion: 65 Resp: 359717
Ion Ratio Lower Upper
65 100
92 70.6 53.5 80.3
138 108.5 84.0 126.0





#47

Dimethylphthalate

Concen: 40.525 ng/ul

RT: 14.069 min Scan# 1866

Delta R.T. 0.006 min

Lab File: BN023556.D

Acq: 04 Jan 2023 17:13

Instrument :

BNA_N

ClientSampleId :

SSTD040220





Tgt Ion:163 Resp: 1474969

Ion Ratio Lower Upper

163 100

194 4.1 3.2 4.8

164 9.9 7.9 11.9



#48

2,6-Dinitrotoluene

Concen: 44.348 ng/ul

RT: 14.187 min Scan# 1887

Delta R.T. -0.000 min

Lab File: BN023556.D

Acq: 04 Jan 2023 17:13




Tgt Ion:165 Resp: 312505


Ion Ratio Lower Upper

165 100

89 60.3 49.9 74.9

121 21.8 19.1 28.7









#50
Acenaphthylene
Concen: 41.216 ng/ul
RT: 14.328 min Scan# 1911
Delta R.T. -0.000 min
Lab File: BN023556.D
Acq: 04 Jan 2023 17:13




Instrument :
BNA_N
ClientSampleId :
SSTD040220


Tgt Ion:152 Resp: 1833385
Ion Ratio Lower Upper
152 100
151 19.9 15.8 23.8
153 12.9 10.2 15.4



#51
3-Nitroaniline
Concen: 40.531 ng/ul
RT: 14.516 min Scan# 1943
Delta R.T. 0.006 min
Lab File: BN023556.D
Acq: 04 Jan 2023 17:13

Tgt Ion:138 Resp: 290812
Ion Ratio Lower Upper
138 100
108 10.8 8.1 12.1
92 116.4 93.2 139.8









#52
Acenaphthene
Concen: 40.740 ng/ul
RT: 14.669 min Scan# 1969
Delta R.T. -0.000 min
Lab File: BN023556.D
Acq: 04 Jan 2023 17:13




Instrument :
BNA_N
ClientSampleId :
SSTD040220


Tgt Ion:153 Resp: 1259860
Ion Ratio Lower Upper
153 100
152 47.5 38.5 57.7
154 88.6 71.5 107.3



#53
2,4-Dinitrophenol
Concen: 40.528 ng/ul
RT: 14.722 min Scan# 1978
Delta R.T. 0.006 min
Lab File: BN023556.D
Acq: 04 Jan 2023 17:13

Tgt Ion:184 Resp: 181981
Ion Ratio Lower Upper
184 100
63 69.2 58.2 87.4
154 60.0 49.4 74.0









#55
4-Nitrophenol
Concen: 40.613 ng/ul
RT: 14.828 min Scan# 1996
Delta R.T. 0.012 min
Lab File: BN023556.D
Acq: 04 Jan 2023 17:13




Instrument :
BNA_N
ClientSampleId :
SSTD040220


Tgt Ion:109 Resp: 241980
Ion Ratio Lower Upper
109 100
139 130.6 104.8 157.2
65 126.3 107.3 160.9



#56
Dibenzofuran
Concen: 40.070 ng/ul
RT: 15.004 min Scan# 2026
Delta R.T. -0.000 min
Lab File: BN023556.D
Acq: 04 Jan 2023 17:13

Tgt Ion:168 Resp: 1725284
Ion Ratio Lower Upper
168 100
139 38.0 29.6 44.4









#57
2,4-Dinitrotoluene
Concen: 43.229 ng/ul
RT: 14.975 min Scan# 2021
Delta R.T. 0.006 min
Lab File: BN023556.D
Acq: 04 Jan 2023 17:13




Instrument :
BNA_N
ClientSampleId :
SSTD040220


Tgt Ion:165 Resp: 449658
Ion Ratio Lower Upper
165 100
63 44.5 37.2 55.8
182 2.8 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 41.864 ng/ul
RT: 15.234 min Scan# 2065
Delta R.T. 0.006 min
Lab File: BN023556.D
Acq: 04 Jan 2023 17:13

Tgt Ion:232 Resp: 375635
Ion Ratio Lower Upper
232 100
131 44.4 39.5 59.3
130 2.4 2.2 3.4
166 30.3 25.1 37.7





#59

Diethylphthalate

Concen: 41.228 ng/ul

RT: 15.428 min Scan# 2098

Delta R.T. -0.000 min

Lab File: BN023556.D

Acq: 04 Jan 2023 17:13

Instrument :

BNA_N

ClientSampleId :

SSTD040220





Tgt Ion:149 Resp: 1492721

Ion Ratio Lower Upper

149 100

177 21.9 17.9 26.9

150 13.1 9.9 14.9



#61

Fluorene

Concen: 40.178 ng/ul

RT: 15.657 min Scan# 2137

Delta R.T. 0.006 min

Lab File: BN023556.D

Acq: 04 Jan 2023 17:13




Tgt Ion:166 Resp: 1400855


Ion Ratio Lower Upper

166 100

165 100.5 77.8 116.6

167 14.1 10.4 15.6





#62

4-Chlorophenyl-phenylether

Concen: 41.053 ng/ul

RT: 15.645 min Scan# 2135

Delta R.T. -0.000 min

Lab File: BN023556.D

Acq: 04 Jan 2023 17:13

Instrument :

BNA_N

ClientSampleId :

SSTD040220





Tgt Ion:204 Resp: 686025

Ion Ratio Lower Upper

204 100

206 31.6 27.0 40.6

141 61.5 51.0 76.4



#63

4-Nitroaniline

Concen: 39.375 ng/ul

RT: 15.681 min Scan# 2141

Delta R.T. 0.006 min

Lab File: BN023556.D

Acq: 04 Jan 2023 17:13




Tgt Ion:138 Resp: 275135


Ion Ratio Lower Upper

138 100

92 51.5 41.3 61.9

108 81.8 64.2 96.2









#66
 4,6-Dinitro-2-methylphenol
 Concen: 43.168 ng/ul
 RT: 15.739 min Scan# 2151
 Delta R.T. 0.006 min
 Lab File: BN023556.D
 Acq: 04 Jan 2023 17:13




Instrument :
 BNA_N
 ClientSampleId :
 SST040220


Tgt Ion:198 Resp: 258616
 Ion Ratio Lower Upper
 198 100
 182 3.9 3.6 5.4
 77 24.5 21.2 31.8



#67
 N-Nitrosodiphenylamine
 Concen: 42.376 ng/ul
 RT: 15.863 min Scan# 2172
 Delta R.T. -0.000 min
 Lab File: BN023556.D
 Acq: 04 Jan 2023 17:13

Tgt Ion:169 Resp: 1222593
 Ion Ratio Lower Upper
 169 100
 168 65.0 53.4 80.2
 167 35.9 29.0 43.4









#68
4-Bromophenyl-phenylether
Concen: 42.824 ng/ul
RT: 16.539 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN023556.D
Acq: 04 Jan 2023 17:13




Instrument :
BNA_N
ClientSampleId :
SSTD040220


Tgt Ion:248 Resp: 430758
Ion Ratio Lower Upper
248 100
250 94.4 76.4 114.6
141 74.8 58.0 87.0



#69
Hexachlorobenzene
Concen: 43.091 ng/ul
RT: 16.657 min Scan# 2307
Delta R.T. -0.000 min
Lab File: BN023556.D
Acq: 04 Jan 2023 17:13

Tgt Ion:284 Resp: 506763
Ion Ratio Lower Upper
284 100
142 38.1 28.9 43.3
249 31.0 24.2 36.2





#70

Atrazine

Concen: 43.474 ng/ul

RT: 16.822 min Scan# 21

Delta R.T. 0.006 min

Lab File: BN023556.D

Acq: 04 Jan 2023 17:13

Instrument :

BNA_N

ClientSampleId :

SSTD040220





Tgt Ion:200 Resp: 458346

Ion Ratio Lower Upper

200 100

173 26.4 21.4 32.2

215 47.6 39.0 58.6



#71

Pentachlorophenol

Concen: 42.426 ng/ul

RT: 17.004 min Scan# 2366

Delta R.T. 0.006 min

Lab File: BN023556.D

Acq: 04 Jan 2023 17:13




Tgt Ion:266 Resp: 329465


Ion Ratio Lower Upper

266 100

264 59.7 51.4 77.0

268 61.5 54.5 81.7









#72
 Phenanthrene
 Concen: 40.655 ng/ul
 RT: 17.398 min Scan# 2432
 Delta R.T. 0.006 min
 Lab File: BN023556.D
 Acq: 04 Jan 2023 17:13

Instrument :
 BNA_N
 ClientSampleId :
 SST040220

Tgt Ion:178 Resp: 2242608




Ion	Ratio	Lower	Upper
178	100		
179	15.0	12.3	18.5
176	19.7	15.2	22.8




#74
 Anthracene
 Concen: 42.327 ng/ul
 RT: 17.486 min Scan# 2448
 Delta R.T. -0.000 min
 Lab File: BN023556.D
 Acq: 04 Jan 2023 17:13

Tgt Ion:178 Resp: 2329663

Ion	Ratio	Lower	Upper
178	100		
179	15.4	12.6	19.0
176	18.1	15.1	22.7





#75

1,2,3,4-Tetrachlorobenzene

Concen: 45.473 ng/uL

RT: 13.404 min Scan# 11

Delta R.T. -0.000 min

Lab File: BN023556.D

Acq: 04 Jan 2023 17:13

Instrument :

BNA_N

ClientSampleId :

SSTD040220





Tgt Ion:216 Resp: 589512

Ion Ratio Lower Upper

216 100

214 71.4 61.0 91.4

218 45.8 39.8 59.8



#76

Pentachlorobenzene

Concen: 43.682 ng/uL

RT: 14.922 min Scan# 2012

Delta R.T. -0.000 min

Lab File: BN023556.D

Acq: 04 Jan 2023 17:13




Tgt Ion:250 Resp: 573851


Ion Ratio Lower Upper

250 100

248 67.1 50.3 75.5

252 64.6 48.9 73.3









#77
 Carbazole
 Concen: 41.114 ng/ul
 RT: 17.763 min Scan# 2495
 Delta R.T. 0.006 min
 Lab File: BN023556.D
 Acq: 04 Jan 2023 17:13




Instrument :
 BNA_N
 ClientSampleId :
 SSTD040220


Tgt Ion:167 Resp: 2083718
 Ion Ratio Lower Upper
 167 100
 166 21.2 17.4 26.2
 139 13.0 11.0 16.6



#78
 Di-n-butylphthalate
 Concen: 43.548 ng/ul
 RT: 18.322 min Scan# 2590
 Delta R.T. -0.000 min
 Lab File: BN023556.D
 Acq: 04 Jan 2023 17:13

Tgt Ion:149 Resp: 2620637
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.4 11.2
 104 5.5 4.3 6.5









#80
Fluoranthene
Concen: 43.761 ng/ul
RT: 19.416 min Scan# 21
Delta R.T. 0.006 min
Lab File: BN023556.D
Acq: 04 Jan 2023 17:13




Instrument :
BNA_N
ClientSampleId :
SSTD040220


Tgt Ion:202 Resp: 2666052
Ion Ratio Lower Upper
202 100
101 11.7 9.5 14.3
100 9.1 7.3 10.9



#82
Pyrene
Concen: 42.565 ng/ul
RT: 19.780 min Scan# 2838
Delta R.T. 0.006 min
Lab File: BN023556.D
Acq: 04 Jan 2023 17:13

Tgt Ion:202 Resp: 2724121
Ion Ratio Lower Upper
202 100
101 13.6 10.9 16.3
100 11.3 8.9 13.3





#83

Butylbenzylphthalate

Concen: 44.755 ng/ul

RT: 20.674 min Scan# 2990

Delta R.T. -0.000 min

Lab File: BN023556.D

Acq: 04 Jan 2023 17:13

Instrument :

BNA_N

ClientSampleId :

SSTD040220





Tgt Ion:149 Resp: 1185291

Ion Ratio Lower Upper

149 100

91 77.2 57.9 86.9

206 21.8 16.8 25.2



#84

3,3'-Dichlorobenzidine

Concen: 42.153 ng/ul

RT: 21.463 min Scan# 3124

Delta R.T. -0.000 min

Lab File: BN023556.D

Acq: 04 Jan 2023 17:13




Tgt Ion:252 Resp: 892019


Ion Ratio Lower Upper

252 100

254 66.8 51.2 76.8

126 12.3 8.8 13.2





#85

Benzo(a)anthracene

Concen: 41.720 ng/ul

RT: 21.527 min Scan# 3135

Delta R.T. -0.000 min

Lab File: BN023556.D

Acq: 04 Jan 2023 17:13

Instrument :

BNA_N

ClientSampleId :

SSTD040220





Tgt Ion:228 Resp: 2691655

Ion Ratio Lower Upper

228 100

229 20.4 15.0 22.4

226 26.5 22.1 33.1



#86

Bis(2-ethylhexyl)phthalate

Concen: 46.059 ng/ul

RT: 21.451 min Scan# 3122

Delta R.T. -0.000 min

Lab File: BN023556.D

Acq: 04 Jan 2023 17:13




Tgt Ion:149 Resp: 1724765


Ion Ratio Lower Upper

149 100

167 26.8 22.0 33.0

279 3.9 3.4 5.2





#87

Chrysene

Concen: 40.778 ng/ul

RT: 21.580 min Scan# 3144

Delta R.T. -0.000 min

Lab File: BN023556.D

Acq: 04 Jan 2023 17:13

Instrument :

BNA_N

ClientSampleId :

SSTD040220





Tgt Ion:228 Resp: 2482805

Ion Ratio Lower Upper

228 100

226 29.5 24.6 37.0

229 18.4 16.8 25.2



#89

Di-n-octyl phthalate

Concen: 48.797 ng/ul




RT: 22.386 min Scan# 3281


Delta R.T. 0.006 min

Lab File: BN023556.D

Acq: 04 Jan 2023 17:13

Tgt Ion:149 Resp: 2909506









#90
Benzo(b)fluoranthene
Concen: 41.982 ng/ul
RT: 23.233 min Scan# 3424
Delta R.T. 0.006 min
Lab File: BN023556.D
Acq: 04 Jan 2023 17:13




Instrument :
BNA_N
ClientSampleId :
SSTD040220


Tgt Ion:252 Resp: 2607116
Ion Ratio Lower Upper
252 100
253 21.2 17.2 25.8
125 11.3 8.6 13.0



#91
Benzo(k)fluoranthene
Concen: 43.377 ng/ul
RT: 23.286 min Scan# 3434
Delta R.T. 0.012 min
Lab File: BN023556.D
Acq: 04 Jan 2023 17:13

Tgt Ion:252 Resp: 2563520
Ion Ratio Lower Upper
252 100
253 19.4 17.8 26.8
125 9.4 8.6 12.8





#93

Benzo(a)pyrene

Concen: 42.144 ng/ul

RT: 23.874 min Scan# 3

Delta R.T. 0.018 min

Lab File: BN023556.D

Acq: 04 Jan 2023 17:13

Instrument :

BNA_N

ClientSampleId :

SSTD040220





Tgt Ion:252 Resp: 2249202

Ion Ratio Lower Upper

252 100

253 22.0 16.6 25.0

125 11.1 9.8 14.6



#94

Indeno(1,2,3-cd)pyrene

Concen: 41.763 ng/ul

RT: 26.515 min Scan# 3983

Delta R.T. 0.012 min

Lab File: BN023556.D

Acq: 04 Jan 2023 17:13




Tgt Ion:276 Resp: 2823627


Ion Ratio Lower Upper

276 100

138 26.2 20.6 30.8

277 25.7 18.9 28.3





#95

Dibenzo(a,h)anthracene

Concen: 42.607 ng/ul

RT: 26.527 min Scan# 3984

Delta R.T. 0.006 min

Lab File: BN023556.D

Acq: 04 Jan 2023 17:13

Instrument :

BNA_N

ClientSampleId :

SSTD040220





Tgt Ion:278 Resp: 2348794

Ion Ratio Lower Upper

278 100

139 17.3 14.1 21.1

279 23.8 18.0 27.0



#96

Benzo(g,h,i)perylene

Concen: 27.294 ng/ul

RT: 27.292 min Scan# 4115

Delta R.T. 0.006 min

Lab File: BN023556.D

Acq: 04 Jan 2023 17:13

Tgt Ion:276 Resp: 1460069

Ion Ratio Lower Upper

276 100

138 24.2 17.0 25.6

277 23.4 17.6 26.4

