

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010523\
 Data File : BN023557.D
 Acq On : 04 Jan 2023 17:53
 Operator : CG/JU
 Sample : SST08057
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SST080221

Quant Time: Jan 05 02:37:57 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN010523.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 05 02:34:34 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.951	152	182215	20.000	ng/u1	0.00
20) Naphthalene-d8	10.769	136	810034	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.610	164	490556	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.357	188	1049553	20.000	ng/u1	0.00
79) Chrysene-d12	21.545	240	851665	20.000	ng/u1	0.00
88) Perylene-d12	23.968	264	922098	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.299	96	128812	29.248	ng/uL	0.00
4) Pyridine-d5	3.722	84	1025812	78.904	ng/u1	0.00
7) Phenol-d5	7.116	99	1323018	78.067	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.281	67	783987	79.495	ng/u1	0.00
11) 2-Chlorophenol-d4	7.481	132	1041371	81.548	ng/u1	0.00
15) 4-Methylphenol-d8	8.681	113	1055160	77.547	ng/u1	0.01
21) Nitrobenzene-d5	9.128	128	537707	85.916	ng/u1	0.00
24) 2-Nitrophenol-d4	9.851	143	609660	92.488	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.392	165	1045000	83.977	ng/u1	0.00
31) 4-Chloroaniline-d4	10.916	131	1531888	76.506	ng/u1	0.00
46) Dimethylphthalate-d6	14.022	166	2936474	79.878	ng/u1	0.00
49) Acenaphthylene-d8	14.304	160	3373630	81.831	ng/u1	0.00
54) 4-Nitrophenol-d4	14.822	143	608277	74.085	ng/u1	0.02
60) Fluorene-d10	15.604	176	2402752	76.728	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.733	200	561467	90.877	ng/u1	0.02
73) Anthracene-d10	17.457	188	3654192	77.688	ng/u1	0.00
81) Pyrene-d10	19.751	212	4057695	87.585	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.821	264	3788069	83.813	ng/u1	0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.334	88	135853	29.764	ng/uL	98
5) Pyridine	3.746	79	1036419	78.019	ng/u1	94
6) Benzaldehyde	7.081	77	486956	80.217	ng/u1	98
8) Phenol	7.146	94	1331509	77.782	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.375	93	1038214	78.086	ng/u1	98
12) 2-Chlorophenol	7.510	128	1050975	80.132	ng/u1	99
13) 2-Methylphenol	8.404	108	1015175	77.941	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.493	45	1383007	82.129	ng/u1	98
16) Acetophenone	8.787	105	1574779	76.492	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.793	70	830651	80.574	ng/u1	99
18) 4-Methylphenol	8.746	108	1110397	76.823	ng/u1	99
19) Hexachloroethane	9.034	117	417646	82.679	ng/u1	95
22) Nitrobenzene	9.175	77	1252158	83.439	ng/u1	98
23) Isophorone	9.710	82	2379549	84.065	ng/u1	99
25) 2-Nitrophenol	9.887	139	629857	87.609	ng/u1	99
26) 2,4-Dimethylphenol	9.945	107	1188082	81.638	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.187	93	1437317	79.379	ng/u1	99
29) 2,4-Dichlorophenol	10.422	162	1018858	82.741	ng/u1	97
30) Naphthalene	10.822	128	3378210	78.064	ng/u1	100
32) 4-Chloroaniline	10.940	127	1521137	76.145	ng/u1	99
33) Hexachlorobutadiene	11.104	225	595553	84.837	ng/u1	98
34) Caprolactam	11.751	113	191394	39.975	ng/u1	98
35) 4-Chloro-3-methylphenol	12.069	107	1132814	81.220	ng/u1	99
36) 2-Methylnaphthalene	12.434	142	2286523	78.396	ng/u1	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010523\
 Data File : BN023557.D
 Acq On : 04 Jan 2023 17:53
 Operator : CG/JU
 Sample : SSTD08057
 Misc :
 ALS Vial : 6 Sample Multiplier: 1


Instrument :
 BNA_N
 ClientSampleId :
 SSTD080221


Quant Time: Jan 05 02:37:57 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN010523.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 05 02:34:34 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.651	142	2299956	76.521	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.804	216	1161645	83.722	ng/ul	97
40) Hexachlorocyclopentadiene	12.781	237	823878	98.716	ng/ul	97
41) 2,4,6-Trichlorophenol	13.045	196	831685	87.144	ng/ul	92
42) 2,4,5-Trichlorophenol	13.116	196	894467	84.681	ng/ul	91
43) 1,1'-Biphenyl	13.445	154	2957983	79.799	ng/ul	99
44) 2-Chloronaphthalene	13.486	162	2333585	80.842	ng/ul	99
45) 2-Nitroaniline	13.698	65	755930	90.367	ng/ul	96
47) Dimethylphthalate	14.075	163	2948307	79.480	ng/ul	100
48) 2,6-Dinitrotoluene	14.198	165	657292	91.520	ng/ul	93
50) Acenaphthylene	14.333	152	3564931	78.633	ng/ul	99
51) 3-Nitroaniline	14.522	138	577981	79.036	ng/ul	97
52) Acenaphthene	14.675	153	2463153	78.150	ng/ul	99
53) 2,4-Dinitrophenol	14.728	184	420649	91.917	ng/ul	96
55) 4-Nitrophenol	14.839	109	472738	77.848	ng/ul	99
56) Dibenzofuran	15.010	168	3339958	76.110	ng/ul	97
57) 2,4-Dinitrotoluene	14.980	165	904195	85.290	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	15.233	232	748609	81.861	ng/ul	99
59) Diethylphthalate	15.433	149	2961353	80.249	ng/ul	99
61) Fluorene	15.657	166	2568943	72.292	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.651	204	1264894	74.268	ng/ul	98
63) 4-Nitroaniline	15.692	138	536197	75.291	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.745	198	540151	89.552	ng/ul	96
67) N-Nitrosodiphenylamine	15.869	169	2293859	78.970	ng/ul	99
68) 4-Bromophenyl-phenylether	16.545	248	843761	83.317	ng/ul	92
69) Hexachlorobenzene	16.663	284	980685	82.826	ng/ul	98
70) Atrazine	16.827	200	882957	83.182	ng/ul	98
71) Pentachlorophenol	17.004	266	672499	86.015	ng/ul	98
72) Phenanthrene	17.398	178	4197641	75.583	ng/ul	99
74) Anthracene	17.498	178	4177673	75.391	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.410	216	1171140	89.728	ng/ul	98
76) Pentachlorobenzene	14.928	250	1129460	85.396	ng/ul	98
77) Carbazole	17.763	167	3935002	77.117	ng/ul	99
78) Di-n-butylphthalate	18.327	149	4975773	82.126	ng/ul	100
80) Fluoranthene	19.416	202	4845140	89.922	ng/ul	99
82) Pyrene	19.786	202	4800267	84.809	ng/ul	99
83) Butylbenzylphthalate	20.674	149	2286544	97.620	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.462	252	1535224	82.029	ng/ul	97
85) Benzo(a)anthracene	21.527	228	4602059	80.653	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.451	149	3084189	93.125	ng/ul	98
87) Chrysene	21.586	228	4218609	78.342	ng/ul	98
89) Di-n-octyl phthalate	22.386	149	5446082	97.529	ng/ul	100
90) Benzo(b)fluoranthene	23.233	252	4703452	80.872	ng/ul	98
91) Benzo(k)fluoranthene	23.292	252	4500105	81.306	ng/ul	99
93) Benzo(a)pyrene	23.874	252	4129097	82.611	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.521	276	3599544	56.847	ng/ul	97
95) Dibenzo(a,h)anthracene	26.539	278	4075469	78.939	ng/ul	99
96) Benzo(g,h,i)perylene	27.303	276	4090278	81.644	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jan 05 02:37:57 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN010523.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Jan 05 02:34:34 2023
Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 29.764 ng/uL
 RT: 3.334 min Scan# 42
 Delta R.T. -0.000 min
 Lab File: BN023557.D
 Acq: 04 Jan 2023 17:53

Instrument :
 BNA_N
 ClientSampleId :
 SST080221

Tgt Ion: 88 Resp: 135853




Ion	Ratio	Lower	Upper
88	100		
43	28.5	24.6	36.8
58	69.5	56.2	84.4




#5
 Pyridine
 Concen: 78.019 ng/uL
 RT: 3.746 min Scan# 112
 Delta R.T. -0.000 min
 Lab File: BN023557.D
 Acq: 04 Jan 2023 17:53

Tgt Ion: 79 Resp: 1036419

Ion	Ratio	Lower	Upper
79	100		
52	69.5	59.9	89.9
51	33.1	28.6	42.8









#6
Benzaldehyde
Concen: 80.217 ng/ul
RT: 7.081 min Scan# 61
Delta R.T. -0.000 min
Lab File: BN023557.D
Acq: 04 Jan 2023 17:53




Instrument :
BNA_N
ClientSampleId :
SSTD080221


Tgt Ion: 77 Resp: 486956
Ion Ratio Lower Upper
77 100
105 96.4 75.1 112.7
106 92.6 72.8 109.2



#8
Phenol
Concen: 77.782 ng/ul
RT: 7.146 min Scan# 690
Delta R.T. 0.006 min
Lab File: BN023557.D
Acq: 04 Jan 2023 17:53

Tgt Ion: 94 Resp: 1331509
Ion Ratio Lower Upper
94 100
65 28.7 23.7 35.5
66 37.9 29.9 44.9









#10
Bis(2-Chloroethyl)ether
Concen: 78.086 ng/ul
RT: 7.375 min Scan# 729
Delta R.T. 0.006 min
Lab File: BN023557.D
Acq: 04 Jan 2023 17:53




Instrument :
BNA_N
ClientSampleId :
SSTD080221


Tgt Ion: 93 Resp: 1038214
Ion Ratio Lower Upper
93 100
63 74.5 61.2 91.8
95 32.2 26.6 40.0



#12
2-Chlorophenol
Concen: 80.132 ng/ul
RT: 7.510 min Scan# 752
Delta R.T. -0.000 min
Lab File: BN023557.D
Acq: 04 Jan 2023 17:53

Tgt Ion:128 Resp: 1050975
Ion Ratio Lower Upper
128 100
64 47.8 37.4 56.2
130 32.4 25.6 38.4





#13

2-Methylphenol

Concen: 77.941 ng/ul

RT: 8.404 min Scan# 904

Delta R.T. 0.006 min

Lab File: BN023557.D

Acq: 04 Jan 2023 17:53

Instrument :

BNA_N

ClientSampleId :





SSTD080221

Tgt Ion:108 Resp: 1015175

Ion Ratio Lower Upper

108 100

107 87.8 72.3 108.5



#14

2,2'-oxybis(1-Chloropropane)

Concen: 82.129 ng/ul

RT: 8.493 min Scan# 919

Delta R.T. 0.006 min

Lab File: BN023557.D

Acq: 04 Jan 2023 17:53




Tgt Ion: 45 Resp: 1383007


Ion Ratio Lower Upper

45 100

77 15.1 11.4 17.2

79 11.6 8.8 13.2









#16
Acetophenone
Concen: 76.492 ng/ul
RT: 8.787 min Scan# 90
Delta R.T. 0.006 min
Lab File: BN023557.D
Acq: 04 Jan 2023 17:53




Instrument :
BNA_N
ClientSampleId :
SSTD080221


Tgt Ion:105 Resp: 1574779
Ion Ratio Lower Upper
105 100
77 82.2 64.2 96.2
51 27.6 22.0 33.0



#17
N-Nitroso-di-n-propylamine
Concen: 80.574 ng/ul
RT: 8.793 min Scan# 970
Delta R.T. 0.023 min
Lab File: BN023557.D
Acq: 04 Jan 2023 17:53

Tgt Ion: 70 Resp: 830651
Ion Ratio Lower Upper
70 100
42 59.1 47.8 71.6
101 9.8 7.8 11.6
130 21.6 16.3 24.5









#18
4-Methylphenol
Concen: 76.823 ng/ul
RT: 8.746 min Scan# 911
Delta R.T. 0.017 min
Lab File: BN023557.D
Acq: 04 Jan 2023 17:53




Instrument :
BNA_N
ClientSampleId :
SSTD080221


Tgt Ion:108 Resp: 1110397
Ion Ratio Lower Upper
108 100
107 112.8 90.7 136.1



#19
Hexachloroethane
Concen: 82.679 ng/ul
RT: 9.034 min Scan# 1011
Delta R.T. -0.000 min
Lab File: BN023557.D
Acq: 04 Jan 2023 17:53

Tgt Ion:117 Resp: 417646
Ion Ratio Lower Upper
117 100
201 91.5 76.9 115.3
199 59.6 44.6 67.0





#22

Nitrobenzene

Concen: 83.439 ng/ul

RT: 9.175 min Scan# 1035

Delta R.T. 0.012 min

Lab File: BN023557.D

Acq: 04 Jan 2023 17:53

Instrument :





BNA_N

ClientSampleId :

SSTD080221

Tgt Ion: 77 Resp: 1252158

Ion	Ratio	Lower	Upper
77	100		
123	46.2	35.6	53.4
65	13.2	11.0	16.4



#23

Isophorone

Concen: 84.065 ng/ul

RT: 9.710 min Scan# 1126




Delta R.T. 0.017 min


Lab File: BN023557.D

Acq: 04 Jan 2023 17:53

Tgt Ion: 82 Resp: 2379549

Ion	Ratio	Lower	Upper
82	100		
95	7.2	5.5	8.3
138	17.9	13.8	20.8





#25

2-Nitrophenol

Concen: 87.609 ng/ul

RT: 9.887 min Scan# 1155

Delta R.T. 0.006 min

Lab File: BN023557.D

Acq: 04 Jan 2023 17:53

Instrument :

BNA_N

ClientSampleId :

SSTD080221





Tgt Ion:139 Resp: 629857

Ion Ratio Lower Upper

139 100

65 49.8 40.3 60.5

109 29.8 24.6 37.0



#26

2,4-Dimethylphenol

Concen: 81.638 ng/ul

RT: 9.945 min Scan# 1166

Delta R.T. 0.006 min

Lab File: BN023557.D

Acq: 04 Jan 2023 17:53




Tgt Ion:107 Resp: 1188082


Ion Ratio Lower Upper

107 100

121 50.9 40.4 60.6

122 86.6 67.0 100.4





#27

Bis(2-Chloroethoxy)methane

Concen: 79.379 ng/ul

RT: 10.187 min Scan# 1206

Delta R.T. 0.006 min

Lab File: BN023557.D

Acq: 04 Jan 2023 17:53

Instrument :

BNA_N

ClientSampleId :

SSTD080221





Tgt Ion: 93 Resp: 1437317

Ion Ratio Lower Upper

93 100

95 32.3 25.5 38.3

123 11.2 9.4 14.2



#29

2,4-Dichlorophenol

Concen: 82.741 ng/ul

RT: 10.422 min Scan# 1247

Delta R.T. 0.006 min

Lab File: BN023557.D

Acq: 04 Jan 2023 17:53




Tgt Ion:162 Resp: 1018858


Ion Ratio Lower Upper

162 100

164 63.9 53.8 80.6

98 35.5 28.8 43.2









#30
Naphthalene
Concen: 78.064 ng/ul
RT: 10.822 min Scan# 11
Delta R.T. 0.006 min
Lab File: BN023557.D
Acq: 04 Jan 2023 17:53




Instrument :
BNA_N
ClientSampleId :
SSTD080221


Tgt Ion:128 Resp: 3378210
Ion Ratio Lower Upper
128 100
129 11.0 9.1 13.7
127 13.5 10.8 16.2



#32
4-Chloroaniline
Concen: 76.145 ng/ul
RT: 10.940 min Scan# 1335
Delta R.T. 0.006 min
Lab File: BN023557.D
Acq: 04 Jan 2023 17:53

Tgt Ion:127 Resp: 1521137
Ion Ratio Lower Upper
127 100
129 32.5 26.4 39.6









#33
Hexachlorobutadiene
Concen: 84.837 ng/ul
RT: 11.104 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN023557.D
Acq: 04 Jan 2023 17:53




Instrument :
BNA_N
ClientSampleId :
SSTD080221


Tgt Ion:225 Resp: 595553
Ion Ratio Lower Upper
225 100
223 65.9 51.7 77.5
227 65.2 50.8 76.2



#34
Caprolactam
Concen: 39.975 ng/ul
RT: 11.751 min Scan# 1473
Delta R.T. 0.041 min
Lab File: BN023557.D
Acq: 04 Jan 2023 17:53

Tgt Ion:113 Resp: 191394
Ion Ratio Lower Upper
113 100
55 162.3 129.1 193.7
56 122.0 94.2 141.4









#35
4-Chloro-3-methylphenol
Concen: 81.220 ng/ul
RT: 12.069 min Scan# 1525
Delta R.T. 0.012 min
Lab File: BN023557.D
Acq: 04 Jan 2023 17:53




Instrument :
BNA_N
ClientSampleId :
SSTD080221


Tgt Ion:107 Resp: 1132814
Ion Ratio Lower Upper
107 100
144 25.8 21.9 32.9
142 80.5 64.0 96.0



#36
2-Methylnaphthalene
Concen: 78.396 ng/ul
RT: 12.434 min Scan# 1589
Delta R.T. 0.006 min
Lab File: BN023557.D
Acq: 04 Jan 2023 17:53

Tgt Ion:142 Resp: 2286523
Ion Ratio Lower Upper
142 100
141 88.8 73.6 110.4





#37

1-Methylnaphthalene

Concen: 76.521 ng/ul

RT: 12.651 min Scan# 1626

Delta R.T. -0.000 min

Lab File: BN023557.D

Acq: 04 Jan 2023 17:53

Instrument :

BNA_N

ClientSampleId :

SSTD080221





Tgt Ion:142 Resp: 2299956

Ion Ratio Lower Upper

142 100

141 90.1 73.8 110.6

116 4.0 3.0 4.4



#39

1,2,4,5-Tetrachlorobenzene

Concen: 83.722 ng/ul

RT: 12.804 min Scan# 1652

Delta R.T. 0.006 min

Lab File: BN023557.D

Acq: 04 Jan 2023 17:53

Tgt Ion:216 Resp: 1161645




Ion Ratio Lower Upper


216 100

214 81.4 62.4 93.6

179 21.5 16.2 24.2

108 17.7 14.6 21.8





#40

Hexachlorocyclopentadiene

Concen: 98.716 ng/ul

RT: 12.781 min Scan# 1648

Delta R.T. 0.006 min

Lab File: BN023557.D

Acq: 04 Jan 2023 17:53

Instrument :

BNA_N

ClientSampleId :

SSTD080221





Tgt Ion:237 Resp: 823878

Ion Ratio Lower Upper

237 100

235 61.7 51.0 76.4

272 13.1 9.8 14.8



#41

2,4,6-Trichlorophenol

Concen: 87.144 ng/ul

RT: 13.045 min Scan# 1693

Delta R.T. 0.006 min

Lab File: BN023557.D

Acq: 04 Jan 2023 17:53




Tgt Ion:196 Resp: 831685


Ion Ratio Lower Upper

196 100

198 97.0 71.4 107.0

200 32.1 22.5 33.7





#42

2,4,5-Trichlorophenol

Concen: 84.681 ng/ul

RT: 13.116 min Scan# 1705

Delta R.T. 0.012 min

Lab File: BN023557.D

Acq: 04 Jan 2023 17:53

Instrument :

BNA_N

ClientSampleId :

SSTD080221





Tgt Ion:196 Resp: 894467

Ion Ratio Lower Upper

196 100

198 95.4 68.1 102.1

200 31.3 24.1 36.1



#43

1,1'-Biphenyl

Concen: 79.799 ng/ul

RT: 13.445 min Scan# 1761

Delta R.T. 0.006 min

Lab File: BN023557.D

Acq: 04 Jan 2023 17:53




Tgt Ion:154 Resp: 2957983


Ion Ratio Lower Upper

154 100

153 40.1 32.5 48.7

76 13.9 11.0 16.6





#44

2-Chloronaphthalene

Concen: 80.842 ng/ul

RT: 13.486 min Scan# 1

Delta R.T. 0.006 min

Lab File: BN023557.D

Acq: 04 Jan 2023 17:53

Instrument :

BNA_N

ClientSampleId :

SSTD080221





Tgt Ion:162 Resp: 2333585

Ion Ratio Lower Upper

162 100

164 33.0 26.2 39.2

127 41.1 31.8 47.8



#45

2-Nitroaniline

Concen: 90.367 ng/ul

RT: 13.698 min Scan# 1804

Delta R.T. 0.012 min

Lab File: BN023557.D

Acq: 04 Jan 2023 17:53




Tgt Ion: 65 Resp: 755930


Ion Ratio Lower Upper

65 100

92 69.9 53.5 80.3

138 109.5 84.0 126.0





#47

Dimethylphthalate

Concen: 79.480 ng/ul

RT: 14.075 min Scan# 1868

Delta R.T. 0.012 min

Lab File: BN023557.D

Acq: 04 Jan 2023 17:53

Instrument :

BNA_N

ClientSampleId :

SSTD080221





Tgt Ion:163 Resp: 2948307

Ion Ratio Lower Upper

163 100

194 4.2 3.2 4.8

164 9.9 7.9 11.9



#48

2,6-Dinitrotoluene

Concen: 91.520 ng/ul

RT: 14.198 min Scan# 1889

Delta R.T. 0.012 min

Lab File: BN023557.D

Acq: 04 Jan 2023 17:53




Tgt Ion:165 Resp: 657292


Ion Ratio Lower Upper

165 100

89 57.1 49.9 74.9

121 20.4 19.1 28.7









#50
 Acenaphthylene
 Concen: 78.633 ng/ul
 RT: 14.333 min Scan# 1912
 Delta R.T. 0.006 min
 Lab File: BN023557.D
 Acq: 04 Jan 2023 17:53

Instrument :
 BNA_N
 ClientSampleId :
 SSTD080221

Tgt Ion:152 Resp: 3564931




Ion	Ratio	Lower	Upper
152	100		
151	20.1	15.8	23.8
153	13.1	10.2	15.4




#51
 3-Nitroaniline
 Concen: 79.036 ng/ul
 RT: 14.522 min Scan# 1944
 Delta R.T. 0.012 min
 Lab File: BN023557.D
 Acq: 04 Jan 2023 17:53

Tgt Ion:138 Resp: 577981

Ion	Ratio	Lower	Upper
138	100		
108	10.1	8.1	12.1
92	113.2	93.2	139.8









#52
Acenaphthene
Concen: 78.150 ng/ul
RT: 14.675 min Scan# 1970
Delta R.T. 0.006 min
Lab File: BN023557.D
Acq: 04 Jan 2023 17:53




Instrument :
BNA_N
ClientSampleId :
SSTD080221


Tgt Ion:153 Resp: 2463153
Ion Ratio Lower Upper
153 100
152 46.3 38.5 57.7
154 89.5 71.5 107.3



#53
2,4-Dinitrophenol
Concen: 91.917 ng/ul
RT: 14.728 min Scan# 1979
Delta R.T. 0.012 min
Lab File: BN023557.D
Acq: 04 Jan 2023 17:53

Tgt Ion:184 Resp: 420649
Ion Ratio Lower Upper
184 100
63 67.7 58.2 87.4
154 61.3 49.4 74.0









#55
4-Nitrophenol
Concen: 77.848 ng/ul
RT: 14.839 min Scan# 1994
Delta R.T. 0.023 min
Lab File: BN023557.D
Acq: 04 Jan 2023 17:53




Instrument :
BNA_N
ClientSampleId :
SSTD080221


Tgt Ion:109 Resp: 472738
Ion Ratio Lower Upper
109 100
139 134.2 104.8 157.2
65 134.5 107.3 160.9



#56
Dibenzofuran
Concen: 76.110 ng/ul
RT: 15.010 min Scan# 2027
Delta R.T. 0.006 min
Lab File: BN023557.D
Acq: 04 Jan 2023 17:53

Tgt Ion:168 Resp: 3339958
Ion Ratio Lower Upper
168 100
139 38.5 29.6 44.4





#57

2,4-Dinitrotoluene

Concen: 85.290 ng/ul

RT: 14.980 min Scan# 2020

Delta R.T. 0.012 min

Lab File: BN023557.D

Acq: 04 Jan 2023 17:53

Instrument :

BNA_N

ClientSampleId :

SSTD080221





Tgt Ion:165 Resp: 904195

Ion Ratio Lower Upper

165 100

63 46.0 37.2 55.8

182 2.7 2.2 3.2



#58

2,3,4,6-Tetrachlorophenol

Concen: 81.861 ng/ul

RT: 15.233 min Scan# 2065

Delta R.T. 0.006 min

Lab File: BN023557.D

Acq: 04 Jan 2023 17:53

Tgt Ion:232 Resp: 748609




Ion Ratio Lower Upper


232 100

131 48.8 39.5 59.3

130 2.6 2.2 3.4

166 31.1 25.1 37.7





#59

Diethylphthalate

Concen: 80.249 ng/ul

RT: 15.433 min Scan# 2099

Delta R.T. 0.006 min

Lab File: BN023557.D

Acq: 04 Jan 2023 17:53

Instrument :

BNA_N

ClientSampleId :

SSTD080221





Tgt Ion:149 Resp: 2961353

Ion Ratio Lower Upper

149 100

177 21.8 17.9 26.9

150 12.6 9.9 14.9



#61

Fluorene

Concen: 72.292 ng/ul

RT: 15.657 min Scan# 2137

Delta R.T. 0.006 min

Lab File: BN023557.D

Acq: 04 Jan 2023 17:53




Tgt Ion:166 Resp: 2568943


Ion Ratio Lower Upper

166 100

165 97.4 77.8 116.6

167 13.5 10.4 15.6









#62
4-Chlorophenyl-phenylether
Concen: 74.268 ng/ul
RT: 15.651 min Scan# 2135
Delta R.T. 0.006 min
Lab File: BN023557.D
Acq: 04 Jan 2023 17:53




Instrument :
BNA_N
ClientSampleId :
SSTD080221


Tgt Ion:204 Resp: 1264894
Ion Ratio Lower Upper
204 100
206 31.9 27.0 40.6
141 62.6 51.0 76.4



#63
4-Nitroaniline
Concen: 75.291 ng/ul
RT: 15.692 min Scan# 2143
Delta R.T. 0.017 min
Lab File: BN023557.D
Acq: 04 Jan 2023 17:53

Tgt Ion:138 Resp: 536197
Ion Ratio Lower Upper
138 100
92 51.1 41.3 61.9
108 81.4 64.2 96.2









#66
4,6-Dinitro-2-methylphenol
Concen: 89.552 ng/ul
RT: 15.745 min Scan# 2150
Delta R.T. 0.012 min
Lab File: BN023557.D
Acq: 04 Jan 2023 17:53




Instrument :
BNA_N
ClientSampleId :
SSTD080221


Tgt Ion:198 Resp: 540151
Ion Ratio Lower Upper
198 100
182 3.7 3.6 5.4
77 28.7 21.2 31.8



#67
N-Nitrosodiphenylamine
Concen: 78.970 ng/ul
RT: 15.869 min Scan# 2173
Delta R.T. 0.006 min
Lab File: BN023557.D
Acq: 04 Jan 2023 17:53

Tgt Ion:169 Resp: 2293859
Ion Ratio Lower Upper
169 100
168 66.0 53.4 80.2
167 35.9 29.0 43.4









#68
4-Bromophenyl-phenylether
Concen: 83.317 ng/ul
RT: 16.545 min Scan# 21
Delta R.T. 0.006 min
Lab File: BN023557.D
Acq: 04 Jan 2023 17:53




Instrument :
BNA_N
ClientSampleId :
SSTD080221


Tgt Ion:248 Resp: 843761
Ion Ratio Lower Upper
248 100
250 102.9 76.4 114.6
141 79.9 58.0 87.0



#69
Hexachlorobenzene
Concen: 82.826 ng/ul
RT: 16.663 min Scan# 2308
Delta R.T. 0.006 min
Lab File: BN023557.D
Acq: 04 Jan 2023 17:53

Tgt Ion:284 Resp: 980685
Ion Ratio Lower Upper
284 100
142 34.0 28.9 43.3
249 30.7 24.2 36.2









#70
 Atrazine
 Concen: 83.182 ng/ul
 RT: 16.827 min Scan# 21
 Delta R.T. 0.012 min
 Lab File: BN023557.D
 Acq: 04 Jan 2023 17:53




Instrument :
 BNA_N
 ClientSampleId :
 SSTD080221


Tgt Ion:200 Resp: 882957
 Ion Ratio Lower Upper
 200 100
 173 27.2 21.4 32.2
 215 50.3 39.0 58.6



#71
 Pentachlorophenol
 Concen: 86.015 ng/ul
 RT: 17.004 min Scan# 2366
 Delta R.T. 0.006 min
 Lab File: BN023557.D
 Acq: 04 Jan 2023 17:53

Tgt Ion:266 Resp: 672499
 Ion Ratio Lower Upper
 266 100
 264 64.7 51.4 77.0
 268 65.6 54.5 81.7









#72
 Phenanthrene
 Concen: 75.583 ng/ul
 RT: 17.398 min Scan# 2432
 Delta R.T. 0.006 min
 Lab File: BN023557.D
 Acq: 04 Jan 2023 17:53

Instrument :
 BNA_N
 ClientSampleId :
 SST080221

Tgt Ion:178 Resp: 4197641




Ion	Ratio	Lower	Upper
178	100		
179	15.8	12.3	18.5
176	19.0	15.2	22.8




#74
 Anthracene
 Concen: 75.391 ng/ul
 RT: 17.498 min Scan# 2450
 Delta R.T. 0.012 min
 Lab File: BN023557.D
 Acq: 04 Jan 2023 17:53

Tgt Ion:178 Resp: 4177673

Ion	Ratio	Lower	Upper
178	100		
179	15.5	12.6	19.0
176	19.0	15.1	22.7





#75

1,2,3,4-Tetrachlorobenzene

Concen: 89.728 ng/uL

RT: 13.410 min Scan# 11

Delta R.T. 0.006 min

Lab File: BN023557.D

Acq: 04 Jan 2023 17:53

Instrument :

BNA_N

ClientSampleId :

SSTD080221





Tgt Ion:216 Resp: 1171140

Ion Ratio Lower Upper

216 100

214 77.3 61.0 91.4

218 47.4 39.8 59.8



#76

Pentachlorobenzene

Concen: 85.396 ng/uL

RT: 14.928 min Scan# 2013

Delta R.T. 0.006 min

Lab File: BN023557.D

Acq: 04 Jan 2023 17:53




Tgt Ion:250 Resp: 1129460


Ion Ratio Lower Upper

250 100

248 62.3 50.3 75.5

252 64.0 48.9 73.3









#77
 Carbazole
 Concen: 77.117 ng/ul
 RT: 17.763 min Scan# 2494
 Delta R.T. 0.006 min
 Lab File: BN023557.D
 Acq: 04 Jan 2023 17:53




Instrument :
 BNA_N
 ClientSampleId :
 SST080221


Tgt Ion:167 Resp: 3935002
 Ion Ratio Lower Upper
 167 100
 166 21.6 17.4 26.2
 139 13.2 11.0 16.6



#78
 Di-n-butylphthalate
 Concen: 82.126 ng/ul
 RT: 18.327 min Scan# 2591
 Delta R.T. 0.006 min
 Lab File: BN023557.D
 Acq: 04 Jan 2023 17:53

Tgt Ion:149 Resp: 4975773
 Ion Ratio Lower Upper
 149 100
 150 9.2 7.4 11.2
 104 5.5 4.3 6.5









#80
Fluoranthene
Concen: 89.922 ng/ul
RT: 19.416 min Scan# 21
Delta R.T. 0.006 min
Lab File: BN023557.D
Acq: 04 Jan 2023 17:53




Instrument :
BNA_N
ClientSampleId :
SSTD080221


Tgt Ion:202 Resp: 4845140
Ion Ratio Lower Upper
202 100
101 12.4 9.5 14.3
100 9.5 7.3 10.9



#82
Pyrene
Concen: 84.809 ng/ul
RT: 19.786 min Scan# 2839
Delta R.T. 0.012 min
Lab File: BN023557.D
Acq: 04 Jan 2023 17:53

Tgt Ion:202 Resp: 4800267
Ion Ratio Lower Upper
202 100
101 13.2 10.9 16.3
100 11.0 8.9 13.3





#83

Butylbenzylphthalate

Concen: 97.620 ng/ul

RT: 20.674 min Scan# 2990

Delta R.T. -0.000 min

Lab File: BN023557.D

Acq: 04 Jan 2023 17:53

Instrument :

BNA_N

ClientSampleId :

SSTD080221





Tgt Ion:149 Resp: 2286544

Ion Ratio Lower Upper

149 100

91 73.4 57.9 86.9

206 20.7 16.8 25.2



#84

3,3'-Dichlorobenzidine

Concen: 82.029 ng/ul

RT: 21.462 min Scan# 3124

Delta R.T. -0.000 min

Lab File: BN023557.D

Acq: 04 Jan 2023 17:53




Tgt Ion:252 Resp: 1535224


Ion Ratio Lower Upper

252 100

254 65.7 51.2 76.8

126 12.9 8.8 13.2





#85

Benzo(a)anthracene

Concen: 80.653 ng/ul

RT: 21.527 min Scan# 3135

Delta R.T. -0.000 min

Lab File: BN023557.D

Acq: 04 Jan 2023 17:53

Instrument :

BNA_N

ClientSampleId :

SSTD080221





Tgt Ion:228 Resp: 4602059

Ion Ratio Lower Upper

228 100

229 20.7 15.0 22.4

226 27.4 22.1 33.1



#86

Bis(2-ethylhexyl)phthalate

Concen: 93.125 ng/ul

RT: 21.451 min Scan# 3122

Delta R.T. -0.000 min

Lab File: BN023557.D

Acq: 04 Jan 2023 17:53




Tgt Ion:149 Resp: 3084189


Ion Ratio Lower Upper

149 100

167 28.7 22.0 33.0

279 4.6 3.4 5.2





#87

Chrysene

Concen: 78.342 ng/ul

RT: 21.586 min Scan# 3144

Delta R.T. 0.006 min

Lab File: BN023557.D

Acq: 04 Jan 2023 17:53

Instrument :

BNA_N

ClientSampleId :

SSTD080221





Tgt Ion:228 Resp: 4218609

Ion Ratio Lower Upper

228 100

226 29.8 24.6 37.0

229 19.9 16.8 25.2



#89

Di-n-octyl phthalate

Concen: 97.529 ng/ul




RT: 22.386 min Scan# 3281


Delta R.T. 0.006 min

Lab File: BN023557.D

Acq: 04 Jan 2023 17:53

Tgt Ion:149 Resp: 5446082





#90

Benzo(b)fluoranthene

Concen: 80.872 ng/ul

RT: 23.233 min Scan# 3424

Delta R.T. 0.006 min

Lab File: BN023557.D

Acq: 04 Jan 2023 17:53

Instrument :

BNA_N

ClientSampleId :

SSTD080221





Tgt Ion:252 Resp: 4703452

Ion Ratio Lower Upper

252 100

253 22.6 17.2 25.8

125 11.2 8.6 13.0



#91

Benzo(k)fluoranthene

Concen: 81.306 ng/ul

RT: 23.292 min Scan# 3435

Delta R.T. 0.017 min

Lab File: BN023557.D

Acq: 04 Jan 2023 17:53




Tgt Ion:252 Resp: 4500105


Ion Ratio Lower Upper

252 100

253 22.1 17.8 26.8

125 10.1 8.6 12.8





#93

Benzo(a)pyrene

Concen: 82.611 ng/ul

RT: 23.874 min Scan# 3

Delta R.T. 0.017 min

Lab File: BN023557.D

Acq: 04 Jan 2023 17:53

Instrument :

BNA_N

ClientSampleId :

SSTD080221





Tgt Ion:252 Resp: 4129097

Ion Ratio Lower Upper

252 100

253 21.7 16.6 25.0

125 11.7 9.8 14.6



#94

Indeno(1,2,3-cd)pyrene

Concen: 56.847 ng/ul

RT: 26.521 min Scan# 3984

Delta R.T. 0.017 min

Lab File: BN023557.D

Acq: 04 Jan 2023 17:53




Tgt Ion:276 Resp: 3599544


Ion Ratio Lower Upper

276 100

138 26.6 20.6 30.8

277 25.3 18.9 28.3





#95

Dibenzo(a,h)anthracene

Concen: 78.939 ng/ul

RT: 26.539 min Scan# 3984

Delta R.T. 0.017 min

Lab File: BN023557.D

Acq: 04 Jan 2023 17:53

Instrument :

BNA_N

ClientSampleId :

SSTD080221





Tgt Ion:278 Resp: 4075469

Ion Ratio Lower Upper

278 100

139 17.2 14.1 21.1

279 22.0 18.0 27.0



#96

Benzo(g,h,i)perylene

Concen: 81.644 ng/ul

RT: 27.303 min Scan# 4117

Delta R.T. 0.017 min

Lab File: BN023557.D

Acq: 04 Jan 2023 17:53

Tgt Ion:276 Resp: 4090278

Ion Ratio Lower Upper

276 100

138 22.9 17.0 25.6

277 22.4 17.6 26.4

