


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010523\
 Data File : BN023560.D
 Acq On : 05 Jan 2023 09:01
 Operator : CG/JU
 Sample : DFTPP266
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP266

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN010523.M
 Title : SVOA CALIBRATION
 Last Update : Thu Jan 05 03:34:47 2023



AutoFind: Scans 2498, 2499, 2500; Background Corrected with Scan 2492

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.8	266048	PASS
68	69	0.00	2	1.7	5047	PASS
69	198	0.00	100	41.3	291072	PASS
70	69	0.00	2	0.5	1570	PASS
127	198	10	80	49.6	349312	PASS
197	198	0.00	2	0.1	1029	PASS
198	198	100	100	100.0	703936	PASS
199	198	5	9	6.7	47349	PASS
275	198	10	60	26.3	185048	PASS
365	198	1	100	3.7	26107	PASS
441	198	0.01	100	15.3	107952	PASS
442	198	50	100	89.7	631765	PASS
443	442	15	24	23.0	145061	PASS

m/z	Abundance
39.90	674.0
43.00	174.0
43.90	737.0
44.90	175.0
48.90	175.0
65.90	178.0
73.00	1071.0
74.90	269.0
76.90	469.0
78.00	201.0
83.80	202.0
96.00	637.0
102.90	194.0
132.90	578.0
134.90	202.0
146.90	572.0
167.00	342.0
176.90	234.0
190.90	633.0
193.00	409.0
207.00	5500.0
208.00	1125.0
209.00	730.0
220.90	550.0
266.90	212.0
281.00	1083.0
281.90	227.0
282.90	178.0
340.90	170.0
355.00	368.0

Instrument :
BNA_N
ClientSampleId :
DFTPP266

m/z	Abundance
35.90	278.0
37.00	1312.0
38.00	3461.0
39.00	17864.0
40.00	1131.0
40.90	587.0
42.90	167.0
43.90	605.0
44.90	471.0
48.00	159.0
49.00	1747.0
50.00	61208.0
51.00	228288.0
52.00	12133.0
53.00	512.0
54.90	1310.0
56.00	6543.0
57.00	14764.0
57.90	689.0
59.00	241.0
60.00	198.0
61.00	3294.0
62.00	3499.0
63.00	9700.0
64.00	1496.0
65.00	4415.0
65.90	414.0
66.90	347.0
68.00	3988.0
68.90	246272.0
69.90	1371.0
73.00	2783.0
74.00	27312.0
75.00	40656.0
76.00	13707.0
77.00	280064.0
78.00	20768.0
79.00	18248.0
80.00	14449.0
81.00	19000.0
82.00	4899.0
83.00	4078.0
83.90	574.0
85.00	3298.0
86.00	6219.0
87.00	2621.0
87.90	1023.0
89.00	511.0
91.00	3904.0
92.00	5339.0
93.00	33232.0
93.90	2330.0
94.90	460.0
96.00	1774.0
97.10	583.0
98.00	26136.0
99.00	20936.0
99.90	1866.0
101.00	10482.0
101.90	602.0
103.00	3780.0
104.00	6662.0
105.00	6341.0
106.00	2017.0
107.00	80472.0
108.00	12572.0
109.00	2359.0
110.00	143936.0
111.00	23160.0
112.00	2968.0
112.90	701.0
114.20	165.0
114.90	415.0
116.00	4597.0
117.00	79648.0
118.00	6626.0
119.00	789.0
120.00	1015.0
121.00	363.0
121.90	5645.0
122.90	8417.0
123.90	3808.0
124.90	3596.0
127.00	278720.0
128.00	21320.0
129.00	106768.0
129.90	9163.0
131.00	1977.0
131.90	953.0
132.90	700.0
134.00	3174.0
135.00	8711.0

Instrument :
BNA_N
ClientSampleId :
DFTPP266

136.00	3561.0
137.00	5389.0
138.00	1079.0
139.00	631.0
140.00	1281.0
140.90	13752.0
142.00	4626.0
142.90	3103.0
143.90	858.0
145.00	706.0
146.00	2651.0
146.90	7280.0
147.90	18656.0
149.00	4152.0
150.00	1269.0
151.00	1948.0
152.90	4189.0
153.90	3280.0
155.00	7468.0
156.00	12392.0
157.00	2169.0
157.90	2629.0
159.00	1933.0
160.00	4432.0
161.00	6484.0
161.90	2012.0
163.00	608.0
164.10	741.0
164.90	5416.0
165.90	3865.0
167.00	30480.0
167.90	20200.0
169.00	2782.0
170.00	842.0
170.90	1240.0
171.90	2446.0
173.00	3172.0
174.00	5837.0
175.00	11258.0
176.00	3050.0
176.90	5096.0
177.90	1647.0
179.00	20584.0
179.90	16357.0
180.90	6484.0
182.00	1086.0
182.90	500.0
183.90	1548.0
185.00	9875.0
186.00	75208.0
187.00	23120.0
188.00	2868.0
188.90	4979.0
189.90	782.0
191.00	2886.0
192.00	6327.0
193.00	7359.0
193.90	1582.0
194.90	862.0
196.00	14596.0
197.00	3088.0
197.90	504128.0
198.90	35232.0
200.00	3009.0
201.40	2122.0
203.00	3914.0
204.00	18704.0
205.00	32456.0
206.00	135296.0
207.00	24128.0
208.00	6169.0
209.00	2120.0
210.00	2219.0
211.00	5936.0
211.80	1116.0
212.90	404.0
213.90	201.0
214.90	1613.0
216.00	3075.0
216.90	36664.0
218.00	4758.0
218.80	403.0
221.00	21904.0
221.70	5435.0
223.00	7777.0
224.00	75192.0
225.00	20344.0
226.00	2836.0
226.90	31480.0
228.00	4526.0
229.00	6713.0
229.90	840.0
231.00	2893.0
231.90	479.0
232.90	559.0
234.00	1997.0

Instrument :
BNA_N
ClientSampleId :
DFTPP266

235.00	2080.0
236.00	1461.0
237.00	4295.0
237.90	613.0
239.00	1048.0
239.90	1128.0
241.00	1527.0
241.90	3915.0
242.90	4514.0
244.00	58608.0
245.00	7801.0
245.90	11459.0
246.90	2315.0
247.90	547.0
248.90	1952.0
249.90	467.0
250.90	481.0
251.90	626.0
253.00	921.0
254.00	1727.0
255.00	286848.0
256.00	43400.0
257.00	3976.0
258.00	16504.0
259.00	2960.0
259.90	523.0
260.90	510.0
263.00	236.0
264.00	490.0
264.90	6841.0
265.90	1119.0
266.90	365.0
269.90	388.0
270.90	728.0
271.90	745.0
273.00	8649.0
274.00	21432.0
275.00	130056.0
276.00	17064.0
276.90	10959.0
277.90	2537.0
279.00	470.0
281.10	1006.0
281.90	497.0
282.90	1377.0
283.90	1017.0
285.00	2055.0
285.90	360.0
288.90	434.0
289.90	284.0
290.80	266.0
292.00	487.0
293.00	2478.0
293.90	886.0
295.00	821.0
296.00	36288.0
296.90	5716.0
298.00	387.0
301.00	551.0
301.90	631.0
303.00	4028.0
304.00	1057.0
304.90	275.0
307.90	495.0
309.00	464.0
309.90	558.0
312.90	390.0
313.90	1675.0
314.90	3856.0
316.00	2292.0
316.90	552.0
320.90	1050.0
322.00	702.0
323.00	10626.0
324.10	2078.0
325.00	248.0
325.90	189.0
326.90	2215.0
327.90	1318.0
329.00	193.0
332.00	760.0
333.00	1141.0
334.00	7624.0
335.00	1911.0
335.90	233.0
339.90	160.0
340.90	1414.0
341.90	460.0
345.80	2317.0
346.90	617.0
351.00	279.0
352.00	3506.0
352.90	2434.0
353.90	3367.0
355.00	1093.0
355.90	186.0

Instrument :
BNA_N
ClientSampleId :
DFTPP266

358.90	296.0
364.90	15897.0
365.90	2344.0
366.90	159.0
369.90	330.0
371.00	965.0
371.90	5804.0
373.00	1377.0
373.90	162.0
382.90	1423.0
384.00	521.0
384.90	235.0
389.90	835.0
390.90	468.0
391.80	335.0
400.90	410.0
401.90	2397.0
402.90	2822.0
403.90	1283.0
404.90	492.0
405.90	176.0
414.80	291.0
421.00	2646.0
422.00	2880.0
423.00	20304.0
424.00	8839.0
425.00	1439.0
439.80	196.0
440.90	56616.0
442.00	351872.0
443.00	78256.0
444.00	6894.0
445.00	545.0

Instrument :
BNA_N
ClientSampleId :
DFTPP266

m/z	Abundance
35.00	300.0
37.00	1515.0
38.00	4296.0
39.00	23280.0
40.00	1379.0
40.90	718.0
42.90	219.0
43.90	711.0
45.00	657.0
47.90	241.0
49.00	2590.0
50.00	81512.0
51.00	298368.0
52.00	14912.0
53.00	634.0
55.00	1641.0
56.00	8645.0
57.00	19512.0
57.90	861.0
58.90	434.0
60.00	384.0
61.00	3861.0
62.00	4931.0
63.00	13139.0
64.00	1943.0
65.00	6190.0
66.00	540.0
66.90	507.0
68.00	5990.0
69.00	327360.0
70.00	1783.0
71.00	177.0
72.00	268.0
73.00	3386.0
74.00	36880.0
75.00	55600.0
76.00	19736.0
77.00	383744.0
78.00	26784.0
79.00	27016.0
80.00	19320.0
81.00	27216.0
81.90	6881.0
83.00	5868.0
84.00	612.0
84.90	5137.0
86.00	8056.0
87.00	3570.0
87.90	1380.0
89.00	652.0
91.00	5785.0
92.00	7473.0
93.00	47032.0
94.00	3217.0
95.00	709.0
96.00	2245.0
98.00	36640.0
99.00	28048.0
100.00	2146.0
101.00	14692.0
101.90	621.0
102.90	4494.0
103.90	9572.0
105.00	9153.0
106.00	2994.0
107.00	118632.0
108.00	16880.0
109.00	3599.0
110.00	202560.0
111.00	31496.0
112.00	3858.0
113.00	1255.0
115.00	515.0
116.00	6376.0
117.00	109096.0
118.00	7903.0
119.00	1053.0
120.00	1828.0
121.00	520.0
121.90	7451.0
122.90	11616.0
124.00	5596.0
124.90	4735.0
127.00	396928.0
128.00	30528.0
129.00	153984.0
130.00	13113.0
131.00	2478.0
132.00	1401.0
133.00	1071.0
134.00	4614.0
135.00	13122.0

Instrument :
BNA_N
ClientSampleId :
DFTPP266

136.00	5151.0
137.00	7227.0
137.90	1502.0
138.90	818.0
139.90	1990.0
140.90	21096.0
142.00	7296.0
142.90	4462.0
144.00	1133.0
145.00	1122.0
146.00	4175.0
147.00	10849.0
148.00	27248.0
149.00	5344.0
150.00	1502.0
151.00	2773.0
152.00	1710.0
153.00	6579.0
153.90	4594.0
155.00	12334.0
156.00	17488.0
157.00	3412.0
157.90	3708.0
158.90	2914.0
160.00	6717.0
161.00	10358.0
161.90	2728.0
163.00	891.0
164.00	1240.0
165.00	7959.0
166.00	6057.0
167.00	45088.0
168.00	25328.0
169.00	3782.0
170.00	1334.0
170.90	1944.0
172.00	3999.0
173.00	5102.0
174.00	8720.0
175.00	16640.0
176.00	4751.0
177.00	7603.0
177.90	2679.0
179.00	31760.0
180.00	23808.0
181.00	10243.0
182.00	1653.0
182.90	672.0
183.90	2490.0
185.00	15222.0
186.00	116256.0
187.00	33928.0
188.00	3738.0
189.00	6741.0
189.90	1250.0
191.00	3868.0
192.00	9436.0
193.00	11955.0
194.00	2536.0
195.00	1220.0
196.00	24088.0
197.90	825792.0
198.90	55720.0
200.00	4560.0
201.60	3224.0
203.00	5933.0
204.00	30152.0
205.00	51544.0
206.00	219200.0
207.00	32384.0
208.00	8854.0
208.90	2999.0
210.10	3496.0
211.00	8568.0
213.00	755.0
214.00	350.0
215.00	2569.0
216.00	4712.0
216.90	56448.0
218.00	7947.0
219.00	789.0
221.00	31280.0
221.80	9110.0
223.00	13691.0
224.00	124136.0
225.00	31336.0
226.00	3831.0
227.00	50856.0
227.90	8272.0
229.00	10968.0
229.90	1498.0
231.00	4296.0
232.00	876.0
232.90	775.0
234.00	3320.0
234.90	3809.0

Instrument :
BNA_N
ClientSampleId :
DFTPP266

235.90	2466.0
237.00	5667.0
238.00	794.0
239.00	1921.0
240.00	2124.0
241.00	2698.0
242.00	6010.0
243.00	6941.0
244.00	92664.0
245.00	12861.0
246.00	20056.0
246.90	3821.0
248.00	963.0
248.90	3785.0
250.00	648.0
250.90	950.0
251.90	909.0
252.90	1991.0
254.00	3963.0
255.00	474432.0
256.00	74528.0
257.00	5690.0
258.00	28528.0
258.90	4258.0
259.90	930.0
260.90	753.0
262.80	349.0
264.00	848.0
265.00	11840.0
266.00	1723.0
267.00	358.0
267.90	196.0
269.90	573.0
271.00	1122.0
271.90	1123.0
273.00	14841.0
274.00	40520.0
275.10	204032.0
276.00	30656.0
277.00	19808.0
278.00	3614.0
279.00	643.0
280.90	837.0
282.00	639.0
283.00	2397.0
284.00	1520.0
284.90	3040.0
285.80	572.0
288.00	216.0
288.90	646.0
289.90	508.0
290.80	484.0
292.00	886.0
293.00	4606.0
293.90	1304.0
296.00	65568.0
297.00	8674.0
297.90	539.0
301.00	937.0
302.00	975.0
303.00	6734.0
303.90	1801.0
305.00	393.0
307.90	816.0
308.90	646.0
309.90	914.0
310.80	186.0
311.90	247.0
312.90	544.0
313.90	2737.0
314.90	6494.0
316.00	4283.0
317.00	566.0
320.00	269.0
320.90	2334.0
322.10	1088.0
323.00	19128.0
324.00	3640.0
325.00	338.0
326.00	427.0
327.00	3672.0
327.90	2020.0
328.90	357.0
330.90	171.0
332.00	1534.0
332.90	1965.0
334.00	13706.0
335.00	3398.0
336.00	404.0
339.00	394.0
339.90	354.0
340.90	2848.0
342.00	743.0
342.80	190.0
346.00	4950.0
347.00	918.0

Instrument :
BNA_N
ClientSampleId :
DFTPP266

350.00	224.0
351.00	357.0
352.00	6341.0
353.00	4594.0
354.00	6353.0
355.00	1570.0
355.90	212.0
357.10	218.0
358.90	390.0
364.90	27592.0
365.90	4385.0
367.00	289.0
370.00	713.0
370.90	1731.0
372.00	10777.0
373.00	2622.0
374.00	344.0
376.90	316.0
383.00	3182.0
384.00	953.0
384.70	294.0
389.90	1461.0
391.00	890.0
392.00	597.0
397.00	195.0
400.90	723.0
401.90	4405.0
403.00	6646.0
403.90	2371.0
404.90	422.0
414.90	264.0
420.90	4899.0
421.90	6073.0
423.00	38144.0
424.00	13059.0
424.90	1603.0
438.90	192.0
440.90	110888.0
442.00	676736.0
443.00	156800.0
444.00	15900.0
445.00	806.0

Instrument :
BNA_N
ClientSampleId :
DFTPP266

m/z	Abundance
36.00	285.0
37.00	1537.0
38.00	3752.0
39.00	20632.0
40.00	1149.0
41.00	615.0
42.90	196.0
43.90	856.0
45.00	657.0
47.90	192.0
49.10	2294.0
50.00	71736.0
51.00	271488.0
52.00	13675.0
53.00	508.0
55.00	1454.0
56.00	8072.0
57.00	17904.0
58.00	907.0
59.00	297.0
59.90	252.0
61.00	3751.0
62.00	4187.0
63.00	11464.0
64.00	1704.0
65.00	5392.0
66.00	436.0
66.90	489.0
68.00	5162.0
69.00	299584.0
69.90	1557.0
70.90	245.0
72.00	220.0
73.00	3600.0
74.00	32304.0
75.00	48288.0
76.00	17928.0
77.00	343744.0
78.00	23992.0
79.00	23576.0
80.00	18144.0
81.00	24464.0
82.00	6096.0
82.90	4934.0
84.00	670.0
85.00	4998.0
86.00	7308.0
86.90	3225.0
88.00	1330.0
88.90	621.0
91.00	5199.0
92.00	6876.0
93.00	41208.0
93.90	3170.0
94.90	625.0
96.00	2228.0
96.80	603.0
98.00	33440.0
99.00	25368.0
99.90	2065.0
101.00	14064.0
102.00	707.0
102.90	4401.0
104.00	8853.0
105.00	8571.0
106.00	2843.0
107.00	108792.0
108.00	16912.0
109.00	3079.0
110.00	183680.0
111.00	28232.0
112.00	3846.0
112.90	1233.0
114.00	239.0
114.70	390.0
115.10	453.0
116.00	6297.0
117.00	101832.0
118.00	7616.0
118.90	909.0
120.00	1365.0
121.10	481.0
121.90	6533.0
123.00	11659.0
124.00	5577.0
125.00	4649.0
126.10	2115.0
127.00	372288.0
128.00	28616.0
129.00	143360.0
130.00	11678.0
131.00	2431.0

Instrument :
BNA_N
ClientSampleId :
DFTPP266

131.90	1289.0
132.90	1078.0
133.90	4328.0
134.90	12060.0
135.90	4882.0
137.00	6930.0
137.80	1255.0
137.90	1251.0
138.90	833.0
140.00	2034.0
141.00	20072.0
142.00	6132.0
143.00	4402.0
144.00	1196.0
145.00	935.0
145.90	3559.0
147.00	11084.0
148.00	26112.0
149.00	5212.0
149.90	1213.0
151.00	2564.0
151.70	1593.0
152.90	6523.0
153.90	4914.0
155.00	10785.0
156.00	16247.0
157.00	3139.0
157.90	3513.0
159.00	2718.0
160.00	5942.0
161.00	9538.0
162.00	2711.0
163.00	927.0
164.00	1173.0
164.90	7711.0
166.00	5869.0
167.00	42712.0
168.00	23640.0
169.00	3701.0
170.00	1706.0
171.00	1662.0
171.90	3892.0
173.00	4494.0
174.00	8280.0
175.00	16472.0
176.00	4440.0
176.90	7674.0
178.00	3199.0
178.90	30024.0
180.00	20040.0
181.00	9684.0
182.00	1345.0
183.00	738.0
184.00	2452.0
185.00	14431.0
186.00	112752.0
187.00	32384.0
188.00	3842.0
188.90	7417.0
189.90	1099.0
191.00	3695.0
191.90	9128.0
193.00	11654.0
194.00	2265.0
195.00	1279.0
196.00	22072.0
197.90	781888.0
198.90	51096.0
200.00	4259.0
201.40	3451.0
203.00	5885.0
204.00	30008.0
205.00	50560.0
206.00	222080.0
207.00	33000.0
208.00	8013.0
209.00	2704.0
210.00	3298.0
211.00	8229.0
213.00	639.0
213.80	194.0
214.10	203.0
215.00	2664.0
216.00	4644.0
217.00	56832.0
218.00	7698.0
219.00	800.0
221.00	32424.0
221.80	8534.0
223.00	13231.0
224.00	119456.0
225.00	31928.0
226.10	3205.0
226.90	51936.0
228.00	7619.0
229.00	11118.0

Instrument :
BNA_N
ClientSampleId :
DFTPP266

230.00	1380.0
230.90	3668.0
231.90	893.0
233.00	955.0
233.90	3520.0
234.90	4021.0
236.00	2678.0
237.00	5057.0
238.00	656.0
238.90	2020.0
239.90	1520.0
240.90	3078.0
242.00	6419.0
243.00	6667.0
244.00	98080.0
245.00	12053.0
246.00	19128.0
246.90	3884.0
247.90	720.0
249.00	3488.0
249.90	680.0
251.00	869.0
251.90	814.0
252.90	2022.0
255.00	502400.0
256.00	72456.0
257.00	5641.0
257.90	29648.0
259.00	4876.0
259.90	957.0
261.00	925.0
261.80	196.0
263.00	382.0
264.00	818.0
265.00	12264.0
265.90	1569.0
267.00	380.0
267.90	206.0
268.90	177.0
269.90	664.0
271.00	1005.0
271.90	1156.0
273.00	14743.0
274.00	42480.0
274.90	221056.0
276.00	31280.0
277.00	20240.0
278.00	3430.0
279.00	1028.0
280.90	1045.0
281.90	648.0
282.90	2473.0
284.00	1423.0
285.10	3494.0
286.00	574.0
287.90	249.0
288.80	870.0
290.00	630.0
290.90	475.0
292.10	949.0
293.00	4466.0
293.90	1211.0
294.90	1632.0
296.00	68192.0
297.00	9568.0
297.80	489.0
301.00	952.0
302.00	1434.0
302.90	7849.0
303.90	1825.0
305.00	411.0
307.90	1003.0
309.00	681.0
310.00	898.0
310.90	208.0
311.90	158.0
313.00	661.0
314.00	3253.0
314.90	7478.0
316.00	4205.0
317.00	793.0
319.90	237.0
321.00	1939.0
322.10	1148.0
323.00	20936.0
324.00	4033.0
325.00	415.0
326.00	528.0
326.90	4186.0
328.00	2279.0
329.00	345.0
331.10	161.0
331.90	1739.0
333.00	2505.0
334.00	14783.0
335.10	3851.0

Instrument :
BNA_N
ClientSampleId :
DFTPP266

335.90	491.0
338.80	377.0
339.90	376.0
340.90	2683.0
342.00	844.0
345.80	4823.0
346.90	974.0
349.90	211.0
351.00	465.0
352.00	7713.0
353.00	5111.0
353.90	6549.0
355.00	1740.0
355.90	224.0
357.00	153.0
358.00	178.0
359.00	536.0
364.90	34832.0
366.00	4887.0
367.00	367.0
370.00	735.0
371.00	1819.0
372.00	12456.0
373.00	2925.0
374.00	347.0
376.80	409.0
383.00	4152.0
384.00	1135.0
384.80	316.0
390.00	1478.0
391.00	1191.0
392.00	788.0
396.90	200.0
400.90	791.0
401.90	5148.0
402.90	7462.0
404.00	2581.0
404.90	607.0
409.90	205.0
414.90	371.0
421.00	7304.0
421.90	6421.0
423.00	49928.0
424.00	13474.0
424.90	1457.0
428.90	177.0
439.20	249.0
441.00	156352.0
442.00	866688.0
443.00	200128.0
444.00	15939.0
445.00	928.0

Instrument :
BNA_N
ClientSampleId :
DFTPP266