

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010722\
 Data File : BN018368.D
 Acq On : 07 Jan 2022 14:02
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN010722

Quant Time: Jan 07 14:42:46 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN010722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 07 14:23:23 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
2	1,4-Dioxane	0.225	0.221	1.8	103	0.00
3	n-Nitrosodimethylamine	0.361	0.366	-1.4	110	0.00
4 S	2-Fluorophenol	0.984	0.994	-1.0	108	0.00
5 S	Phenol-d6	0.986	0.988	-0.2	109	0.00
6	bis(2-Chloroethyl)ether	1.068	1.086	-1.7	105	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	103	0.00
8 S	Nitrobenzene-d5	0.255	0.226	11.4	111	0.00
9	Naphthalene	1.016	1.012	0.4	102	0.00
10	Hexachlorobutadiene	0.189	0.188	0.5	103	0.00
11 SURR	2-Methylnaphthalene-d10	0.626	0.622	0.6	103	0.00
12	2-Methylnaphthalene	0.622	0.632	-1.6	103	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	103	0.00
14 S	2,4,6-Tribromophenol	0.158	0.123	22.2	113	0.00
15 S	2-Fluorobiphenyl	1.444	1.461	-1.2	103	0.00
16	Acenaphthylene	1.588	1.534	3.4	106	0.01
17	Acenaphthene	1.109	1.101	0.7	102	0.00
18	Fluorene	1.246	1.249	-0.2	104	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	105	0.00
20	4,6-Dinitro-2-methylphenol	0.028	0.013	53.6#	124	-0.01
21	4-Bromophenyl-phenylether	0.220	0.214	2.7	106	0.00
22	Hexachlorobenzene	0.293	0.289	1.4	103	0.00
23	Atrazine	0.146	0.128	12.3	116	0.00
24	Pentachlorophenol	0.071	0.044	38.0#	133	0.00
25	Phenanthrene	1.124	1.133	-0.8	106	0.00
26	Anthracene	0.929	0.887	4.5	107	0.00
27 SURR	Fluoranthene-d10	1.142	1.124	1.6	106	0.00
28	Fluoranthene	1.194	1.217	-1.9	105	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	107	0.00
30	Pyrene	1.416	1.414	0.1	106	0.00
31 S	Terphenyl-d14	0.866	0.869	-0.3	107	0.00
32	Benzo(a)anthracene	1.178	1.175	0.3	118	0.00
33	Chrysene	1.338	1.288	3.7	100	0.00
34	Bis(2-ethylhexyl)phthalate	0.707	0.536	24.2	118	0.00
35 I	Perylene-d12	1.000	1.000	0.0	107	0.00
36	Indeno(1,2,3-cd)pyrene	1.271	1.189	6.5	111	0.00
37	Benzo(b)fluoranthene	1.297	1.304	-0.5	105	0.00
38	Benzo(k)fluoranthene	1.268	1.316	-3.8	111	0.00
39 C	Benzo(a)pyrene	1.199	1.174	2.1	108	0.00
40	Dibenzo(a,h)anthracene	0.951	0.876	7.9	115	0.00
41	Benzo(g,h,i)perylene	1.163	1.115	4.1	107	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
ClientSampleId :
ICVBN010722