

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN010820\
 Data File : BN009327.D
 Acq On : 08 Jan 2020 12:29
 Operator : JU
 Sample : K6467-02
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :

Quant Time: Jan 08 13:13:19 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM01-EPA-BN010620.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jan 06 17:39:39 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.47	152	235781	20.00	ng/ul	-0.01
18) Naphthalene-d8	10.23	136	986340	20.00	ng/ul	-0.01
36) Acenaphthene-d10	14.11	164	643731	20.00	ng/ul	0.00
62) Phenanthrene-d10	16.86	188	1392854	20.00	ng/ul	0.00
78) Chrysene-d12	21.07	240	1271881	20.00	ng/ul	0.00
86) Perylene-d12	23.19	264	1233866	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.10	96	42627	7.65	ng/uL	0.00
5) Phenol-d5	6.67	99	889156	42.35	ng/ul	-0.01
7) Bis-(2-Chloroethyl)ether-d	6.82	67	621212	43.00	ng/ul	-0.01
9) 2-Chlorophenol-d4	7.02	132	662860	43.27	ng/ul	0.00
13) 4-Methylphenol-d8	8.18	113	736606	42.53	ng/ul	-0.01
19) Nitrobenzene-d5	8.62	128	345492	47.80	ng/ul	0.00
22) 2-Nitrophenol-d4	9.32	143	388181	48.18	ng/ul	-0.02
26) 2,4-Dichlorophenol-d3	9.86	165	716086	46.54	ng/ul	-0.01
29) 4-Chloroaniline-d4	10.37	131	958376	53.67	ng/ul	-0.01
44) Dimethylphthalate-d6	13.53	166	2266889	47.65	ng/ul	0.00
47) Acenaphthylene-d8	13.80	160	2667403	46.87	ng/ul	0.00
52) 4-Nitrophenol-d4	14.34	143	418528	46.45	ng/ul	0.00
58) Fluorene-d10	15.12	176	1963983	47.44	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.25	200	401368	44.85	ng/ul	0.00
71) Anthracene-d10	16.96	188	3033346	47.97	ng/ul	0.00
79) Pyrene-d10	19.27	212	3349975	48.45	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.06	264	3060171	49.76	ng/ul	0.00

Target Compounds

					Ovalue	
6) Phenol	6.70	94	507117	23.114	ng/ul	99
10) 2-Chlorophenol	7.05	128	378011	24.152	ng/ul	98
12) 2,2'-oxybis(1-Chloropropan	8.00	45	1369651	32.865	ng/ul	99
15) N-Nitroso-di-n-propylamine	8.18	70	20667	1.357	ng/ul#	1
17) Hexachloroethane	8.53	117	163435	22.838	ng/ul	99
27) 2,4-Dichlorophenol	9.89	162	394022	26.035	ng/ul	96
28) Naphthalene	10.28	128	1288981	24.611	ng/ul	99
31) Hexachlorobutadiene	10.57	225	354424	36.034	ng/ul	97
46) 2,6-Dinitrotoluene	13.70	165	368313	37.539	ng/ul	98
50) Acenaphthene	14.17	153	1028964	25.181	ng/ul	96
55) 2,4-Dinitrotoluene	14.49	165	545544	38.180	ng/ul	98
57) Diethylphthalate	14.96	149	1570259	30.864	ng/ul	99
59) Fluorene	15.17	166	1217920	25.572	ng/ul	99
61) 4-Nitroaniline	15.17	138	13859	1.396	ng/ul#	16
67) Hexachlorobenzene	16.17	284	482791	30.278	ng/ul	95
69) Pentachlorophenol	16.53	266	502860	52.241	ng/ul	95
70) Phenanthrene	17.00	178	1724542	22.402	ng/ul	99
72) Anthracene	17.00	178	1724542	22.158	ng/ul	99
80) Pyrene	19.30	202	2354617	25.698	ng/ul	99
81) Butylbenzylphthalate	20.23	149	2148052	55.903	ng/ul	99
83) Benzo(a)anthracene	21.06	228	2156584	24.939	ng/ul	98
84) Bis(2-ethylhexyl)phthalate	21.02	149	2473032	42.668	ng/ul	99
85) Chrysene	21.11	228	2146189	25.505	ng/ul	98

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN010820\
Data File : BN009327.D
Acq On : 08 Jan 2020 12:29
Operator : JU
Sample : K6467-02
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :

Quant Time: Jan 08 13:13:19 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM01-EPA-BN010620.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Jan 06 17:39:39 2020
Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
88) Benzo(b)fluoranthene	22.57	252	3343169	42.987	ng/ul	99
89) Benzo(k)fluoranthene	22.61	252	1876147	24.430	ng/ul	98
91) Benzo(a)pyrene	23.10	252	1732093	25.101	ng/ul	97
92) Indeno(1,2,3-cd)pyrene	25.28	276	2682613	35.875	ng/ul	97
93) Dibenzo(a,h)anthracene	25.29	278	1971979	31.390	ng/ul	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN010820\
Data File : BN009327.D
Acq On : 08 Jan 2020 12:29
Operator : JU
Sample : K6467-02
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampled :

Quant Time: Jan 08 13:13:19 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM01-EPA-BN010620.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Jan 06 17:39:39 2020
Response via : Initial Calibration









#6

Phenol

Concen: 23.114 ng/ul
 RT: 6.70 min Scan# 631
 Delta R.T. -0.01 min
 Lab File: BN009327.D
 Acq: 08 Jan 2020 12:29

Instrument :
 BNA_N
 ClientSampled :

Tot Ion	94	Resp	507117
Ion	Ratio	Lower	Upper
94	100		
65	34.1	28.6	42.8
66	47.0	37.7	56.5







#10

2-Chlorophenol

Concen: 24.152 ng/ul
 RT: 7.05 min Scan# 690
 Delta R.T. -0.01 min
 Lab File: BN009327.D
 Acq: 08 Jan 2020 12:29

Tgt Ion	128	Resp	378011
Ion	Ratio	Lower	Upper
128	100		
64	66.5	52.3	78.5
130	32.5	27.4	41.0









#12
2,2'-oxvbis(1-Chloropropane)
Concen: 32.865 ng/ul
RT: 8.00 min Scan# 853
Delta R.T. -0.01 min
Lab File: BN009327.D
Acq: 08 Jan 2020 12:29

Instrument :
BNA_N
ClientSampleId :

Tot Ion: 45 Resp: 1369651
Ion Ratio Lower Upper
45 100
77 9.2 7.0 10.6
79 7.1 5.2 7.8




Abundance Ion 45.00 (44.70 to 45.70): BNC
Ion 77.00 (76.70 to 77.70): BNC
Ion 79.00 (78.70 to 79.70): BNC






#15
N-Nitroso-di-n-propylamine
Concen: 1.357 ng/ul
RT: 8.18 min Scan# 883
Delta R.T. -0.11 min
Lab File: BN009327.D
Acq: 08 Jan 2020 12:29

Tgt Ion: 70 Resp: 20667
Ion Ratio Lower Upper
70 100
42 562.8 64.6 97.0#
101 0.0 7.6 11.4#
130 0.0 14.5 21.7#



Abundance Ion 70.00 (69.70 to 70.70): BNC
Ion 42.00 (41.70 to 42.70): BNC
Ion 101.00 (100.70 to 101.70): BNC
Ion 130.00 (129.70 to 130.70): BNC









#17
Hexachloroethane
Concen: 22.838 ng/ul
RT: 8.53 min Scan# 942
Delta R.T. -0.01 min
Lab File: BN009327.D
Acq: 08 Jan 2020 12:29

Instrument :
BNA_N
ClientSampleId :

Tot Ion:117 Resp: 163435
Ion Ratio Lower Upper
117 100
201 96.2 78.4 117.6
199 62.6 50.2 75.4




Abundance Ion 117.00 (116.70 to 117.70): E
Ion 201.00 (200.70 to 201.70): E
Ion 199.00 (198.70 to 199.70): E






#27
2,4-Dichlorophenol
Concen: 26.035 ng/ul
RT: 9.89 min Scan# 1173
Delta R.T. -0.01 min
Lab File: BN009327.D
Acq: 08 Jan 2020 12:29

Tgt Ion:162 Resp: 394022
Ion Ratio Lower Upper
162 100
164 66.1 50.6 75.8
98 41.1 31.2 46.8



Abundance Ion 162.00 (161.70 to 162.70): E
Ion 164.00 (163.70 to 164.70): E
Ion 98.00 (97.70 to 98.70): B





#28

Naphthalene

Concen: 24.611 ng/ul

RT: 10.28 min Scan# 1239

Delta R.T. -0.01 min

Lab File: BN009327.D

Acq: 08 Jan 2020 12:29

Instrument :

BNA_N

ClientSampled :


Tot Ion:128 Resp: 1288981

Ion Ratio Lower Upper




128 100

129 11.0 9.0 13.4

127 12.7 10.8 16.2



Abundance Ion 128.00 (127.70 to 128.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 127.00 (126.70 to 127.70): E



#31

Hexachlorobutadiene

Concen: 36.034 ng/ul

RT: 10.57 min Scan# 1289

Delta R.T. -0.01 min

Lab File: BN009327.D

Acq: 08 Jan 2020 12:29


Tgt Ion:225 Resp: 354424

Ion Ratio Lower Upper



225 100


223 62.6 48.2 72.4

227 62.0 51.0 76.6



Abundance Ion 225.00 (224.70 to 225.70): E
Ion 223.00 (222.70 to 223.70): E
Ion 227.00 (226.70 to 227.70): E









#46
 2,6-Dinitrotoluene
 Concen: 37.539 ng/ul
 RT: 13.70 min Scan# 1821
 Delta R.T. -0.01 min
 Lab File: BN009327.D
 Acq: 08 Jan 2020 12:29




Instrument :
 BNA_N
 ClientSampleId :


Tot Ion	Ratio	Lower	Upper
165	100		
89	64.2	50.5	75.7
121	21.0	17.8	26.8



#50
 Acenaphthene
 Concen: 25.181 ng/ul
 RT: 14.17 min Scan# 1902
 Delta R.T. -0.01 min
 Lab File: BN009327.D
 Acq: 08 Jan 2020 12:29

Tgt Ion	Ratio	Lower	Upper
153	100		
152	48.4	37.2	55.8
154	91.6	69.8	104.6









#55
2,4-Dinitrotoluene
Concen: 38.180 ng/ul
RT: 14.49 min Scan# 1956
Delta R.T. -0.01 min
Lab File: BN009327.D
Acq: 08 Jan 2020 12:29

Instrument :
BNA_N
ClientSampleId :

Tot Ion:165 Resp: 545544
Ion Ratio Lower Upper
165 100
63 60.8 49.8 74.6
182 3.1 2.2 3.4




Abundance Ion 165.00 (164.70 to 165.70): E
Ion 63.00 (62.70 to 63.70): BNO
Ion 182.00 (181.70 to 182.70): E






#57
Diethylphthalate
Concen: 30.864 ng/ul
RT: 14.96 min Scan# 2036
Delta R.T. -0.01 min
Lab File: BN009327.D
Acq: 08 Jan 2020 12:29

Tgt Ion:149 Resp: 1570259
Ion Ratio Lower Upper
149 100
177 21.6 17.1 25.7
150 12.5 9.8 14.6



Abundance Ion 149.00 (148.70 to 149.70): E
Ion 177.00 (176.70 to 177.70): E
Ion 150.00 (149.70 to 150.70): E









#59
Fluorene
 Concen: 25.572 ng/ul
 RT: 15.17 min Scan# 2071
 Delta R.T. -0.01 min
 Lab File: BN009327.D
 Acq: 08 Jan 2020 12:29

Instrument :
 BNA_N
 ClientSampleId :

Tot Ion:166 Resp: 1217920
 Ion Ratio Lower Upper
 166 100
 165 92.8 75.0 112.6
 167 12.4 10.5 15.7




Abundance Ion 166.00 (165.70 to 166.70): E
 Ion 165.00 (164.70 to 165.70): E
 Ion 167.00 (166.70 to 167.70): E






#61
4-Nitroaniline
 Concen: 1.396 ng/ul
 RT: 15.17 min Scan# 2071
 Delta R.T. -0.03 min
 Lab File: BN009327.D
 Acq: 08 Jan 2020 12:29

Tgt Ion:138 Resp: 13859
 Ion Ratio Lower Upper
 138 100
 92 0.0 45.0 67.6#
 108 2.9 65.6 98.4#



Abundance Ion 138.00 (137.70 to 138.70): E
 Ion 92.00 (91.70 to 92.70): BNC
 Ion 108.00 (107.70 to 108.70): E





#67

Hexachlorobenzene

Concen: 30.278 ng/ul

RT: 16.17 min Scan# 2242

Delta R.T. -0.01 min

Lab File: BN009327.D

Acq: 08 Jan 2020 12:29

Instrument :

BNA_N

ClientSampled :


Tot Ion:284 Resp: 482791

Ion Ratio Lower Upper

284 100

142 36.6 31.6 47.4




249 29.7 25.7 38.5



Abundance Ion 284.00 (283.70 to 284.70): E

Ion 142.00 (141.70 to 142.70): E

Ion 249.00 (248.70 to 249.70): E



#69

Pentachlorophenol

Concen: 52.241 ng/ul

RT: 16.53 min Scan# 2302

Delta R.T. -0.00 min

Lab File: BN009327.D

Acq: 08 Jan 2020 12:29


Tgt Ion:266 Resp: 502860

Ion Ratio Lower Upper

266 100

264 68.0 49.5 74.3



268 68.8 53.7 80.5




Abundance Ion 265.70 (265.40 to 266.40): E

Ion 264.00 (263.70 to 264.70): E

Ion 268.00 (267.70 to 268.70): E





#70

Phenanthrene

Concen: 22.402 ng/ul

RT: 17.00 min Scan# 2382

Delta R.T. 0.09 min

Lab File: BN009327.D

Acq: 08 Jan 2020 12:29

Instrument :

BNA_N

ClientSampleId :


Tot Ion:178 Resp: 1724542

Ion Ratio Lower Upper

178 100

179 15.1 12.2 18.4




176 18.1 15.2 22.8



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 176.00 (175.70 to 176.70): E



#72

Anthracene

Concen: 22.158 ng/ul

RT: 17.00 min Scan# 2382

Delta R.T. -0.01 min

Lab File: BN009327.D

Acq: 08 Jan 2020 12:29


Tgt Ion:178 Resp: 1724542

Ion Ratio Lower Upper

178 100

179 15.1 12.2 18.2



176 18.1 14.8 22.2




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 176.00 (175.70 to 176.70): E









#80
Pvrene
Concen: 25.698 ng/ul
RT: 19.30 min Scan# 2773
Delta R.T. -0.00 min
Lab File: BN009327.D
Acq: 08 Jan 2020 12:29




Instrument :
BNA_N
ClientSampleId :


Tot Ion:202 Resp: 2354617
Ion Ratio Lower Upper
202 100
101 13.2 10.4 15.6
100 11.3 8.6 12.8



#81
Butylbenzylphthalate
Concen: 55.903 ng/ul
RT: 20.23 min Scan# 2931
Delta R.T. -0.00 min
Lab File: BN009327.D
Acq: 08 Jan 2020 12:29

Tgt Ion:149 Resp: 2148052
Ion Ratio Lower Upper
149 100
91 77.8 61.9 92.9
206 21.6 16.8 25.2





#83

Benzo(a)anthracene

Concen: 24.939 ng/ul

RT: 21.06 min Scan# 3072

Delta R.T. -0.00 min

Lab File: BN009327.D

Acq: 08 Jan 2020 12:29

Instrument :

BNA_N

ClientSampled :


Tot Ion:228 Resp: 2156584

Ion	Ratio	Lower	Upper
228	100		
229	20.5	15.4	23.0
226	27.3	21.5	32.3

228 100

229 20.5 15.4 23.0

226 27.3 21.5 32.3






Abundance

Ion 228.00 (227.70 to 228.70): E

Ion 229.00 (228.70 to 229.70): E

Ion 226.00 (225.70 to 226.70): E



#84

Bis(2-ethylhexyl)phthalate

Concen: 42.668 ng/ul

RT: 21.02 min Scan# 3066

Delta R.T. 0.01 min

Lab File: BN009327.D

Acq: 08 Jan 2020 12:29


Tgt Ion:149 Resp: 2473032

Ion	Ratio	Lower	Upper
149	100		
167	28.1	22.2	33.2
279	4.9	3.4	5.0

149 100

167 28.1 22.2 33.2

279 4.9 3.4 5.0






Abundance

Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E









#85
Chrysene
Concen: 25.505 ng/ul
RT: 21.11 min Scan# 3081
Delta R.T. -0.00 min
Lab File: BN009327.D
Acq: 08 Jan 2020 12:29

Instrument :
BNA_N
ClientSampled :

Tot Ion:228 Resp: 2146189
Ion Ratio Lower Upper
228 100
226 29.9 23.4 35.2
229 19.5 14.9 22.3




Abundance Ion 228.00 (227.70 to 228.70): E
Ion 226.00 (225.70 to 226.70): E
Ion 229.00 (228.70 to 229.70): E






#88
Benzo(b)fluoranthene
Concen: 42.987 ng/ul
RT: 22.57 min Scan# 3330
Delta R.T. 0.01 min
Lab File: BN009327.D
Acq: 08 Jan 2020 12:29

Tgt Ion:252 Resp: 3343169
Ion Ratio Lower Upper
252 100
253 22.1 17.0 25.6
125 9.9 8.0 12.0



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E





#89

Benzo(k)fluoranthene

Concen: 24.430 ng/ul

RT: 22.61 min Scan# 3336

Delta R.T. -0.00 min

Lab File: BN009327.D

Acq: 08 Jan 2020 12:29

Instrument :

BNA_N

ClientSampled :


Tot Ion:252 Resp: 1876147

Ion Ratio Lower Upper

252 100

253 21.0 17.8 26.8




125 10.3 8.2 12.2



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#91

Benzo(a)pyrene

Concen: 25.101 ng/ul

RT: 23.10 min Scan# 3420

Delta R.T. -0.00 min

Lab File: BN009327.D

Acq: 08 Jan 2020 12:29


Tgt Ion:252 Resp: 1732093

Ion Ratio Lower Upper

252 100

253 22.7 16.5 24.7



125 11.7 9.3 13.9




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E









#92
 Indeno(1,2,3-cd)pyrene
 Concen: 35.875 ng/ul
 RT: 25.28 min Scan# 3790
 Delta R.T. 0.01 min
 Lab File: BN009327.D
 Acq: 08 Jan 2020 12:29

Instrument :
 BNA_N
 ClientSampleId :

Tot Ion:276 Resp: 2682613
 Ion Ratio Lower Upper
 276 100
 138 24.9 18.4 27.6
 277 25.2 19.1 28.7



#93
 Dibenzo(a,h)anthracene
 Concen: 31.390 ng/ul
 RT: 25.29 min Scan# 3792
 Delta R.T. -0.00 min
 Lab File: BN009327.D
 Acq: 08 Jan 2020 12:29

Tgt Ion:278 Resp: 1971979
 Ion Ratio Lower Upper
 278 100
 139 16.2 12.1 18.1
 279 23.8 17.8 26.6

