


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010919\
Data File : BN004321.D
Acq On : 09 Jan 2019 08:39
Operator : JU/SJ
Sample : DFTPP55
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP55

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN010919MA.M
Title : SVOA CALIBRATION
Last Update : Wed Jan 09 12:45:02 2019



AutoFind: Scans 2480, 2481, 2482; Background Corrected with Scan 2475

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	28.1	28234	PASS
68	69	0.00	2	1.8	548	PASS
69	198	0.00	100	30.5	30677	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	45.9	46101	PASS
197	198	0.00	2	0.6	604	PASS
198	198	100	100	100.0	100421	PASS
199	198	5	9	6.9	6915	PASS
275	198	10	60	28.8	28904	PASS
365	198	1	100	3.8	3767	PASS
441	198	0.01	100	11.2	11224	PASS
442	198	50	100	71.2	71480	PASS
443	442	15	24	19.3	13805	PASS

m/z	Abundance
44.00	1013.0
207.00	973.0

Instrument :
BNA_N
ClientSampleId :
DFTPP55

m/z	Abundance
38.00	337.0
39.00	1873.0
43.90	855.0
50.00	7016.0
51.00	26784.0
52.00	1372.0
55.90	877.0
57.00	1871.0
60.90	331.0
61.90	370.0
63.00	1128.0
65.00	521.0
68.00	529.0
69.00	28856.0
72.90	311.0
74.00	3067.0
75.00	4969.0
76.00	1673.0
77.00	34072.0
78.00	2519.0
79.00	2318.0
80.00	1758.0
80.90	2648.0
81.90	635.0
83.00	556.0
85.00	432.0
85.90	806.0
87.00	387.0
90.90	582.0
92.00	744.0
93.00	4226.0
94.00	310.0
98.00	3345.0
99.00	2887.0
101.00	1444.0
102.90	526.0
103.90	903.0
104.90	946.0
106.00	332.0
107.00	11138.0
108.00	1693.0
109.00	313.0
110.00	19864.0
111.00	3234.0
111.90	411.0
115.90	609.0
117.00	9839.0
118.00	768.0
121.90	761.0
122.90	1222.0
123.90	508.0
125.00	543.0
127.00	40240.0
128.00	3102.0
129.00	15909.0
130.00	1345.0
133.90	422.0
135.00	1299.0
135.90	509.0
136.90	750.0
141.00	2099.0
142.00	723.0
142.90	437.0
146.00	389.0
147.00	1078.0
148.00	2402.0
149.00	608.0
152.90	656.0
154.00	424.0
155.00	1186.0
156.00	1792.0
157.00	306.0
157.90	374.0
158.90	306.0
160.00	597.0
160.90	982.0
161.90	313.0
165.00	798.0
166.00	598.0
167.00	3992.0
168.00	2726.0
169.00	377.0
171.90	378.0
172.90	427.0
173.90	879.0
175.00	1538.0
176.00	444.0
177.00	803.0
178.90	3208.0
180.00	2261.0
181.00	1083.0

Instrument :
BNA_N
ClientSampleId :
DFTPP55

185.00	1434.0
186.00	11878.0
187.00	3530.0
188.00	463.0
188.90	810.0
191.00	409.0
192.00	1041.0
193.00	1208.0
196.00	2270.0
197.00	514.0
198.00	81576.0
199.00	5585.0
199.90	406.0
201.30	317.0
202.90	678.0
204.00	3024.0
205.00	5348.0
206.00	21872.0
207.00	3726.0
208.00	1045.0
209.00	377.0
210.20	405.0
211.10	826.0
215.90	443.0
217.00	5799.0
217.90	910.0
221.00	3264.0
221.90	839.0
223.00	1430.0
224.00	12234.0
225.00	3154.0
226.00	407.0
227.00	5305.0
228.00	800.0
229.00	1030.0
231.00	479.0
233.90	347.0
235.00	368.0
237.00	530.0
240.90	305.0
241.90	608.0
243.00	679.0
244.00	9432.0
245.00	1287.0
246.00	1892.0
246.90	360.0
249.00	370.0
254.10	356.0
255.00	48648.0
256.00	7245.0
256.90	652.0
258.00	2939.0
259.00	409.0
265.00	1162.0
273.00	1400.0
274.00	3570.0
275.00	21408.0
276.00	2645.0
277.00	1850.0
278.00	367.0
284.90	323.0
293.00	429.0
296.00	5767.0
297.00	886.0
303.00	703.0
314.90	673.0
323.00	1751.0
324.00	366.0
327.00	333.0
334.00	1208.0
345.90	394.0
352.00	501.0
352.90	372.0
354.00	550.0
364.90	2392.0
372.00	840.0
402.00	347.0
403.00	476.0
421.00	345.0
422.00	369.0
423.00	2449.0
424.00	798.0
441.00	6414.0
442.00	41752.0
443.00	8577.0
444.00	779.0

Instrument :
BNA_N
ClientSampleId :
DFTPP55

m/z	Abundance
38.00	376.0
39.00	2309.0
44.00	822.0
50.00	8632.0
51.00	33864.0
52.00	1721.0
55.90	1058.0
57.00	2288.0
61.00	528.0
61.90	487.0
63.00	1473.0
65.00	749.0
68.00	656.0
69.00	36408.0
73.00	466.0
74.00	4016.0
75.00	6155.0
76.00	2138.0
77.00	43808.0
78.00	3069.0
79.00	3029.0
80.00	2361.0
81.00	3429.0
81.90	695.0
83.00	740.0
85.00	606.0
86.00	972.0
87.00	483.0
90.90	736.0
92.00	847.0
93.00	5503.0
94.00	343.0
98.00	4328.0
99.00	3580.0
101.00	1919.0
102.90	603.0
104.00	1231.0
105.00	1121.0
106.00	393.0
107.00	15146.0
108.00	2236.0
109.00	403.0
110.00	27568.0
111.00	4198.0
111.90	550.0
116.00	862.0
117.00	13180.0
118.00	922.0
122.00	1025.0
123.00	1631.0
124.00	715.0
124.90	770.0
127.00	55560.0
128.00	4332.0
129.00	21520.0
130.00	1856.0
130.90	366.0
134.00	651.0
135.00	1783.0
136.00	759.0
137.00	958.0
141.00	2887.0
141.90	1009.0
142.90	650.0
146.00	526.0
147.00	1444.0
148.00	3467.0
148.90	641.0
151.20	367.0
152.90	846.0
154.00	687.0
155.00	1548.0
156.00	2376.0
157.00	423.0
157.90	512.0
159.00	371.0
160.00	1011.0
161.00	1382.0
162.00	399.0
165.00	1049.0
166.00	840.0
167.00	5923.0
168.00	3268.0
169.00	567.0
171.90	575.0
173.00	706.0
174.00	1258.0
175.00	2345.0
175.90	607.0
176.90	1164.0
178.00	399.0

Instrument :
BNA_N
ClientSampleId :
DFTPP55

179.00	4735.0
180.00	3331.0
181.00	1451.0
183.90	387.0
185.00	2102.0
186.00	17232.0
187.00	5004.0
188.00	579.0
189.00	1247.0
191.00	579.0
192.00	1473.0
193.00	1778.0
194.00	384.0
196.00	3402.0
197.00	677.0
198.00	119784.0
199.00	8313.0
199.90	619.0
201.50	536.0
203.00	825.0
204.00	4684.0
205.00	7820.0
206.00	33680.0
207.00	5369.0
208.00	1261.0
209.00	485.0
210.10	574.0
211.00	1209.0
214.90	354.0
215.90	732.0
217.00	8840.0
218.00	1252.0
221.00	5033.0
221.80	1309.0
223.00	2076.0
224.00	18976.0
225.00	4889.0
226.00	551.0
227.00	8010.0
228.00	1157.0
229.00	1540.0
231.00	768.0
234.00	570.0
234.90	549.0
235.90	383.0
237.00	779.0
238.90	307.0
241.00	572.0
242.00	1029.0
243.00	1161.0
244.00	14481.0
245.00	1984.0
246.00	2988.0
246.90	622.0
249.00	487.0
253.00	311.0
254.10	626.0
255.00	76136.0
256.00	11020.0
257.00	881.0
258.00	4246.0
258.90	759.0
265.00	1969.0
273.00	2332.0
274.00	6101.0
275.00	34232.0
276.00	4584.0
277.00	3059.0
278.00	531.0
283.00	334.0
285.00	533.0
293.00	692.0
296.00	9807.0
297.00	1402.0
303.00	1200.0
314.00	474.0
315.00	1158.0
316.00	570.0
321.00	341.0
323.00	3132.0
324.00	544.0
326.90	527.0
328.00	305.0
334.00	2183.0
335.00	551.0
341.00	407.0
346.00	710.0
352.00	925.0
353.00	583.0
354.00	974.0
365.00	4451.0
366.00	658.0
372.00	1699.0
373.00	420.0
383.00	488.0

Instrument :

BNA_N

ClientSampleId :

DFTPP55

402.00	674.0
403.00	909.0
404.00	353.0
421.00	726.0
422.00	684.0
423.00	5148.0
424.00	1245.0
441.00	13118.0
442.00	83552.0
443.00	16104.0
444.00	1609.0

Instrument :
BNA_N
ClientSampleId :
DFTPP55

m/z	Abundance
38.00	327.0
39.00	1669.0
44.00	883.0
50.00	6079.0
51.00	24056.0
52.00	1254.0
55.90	710.0
57.00	1700.0
60.90	337.0
61.90	357.0
63.00	1014.0
64.90	557.0
68.00	459.0
69.00	26768.0
73.00	319.0
74.00	2786.0
75.00	4423.0
76.00	1539.0
77.00	32056.0
78.00	2178.0
79.00	2275.0
79.90	1741.0
81.00	2302.0
82.00	517.0
82.90	516.0
85.00	469.0
85.90	730.0
87.00	313.0
91.00	539.0
91.90	612.0
93.00	3935.0
98.00	3203.0
99.00	2675.0
101.00	1400.0
102.90	447.0
103.90	870.0
104.90	897.0
106.00	304.0
107.00	11553.0
108.00	1664.0
109.00	317.0
110.00	20800.0
111.00	3006.0
111.90	440.0
115.90	638.0
117.00	10240.0
118.00	805.0
121.90	817.0
123.00	1221.0
124.00	514.0
124.90	509.0
127.00	42504.0
128.00	3113.0
129.00	16696.0
130.00	1469.0
133.90	412.0
135.00	1343.0
136.00	489.0
137.00	764.0
141.00	2206.0
141.90	715.0
142.90	578.0
146.00	380.0
147.00	1268.0
148.00	2838.0
148.90	602.0
153.00	783.0
154.00	527.0
155.00	1283.0
156.00	1940.0
157.00	358.0
157.90	422.0
159.00	319.0
159.90	702.0
161.00	1034.0
162.00	395.0
164.90	820.0
166.00	681.0
167.00	4617.0
168.00	2501.0
169.00	422.0
172.00	469.0
173.00	546.0
174.00	997.0
175.00	2025.0
176.00	610.0
176.90	932.0
179.00	3696.0
180.00	2557.0
181.00	1212.0
185.00	1757.0

Instrument :
BNA_N
ClientSampleId :
DFTPP55

186.00	13769.0
187.00	4053.0
188.00	398.0
189.00	907.0
191.00	649.0
192.00	1240.0
193.00	1484.0
194.00	307.0
196.00	2717.0
197.10	623.0
198.00	99904.0
199.00	6847.0
200.00	573.0
201.50	501.0
202.90	712.0
204.00	3891.0
205.00	6478.0
206.00	27328.0
207.00	4552.0
208.00	1048.0
209.00	376.0
211.00	1197.0
215.00	304.0
216.00	634.0
217.00	7472.0
218.00	903.0
221.00	4306.0
221.90	1108.0
223.00	1650.0
224.00	15843.0
225.00	4006.0
226.00	340.0
227.00	6959.0
228.00	929.0
229.00	1524.0
231.00	589.0
233.90	441.0
235.00	449.0
236.00	322.0
237.00	588.0
241.00	328.0
242.00	809.0
243.00	956.0
244.00	12879.0
245.00	1708.0
246.00	2648.0
247.00	493.0
249.00	450.0
254.10	547.0
255.00	68176.0
256.00	9541.0
256.90	721.0
258.00	3899.0
259.00	640.0
265.00	1643.0
273.00	2023.0
274.00	5415.0
275.00	31072.0
276.00	4179.0
277.00	2679.0
277.90	448.0
283.00	346.0
285.00	452.0
293.00	592.0
296.00	9482.0
297.00	1336.0
303.00	1006.0
313.90	455.0
315.00	1062.0
316.00	590.0
320.90	314.0
323.00	2781.0
324.00	576.0
326.90	509.0
333.00	310.0
334.00	1986.0
335.00	527.0
341.00	341.0
345.90	653.0
352.00	1016.0
353.00	752.0
354.00	965.0
364.90	4458.0
365.90	608.0
372.00	1730.0
373.00	404.0
382.90	399.0
402.00	676.0
403.00	878.0
403.90	358.0
421.00	774.0
422.00	783.0
423.00	4824.0
424.00	1152.0
441.00	14142.0

Instrument :
BNA_N
ClientSampleId :
DFTPP55

442.00	89136.0
443.00	16736.0
444.00	1559.0

Instrument :
BNA_N
ClientSampleId :
DFTPP55