

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010924\
 Data File : BN029230.D
 Acq On : 09 Jan 2024 16:20
 Operator : MA/JU
 Sample : P1042-03
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SP-303-20240104

Quant Time: Jan 09 16:56:43 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN122723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 28 01:20:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.724	152	718	0.400	ng	0.00
7) Naphthalene-d8	10.520	136	1352	0.400	ng	# 0.00
13) Acenaphthene-d10	14.372	164	777	0.400	ng	0.00
19) Phenanthrene-d10	17.131	188	1156	0.400	ng	# 0.00
29) Chrysene-d12	21.349	240	363	0.400	ng	# 0.00
35) Perylene-d12	23.636	264	512	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.334	112	12	0.007	ng	0.00
5) Phenol-d6	6.930	99	5	0.003	ng	0.00
8) Nitrobenzene-d5	8.918	82	412	0.243	ng	0.01
11) 2-Methylnaphthalene-d10	12.125	152	573	0.245	ng	0.00
14) 2,4,6-Tribromophenol	15.903	330	29	0.048	ng	0.02
15) 2-Fluorobiphenyl	13.014	172	1076	0.258	ng	0.01
27) Fluoranthene-d10	19.173	212	590	0.251	ng	0.00
31) Terphenyl-d14	19.787	244	357	0.317	ng	0.00
Target Compounds						
34) Bis(2-ethylhexyl)phtha...	21.268	149	721	0.237	ng	# 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010924\
Data File : BN029230.D
Acq On : 09 Jan 2024 16:20
Operator : MA/JU
Sample : P1042-03
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SP-303-20240104

Quant Time: Jan 09 16:56:43 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN122723.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 28 01:20:47 2023
Response via : Initial Calibration

