


Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN011020\
 Data File : BN009329.D
 Acq On : 10 Jan 2020 09:52
 Operator : JU
 Sample : DFTPP04
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP04

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN011020.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Jan 10 16:45:39 2020



AutoFind: Scans 2438, 2439, 2440; Background Corrected with Scan 2432

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	58.3	149348	PASS
68	69	0.00	2	1.8	2174	PASS
69	198	0.00	100	46.0	117842	PASS
70	69	0.00	2	0.6	662	PASS
127	198	10	80	49.3	126338	PASS
197	198	0.00	2	0.4	1013	PASS
198	198	100	100	100.0	256341	PASS
199	198	5	9	6.9	17710	PASS
275	198	10	60	25.4	65128	PASS
365	198	1	100	3.2	8270	PASS
441	198	0.01	100	10.0	25557	PASS
442	198	50	100	68.1	174682	PASS
443	442	15	24	18.3	31925	PASS

m/z	Abundance
39.90	1430.0
42.90	366.0
44.00	7250.0
49.00	754.0
51.00	348.0
72.90	577.0
83.80	304.0
85.90	322.0
147.00	323.0
207.00	1631.0
207.90	308.0
281.10	559.0

Instrument :
BNA_N
ClientSampleId :
DFTPP04

m/z	Abundance
37.00	644.0
38.00	1342.0
39.10	9578.0
40.00	1893.0
41.00	529.0
44.00	6045.0
44.90	672.0
46.80	329.0
49.00	1446.0
50.00	30808.0
51.00	132800.0
52.00	7619.0
53.00	520.0
55.10	592.0
56.00	3794.0
57.00	9048.0
58.00	386.0
61.00	1870.0
62.00	1720.0
63.00	4883.0
64.00	563.0
65.00	2642.0
68.10	1839.0
69.00	103552.0
70.00	482.0
73.00	1711.0
74.00	10669.0
75.00	16070.0
76.10	4894.0
77.00	110880.0
78.10	7487.0
79.00	8346.0
80.00	6554.0
81.00	8591.0
82.00	1986.0
83.00	1894.0
83.90	637.0
85.00	1420.0
86.00	2471.0
87.00	1064.0
87.90	374.0
90.90	2032.0
92.00	2541.0
93.00	12491.0
94.00	1064.0
96.10	703.0
98.00	9030.0
99.00	8458.0
100.00	809.0
101.00	4043.0
102.90	1587.0
104.00	3010.0
105.00	2960.0
106.00	1066.0
107.00	34344.0
108.00	4693.0
108.90	894.0
110.00	55432.0
111.00	9708.0
112.00	1386.0
113.00	601.0
115.90	1819.0
117.00	26488.0
117.90	2288.0
119.10	466.0
122.00	2131.0
123.00	2508.0
124.00	1773.0
124.90	1179.0
126.10	860.0
127.00	100408.0
128.00	7482.0
129.00	38960.0
129.90	3638.0
130.90	692.0
133.90	1470.0
135.00	3462.0
135.90	1141.0
137.00	1690.0
137.90	459.0
139.90	499.0
140.90	4747.0
142.00	1520.0
142.90	1305.0
144.00	336.0
146.00	945.0
147.00	3285.0
148.00	6650.0
148.90	1750.0
150.10	464.0
151.00	820.0
152.10	403.0

Instrument :
BNA_N
ClientSampleId :
DFTPP04

152.90	1805.0
153.90	1228.0
155.00	3148.0
156.00	4509.0
156.90	1122.0
158.00	1051.0
158.90	1307.0
160.00	1511.0
161.00	2464.0
162.00	685.0
165.00	2190.0
165.90	1763.0
167.00	10227.0
168.00	8202.0
169.00	1267.0
169.90	331.0
171.10	366.0
171.90	937.0
173.00	1258.0
174.00	2271.0
175.00	3612.0
176.00	649.0
177.00	2297.0
177.80	454.0
178.90	7112.0
179.90	6175.0
181.00	2712.0
183.90	525.0
185.00	3682.0
186.00	27840.0
187.00	8138.0
188.00	1024.0
189.00	1499.0
189.90	428.0
191.10	917.0
191.80	2231.0
193.00	2871.0
194.00	948.0
195.10	348.0
195.90	4719.0
197.10	1351.0
197.90	194176.0
198.90	13172.0
200.00	863.0
201.50	838.0
203.00	1457.0
204.00	6222.0
205.00	11736.0
206.00	50336.0
207.00	8265.0
207.90	2020.0
208.90	878.0
210.00	826.0
211.00	2243.0
214.90	868.0
215.90	1121.0
217.00	13139.0
217.90	1592.0
219.00	352.0
220.10	337.0
221.00	7948.0
222.00	1555.0
222.90	2846.0
224.00	26144.0
225.00	7325.0
226.00	1328.0
227.00	11710.0
227.90	1459.0
229.00	2913.0
230.00	305.0
231.00	1191.0
232.80	358.0
234.00	779.0
234.90	1019.0
236.00	633.0
237.00	1554.0
238.90	486.0
239.70	331.0
241.00	503.0
242.00	1868.0
243.00	1307.0
244.00	18376.0
245.10	2818.0
246.00	3951.0
246.90	823.0
249.00	643.0
252.90	647.0
255.00	100112.0
256.00	14829.0
256.90	1900.0
258.00	6306.0
259.00	1080.0
265.00	2688.0
266.00	416.0
272.00	370.0
273.00	3025.0

Instrument :
BNA_N
ClientSampleId :
DFTPP04

274.00	7370.0
275.00	44096.0
275.90	5582.0
277.00	3677.0
277.90	783.0
280.90	647.0
282.90	801.0
283.80	426.0
285.00	612.0
292.00	319.0
292.90	816.0
294.00	610.0
296.00	13775.0
297.00	2050.0
302.10	304.0
303.00	1427.0
304.00	459.0
313.90	752.0
315.00	1043.0
316.00	627.0
322.00	403.0
323.00	3289.0
323.90	638.0
327.00	609.0
331.90	413.0
333.00	409.0
334.00	2444.0
335.00	722.0
340.80	335.0
346.00	892.0
351.90	824.0
353.00	738.0
353.90	1322.0
355.00	438.0
365.00	3948.0
366.00	711.0
371.90	1873.0
372.90	672.0
383.00	474.0
389.80	479.0
401.80	548.0
402.90	920.0
404.00	416.0
421.10	1154.0
421.90	776.0
422.90	6255.0
424.00	2785.0
425.10	487.0
441.00	13655.0
442.00	92496.0
443.00	17440.0
444.00	1881.0

Instrument :
BNA_N
ClientSampleId :
DFTPP04

m/z	Abundance
37.00	747.0
38.10	2012.0
39.10	14297.0
40.00	1994.0
41.10	764.0
44.00	5638.0
45.00	428.0
48.00	376.0
49.10	1656.0
50.00	44040.0
51.00	180672.0
52.10	9578.0
53.00	506.0
55.00	983.0
56.00	4612.0
57.00	12313.0
57.90	657.0
61.00	2452.0
62.00	2590.0
63.00	6816.0
64.10	949.0
65.00	3481.0
67.00	308.0
68.10	2348.0
69.00	139328.0
70.00	857.0
73.00	1893.0
74.00	15747.0
75.00	23416.0
76.00	7065.0
77.00	161472.0
78.00	11853.0
79.00	11583.0
80.00	6929.0
81.00	12336.0
81.90	2474.0
83.00	2932.0
83.90	797.0
85.00	1869.0
86.00	3496.0
87.00	1401.0
88.00	382.0
91.00	2639.0
92.00	3381.0
93.00	19376.0
93.90	921.0
96.00	716.0
97.00	550.0
98.00	14022.0
99.00	12281.0
100.00	976.0
101.00	6463.0
102.90	1841.0
104.00	4612.0
105.00	3929.0
106.10	1411.0
107.00	47144.0
108.00	7309.0
109.10	1422.0
110.00	84648.0
111.00	13082.0
112.00	1476.0
113.00	351.0
116.00	2694.0
117.00	39824.0
117.90	3008.0
119.00	714.0
119.90	808.0
121.90	3726.0
122.90	4345.0
124.00	2477.0
125.00	2134.0
127.00	149696.0
128.00	11457.0
129.00	59160.0
130.00	4351.0
131.10	986.0
132.00	483.0
133.00	336.0
134.00	1901.0
135.00	4723.0
136.00	1971.0
137.00	2569.0
138.00	404.0
139.10	424.0
139.90	625.0
141.00	8190.0
142.00	2446.0
143.00	1666.0
144.00	431.0
145.00	544.0
145.90	1210.0

Instrument :
BNA_N
ClientSampleId :
DFTPP04

147.00	3821.0
148.00	9063.0
148.90	2714.0
149.90	539.0
151.10	1603.0
151.90	544.0
152.90	2411.0
154.00	1934.0
155.00	5034.0
156.00	6583.0
157.00	1388.0
158.00	1909.0
158.90	715.0
159.00	717.0
160.00	2617.0
161.00	3462.0
161.90	1350.0
163.90	573.0
164.90	3265.0
166.00	2292.0
167.00	16936.0
168.00	9245.0
169.10	1630.0
169.90	609.0
170.70	670.0
172.00	1425.0
173.00	1316.0
174.00	3423.0
175.00	6497.0
175.90	1473.0
176.90	3021.0
178.00	1316.0
178.90	11201.0
180.00	9056.0
181.10	3832.0
182.10	708.0
184.00	841.0
185.00	5281.0
186.00	42920.0
187.00	13001.0
187.90	1875.0
188.90	3263.0
189.90	779.0
191.00	1528.0
192.00	4183.0
193.00	3992.0
194.00	992.0
195.10	396.0
196.00	9041.0
196.80	1827.0
197.90	299456.0
199.00	21600.0
199.90	1776.0
201.40	1356.0
202.90	2639.0
204.00	11547.0
205.00	18968.0
206.00	80192.0
207.00	12651.0
208.00	2728.0
209.00	1415.0
210.00	1524.0
210.90	3008.0
215.10	871.0
216.00	1448.0
217.00	22048.0
218.00	3002.0
220.20	568.0
221.00	10583.0
221.90	2646.0
222.90	4103.0
224.00	43680.0
225.00	12997.0
226.00	1709.0
227.00	19152.0
228.00	2479.0
228.90	3873.0
230.00	752.0
231.00	1225.0
234.00	1255.0
234.90	1493.0
235.90	836.0
237.00	2128.0
239.10	705.0
240.00	577.0
241.00	963.0
242.00	2340.0
243.00	2776.0
244.00	30720.0
245.00	4071.0
246.00	6648.0
247.00	1373.0
248.90	1553.0
250.70	371.0
252.90	614.0
255.00	170112.0

Instrument :
BNA_N
ClientSampleId :
DFTPP04

256.00	28928.0
256.90	2205.0
258.00	10890.0
259.00	1688.0
259.90	319.0
265.00	4247.0
266.00	816.0
271.20	336.0
272.00	541.0
273.00	5461.0
274.00	14162.0
275.00	75888.0
276.00	9075.0
277.00	7139.0
278.00	1860.0
278.90	344.0
280.90	424.0
283.00	1016.0
284.00	393.0
285.00	1684.0
286.00	362.0
292.00	345.0
293.00	1329.0
294.00	663.0
296.00	21936.0
297.00	3287.0
297.90	360.0
302.00	649.0
303.00	2632.0
304.00	1136.0
313.90	879.0
314.90	2228.0
316.00	1818.0
317.00	364.0
320.90	781.0
321.90	437.0
323.00	6664.0
323.90	1502.0
327.00	1092.0
328.00	559.0
332.00	472.0
333.00	670.0
334.00	3978.0
334.90	986.0
341.00	942.0
346.00	1852.0
351.90	2039.0
353.00	1215.0
354.00	1580.0
354.90	657.0
364.90	9416.0
366.00	1307.0
369.90	341.0
372.00	3117.0
373.00	1239.0
383.00	741.0
402.00	1223.0
402.90	1566.0
404.10	718.0
421.00	1779.0
422.00	1732.0
423.00	12387.0
423.90	4146.0
425.10	761.0
440.10	363.0
441.00	26104.0
442.00	195008.0
443.00	36288.0
444.00	3487.0

Instrument :
BNA_N
ClientSampleId :
DFTPP04

m/z	Abundance
37.00	904.0
38.00	1822.0
39.10	9864.0
39.90	2315.0
41.10	506.0
44.00	6417.0
45.00	493.0
49.10	1829.0
50.00	34224.0
51.00	135616.0
52.00	7199.0
53.10	447.0
55.00	1148.0
55.90	4273.0
57.00	9409.0
58.00	406.0
61.00	1879.0
62.00	1690.0
63.10	5448.0
64.10	579.0
65.00	2425.0
68.10	2337.0
69.00	110648.0
70.10	648.0
73.10	1434.0
74.00	10749.0
75.00	18616.0
76.00	5924.0
77.00	121864.0
78.00	8189.0
79.00	9026.0
80.00	6682.0
81.00	9069.0
81.90	2340.0
83.00	2383.0
83.90	384.0
85.00	1444.0
86.00	2800.0
87.00	1033.0
87.90	584.0
90.90	2009.0
92.00	2069.0
93.00	14421.0
93.90	1004.0
95.00	609.0
96.00	915.0
96.90	369.0
98.00	11928.0
99.00	9681.0
100.00	748.0
100.90	4416.0
101.90	314.0
102.90	1770.0
104.00	3588.0
105.00	2845.0
106.10	1024.0
107.00	36928.0
108.00	6560.0
109.20	1214.0
110.00	66256.0
111.00	10732.0
111.90	1053.0
113.00	398.0
116.00	2033.0
117.00	31488.0
117.90	2059.0
119.00	418.0
120.00	653.0
121.90	2262.0
123.00	3541.0
124.00	1674.0
125.10	1561.0
127.00	128912.0
128.00	9624.0
129.00	48288.0
130.00	4253.0
130.90	683.0
132.00	461.0
132.80	428.0
133.90	1651.0
135.00	3612.0
136.10	1808.0
137.00	2439.0
138.10	615.0
138.80	354.0
139.90	807.0
141.00	6677.0
142.00	1694.0
143.00	1601.0
144.00	398.0
145.00	325.0
146.00	1095.0

Instrument :
BNA_N
ClientSampleId :
DFTPP04

147.00	3568.0
148.00	8220.0
149.00	2025.0
149.90	446.0
151.00	677.0
151.90	464.0
153.00	2383.0
153.90	1617.0
155.00	3762.0
156.00	5678.0
157.10	1755.0
157.80	949.0
158.00	883.0
158.90	1168.0
159.90	2124.0
161.00	3072.0
162.00	703.0
164.00	436.0
165.00	2559.0
166.00	2514.0
167.00	13892.0
168.00	8903.0
169.00	1237.0
171.00	596.0
172.00	1242.0
173.00	1400.0
174.00	3360.0
175.00	6266.0
176.00	1365.0
177.00	3090.0
178.10	897.0
178.90	10941.0
180.00	7746.0
180.90	3403.0
182.00	675.0
182.90	524.0
184.00	433.0
184.90	4741.0
186.00	38576.0
187.00	10637.0
188.00	1396.0
189.00	2251.0
190.90	1350.0
191.90	3356.0
193.00	3977.0
194.00	1153.0
194.70	473.0
196.00	7366.0
197.00	1690.0
198.00	275392.0
199.00	18360.0
199.90	1471.0
201.40	1619.0
203.00	2008.0
204.00	9896.0
205.00	17848.0
206.00	74792.0
207.00	10663.0
207.90	3289.0
209.00	1192.0
210.00	1331.0
210.90	2637.0
214.90	678.0
216.00	1582.0
216.90	19512.0
217.90	2210.0
221.00	9706.0
221.70	2535.0
223.00	4136.0
224.00	43480.0
225.00	10344.0
226.00	1029.0
226.90	18152.0
228.10	1869.0
229.00	4635.0
230.00	644.0
231.00	1408.0
233.90	836.0
235.00	1219.0
235.90	954.0
237.00	1476.0
238.90	765.0
240.00	507.0
241.00	877.0
241.90	2456.0
243.00	2547.0
244.00	29112.0
245.00	3345.0
246.00	5989.0
247.00	1523.0
247.90	323.0
249.00	1241.0
251.00	321.0
252.00	352.0
253.10	503.0
254.20	1920.0

Instrument :
BNA_N
ClientSampleId :
DFTPP04

255.00	166208.0
256.00	22960.0
257.00	2559.0
257.90	10533.0
259.00	1777.0
260.90	363.0
265.00	4266.0
265.90	405.0
270.90	538.0
272.10	362.0
273.00	5553.0
274.00	13974.0
275.00	75400.0
276.00	10625.0
276.90	5914.0
278.00	1027.0
281.00	758.0
282.00	372.0
282.90	926.0
283.90	477.0
285.00	1038.0
289.00	334.0
289.90	501.0
292.10	305.0
293.00	2025.0
293.90	400.0
295.00	336.0
296.00	22424.0
296.90	2365.0
302.90	2802.0
304.10	475.0
309.10	372.0
314.00	1099.0
315.00	2791.0
315.90	996.0
321.00	625.0
321.80	433.0
323.10	6653.0
324.00	1362.0
327.00	1732.0
328.00	839.0
331.80	490.0
332.90	741.0
334.00	4759.0
335.00	1600.0
341.00	1103.0
346.00	973.0
352.00	2226.0
352.90	1684.0
354.00	2258.0
355.10	517.0
364.90	11448.0
365.90	1393.0
370.90	536.0
372.00	3606.0
373.10	960.0
382.90	1042.0
383.90	355.0
390.10	465.0
390.90	319.0
391.90	492.0
401.90	1813.0
403.00	2554.0
404.00	901.0
421.00	1727.0
422.00	1975.0
423.00	14439.0
424.00	3598.0
424.90	605.0
441.00	36912.0
442.00	236544.0
443.00	42048.0
444.10	3846.0

Instrument :
BNA_N
ClientSampleId :
DFTPP04