




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011223\
Data File : BN023615.D
Acq On : 11 Jan 2023 17:07
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jan 12 04:02:32 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 12 04:02:28 2023
Response via : Initial Calibration



TIC: BN023615.D\data.ms

(70) Pentachlorophenol (C)




16.992min (0.000) 20863.53 ng

response 356444

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	64.49
264.00	61.60	67.16
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011223\
Data File : BN023615.D
Acq On : 11 Jan 2023 17:07
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jan 12 04:02:32 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 12 04:02:28 2023
Response via : Initial Calibration



TIC: BN023615.D\data.ms

(77) Benzidine

19.598min (0.000) 0.00 ng

response 1953117

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.98
183.00	13.20	11.93
0.00	0.00	0.00