

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011322\
 Data File : BN018400.D
 Acq On : 13 Jan 2022 19:12
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 13 20:21:06 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN010722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 10 13:58:28 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	84	0.00
2	1,4-Dioxane	0.400	0.273	31.8#	59	-0.02
3	n-Nitrosodimethylamine	0.400	0.427	-6.7	96	0.00
4 S	2-Fluorophenol	0.400	0.417	-4.2	92	0.00
5 S	Phenol-d6	0.400	0.416	-4.0	93	0.00
6	bis(2-Chloroethyl)ether	0.400	0.413	-3.2	88	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	85	-0.03
8 S	Nitrobenzene-d5	0.400	0.436	-9.0	96	-0.01
9	Naphthalene	0.400	0.402	-0.5	85	-0.01
10	Hexachlorobutadiene	0.400	0.415	-3.7	89	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.400	0.0	86	-0.01
12	2-Methylnaphthalene	0.400	0.411	-2.7	86	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	88	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.441	-10.2	106	-0.01
15 S	2-Fluorobiphenyl	0.400	0.412	-3.0	90	-0.01
16	Acenaphthylene	0.400	0.380	5.0	89	0.00
17	Acenaphthene	0.400	0.402	-0.5	89	-0.01
18	Fluorene	0.400	0.414	-3.5	92	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	95	-0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.284	29.0#	57	-0.01
21	4-Bromophenyl-phenylether	0.400	0.387	3.3	95	-0.01
22	Hexachlorobenzene	0.400	0.390	2.5	93	-0.01
23	Atrazine	0.400	0.371	7.3	107	-0.01
24	Pentachlorophenol	0.400	0.479	-19.7	120	0.00
25	Phenanthrene	0.400	0.397	0.8	95	0.00
26	Anthracene	0.400	0.389	2.8	98	-0.01
27 SURR	Fluoranthene-d10	0.400	0.404	-1.0	98	0.00
28	Fluoranthene	0.400	0.419	-4.7	98	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	101	0.00
30	Pyrene	0.400	0.402	-0.5	101	0.00
31 S	Terphenyl-d14	0.400	0.403	-0.8	101	0.00
32	Benzo(a)anthracene	0.400	0.401	-0.3	113	-0.01
33	Chrysene	0.400	0.404	-1.0	99	-0.01
34	Bis(2-ethylhexyl)phthalate	0.400	0.431	-7.7	122	-0.01
35 I	Perylene-d12	0.400	0.400	0.0	112	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.363	9.3	113	-0.01
37	Benzo(b)fluoranthene	0.400	0.397	0.8	108	0.00
38	Benzo(k)fluoranthene	0.400	0.406	-1.5	113	0.00
39 C	Benzo(a)pyrene	0.400	0.395	1.3	114	-0.01
40	Dibenzo(a,h)anthracene	0.400	0.345	13.8	116	-0.01
41	Benzo(g,h,i)perylene	0.400	0.368	8.0	107	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4