


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011522\
 Data File : BN018406.D
 Acq On : 15 Jan 2022 09:47
 Operator : CG/JU
 Sample : DFTPP223
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP223

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN010622.M
 Title : SVOA CALIBRATION
 Last Update : Fri Jan 07 01:31:51 2022



AutoFind: Scans 2457, 2458, 2459; Background Corrected with Scan 2451

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	55.0	535347	PASS
68	69	0.00	2	1.9	9955	PASS
69	198	0.00	100	53.6	521771	PASS
70	69	0.00	2	0.5	2493	PASS
127	198	10	80	58.8	572181	PASS
197	198	0.00	2	0.2	1707	PASS
198	198	100	100	100.0	973611	PASS
199	198	5	9	6.7	65563	PASS
275	198	10	60	25.8	251008	PASS
365	198	1	100	3.5	34096	PASS
441	198	0.01	100	11.9	115379	PASS
442	198	50	100	81.9	797632	PASS
443	442	15	24	19.5	155637	PASS

m/z	Abundance
38.90	418.0
40.00	982.0
41.10	386.0
43.10	814.0
44.10	901.0
48.90	617.0
51.10	504.0
55.00	325.0
57.10	430.0
73.00	1663.0
75.00	309.0
76.90	579.0
79.00	433.0
83.90	577.0
107.10	324.0
133.10	450.0
147.00	394.0
167.00	372.0
207.10	1886.0
207.90	506.0
208.90	442.0
221.10	406.0
281.10	429.0

Instrument :
BNA_N
ClientSampleId :
DFTPP223

m/z	Abundance
36.00	789.0
37.00	2527.0
38.00	5748.0
39.00	31312.0
40.00	2645.0
41.00	1509.0
43.00	1006.0
43.90	1155.0
44.90	639.0
47.80	533.0
49.00	4180.0
50.00	115576.0
51.00	445760.0
52.00	23800.0
53.00	602.0
55.00	2871.0
55.90	11682.0
57.00	26008.0
58.00	1510.0
58.90	596.0
60.10	463.0
61.00	5508.0
62.00	6565.0
63.00	17896.0
64.00	2404.0
65.00	8013.0
66.10	503.0
67.00	804.0
68.00	8477.0
69.00	425728.0
70.00	2290.0
71.00	651.0
73.10	5598.0
74.00	44624.0
75.00	71360.0
76.10	26384.0
77.00	501888.0
78.00	36336.0
79.00	33264.0
80.00	23384.0
81.00	35144.0
82.00	8785.0
83.00	7535.0
84.00	1525.0
85.00	7942.0
86.00	10437.0
87.00	4056.0
88.00	1547.0
89.00	883.0
91.00	7719.0
92.00	9691.0
93.00	59688.0
93.90	2960.0
95.10	1475.0
96.00	2590.0
97.00	1235.0
98.00	43232.0
99.00	35688.0
100.00	2606.0
101.00	20056.0
102.00	1432.0
103.00	5454.0
104.00	11351.0
105.00	11358.0
105.90	4109.0
107.00	141632.0
108.00	22048.0
109.10	3869.0
110.00	239744.0
111.00	38112.0
111.90	6027.0
113.00	1666.0
114.20	524.0
115.00	947.0
116.00	7463.0
117.00	128784.0
118.00	10146.0
119.00	1482.0
120.00	1649.0
121.20	658.0
122.00	8839.0
123.00	13129.0
124.00	6211.0
125.00	6031.0
127.00	446272.0
128.10	35200.0
129.00	164032.0
130.00	13411.0
131.10	2967.0
132.10	1831.0
133.00	1286.0
134.00	5255.0

Instrument :
BNA_N
ClientSampleId :
DFTPP223

135.00	14276.0
136.00	6004.0
137.00	7797.0
138.10	1530.0
139.00	982.0
140.10	2145.0
141.00	22144.0
142.00	7039.0
143.00	5327.0
144.00	1782.0
145.00	1059.0
146.00	4297.0
147.00	11952.0
148.00	27672.0
149.00	7257.0
150.00	2283.0
151.00	2485.0
151.70	1535.0
152.10	1519.0
153.00	6261.0
154.00	4695.0
155.00	12379.0
156.00	16880.0
157.00	3536.0
158.00	3950.0
158.90	3002.0
160.00	6022.0
161.00	8131.0
162.10	2763.0
162.90	863.0
163.90	945.0
165.00	7116.0
166.00	6030.0
167.00	41144.0
168.00	29072.0
169.10	4422.0
170.10	1265.0
171.10	1513.0
172.00	3893.0
173.00	5307.0
174.00	8278.0
175.10	16552.0
175.90	3016.0
177.00	5805.0
178.00	2810.0
179.00	29776.0
180.00	21480.0
181.00	9243.0
182.00	1695.0
183.20	1072.0
184.00	2676.0
185.10	14342.0
186.10	103352.0
187.00	31064.0
188.00	4211.0
189.00	7129.0
189.90	1356.0
191.10	2948.0
192.00	8107.0
193.00	10735.0
194.00	2057.0
195.00	1243.0
196.00	22408.0
196.70	5122.0
198.00	683392.0
199.00	51312.0
200.00	4542.0
201.60	3736.0
202.10	1062.0
203.00	4491.0
204.00	25728.0
205.10	47056.0
206.10	185216.0
207.10	27664.0
208.10	7245.0
209.00	2606.0
210.20	3303.0
211.10	6551.0
212.00	1268.0
212.90	305.0
215.00	2265.0
216.00	3446.0
217.00	43792.0
218.00	7319.0
218.80	703.0
220.00	1007.0
221.10	28296.0
221.80	6754.0
223.10	10399.0
224.10	95728.0
225.00	23664.0
226.10	3542.0
227.00	40472.0
228.00	5794.0
229.00	7596.0
230.00	1174.0

Instrument :
BNA_N
ClientSampleId :
DFTPP223

231.00	3508.0
232.00	869.0
232.90	357.0
233.10	358.0
234.00	3225.0
235.00	2784.0
236.10	2449.0
237.00	5093.0
237.90	746.0
238.90	1621.0
240.00	1538.0
241.00	2530.0
242.10	5403.0
243.10	5434.0
244.10	71520.0
245.10	9847.0
246.00	13355.0
247.00	4082.0
248.00	616.0
249.00	2823.0
250.10	699.0
251.00	976.0
252.10	866.0
253.00	1385.0
255.00	371968.0
256.00	56680.0
257.00	5605.0
258.00	22056.0
258.90	2845.0
260.10	644.0
261.00	763.0
264.00	693.0
265.00	10029.0
266.00	1539.0
267.00	654.0
270.90	568.0
272.00	1525.0
273.00	10468.0
274.00	31264.0
275.00	162176.0
276.00	21512.0
277.00	14794.0
278.00	2812.0
278.90	455.0
281.10	801.0
282.00	388.0
283.00	1532.0
284.10	892.0
285.10	2352.0
286.10	678.0
289.00	797.0
290.10	712.0
292.10	642.0
293.00	2529.0
293.90	903.0
295.10	1662.0
296.00	45304.0
297.00	6663.0
298.00	598.0
301.10	531.0
302.00	852.0
303.10	5138.0
303.90	1346.0
305.00	315.0
307.90	879.0
309.00	672.0
310.10	630.0
313.10	543.0
314.10	2535.0
315.00	5610.0
316.00	2417.0
317.00	780.0
319.90	402.0
321.00	1595.0
322.00	1009.0
323.00	13570.0
324.10	3600.0
325.10	502.0
326.10	316.0
327.00	3330.0
327.90	1602.0
329.00	352.0
332.00	908.0
333.10	1465.0
334.00	11928.0
335.10	2599.0
336.10	553.0
341.00	1763.0
341.90	409.0
346.10	2938.0
347.00	987.0
352.00	3740.0
353.00	3973.0
354.00	5232.0
355.00	1406.0
358.10	309.0

Instrument :
BNA_N
ClientSampleId :
DFTPP223

364.00	531.0
365.00	20376.0
366.10	3374.0
366.80	417.0
370.00	489.0
370.90	968.0
372.00	6848.0
373.10	2003.0
374.00	339.0
383.00	1992.0
383.80	619.0
384.90	350.0
390.10	704.0
391.00	1092.0
392.00	373.0
400.80	443.0
401.90	2504.0
403.10	4521.0
404.00	1275.0
404.90	315.0
415.00	495.0
421.00	3457.0
422.00	3063.0
423.00	25944.0
424.00	7904.0
425.00	1547.0
441.10	67984.0
442.10	416448.0
443.00	84768.0
444.00	6040.0
445.00	577.0

Instrument :
BNA_N
ClientSampleId :
DFTPP223

m/z	Abundance
36.00	816.0
37.10	2826.0
38.00	7864.0
39.00	45712.0
40.00	3050.0
41.00	1719.0
42.10	380.0
43.00	1026.0
44.00	1537.0
44.90	1157.0
46.90	369.0
48.10	683.0
49.00	5528.0
50.00	168768.0
51.00	622848.0
52.00	31024.0
53.00	1438.0
55.00	3777.0
56.00	16034.0
57.00	39144.0
57.90	1915.0
59.00	499.0
59.90	532.0
61.00	6854.0
62.00	8912.0
63.00	22504.0
64.10	3475.0
65.00	10999.0
66.00	886.0
67.00	1654.0
68.10	11977.0
69.00	601216.0
70.00	2544.0
70.90	536.0
73.10	7511.0
74.00	64712.0
75.00	99600.0
76.10	36648.0
77.00	700096.0
78.00	50504.0
79.00	48384.0
80.00	36248.0
81.00	53496.0
82.00	11644.0
83.00	10626.0
84.00	1634.0
85.00	12232.0
86.00	14596.0
87.00	6099.0
88.00	2360.0
88.90	1363.0
90.10	305.0
91.00	9348.0
92.00	12887.0
93.00	84272.0
94.00	5886.0
95.00	1693.0
96.00	3688.0
97.10	1966.0
98.00	64816.0
99.00	54272.0
100.00	3562.0
101.00	27824.0
102.00	1505.0
103.00	7783.0
104.00	18112.0
105.00	16768.0
106.10	5565.0
107.00	208000.0
108.00	30376.0
109.00	5659.0
110.00	343936.0
111.00	51552.0
112.00	6696.0
113.00	2362.0
114.00	795.0
115.10	1387.0
116.00	10261.0
117.00	189888.0
118.00	13922.0
119.00	1814.0
120.00	2691.0
121.10	1305.0
122.00	12484.0
123.00	22624.0
124.00	9804.0
125.00	8402.0
127.00	669056.0
128.00	53344.0
129.00	254784.0
130.00	21416.0
131.10	4284.0

Instrument :
BNA_N
ClientSampleId :
DFTPP223

131.90	2197.0
133.00	932.0
134.00	7665.0
135.00	21280.0
136.00	7649.0
137.00	11567.0
137.90	2147.0
139.00	1514.0
140.00	3275.0
141.00	32616.0
142.00	10399.0
143.10	7281.0
144.00	1762.0
145.00	1690.0
146.00	6385.0
147.00	15967.0
148.00	44840.0
149.00	8305.0
150.00	2258.0
151.10	3853.0
151.80	2431.0
153.00	8870.0
154.00	8495.0
155.00	19504.0
156.00	25608.0
157.10	5113.0
157.90	4692.0
158.90	4051.0
160.00	9862.0
161.00	14046.0
162.00	3726.0
162.90	1219.0
164.00	1443.0
165.00	11655.0
166.10	8635.0
167.00	70496.0
168.00	39920.0
169.00	5254.0
169.90	1897.0
171.00	2298.0
172.00	5799.0
173.00	7135.0
174.00	13548.0
175.10	24880.0
176.00	5637.0
177.00	11241.0
178.10	4541.0
179.00	45168.0
180.00	31016.0
181.00	14922.0
181.90	2131.0
183.00	1047.0
184.10	4874.0
185.10	21408.0
186.00	170176.0
187.00	49384.0
188.00	4747.0
189.00	10648.0
189.90	2391.0
191.00	5158.0
192.00	14561.0
193.10	15793.0
194.00	3577.0
195.00	1852.0
196.00	31736.0
198.00	1149952.0
199.00	73296.0
200.00	6067.0
201.50	5485.0
203.00	7608.0
204.00	38744.0
205.00	75784.0
206.10	317824.0
207.10	40576.0
208.10	9801.0
209.00	3312.0
210.00	4690.0
210.90	10584.0
213.00	745.0
215.00	3651.0
216.00	6633.0
217.00	75680.0
218.00	10101.0
219.10	1065.0
221.00	41336.0
223.00	17480.0
224.00	161984.0
225.10	43256.0
226.10	5537.0
227.00	67264.0
228.00	10221.0
229.00	12999.0
230.00	1925.0
231.10	4682.0
231.90	935.0
233.00	1079.0

Instrument :
BNA_N
ClientSampleId :
DFTPP223

234.10	4582.0
235.00	5610.0
236.00	3962.0
237.00	6925.0
238.20	1265.0
239.00	3010.0
240.00	2486.0
241.00	3256.0
242.00	8170.0
243.10	9382.0
244.10	123176.0
245.00	17328.0
246.00	26736.0
247.00	5650.0
248.00	1196.0
249.00	4067.0
250.10	913.0
251.00	1243.0
252.00	1202.0
253.00	3429.0
255.00	660736.0
256.00	95368.0
257.00	7701.0
258.00	36064.0
259.00	5525.0
260.00	748.0
261.10	1286.0
263.10	391.0
264.10	1074.0
265.00	13428.0
266.00	2239.0
266.80	490.0
269.00	350.0
270.30	771.0
271.10	1516.0
272.00	1866.0
273.00	20704.0
274.00	52320.0
275.10	292032.0
276.00	41728.0
277.00	25104.0
278.10	4234.0
279.00	1490.0
281.10	869.0
282.10	829.0
283.00	2302.0
284.00	2110.0
285.00	4243.0
285.80	630.0
286.10	723.0
288.90	920.0
290.00	571.0
291.10	397.0
292.10	1507.0
293.00	6343.0
294.10	1948.0
295.00	2157.0
296.00	81184.0
297.10	10356.0
298.00	697.0
301.00	1306.0
302.10	1709.0
303.00	10237.0
304.10	2472.0
305.00	370.0
308.00	1323.0
308.90	840.0
310.00	1153.0
311.00	557.0
313.00	1189.0
313.90	3225.0
314.90	7290.0
316.00	4995.0
317.00	752.0
321.00	2873.0
321.90	1521.0
323.10	29232.0
324.10	5010.0
325.10	524.0
326.20	614.0
327.00	5351.0
327.90	2761.0
329.00	685.0
331.90	1750.0
333.00	2492.0
334.00	16448.0
335.10	4874.0
336.10	495.0
340.00	580.0
341.10	3324.0
342.10	1164.0
346.00	7126.0
347.00	1068.0
349.80	308.0
350.90	496.0
352.00	7618.0

Instrument :
BNA_N
ClientSampleId :
DFTPP223

353.00	5223.0
354.00	10208.0
355.10	1731.0
356.10	471.0
359.00	435.0
364.10	784.0
365.00	35560.0
365.90	6673.0
367.00	553.0
370.10	677.0
371.00	2459.0
372.00	15714.0
373.10	4067.0
374.20	487.0
377.10	304.0
382.30	362.0
383.00	4454.0
384.00	1367.0
384.70	462.0
390.10	1759.0
391.10	1148.0
391.90	802.0
401.00	1273.0
402.00	7130.0
403.00	9111.0
404.00	3319.0
405.10	742.0
415.10	340.0
421.00	7992.0
422.00	6752.0
423.00	48784.0
424.00	13961.0
425.00	1662.0
438.50	392.0
439.00	506.0
441.00	122568.0
442.00	867456.0
443.00	169856.0
444.00	14629.0
444.90	851.0

Instrument :

BNA_N

ClientSampleId :

DFTPP223

m/z	Abundance
35.90	904.0
37.00	2650.0
38.00	6727.0
39.00	39576.0
40.00	2036.0
41.10	1666.0
42.10	312.0
43.00	964.0
44.00	1414.0
45.00	1104.0
46.90	308.0
47.90	634.0
49.10	4504.0
50.00	140672.0
51.00	538944.0
52.10	26616.0
53.00	1129.0
55.00	2803.0
56.00	14342.0
57.00	33728.0
58.00	1868.0
58.90	416.0
60.00	444.0
61.00	7072.0
62.00	8285.0
63.00	21632.0
64.00	2478.0
65.10	9777.0
66.00	1089.0
66.90	1031.0
68.00	9412.0
69.00	538368.0
70.00	2644.0
71.00	338.0
71.90	532.0
73.00	6988.0
74.00	55680.0
75.00	90944.0
76.00	34576.0
77.00	638208.0
78.00	43104.0
79.00	41632.0
80.00	30584.0
81.00	45504.0
82.00	11276.0
83.00	8608.0
84.00	1404.0
85.00	11230.0
86.00	13859.0
87.00	5667.0
88.00	2018.0
89.00	933.0
91.00	9369.0
92.00	11699.0
93.00	72800.0
94.00	4524.0
95.00	1538.0
96.00	3072.0
98.00	59952.0
99.00	44920.0
100.00	3993.0
101.00	23904.0
102.00	1330.0
103.00	6734.0
104.00	15124.0
105.00	14221.0
106.00	4649.0
107.00	202176.0
108.00	26800.0
109.00	6051.0
110.00	329984.0
111.00	52224.0
112.00	6000.0
113.00	1947.0
114.10	714.0
115.00	1187.0
116.00	11184.0
117.00	183232.0
118.00	13245.0
118.90	1679.0
120.00	2048.0
121.10	1259.0
122.00	12616.0
123.00	17600.0
124.00	7360.0
125.00	8886.0
127.00	601216.0
128.00	48848.0
129.00	240064.0
130.00	19896.0
131.00	3871.0
132.00	1918.0

Instrument :
BNA_N
ClientSampleId :
DFTPP223

133.10	1122.0
134.00	5295.0
135.00	19264.0
136.00	8444.0
137.00	10313.0
137.70	1930.0
139.00	1353.0
140.10	2915.0
141.00	30792.0
142.00	9997.0
143.00	7130.0
144.00	1770.0
145.10	1916.0
146.00	5519.0
147.00	17664.0
148.00	38584.0
149.00	8592.0
150.00	2353.0
151.00	3938.0
151.80	2911.0
153.00	9480.0
154.00	6304.0
155.00	15606.0
156.00	23120.0
157.00	3937.0
158.00	5378.0
159.00	3892.0
160.10	10735.0
161.00	14649.0
162.00	4415.0
163.00	1438.0
164.20	1291.0
165.00	10527.0
166.00	9745.0
167.00	62344.0
168.00	39672.0
169.00	5113.0
170.10	2121.0
171.10	2224.0
172.00	5333.0
173.10	6503.0
174.00	12375.0
175.00	23344.0
176.00	5710.0
177.00	11520.0
178.10	3420.0
179.00	43208.0
180.00	30568.0
181.00	14851.0
182.10	2213.0
183.10	1326.0
184.00	3927.0
185.10	20496.0
186.00	167168.0
187.00	47128.0
188.00	5374.0
189.00	8807.0
190.00	1391.0
191.00	4703.0
192.00	15589.0
193.10	17240.0
194.00	4214.0
195.10	2181.0
196.00	32848.0
198.00	1087488.0
199.00	72080.0
200.10	5859.0
201.10	3025.0
201.60	3336.0
203.00	7754.0
204.10	41488.0
205.10	68928.0
206.10	314816.0
207.10	39552.0
208.00	10685.0
209.00	3211.0
210.20	5761.0
211.10	11146.0
212.00	2329.0
213.10	1015.0
214.00	344.0
215.00	2499.0
216.00	7079.0
217.00	75424.0
218.00	9638.0
219.20	806.0
221.00	42976.0
221.90	11072.0
223.10	18112.0
224.10	167744.0
225.10	42360.0
226.10	5035.0
227.00	68768.0
228.00	9282.0
229.00	13248.0
230.10	2113.0

Instrument :
BNA_N
ClientSampleId :
DFTPP223

231.00	5552.0
231.90	1035.0
233.00	1434.0
234.00	4361.0
235.00	5984.0
236.10	3137.0
237.00	6149.0
238.00	1153.0
239.00	2398.0
240.00	2047.0
241.10	3980.0
242.00	9037.0
243.10	10755.0
244.00	127480.0
245.00	17024.0
246.00	24648.0
247.00	5169.0
248.00	1014.0
249.00	4733.0
250.00	939.0
251.00	1524.0
252.10	1471.0
253.00	3201.0
255.00	646080.0
256.00	91600.0
257.00	7942.0
258.00	40800.0
259.00	5893.0
260.00	1149.0
261.00	670.0
262.90	346.0
264.00	1503.0
265.00	16126.0
265.90	2484.0
267.00	309.0
269.80	506.0
271.00	1675.0
272.00	2501.0
273.00	22048.0
274.10	57816.0
275.10	298816.0
276.00	41960.0
277.10	25800.0
278.00	5448.0
278.90	899.0
280.90	911.0
282.00	577.0
283.10	3404.0
284.10	2418.0
285.00	4953.0
286.00	900.0
289.10	991.0
289.90	886.0
291.00	758.0
291.90	1587.0
293.00	5356.0
294.00	1603.0
296.00	85424.0
297.00	11264.0
297.90	668.0
301.00	1321.0
302.00	1679.0
303.10	10362.0
304.10	3104.0
306.00	327.0
308.00	1319.0
309.00	1019.0
310.10	1026.0
311.00	563.0
311.80	394.0
313.00	950.0
314.00	5332.0
315.00	9798.0
316.00	5361.0
317.10	878.0
320.00	383.0
321.00	3420.0
321.90	1389.0
323.10	29320.0
324.10	4729.0
325.10	763.0
326.10	955.0
327.00	6081.0
328.00	3395.0
329.10	526.0
332.00	2195.0
333.00	2944.0
334.00	22208.0
335.00	5368.0
336.00	802.0
339.00	640.0
339.90	623.0
341.00	3682.0
341.90	615.0
345.90	7447.0
347.00	1388.0

Instrument :
BNA_N
ClientSampleId :
DFTPP223

350.20	316.0
350.90	632.0
352.00	9317.0
353.10	7139.0
354.10	10648.0
355.00	1939.0
358.80	796.0
363.80	543.0
365.00	46352.0
366.00	6595.0
366.90	481.0
370.00	1075.0
371.00	2763.0
372.00	17080.0
373.00	4398.0
373.80	544.0
374.00	510.0
377.10	358.0
383.00	4607.0
384.00	1274.0
384.80	504.0
390.00	2406.0
391.00	1697.0
392.10	1711.0
401.00	1084.0
402.00	8562.0
403.00	10171.0
404.00	3454.0
405.00	832.0
415.10	367.0
421.00	8596.0
422.00	8610.0
423.00	67560.0
424.10	15351.0
424.90	1671.0
437.70	435.0
438.90	798.0
441.10	155584.0
442.00	1108992.0
443.10	212288.0
444.00	18312.0
445.00	1183.0

Instrument :

BNA_N

ClientSampleId :

DFTPP223