

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN011720\
 Data File : BN009394.D
 Acq On : 17 Jan 2020 14:03
 Operator : JU
 Sample : L1071-08
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 X2L46

Quant Time: Jan 17 14:37:16 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN011020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 17 12:28:30 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.48	152	1120	0.40	ng/ul	0.00
2) Naphthalene-d8	10.23	136	3642	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.10	164	2151	0.40	ng/ul	0.00
10) Phenanthrene-d10	16.86	188	4347	0.40	ng/ul	0.00
16) Chrysene-d12	21.07	240	3662	0.40	ng/ul	0.00
20) Perylene-d12	23.19	264	3931	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	11.84	152	1536	0.26	ng/ul	0.01
14) Fluoranthene-d10	18.89	212	4113	0.31	ng/ul	0.00
Target Compounds				Ovalue		
3) Naphthalene	10.27	128	10186	0.930	ng/ul	99
7) Acenaphthylene	13.82	152	4363	0.398	ng/ul	98
8) Acenaphthene	14.17	153	4149	0.470	ng/ul	99
9) Fluorene	15.17	166	3705	0.365	ng/ul	100
11) Pentachlorophenol	16.54	266	243	0.204	ng/ul	98
12) Phenanthrene	16.90	178	6814	0.454	ng/ul	100
13) Anthracene	17.00	178	6036	0.453	ng/ul	99
15) Fluoranthene	18.93	202	7720	0.434	ng/ul	99
17) Pyrene	19.29	202	10568	0.601	ng/ul	99
18) Benzo(a)anthracene	21.05	228	7933	0.494	ng/ul	99
21) Benzo(b)fluoranthene	22.56	252	8457	0.485	ng/ul	95
24) Indeno(1,2,3-cd)pyrene	25.25	276	12604	0.647	ng/ul#	95
25) Dibenzo(a,h)anthracene	25.26	278	9522	0.619	ng/ul	98
26) Benzo(g,h,i)perylene	25.88	276	11241	0.659	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN011720\
Data File : BN009394.D
Acq On : 17 Jan 2020 14:03
Operator : JU
Sample : L1071-08
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
X2L46

Quant Time: Jan 17 14:37:16 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN011020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 17 12:28:30 2020
Response via : Initial Calibration

