


Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN011720\
 Data File : BN009389.D
 Acq On : 17 Jan 2020 11:05
 Operator : JU
 Sample : DFTPP06
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP06

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN011020.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Jan 17 12:28:30 2020



AutoFind: Scans 2437, 2438, 2439; Background Corrected with Scan 2430

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	62.0	97142	PASS
68	69	0.00	2	1.5	1077	PASS
69	198	0.00	100	45.8	71730	PASS
70	69	0.00	2	0.4	300	PASS
127	198	10	80	48.4	75821	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	156664	PASS
199	198	5	9	6.7	10512	PASS
275	198	10	60	24.2	37866	PASS
365	198	1	100	3.7	5832	PASS
441	198	0.01	100	9.9	15468	PASS
442	198	50	100	64.4	100882	PASS
443	442	15	24	17.1	17246	PASS

Scan 2430 (17.281 min): BN009389.D
DFTPP06

m/z	Abundance
40.00	1745.0
41.00	409.0
43.10	327.0
44.00	7041.0
49.00	773.0
50.90	324.0
72.90	701.0
84.00	388.0
85.90	311.0
207.00	1148.0

Instrument :
BNA_N
ClientSampleId :
DFTPP06

m/z	Abundance
36.90	595.0
38.00	1871.0
39.00	11056.0
40.00	1893.0
40.90	479.0
42.90	339.0
44.00	6792.0
44.90	417.0
49.00	1160.0
50.00	34184.0
51.00	131840.0
52.00	6375.0
53.00	523.0
55.00	1192.0
56.00	4032.0
57.00	8828.0
58.00	510.0
61.00	1715.0
62.10	2058.0
63.00	4739.0
64.00	619.0
65.00	2103.0
68.10	2064.0
69.00	95616.0
70.00	406.0
73.00	1709.0
74.00	10200.0
75.00	16448.0
76.00	4680.0
77.00	103960.0
78.00	6644.0
79.00	7226.0
80.00	5402.0
81.00	7661.0
81.90	1663.0
83.00	1776.0
83.90	531.0
84.90	1158.0
85.90	2322.0
87.00	847.0
88.00	330.0
89.10	407.0
91.00	1598.0
92.00	2185.0
93.00	12827.0
94.00	1027.0
95.90	874.0
98.00	9761.0
99.00	9070.0
99.90	918.0
101.00	4819.0
103.00	923.0
103.90	2948.0
105.00	2620.0
106.00	1021.0
107.00	30680.0
108.00	5022.0
108.90	827.0
110.00	52048.0
111.00	7884.0
111.90	951.0
113.00	399.0
115.90	1599.0
117.00	29976.0
118.00	1939.0
119.00	652.0
119.80	415.0
121.90	2113.0
123.00	3214.0
124.00	1586.0
125.00	994.0
126.00	321.0
127.00	94744.0
128.00	6898.0
129.00	38464.0
130.00	3091.0
131.00	708.0
131.90	336.0
133.90	832.0
135.00	2702.0
135.90	1255.0
136.90	2456.0
138.00	309.0
139.90	808.0
140.90	5116.0
141.90	1616.0
143.00	1302.0
144.00	373.0
144.90	360.0
146.00	1006.0
147.00	2753.0
148.00	6643.0

Instrument :
BNA_N
ClientSampleId :
DFTPP06

149.00	1676.0
151.10	586.0
151.80	402.0
153.00	1782.0
153.80	1262.0
155.00	2631.0
156.00	4302.0
157.00	639.0
157.90	941.0
159.00	609.0
159.90	1293.0
161.00	2167.0
161.90	586.0
164.10	397.0
165.00	1912.0
166.00	1939.0
167.00	11092.0
167.90	7052.0
169.00	1153.0
169.90	748.0
171.00	403.0
171.80	1030.0
172.90	1256.0
174.00	2514.0
175.00	3887.0
176.10	1037.0
176.90	1397.0
177.90	742.0
178.90	8276.0
180.00	5736.0
180.90	2680.0
181.90	505.0
184.00	587.0
184.90	2891.0
186.00	27832.0
187.00	7649.0
187.90	1323.0
188.90	1930.0
190.00	330.0
191.00	1189.0
191.90	2520.0
193.00	3256.0
194.00	795.0
194.80	477.0
196.00	5041.0
197.90	181312.0
198.90	12916.0
199.80	966.0
201.30	731.0
203.00	1448.0
204.00	6822.0
205.00	11050.0
206.00	49512.0
207.00	8528.0
207.90	2389.0
209.00	1085.0
209.90	940.0
211.00	2025.0
211.80	556.0
215.00	799.0
215.90	1335.0
216.90	14334.0
217.90	1834.0
221.00	6631.0
221.90	1596.0
223.00	2586.0
224.00	26600.0
225.00	6116.0
225.90	1057.0
227.00	12149.0
227.90	2012.0
229.00	2268.0
229.90	419.0
231.00	936.0
233.90	997.0
235.00	955.0
235.90	616.0
237.00	1558.0
238.90	719.0
239.80	433.0
240.90	474.0
241.90	1591.0
243.00	1277.0
244.00	19256.0
245.10	2892.0
245.90	4239.0
247.10	858.0
249.00	633.0
253.00	600.0
255.00	97768.0
256.00	14622.0
256.90	1471.0
258.00	6174.0
258.90	1112.0
264.90	2264.0
271.00	443.0

Instrument :
BNA_N
ClientSampleId :
DFTPP06

272.90	3450.0
274.00	6629.0
275.00	39408.0
276.00	5008.0
276.90	4219.0
277.90	968.0
280.90	544.0
283.00	326.0
285.00	687.0
292.90	864.0
294.90	352.0
296.00	12673.0
297.00	1687.0
303.00	1384.0
303.90	371.0
314.00	524.0
314.90	1343.0
315.90	901.0
320.90	385.0
323.00	3932.0
323.90	671.0
326.80	678.0
327.90	683.0
331.80	411.0
333.10	607.0
334.00	2589.0
335.00	518.0
341.10	437.0
345.90	738.0
346.90	411.0
351.90	1375.0
352.80	762.0
354.00	1088.0
364.90	5468.0
365.90	615.0
370.80	352.0
371.90	2572.0
382.80	638.0
402.00	867.0
402.90	1046.0
403.90	490.0
420.90	532.0
421.90	894.0
422.90	5970.0
423.90	2669.0
425.00	542.0
441.00	14029.0
442.00	88904.0
443.00	15866.0
444.00	1804.0

Instrument :
BNA_N
ClientSampleId :
DFTPP06

m/z	Abundance
36.90	597.0
38.10	1072.0
39.10	7955.0
40.00	1993.0
41.10	749.0
43.10	486.0
44.00	6412.0
45.10	413.0
47.80	368.0
49.10	1768.0
50.00	28792.0
51.00	110648.0
52.00	5869.0
53.10	438.0
55.00	1262.0
56.00	3697.0
57.00	7635.0
58.00	420.0
61.00	1232.0
62.00	1419.0
63.00	4119.0
64.00	637.0
65.10	1691.0
67.80	1267.0
69.00	82080.0
69.90	496.0
73.00	1406.0
74.00	8264.0
75.00	13614.0
76.00	4564.0
77.00	94552.0
78.00	6986.0
79.00	6363.0
79.90	5563.0
81.00	6401.0
82.00	1676.0
82.90	1315.0
84.00	333.0
85.00	1198.0
86.00	1934.0
87.00	988.0
88.90	312.0
90.90	1861.0
91.90	1562.0
93.00	10881.0
94.00	921.0
94.90	390.0
96.00	738.0
97.00	376.0
98.00	8338.0
99.00	8599.0
99.90	773.0
100.90	3437.0
102.90	1502.0
104.00	2453.0
104.90	2894.0
106.00	764.0
107.00	28208.0
108.00	4587.0
109.00	795.0
110.00	48424.0
111.00	7448.0
112.00	1498.0
113.00	431.0
115.90	1541.0
117.00	28128.0
117.90	2258.0
119.90	460.0
122.00	1979.0
122.90	3396.0
124.00	1218.0
125.00	1441.0
127.00	90416.0
128.00	6728.0
129.00	36384.0
130.00	2502.0
131.00	596.0
132.10	355.0
133.00	310.0
133.90	918.0
134.90	3529.0
135.90	1109.0
136.90	1865.0
137.90	407.0
140.00	603.0
141.00	4941.0
141.90	1228.0
143.00	1192.0
144.00	328.0
145.00	343.0
145.90	1315.0
146.90	3249.0

Instrument :
BNA_N
ClientSampleId :
DFTPP06

148.00	6216.0
149.00	1359.0
150.00	311.0
151.10	615.0
151.80	433.0
153.00	1591.0
154.00	1102.0
154.90	3074.0
156.00	4281.0
157.00	1002.0
158.00	1132.0
158.90	702.0
160.00	1320.0
161.00	2745.0
161.80	981.0
163.90	317.0
164.90	2034.0
166.00	1245.0
167.00	10572.0
167.90	7171.0
169.00	648.0
170.00	605.0
170.50	323.0
171.00	568.0
171.80	1209.0
173.00	1218.0
174.00	2251.0
175.00	4735.0
176.00	840.0
177.00	2227.0
177.90	798.0
178.90	8396.0
180.00	6207.0
181.00	2689.0
182.10	518.0
183.90	737.0
185.00	3574.0
186.00	27344.0
187.00	7953.0
188.00	1102.0
189.00	2354.0
189.90	464.0
191.10	1138.0
192.00	2506.0
193.00	2758.0
193.90	494.0
195.00	391.0
195.90	5251.0
197.90	190592.0
199.00	12291.0
200.00	829.0
201.30	843.0
203.00	1532.0
204.00	6799.0
205.00	12713.0
206.00	53072.0
207.00	7396.0
208.00	2447.0
209.10	1001.0
210.10	1253.0
211.00	1855.0
211.90	505.0
214.90	584.0
215.90	1423.0
216.90	14178.0
218.00	2160.0
221.00	7211.0
222.00	1674.0
223.00	3284.0
224.00	28704.0
225.00	6915.0
226.00	1036.0
227.00	14567.0
227.90	1982.0
228.90	2862.0
229.90	346.0
231.00	773.0
233.90	920.0
234.90	872.0
235.90	841.0
237.00	1573.0
238.90	623.0
239.90	557.0
240.90	878.0
241.90	1409.0
243.10	1624.0
244.00	21432.0
245.00	2629.0
245.90	4535.0
246.90	1283.0
249.00	673.0
253.00	581.0
254.00	857.0
255.00	106864.0
256.00	15884.0
256.90	1567.0

Instrument :
BNA_N
ClientSampleId :
DFTPP06

257.90	7430.0
259.00	1453.0
265.00	2560.0
265.90	442.0
271.20	443.0
272.90	3553.0
274.00	9322.0
274.90	45024.0
276.00	6517.0
277.00	4478.0
277.90	1061.0
280.90	651.0
282.90	617.0
285.00	742.0
291.80	365.0
292.90	1016.0
296.00	14821.0
296.90	1826.0
303.00	2106.0
303.90	564.0
314.00	737.0
314.90	1697.0
315.90	774.0
320.90	762.0
323.00	4141.0
324.00	823.0
327.00	958.0
328.00	516.0
332.00	402.0
333.00	565.0
334.00	2867.0
335.00	917.0
341.00	515.0
346.00	998.0
351.90	1207.0
352.90	810.0
354.00	1414.0
365.00	7062.0
366.00	713.0
371.00	383.0
372.00	2897.0
373.00	662.0
382.90	633.0
400.90	357.0
402.00	987.0
402.90	1330.0
403.90	664.0
420.90	771.0
421.90	977.0
422.90	7807.0
423.90	2695.0
424.90	392.0
440.90	18888.0
442.00	127808.0
443.00	21768.0
444.00	2144.0

Instrument :
BNA_N
ClientSampleId :
DFTPP06

m/z	Abundance
37.90	698.0
39.10	4325.0
39.90	1490.0
41.00	356.0
43.20	680.0
44.10	6243.0
45.00	326.0
49.00	815.0
50.00	13405.0
51.00	49912.0
52.00	2691.0
55.00	758.0
55.90	2043.0
56.90	3651.0
61.10	616.0
61.90	732.0
63.10	2025.0
64.00	352.0
65.00	966.0
68.10	1168.0
69.00	37496.0
73.00	1254.0
74.00	4181.0
75.00	5905.0
76.10	2404.0
77.00	42752.0
78.00	2772.0
78.90	2464.0
80.00	2269.0
80.90	3590.0
82.00	737.0
82.90	893.0
83.90	337.0
85.00	447.0
86.00	940.0
86.90	540.0
91.00	962.0
92.00	993.0
93.00	5448.0
93.80	510.0
96.00	619.0
98.00	4346.0
99.00	3296.0
99.90	414.0
101.00	1553.0
102.90	615.0
104.00	1161.0
105.00	1259.0
106.10	394.0
107.00	12684.0
107.90	1776.0
109.10	330.0
109.90	20256.0
110.90	3562.0
111.90	606.0
116.20	667.0
117.00	12651.0
118.00	1045.0
122.00	932.0
123.00	1384.0
124.00	488.0
124.90	604.0
126.20	477.0
127.00	42304.0
128.00	3083.0
129.00	17808.0
130.00	1260.0
130.80	448.0
132.80	344.0
133.90	570.0
134.90	1171.0
136.00	619.0
136.80	742.0
140.90	2635.0
141.90	944.0
143.00	696.0
145.90	732.0
147.00	1618.0
147.90	3093.0
149.00	853.0
151.10	637.0
153.00	818.0
154.00	490.0
154.90	1407.0
156.00	2539.0
156.90	377.0
157.80	396.0
160.00	961.0
160.80	1009.0
161.90	438.0
165.00	947.0
166.00	665.0

Instrument :
BNA_N
ClientSampleId :
DFTPP06

167.00	4652.0
168.00	3619.0
168.90	629.0
171.90	379.0
173.10	749.0
174.00	1248.0
175.00	2222.0
176.00	1019.0
176.90	1295.0
178.90	4204.0
179.90	3103.0
181.00	1881.0
184.00	335.0
185.00	1842.0
186.00	13892.0
187.00	3917.0
188.00	418.0
188.90	1246.0
191.00	716.0
191.90	1377.0
192.90	1826.0
193.90	485.0
196.00	2653.0
197.90	98088.0
199.00	6331.0
200.00	715.0
201.60	619.0
203.00	701.0
203.90	4096.0
205.00	6989.0
206.00	28528.0
207.00	4973.0
208.00	1133.0
209.00	646.0
210.20	627.0
211.00	1072.0
211.70	327.0
214.80	339.0
215.90	581.0
217.00	8626.0
217.90	1067.0
221.00	4059.0
222.00	970.0
223.00	1915.0
224.00	17056.0
225.00	4090.0
226.00	440.0
227.00	6791.0
227.90	1138.0
229.00	1420.0
231.10	482.0
234.00	705.0
234.90	775.0
235.90	534.0
236.80	845.0
240.90	686.0
242.00	1103.0
243.00	940.0
244.00	12518.0
245.00	1456.0
245.90	2943.0
246.90	868.0
252.80	333.0
254.10	622.0
255.00	68752.0
256.00	8985.0
257.00	791.0
258.00	4110.0
258.80	1017.0
265.00	1881.0
265.90	386.0
272.90	2036.0
274.00	4587.0
275.00	29168.0
275.90	3226.0
276.90	2991.0
278.00	994.0
281.00	614.0
283.00	489.0
283.90	340.0
285.00	563.0
293.00	691.0
296.00	9046.0
297.00	1348.0
301.90	303.0
302.90	881.0
304.00	454.0
314.00	601.0
315.00	1299.0
315.90	404.0
320.90	484.0
323.00	2897.0
323.90	349.0
326.70	523.0
328.00	372.0
333.00	321.0

Instrument :
BNA_N
ClientSampleId :
DFTPP06

334.00	2090.0
335.00	375.0
340.80	550.0
345.80	518.0
352.00	1009.0
353.00	657.0
353.90	850.0
364.90	4966.0
366.00	724.0
371.20	305.0
372.00	1578.0
373.10	447.0
382.90	756.0
401.90	961.0
402.80	760.0
404.00	548.0
420.90	647.0
421.90	785.0
423.00	5046.0
424.00	1904.0
441.00	13487.0
442.00	85936.0
443.00	14105.0
443.90	1240.0

Instrument :
BNA_N
ClientSampleId :
DFTPP06