

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012125\
 Data File : BN036007.D
 Acq On : 21 Jan 2025 15:35
 Operator : RC/JU
 Sample : SSTD3.260
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD3.2249

Quant Time: Jan 21 23:46:36 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN012125.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 21 23:44:46 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.816	152	2166	0.400	ng/ul	0.00
4) Naphthalene-d8	10.609	136	4281	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.449	164	2258	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.191	188	4416	0.400	ng/ul	0.00
17) Chrysene-d12	21.372	240	4305	0.400	ng/ul	0.00
23) Perylene-d12	23.666	264	4121	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.276	96	8018	3.381	ng/ul	0.00
6) 2-Methylnaphthalene-d10	0.000	152	0d	0.000	ng/ul	
18) Fluoranthene-d10	0.000	212	0d	0.000	ng/ul	
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.310	88	8857	3.499	ng/ul	88
14) Pentachlorophenol	16.841	266	5540	3.581	ng/ul	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012125\
Data File : BN036007.D
Acq On : 21 Jan 2025 15:35
Operator : RC/JU
Sample : SSTD3.260
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD3.2249

Quant Time: Jan 21 23:46:36 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN012125.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 21 23:44:46 2025
Response via : Initial Calibration

