

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN012221\
 Data File : BN013459.D
 Acq On : 22 Jan 2021 15:23
 Operator : CG/JU
 Sample : M1050-07
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 X2X39

Quant Time: Jan 22 15:57:55 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-BN012221.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jan 22 09:47:41 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.58	152	487875	20.00	ng/ul	0.00
20) Naphthalene-d8	10.35	136	1900067	20.00	ng/ul	0.00
38) Acenaphthene-d10	14.22	164	998753	20.00	ng/ul	0.00
64) Phenanthrene-d10	16.97	188	1768966	20.00	ng/ul	0.00
79) Chrysene-d12	21.17	240	1403024	20.00	ng/ul	0.00
88) Perylene-d12	23.36	264	1450042	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.18	96	42905	3.59	ng/uL	0.00
4) Pyridine-d5	3.56	84	622993	19.95	ng/ul	0.00
7) Phenol-d5	6.76	99	852072	22.21	ng/ul	0.00
9) Bis-(2-Chloroethyl)ether-d	6.93	67	497006	22.04	ng/ul	0.00
11) 2-Chlorophenol-d4	7.12	132	747454	22.27	ng/ul	0.00
15) 4-Methylphenol-d8	8.28	113	645352	21.20	ng/ul	0.00
21) Nitrobenzene-d5	8.72	128	327745	21.98	ng/ul	0.00
24) 2-Nitrophenol-d4	9.43	143	349900	22.86	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	9.97	165	624204	21.30	ng/ul	0.00
31) 4-Chloroaniline-d4	10.48	131	720899	17.43	ng/ul	0.00
46) Dimethylphthalate-d6	13.63	166	2113158	28.52	ng/ul	0.00
49) Acenaphthylene-d8	13.90	160	2397681	25.44	ng/ul	0.00
54) 4-Nitrophenol-d4	14.42	143	371706	26.79	ng/ul	0.00
60) Fluorene-d10	15.22	176	1756453	27.26	ng/ul	0.00
65) 4,6-Dinitro-2-methylphenol	15.33	200	232096	22.52	ng/ul	0.00
73) Anthracene-d10	17.07	188	2839636	33.22	ng/ul	0.00
81) Pyrene-d10	19.37	212	2810365	33.69	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.22	264	2784611	36.81	ng/ul	0.00

Target Compounds

					Ovalue
8) Phenol	6.79	94	1755238	44.299	ng/ul 99
22) Nitrobenzene	8.76	77	1215813	35.930	ng/ul 99
23) Isophorone	9.28	82	2573508	44.258	ng/ul 100
30) Naphthalene	10.39	128	5182088	48.310	ng/ul 100
36) 2-Methylnaphthalene	12.02	142	2727572	38.034	ng/ul 99
41) 2,4,6-Trichlorophenol	12.63	196	1059848	54.775	ng/ul 98
43) 1,1'-Biophenyl	13.05	154	2772391	32.902	ng/ul 97
50) Acenaphthylene	13.93	152	2271534	23.394	ng/ul 98
56) Dibenzofuran	14.62	168	4298733	46.325	ng/ul 98
59) Diethylphthalate	15.06	149	3054660	41.032	ng/ul 99
69) Hexachlorobenzene	16.27	284	868033	38.004	ng/ul 97
71) Pentachlorophenol	16.62	266	795292	65.240	ng/ul 98
74) Anthracene	17.10	178	5178235	48.839	ng/ul 98
77) Carbazole	17.37	167	5650331	63.774	ng/ul 99
80) Fluoranthene	19.04	202	5112626	47.208	ng/ul 99
85) Benzo(a)anthracene	21.16	228	3861901	41.435	ng/ul 99
86) Bis(2-ethylhexyl)phthalate	21.12	149	3022049	58.069	ng/ul 100
89) Di-n-octyl phthalate	21.98	149	5621472	62.968	ng/ul 100
94) Indeno(1,2,3-cd)pyrene	25.54	276	5488253	50.756	ng/ul 99
95) Dibenzo(a,h)anthracene	25.56	278	4851885	53.713	ng/ul 98
96) Benzo(a,h,i)perylene	26.21	276	4967726	55.268	ng/ul 94

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN012221\
Data File : BN013459.D
Acq On : 22 Jan 2021 15:23
Operator : CG/JU
Sample : M1050-07
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
X2X39

Quant Time: Jan 22 15:57:55 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-BN012221.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Jan 22 09:47:41 2021
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
--------------------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN012221\
Data File : BN013459.D
Acq On : 22 Jan 2021 15:23
Operator : CG/JU
Sample : M1050-07
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
X2X39

Quant Time: Jan 22 15:57:55 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-BN012221.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Jan 22 09:47:41 2021
Response via : Initial Calibration

