


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012319\
Data File : BN004335.D
Acq On : 23 Jan 2019 09:53
Operator : JU/SJ
Sample : DFTPP57
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN010919MA.M
Title : SVOA CALIBRATION
Last Update : Wed Jan 09 12:45:02 2019



AutoFind: Scans 2479, 2480, 2481; Background Corrected with Scan 2474

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	25.3	40088	PASS
68	69	0.00	2	1.8	819	PASS
69	198	0.00	100	28.5	45125	PASS
70	69	0.00	2	0.2	102	PASS
127	198	10	80	43.0	68066	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	158378	PASS
199	198	5	9	6.8	10712	PASS
275	198	10	60	29.0	45928	PASS
365	198	1	100	3.8	5982	PASS
441	198	0.01	100	11.9	18810	PASS
442	198	50	100	74.7	118314	PASS
443	442	15	24	19.7	23357	PASS

m/z	Abundance
38.00	461.0
39.00	2769.0
44.00	977.0
49.00	324.0
50.00	10457.0
51.00	40208.0
52.00	1955.0
56.00	1224.0
57.00	2829.0
61.00	567.0
62.00	716.0
63.00	1702.0
65.00	825.0
68.00	819.0
69.00	44312.0
73.00	480.0
74.00	4763.0
75.00	7265.0
76.00	2587.0
77.00	53048.0
78.00	3821.0
79.00	3632.0
80.00	2636.0
81.00	3897.0
82.00	926.0
83.00	865.0
85.00	661.0
85.90	1168.0
87.00	501.0
91.00	904.0
92.00	1074.0
93.00	6087.0
94.00	425.0
95.90	307.0
98.00	4901.0
99.00	4553.0
100.00	417.0
101.00	2168.0
102.90	830.0
104.00	1369.0
105.00	1283.0
106.00	376.0
107.00	16952.0
108.00	2709.0
109.10	533.0
110.00	32536.0
111.00	4886.0
111.90	546.0
115.90	912.0
117.00	14482.0
118.00	1168.0
122.00	1183.0
123.00	1884.0
124.00	806.0
124.90	724.0
127.00	63392.0
128.00	4903.0
129.00	24248.0
130.00	2090.0
131.00	420.0
134.00	665.0
135.00	2063.0
136.00	812.0
137.00	1239.0
141.00	3109.0
142.00	1063.0
142.90	813.0
146.00	603.0
147.00	1543.0
148.00	3620.0
149.00	849.0
151.10	423.0
153.00	1129.0
154.00	805.0
155.00	1853.0
156.00	2652.0
157.00	564.0
158.00	572.0
158.90	474.0
160.00	975.0
161.00	1634.0
162.00	467.0
165.00	1150.0
166.00	1069.0
167.00	6390.0
168.00	4522.0
169.00	641.0
172.00	619.0
173.00	782.0
174.00	1262.0
175.00	2456.0

Instrument :
BNA_N
ClientSampleId :

176.00	775.0
176.90	1293.0
178.00	421.0
179.00	4825.0
180.00	3759.0
181.00	1530.0
183.90	361.0
185.00	2357.0
186.00	18352.0
187.00	5497.0
188.00	662.0
189.00	1161.0
191.00	600.0
192.00	1585.0
193.00	1798.0
194.00	342.0
196.00	3918.0
198.00	134464.0
199.00	9808.0
200.00	647.0
201.60	601.0
203.00	949.0
204.00	4806.0
205.00	8399.0
206.00	35552.0
207.00	5716.0
208.00	1437.0
209.00	553.0
210.20	624.0
211.00	1463.0
214.90	356.0
216.00	731.0
217.00	9032.0
218.00	1309.0
221.00	5874.0
223.00	2104.0
224.00	19744.0
225.00	5126.0
226.00	791.0
227.00	8354.0
228.00	1216.0
229.00	1835.0
231.00	761.0
234.00	494.0
235.00	550.0
235.90	363.0
237.00	964.0
239.00	331.0
239.90	316.0
241.00	533.0
242.00	1122.0
243.00	1094.0
244.00	15281.0
245.00	2185.0
246.00	3104.0
247.00	670.0
248.90	545.0
255.00	78728.0
256.00	11623.0
257.00	1174.0
258.00	4508.0
259.00	769.0
265.00	1894.0
273.00	2406.0
274.00	6271.0
275.00	36528.0
276.00	5003.0
277.00	2960.0
278.00	596.0
283.00	347.0
285.00	551.0
293.00	638.0
296.00	9703.0
297.00	1436.0
303.00	1070.0
304.00	303.0
314.00	481.0
315.00	1064.0
316.00	606.0
323.00	3034.0
324.00	602.0
326.90	655.0
328.00	338.0
334.00	2055.0
335.00	556.0
340.90	378.0
346.00	673.0
352.00	932.0
353.00	605.0
354.00	1000.0
365.00	4253.0
365.90	658.0
372.00	1562.0
373.00	428.0
383.00	453.0

Instrument :
BNA_N
ClientSampleId :

402.00	611.0
403.00	977.0
404.00	379.0
421.00	724.0
422.00	744.0
423.00	4897.0
424.00	1666.0
441.00	11648.0
442.00	74816.0
443.00	15640.0
444.10	1378.0

Instrument :
BNA_N
ClientSampleId :

m/z	Abundance
38.00	559.0
39.00	3301.0
44.00	1059.0
49.00	398.0
50.00	12229.0
51.00	47888.0
52.00	2374.0
56.00	1560.0
57.00	3279.0
61.00	614.0
62.00	749.0
63.00	2022.0
64.00	318.0
65.00	1011.0
68.00	977.0
69.00	54200.0
70.00	308.0
73.00	474.0
74.00	5587.0
75.00	9082.0
76.10	3000.0
77.00	64416.0
78.00	4461.0
79.00	4391.0
80.00	3216.0
81.00	4678.0
82.00	1068.0
83.00	1011.0
85.00	838.0
86.00	1354.0
86.90	629.0
91.00	1091.0
92.00	1235.0
93.00	7861.0
93.90	541.0
96.00	382.0
98.00	5798.0
99.00	5084.0
100.00	415.0
101.00	2948.0
102.90	951.0
104.00	1894.0
105.00	1667.0
106.00	558.0
107.00	21896.0
108.00	3462.0
109.00	618.0
110.00	40888.0
111.00	6222.0
112.00	759.0
116.00	1102.0
117.00	18736.0
118.00	1393.0
122.00	1606.0
123.00	2329.0
124.00	1078.0
125.00	899.0
127.00	81328.0
128.00	6174.0
129.00	31288.0
130.00	2700.0
131.00	550.0
132.00	305.0
134.00	866.0
135.00	2572.0
136.00	1090.0
137.00	1562.0
139.90	377.0
141.00	4034.0
142.00	1319.0
143.00	907.0
144.00	310.0
145.90	637.0
147.00	2185.0
148.00	4737.0
149.00	1016.0
151.10	530.0
151.80	352.0
153.00	1358.0
153.90	1039.0
155.00	2300.0
156.00	3564.0
157.00	687.0
158.00	806.0
159.00	681.0
160.00	1425.0
161.00	2137.0
161.90	676.0
165.00	1547.0
166.00	1320.0
167.00	8952.0

Instrument :
BNA_N
ClientSampleId :

168.00	4962.0
169.00	843.0
171.00	400.0
171.90	785.0
173.00	996.0
174.00	1818.0
175.00	3694.0
176.00	939.0
176.90	1709.0
177.90	579.0
179.00	6497.0
180.00	4677.0
181.00	2182.0
182.00	341.0
184.00	599.0
185.00	3288.0
186.00	25056.0
187.00	7322.0
188.00	828.0
189.00	1634.0
191.10	901.0
192.00	2238.0
193.00	2483.0
194.00	479.0
196.00	5658.0
198.00	192192.0
199.00	12803.0
200.00	1006.0
201.40	781.0
203.00	1318.0
204.00	6969.0
205.00	11966.0
206.10	50648.0
207.00	7633.0
208.00	1805.0
209.00	649.0
210.30	899.0
211.00	2012.0
211.80	503.0
214.90	601.0
216.10	1038.0
217.00	13094.0
218.00	1707.0
221.00	8518.0
221.80	2149.0
223.00	3063.0
224.00	28768.0
225.00	7651.0
226.00	899.0
227.00	11715.0
228.00	1742.0
229.00	2606.0
231.00	974.0
234.00	713.0
235.00	888.0
236.00	531.0
237.00	1246.0
238.90	445.0
240.00	384.0
241.00	715.0
242.00	1546.0
243.00	1608.0
244.10	23128.0
245.00	3160.0
246.00	4724.0
247.00	1035.0
248.90	844.0
252.90	479.0
255.00	120128.0
256.00	17776.0
257.00	1407.0
258.00	6485.0
259.00	1084.0
265.00	2732.0
265.90	456.0
272.00	311.0
273.00	3713.0
274.00	9482.0
275.00	55040.0
276.00	7533.0
277.00	4617.0
278.00	820.0
283.00	567.0
284.00	433.0
285.00	763.0
293.00	960.0
295.10	386.0
296.00	15836.0
297.00	2181.0
303.00	1848.0
304.00	508.0
314.00	729.0
315.00	1711.0
316.00	883.0
320.90	463.0

Instrument :
BNA_N
ClientSampleId :

323.00	4962.0
324.10	850.0
326.90	992.0
328.00	516.0
332.00	353.0
333.00	508.0
334.00	3379.0
335.00	928.0
340.90	556.0
345.90	1115.0
352.00	1551.0
353.00	1085.0
354.00	1670.0
355.00	324.0
365.00	7136.0
366.00	1046.0
371.00	424.0
372.00	2634.0
373.00	659.0
383.00	747.0
390.00	359.0
402.00	1060.0
403.00	1488.0
404.00	604.0
421.00	1269.0
422.00	1196.0
423.00	8565.0
424.00	2359.0
441.00	22600.0
442.00	139200.0
443.10	27912.0
444.00	2562.0

Instrument :
BNA_N
ClientSampleId :

m/z	Abundance
38.00	349.0
39.00	2212.0
44.00	1047.0
50.00	8160.0
51.00	32168.0
52.00	1680.0
55.90	975.0
57.00	2121.0
61.00	455.0
61.90	463.0
63.00	1389.0
65.00	646.0
68.10	661.0
69.00	36864.0
73.00	341.0
74.00	3962.0
75.00	6065.0
76.00	2041.0
77.00	44368.0
78.00	2984.0
79.00	2985.0
80.00	2383.0
81.00	3268.0
82.00	780.0
83.00	746.0
85.00	603.0
86.00	1006.0
87.00	422.0
91.00	746.0
91.90	846.0
93.00	5214.0
93.90	355.0
95.90	305.0
98.00	4347.0
99.00	3619.0
99.90	327.0
101.00	2015.0
103.00	649.0
104.00	1207.0
105.00	1170.0
106.00	404.0
107.00	15464.0
108.00	2337.0
108.90	388.0
110.00	28192.0
111.00	4238.0
112.00	556.0
116.00	836.0
117.00	12996.0
117.90	998.0
122.00	1070.0
123.00	1622.0
124.00	824.0
125.00	724.0
127.00	59480.0
128.00	4187.0
129.00	22448.0
130.00	1932.0
130.90	320.0
133.90	745.0
135.00	1958.0
136.00	817.0
137.00	993.0
141.00	2918.0
141.90	956.0
143.00	753.0
146.00	527.0
147.00	1583.0
148.00	3473.0
148.90	739.0
151.10	396.0
153.00	1100.0
153.90	730.0
155.00	1664.0
156.00	2687.0
157.00	513.0
158.00	650.0
158.90	512.0
160.00	874.0
161.00	1568.0
162.00	456.0
165.00	1232.0
166.00	947.0
167.00	6461.0
168.00	3517.0
169.00	540.0
171.90	584.0
173.00	715.0
174.00	1408.0
175.00	2735.0
176.00	800.0

Instrument :
BNA_N
ClientSampleId :

177.00	1212.0
178.00	423.0
179.00	4833.0
180.00	3671.0
181.00	1741.0
183.90	410.0
185.00	2546.0
186.00	19456.0
187.00	5787.0
188.00	614.0
189.00	1165.0
191.00	636.0
192.00	1654.0
193.00	1823.0
193.90	403.0
196.00	4246.0
198.00	148480.0
199.00	9525.0
200.00	831.0
201.60	641.0
203.00	974.0
204.00	5325.0
205.00	9160.0
206.00	39064.0
207.00	5948.0
208.00	1463.0
209.00	608.0
210.20	655.0
211.00	1576.0
215.00	458.0
216.00	875.0
217.00	10415.0
218.00	1317.0
221.00	7104.0
221.80	1459.0
223.00	2324.0
224.00	23032.0
225.00	5979.0
226.00	610.0
227.00	9254.0
228.00	1324.0
229.00	1956.0
231.00	847.0
234.00	591.0
235.00	741.0
235.90	465.0
237.00	867.0
238.90	384.0
239.90	329.0
240.90	526.0
242.00	1175.0
243.00	1373.0
244.00	18312.0
245.10	2677.0
246.00	3668.0
247.00	752.0
248.90	740.0
253.00	414.0
255.00	97136.0
256.00	13657.0
257.00	1160.0
258.00	5398.0
259.00	908.0
265.00	2577.0
265.90	341.0
273.00	3120.0
274.00	7738.0
275.00	46216.0
276.00	6074.0
277.00	3725.0
278.00	668.0
283.00	511.0
285.00	719.0
293.00	858.0
296.00	13190.0
297.00	1825.0
303.00	1519.0
304.00	376.0
314.00	580.0
315.00	1526.0
316.00	871.0
321.00	449.0
323.00	4180.0
324.00	780.0
326.90	936.0
328.00	481.0
333.00	392.0
334.00	3016.0
335.00	761.0
341.00	505.0
346.00	983.0
352.00	1423.0
353.00	979.0
354.00	1502.0
365.00	6559.0

Instrument :
BNA_N
ClientSampleId :

366.00	913.0
371.00	380.0
372.00	2576.0
373.00	596.0
383.00	640.0
390.00	346.0
402.00	945.0
403.00	1466.0
403.90	462.0
421.00	1236.0
422.00	1187.0
423.00	8986.0
424.00	1986.0
441.10	22184.0
442.00	140928.0
443.10	26520.0
444.00	2478.0

Instrument :
BNA_N
ClientSampleId :

m/z	Abundance
44.00	1253.0
207.00	1020.0

Instrument :
BNA_N
ClientSampleId :