

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN013123\
 Data File : BN023878.D
 Acq On : 01 Feb 2023 01:24
 Operator : CG/JU
 Sample : 01278-02
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 EX9G9

Quant Time: Feb 01 02:50:57 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN013023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 01 01:00:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.859	152	4220	0.400	ng/ul	0.00
4) Naphthalene-d8	10.667	136	11944	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.507	164	7698	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.256	188	18107	0.400	ng/ul	# 0.00
17) Chrysene-d12	21.447	240	17661	0.400	ng/ul	# 0.00
23) Perylene-d12	23.844	264	16119	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.243	96	17906	3.958	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.256	152	6433	0.344	ng/ul	0.00
18) Fluoranthene-d10	19.286	212	19397	0.409	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.281	88	267	0.056	ng/ul#	63

(#) = qualifier out of range (m) = manual integration (+) = signals summed

```
Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN013123\  
Data File : BN023878.D  
Acq On    : 01 Feb 2023   01:24  
Operator   : CG/JU  
Sample     : 01278-02  
Misc       :  
ALS Vial   : 26      Sample Multiplier: 1
```

Instrument :
BNA_N
ClientSampleId :
EX9G9

Quant Time: Feb 01 02:50:57 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN013023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Qlast Update : Wed Feb 01 01:00:14 2023
Response via : Initial Calibration

