

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN013123\
 Data File : BN023886.D
 Acq On : 01 Feb 2023 10:50
 Operator : CG/JU
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD0.4376

Quant Time: Feb 01 11:22:55 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN013023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 01 01:00:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.855	152	5327	0.400	ng/ul	0.00
4) Naphthalene-d8	10.660	136	13846	0.400	ng/ul	-0.01
9) Acenaphthene-d10	14.505	164	8509	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.254	188	19474	0.400	ng/ul	# 0.00
17) Chrysene-d12	21.441	240	17966	0.400	ng/ul	#-0.01
23) Perylene-d12	23.823	264	14598	0.400	ng/ul	-0.03
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.244	96	2034	0.356	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.255	152	7162	0.331	ng/ul	0.00
18) Fluoranthene-d10	19.284	212	17897	0.371	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.281	88	2365	0.391	ng/ul#	65
5) Naphthalene	10.710	128	13231	0.349	ng/ul	99
7) 2-Methylnaphthalene	12.332	142	8502	0.345	ng/ul	100
8) 1-Methylnaphthalene	12.546	142	8683	0.344	ng/ul	100
10) Acenaphthylene	14.228	152	11032	0.339	ng/ul	99
11) Acenaphthene	14.570	153	9820	0.343	ng/ul	99
12) Fluorene	15.555	166	11568	0.334	ng/ul	97
14) Pentachlorophenol	16.908	266	1944	0.303	ng/ul	99
15) Phenanthrene	17.296	178	19666	0.339	ng/ul	100
16) Anthracene	17.389	178	16343	0.336	ng/ul	99
19) Fluoranthene	19.317	202	23711	0.397	ng/ul	99
20) Pyrene	19.679	202	24081	0.389	ng/ul	98
21) Benzo(a)anthracene	21.427	228	21921	0.359	ng/ul	100
22) Chrysene	21.479	228	23196	0.361	ng/ul	100
24) Benzo(b)fluoranthene	23.104	252	23742	0.368	ng/ul	99
25) Benzo(k)fluoranthene	23.151	252	22693	0.360	ng/ul	99
26) Benzo(a)pyrene	23.724	252	19384	0.361	ng/ul	97
27) Indeno(1,2,3-cd)pyrene	26.292	276	24744	0.328	ng/ul#	97
28) Dibenzo(a,h)anthracene	26.309	278	19403	0.322	ng/ul	97
29) Benzo(g,h,i)perylene	27.060	276	20425	0.329	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN013123\
Data File : BN023886.D
Acq On : 01 Feb 2023 10:50
Operator : CG/JU
Sample : SSTDCCC0.4EC
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.4376

Quant Time: Feb 01 11:22:55 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN013023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 01 01:00:14 2023
Response via : Initial Calibration

