

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN020620\
 Data File : BN009626.D
 Acq On : 07 Feb 2020 01:11
 Operator : CG/JU
 Sample : TESTBL-01
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 TESTBL-01

Quant Time: Feb 07 02:08:19 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN020320MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Feb 03 16:39:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.43	152	136660	20.00	ng/ul	-0.01
18) Naphthalene-d8	10.19	136	569618	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.07	164	358490	20.00	ng/ul	0.00
61) Phenanthrene-d10	16.83	188	734253	20.00	ng/ul	0.00
77) Chrysene-d12	21.04	240	656227	20.00	ng/ul	0.00
85) Perylene-d12	23.15	264	688778	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.08	96	9240	2.77	ng/uL	0.00
5) Phenol-d5	6.63	99	202814	17.02	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.79	67	271457	33.91	ng/ul	0.00
9) 2-Chlorophenol-d4	6.98	132	282112	32.16	ng/ul	-0.01
13) 4-Methylphenol-d8	8.15	113	242786	25.58	ng/ul	-0.01
19) Nitrobenzene-d5	8.58	128	153964	36.34	ng/ul	-0.01
22) 2-Nitrophenol-d4	9.29	143	171262	36.23	ng/ul	-0.01
26) 2,4-Dichlorophenol-d3	9.82	165	294076	33.20	ng/ul	0.00
29) 4-Chloroaniline-d4	10.34	131	46918	4.93	ng/ul	0.00
43) Dimethylphthalate-d6	13.50	166	938007	35.84	ng/ul	0.00
46) Acenaphthylene-d8	13.76	160	1054352	33.34	ng/ul	-0.01
51) 4-Nitrophenol-d4	14.31	143	69079	13.98	ng/ul	0.00
57) Fluorene-d10	15.07	176	815319	35.43	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.22	200	130204	28.39	ng/ul	0.00
70) Anthracene-d10	16.93	188	1077198	32.29	ng/ul	0.00
78) Pyrene-d10	19.23	212	1284281	37.23	ng/ul	0.00
89) Benzo(a)pyrene-d12	23.02	264	1039826	30.66	ng/ul	0.00

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN020620\
Data File : BN009626.D
Acq On : 07 Feb 2020 01:11
Operator : CG/JU
Sample : TESTBL-01
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
TESTBL-01

Quant Time: Feb 07 02:08:19 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN020320MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Feb 03 16:39:56 2020
Response via : Initial Calibration

