

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN020620\
 Data File : BN009662.D
 Acq On : 08 Feb 2020 00:37
 Operator : CG/JU
 Sample : L1401-07
 Misc :
 ALS Vial : 57 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 CB6K0

Quant Time: Feb 08 01:29:52 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN020320MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Feb 07 23:20:27 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.43	152	143249	20.00	ng/ul	0.00
18) Naphthalene-d8	10.18	136	609846	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.07	164	367950	20.00	ng/ul	0.00
61) Phenanthrene-d10	16.82	188	758251	20.00	ng/ul	0.00
77) Chrysene-d12	21.03	240	709975	20.00	ng/ul	-0.01
85) Perylene-d12	23.15	264	722239	20.00	ng/ul	-0.01

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.08	96	11467	3.28	ng/uL	0.00
5) Phenol-d5	6.63	99	234094	18.74	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.79	67	165510	19.72	ng/ul	0.00
9) 2-Chlorophenol-d4	6.98	132	188482	20.50	ng/ul	0.00
13) 4-Methylphenol-d8	8.15	113	187659	18.86	ng/ul	0.00
19) Nitrobenzene-d5	8.58	128	94562	20.85	ng/ul	0.00
22) 2-Nitrophenol-d4	9.29	143	103089	20.37	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.82	165	184630	19.47	ng/ul	0.00
29) 4-Chloroaniline-d4	10.33	131	242363	23.78	ng/ul	0.00
43) Dimethylphthalate-d6	13.49	166	569230	21.19	ng/ul	0.00
46) Acenaphthylene-d8	13.76	160	707301	21.79	ng/ul	0.00
51) 4-Nitrophenol-d4	14.31	143	71844	14.16	ng/ul	0.00
57) Fluorene-d10	15.07	176	512772	21.71	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.22	200	72468	15.30	ng/ul	0.00
70) Anthracene-d10	16.92	188	762527	22.13	ng/ul	0.00
78) Pyrene-d10	19.23	212	845297	22.65	ng/ul	0.00
89) Benzo(a)pyrene-d12	23.02	264	803361	22.59	ng/ul	0.00

Target Compounds

					Ovalue	
44) Dimethylphthalate	13.54	163	54260	1.903	ng/ul#	95
69) Phenanthrene	16.86	178	265995	6.151	ng/ul	99
71) Anthracene	16.96	178	60827	1.386	ng/ul	99
76) Fluoranthene	18.89	202	466696	9.456	ng/ul	100
79) Pyrene	19.26	202	409023	7.892	ng/ul	98
82) Benzo(a)anthracene	21.02	228	217464	4.436	ng/ul	97
83) Bis(2-ethylhexyl)phthalate	20.98	149	116893	3.736	ng/ul	99
84) Chrysene	21.07	228	225866	4.628	ng/ul	97
87) Benzo(b)fluoranthene	22.53	252	276266	6.114	ng/ul#	98
88) Benzo(k)fluoranthene	22.57	252	91005	2.045	ng/ul#	92
90) Benzo(a)pyrene	23.06	252	177640	4.243	ng/ul#	94
91) Indeno(1,2,3-cd)pyrene	25.22	276	117019	2.351	ng/ul	92
93) Benzo(g,h,i)perylene	25.84	276	108480	2.600	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN020620\
Data File : BN009662.D
Acq On : 08 Feb 2020 00:37
Operator : CG/JU
Sample : L1401-07
Misc :
ALS Vial : 57 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
CB6K0

Quant Time: Feb 08 01:29:52 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN020320MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Feb 07 23:20:27 2020
Response via : Initial Calibration

