


Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN020620\
 Data File : BN009605.D
 Acq On : 06 Feb 2020 11:05
 Operator : CG/JU
 Sample : DFTPP56
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP56

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN020320MA.M
 Title : SVOA CALIBRATION
 Last Update : Fri Feb 07 15:21:46 2020



AutoFind: Scans 2433, 2434, 2435; Background Corrected with Scan 2428

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	60.4	68021	PASS
68	69	0.00	2	1.4	757	PASS
69	198	0.00	100	47.8	53890	PASS
70	69	0.00	2	0.3	139	PASS
127	198	10	80	51.3	57808	PASS
197	198	0.00	2	0.3	307	PASS
198	198	100	100	100.0	112629	PASS
199	198	5	9	7.0	7930	PASS
275	198	10	60	26.6	29922	PASS
365	198	1	100	3.7	4124	PASS
441	198	0.01	100	11.0	12408	PASS
442	198	50	100	65.2	73397	PASS
443	442	15	24	18.8	13787	PASS

Scan 2428 (17.269 min): BN009605.D
DFTPP56

m/z	Abundance
39.90	1248.0
44.00	3490.0
48.90	599.0
73.00	642.0
207.00	1067.0
207.80	305.0

Instrument :
BNA_N
ClientSampleId :
DFTPP56

m/z	Abundance
37.10	314.0
38.10	1125.0
39.00	5704.0
40.10	1142.0
44.00	3378.0
48.90	1034.0
50.00	19048.0
51.00	78488.0
52.00	4335.0
54.90	462.0
56.00	2343.0
57.00	6173.0
57.90	342.0
60.90	1017.0
62.10	861.0
63.00	2738.0
63.90	378.0
65.10	1375.0
69.00	62328.0
70.10	419.0
72.90	1089.0
74.00	6021.0
75.00	9493.0
76.10	3310.0
77.00	65336.0
78.00	4513.0
79.00	5269.0
80.00	3158.0
81.00	4860.0
82.00	1141.0
82.90	984.0
84.90	646.0
85.90	1643.0
87.00	893.0
88.10	323.0
91.10	1072.0
91.90	1363.0
93.00	8301.0
93.90	410.0
95.90	374.0
98.00	6400.0
99.00	5083.0
100.10	402.0
101.00	2149.0
103.10	1089.0
103.90	1747.0
105.00	1752.0
106.00	780.0
107.00	18288.0
108.00	3010.0
110.00	33496.0
111.00	5352.0
112.00	711.0
115.90	1086.0
117.00	17296.0
118.00	1907.0
120.00	307.0
121.90	1211.0
123.00	1990.0
123.90	1004.0
124.90	464.0
127.00	65264.0
127.90	5444.0
129.00	26280.0
130.00	2226.0
131.00	349.0
133.10	536.0
133.80	619.0
134.90	2067.0
136.00	700.0
137.10	1554.0
138.00	309.0
141.00	3809.0
142.00	1072.0
142.90	568.0
143.90	316.0
145.90	492.0
147.00	2065.0
148.00	3745.0
149.00	992.0
150.90	319.0
151.40	451.0
152.00	432.0
153.00	1243.0
153.90	653.0
154.90	1839.0
155.90	2630.0
157.10	645.0
158.10	511.0
158.90	400.0
159.90	1261.0
161.00	1542.0

Instrument :
BNA_N
ClientSampleId :
DFTPP56

161.90	533.0
164.90	1552.0
165.90	963.0
167.00	5777.0
168.00	3814.0
168.90	688.0
171.00	333.0
171.90	622.0
172.90	872.0
173.90	1490.0
175.00	2967.0
175.90	710.0
177.00	1291.0
178.00	450.0
179.00	3998.0
180.00	3492.0
181.00	1953.0
183.80	488.0
184.90	2083.0
186.00	16760.0
187.00	4591.0
187.90	760.0
188.90	1202.0
190.90	662.0
192.00	1714.0
193.00	2277.0
194.00	331.0
194.80	359.0
196.00	2815.0
197.10	921.0
197.90	111424.0
198.90	8566.0
199.80	753.0
201.20	584.0
202.00	333.0
203.00	911.0
204.00	4663.0
205.00	7136.0
206.00	30568.0
207.00	6459.0
207.90	1677.0
208.90	561.0
210.00	626.0
211.00	1118.0
214.90	357.0
215.90	676.0
216.90	9457.0
218.00	1355.0
221.00	4231.0
221.70	1522.0
223.00	1688.0
224.00	16346.0
225.00	4786.0
226.00	718.0
227.00	6796.0
228.00	1033.0
229.00	1854.0
229.90	352.0
231.00	646.0
233.90	435.0
234.80	333.0
235.80	311.0
237.00	727.0
238.90	444.0
241.90	1019.0
242.80	621.0
244.00	13079.0
245.00	1459.0
245.90	2692.0
247.00	515.0
248.90	882.0
253.00	461.0
253.90	547.0
255.00	67600.0
256.00	9132.0
256.90	1007.0
258.00	3797.0
259.00	594.0
264.90	1597.0
272.90	1850.0
274.00	4944.0
275.00	28688.0
276.00	3477.0
277.00	2538.0
277.90	422.0
280.90	461.0
283.00	409.0
284.90	467.0
293.00	788.0
296.00	7541.0
296.80	791.0
302.90	648.0
314.80	862.0
315.90	358.0
323.00	2290.0
323.90	351.0

Instrument :
BNA_N
ClientSampleId :
DFTPP56

326.90	479.0
334.00	1587.0
341.00	329.0
345.90	353.0
352.00	699.0
352.90	484.0
353.90	686.0
364.90	3324.0
366.00	404.0
371.90	1011.0
373.00	362.0
382.80	332.0
401.80	422.0
402.90	771.0
404.00	304.0
421.00	651.0
421.80	369.0
423.00	3712.0
424.00	1910.0
425.00	490.0
441.00	9588.0
442.00	55808.0
443.00	10663.0
443.90	1056.0

Instrument :
BNA_N
ClientSampleId :
DFTPP56

m/z	Abundance
37.00	580.0
38.00	920.0
39.00	5840.0
40.00	1316.0
41.10	306.0
44.00	3551.0
49.10	949.0
50.00	20672.0
51.00	81256.0
52.10	4364.0
54.80	518.0
56.00	2648.0
57.00	5559.0
61.00	1078.0
62.00	1127.0
63.00	3018.0
64.00	612.0
65.00	1671.0
68.00	1453.0
69.00	62104.0
72.90	1040.0
74.00	7190.0
75.00	9521.0
76.00	2671.0
77.00	69832.0
78.00	5341.0
79.00	4987.0
80.00	3986.0
81.00	5398.0
82.00	1029.0
82.90	873.0
84.00	387.0
84.90	858.0
86.00	1531.0
86.80	726.0
90.90	934.0
92.00	1611.0
93.00	8168.0
94.00	604.0
96.10	459.0
98.00	5683.0
99.00	6147.0
99.90	554.0
101.00	2995.0
102.90	665.0
104.00	2326.0
105.00	1388.0
106.10	824.0
107.00	21552.0
108.00	3236.0
108.90	813.0
110.00	35968.0
111.00	5664.0
111.80	652.0
115.90	1418.0
117.00	19280.0
117.90	1204.0
119.00	341.0
120.00	347.0
121.90	1236.0
123.00	2123.0
124.00	1197.0
125.00	756.0
127.00	66512.0
128.00	4663.0
129.00	27800.0
130.00	2396.0
131.00	384.0
133.80	1056.0
135.00	1652.0
136.10	652.0
137.10	1150.0
139.80	444.0
140.90	4089.0
141.90	1554.0
143.00	675.0
145.90	678.0
147.00	2027.0
147.90	4494.0
148.90	1155.0
151.20	589.0
152.90	1349.0
154.00	867.0
155.00	1890.0
156.00	2899.0
156.90	684.0
157.90	461.0
159.10	467.0
160.00	1388.0
160.90	1458.0
161.60	551.0
164.90	1460.0

Instrument :
BNA_N
ClientSampleId :
DFTPP56

166.00	1293.0
167.00	7511.0
168.00	4055.0
168.90	941.0
170.00	321.0
171.00	645.0
171.80	767.0
172.90	821.0
173.90	1902.0
175.00	2623.0
176.00	647.0
176.90	1258.0
177.90	515.0
179.00	5421.0
179.90	4132.0
181.00	1835.0
181.90	314.0
183.90	304.0
184.90	2495.0
186.00	19248.0
187.00	6243.0
188.00	500.0
188.90	1105.0
190.00	337.0
190.80	829.0
192.00	1405.0
192.90	2578.0
194.00	869.0
195.90	3447.0
197.90	138560.0
198.90	9190.0
199.90	718.0
201.60	610.0
202.90	1214.0
204.00	4719.0
205.00	9819.0
206.00	38128.0
207.00	5820.0
207.90	1128.0
209.00	952.0
209.90	712.0
210.90	1411.0
211.80	370.0
215.00	375.0
216.00	789.0
216.90	10393.0
217.90	1042.0
220.90	5093.0
221.70	1455.0
223.00	2273.0
224.00	21680.0
225.00	4979.0
225.90	658.0
226.90	8359.0
227.90	1370.0
228.90	1654.0
229.90	324.0
230.90	930.0
233.80	501.0
234.90	694.0
236.00	377.0
236.90	1076.0
238.90	393.0
241.00	610.0
241.90	1198.0
243.00	1140.0
244.00	16357.0
245.00	1871.0
245.90	3775.0
247.00	816.0
249.00	839.0
252.90	376.0
254.20	568.0
255.00	78928.0
256.00	12050.0
257.00	1076.0
257.90	4975.0
259.00	649.0
265.00	1191.0
265.90	351.0
272.90	2179.0
274.00	6257.0
275.00	35576.0
275.90	4990.0
276.90	3053.0
278.00	565.0
281.00	315.0
285.00	716.0
292.90	977.0
293.90	327.0
296.00	9959.0
296.90	1311.0
303.00	1403.0
303.80	316.0
314.10	480.0
314.90	1460.0

Instrument :
BNA_N
ClientSampleId :
DFTPP56

316.00	488.0
323.00	3164.0
323.90	602.0
326.90	752.0
333.00	471.0
334.10	1445.0
335.00	403.0
341.00	424.0
346.00	586.0
352.00	1338.0
352.90	878.0
354.00	1159.0
355.00	323.0
364.90	5119.0
365.90	506.0
372.00	1695.0
382.80	348.0
390.90	334.0
402.00	647.0
403.00	877.0
403.80	356.0
420.90	847.0
421.90	634.0
423.00	5689.0
424.00	2105.0
441.00	14864.0
442.00	86992.0
443.00	17840.0
444.00	1525.0

Instrument :
BNA_N
ClientSampleId :
DFTPP56

m/z	Abundance
37.00	392.0
37.90	536.0
39.00	3279.0
39.90	1341.0
44.00	3572.0
49.00	627.0
50.00	11200.0
51.00	44320.0
52.00	2360.0
55.00	344.0
56.00	1408.0
57.00	2268.0
60.90	424.0
62.00	675.0
63.00	1689.0
65.00	958.0
68.00	818.0
69.00	37240.0
73.00	999.0
73.90	3881.0
75.00	6157.0
76.00	1595.0
77.00	39240.0
78.00	3090.0
79.00	2823.0
80.00	2247.0
81.00	2969.0
81.90	577.0
82.90	799.0
83.90	339.0
84.90	652.0
86.00	942.0
87.00	342.0
91.00	514.0
92.00	888.0
93.00	4983.0
98.00	4249.0
99.00	3176.0
100.00	314.0
101.00	1892.0
102.90	631.0
104.00	1222.0
105.00	940.0
107.00	12007.0
107.90	1653.0
110.00	21416.0
110.90	3051.0
111.90	338.0
116.00	651.0
117.00	11434.0
118.00	1231.0
122.00	1200.0
122.90	1517.0
123.80	369.0
124.90	451.0
127.00	41648.0
128.00	2848.0
129.00	17248.0
129.90	864.0
133.80	403.0
135.00	1625.0
136.00	516.0
137.00	917.0
141.00	2256.0
141.80	617.0
143.00	569.0
146.00	583.0
146.90	1317.0
148.00	2480.0
149.00	780.0
153.00	1111.0
153.80	690.0
154.90	1067.0
156.00	2194.0
157.00	479.0
158.00	600.0
159.10	373.0
159.80	618.0
160.90	1280.0
161.80	405.0
164.90	897.0
165.90	837.0
167.00	4345.0
167.90	3013.0
170.80	307.0
171.90	566.0
173.00	625.0
173.90	576.0
175.00	1715.0
175.90	523.0
176.90	1242.0
179.00	4017.0

Instrument :
BNA_N
ClientSampleId :
DFTPP56

179.90	3001.0
181.00	1267.0
185.00	1428.0
186.00	13443.0
187.00	3876.0
188.90	1052.0
191.00	659.0
191.90	995.0
193.00	1269.0
195.90	2358.0
197.90	87904.0
198.90	6035.0
199.90	325.0
202.90	767.0
204.00	3621.0
205.00	6843.0
206.00	23720.0
207.00	4017.0
207.90	1132.0
208.90	363.0
210.30	648.0
211.00	920.0
215.90	544.0
217.00	6423.0
218.00	776.0
221.00	4428.0
222.90	1594.0
224.00	14208.0
225.00	3993.0
226.00	489.0
226.90	5533.0
228.00	740.0
228.90	1016.0
231.00	589.0
234.00	525.0
235.00	355.0
235.90	362.0
236.90	741.0
240.00	439.0
240.90	492.0
241.90	516.0
242.10	561.0
243.00	956.0
244.00	10962.0
245.00	1612.0
245.90	1949.0
246.90	454.0
249.00	505.0
250.80	300.0
255.00	56384.0
256.00	8509.0
257.00	782.0
258.00	3433.0
258.90	493.0
265.00	1574.0
273.00	1407.0
273.90	4566.0
275.00	25504.0
276.00	3201.0
277.00	2360.0
283.00	435.0
284.00	400.0
292.90	1028.0
294.90	392.0
296.00	8300.0
297.00	1205.0
302.90	836.0
314.20	357.0
314.90	820.0
316.10	515.0
320.90	313.0
323.00	2794.0
323.90	692.0
326.90	581.0
334.00	1785.0
334.90	379.0
341.00	309.0
345.90	483.0
351.80	573.0
352.90	554.0
354.00	991.0
365.00	3930.0
366.00	384.0
372.00	912.0
372.90	448.0
382.90	401.0
401.90	515.0
402.90	797.0
421.00	501.0
422.00	723.0
422.90	4576.0
424.00	1061.0
441.00	12773.0
442.00	77392.0
443.00	12860.0
443.90	1454.0

Instrument :
BNA_N
ClientSampleId :
DFTPP56