

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN020819\
 Data File : BN004367.D
 Acq On : 08 Feb 2019 18:38
 Operator : JU/SJ
 Sample : K1428-02
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BEJH0

Quant Time: Feb 09 04:45:25 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN020719MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Feb 09 04:41:24 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.75	152	29541	20.00	ng/ul	0.00
18) Naphthalene-d8	10.54	136	146349	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.39	164	94776	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.13	188	234379	20.00	ng/ul	0.00
78) Chrysene-d12	21.33	240	310273	20.00	ng/ul	-0.01
86) Perylene-d12	23.60	264	344495	20.00	ng/ul	-0.01

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.25	96	2431	4.95	ng/uL	0.00
5) Phenol-d5	6.90	99	19240	7.31	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.09	67	47951	36.30	ng/ul	0.00
9) 2-Chlorophenol-d4	7.28	132	62764	29.40	ng/ul	0.00
13) 4-Methylphenol-d8	8.44	113	41287	18.24	ng/ul	0.00
19) Nitrobenzene-d5	8.91	128	39740	41.02	ng/ul	0.00
22) 2-Nitrophenol-d4	9.62	143	44547	44.05	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.15	165	84996	37.07	ng/ul	0.00
29) 4-Chloroaniline-d4	10.68	131	1621	0.64	ng/ul	0.00
44) Dimethylphthalate-d6	13.80	166	329341	40.97	ng/ul	0.00
47) Acenaphthylene-d8	14.08	160	381664	40.08	ng/ul	0.00
52) 4-Nitrophenol-d4	14.58	143	11347	7.37	ng/ul	0.00
58) Fluorene-d10	15.38	176	282376	39.91	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.50	200	48073	35.17	ng/ul	0.00
71) Anthracene-d10	17.23	188	459336	40.89	ng/ul	0.00
79) Pyrene-d10	19.53	212	577194	41.50	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.46	264	755800	41.80	ng/ul	-0.01

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN020819\
Data File : BN004367.D
Acq On : 08 Feb 2019 18:38
Operator : JU/SJ
Sample : K1428-02
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BEJH0

Quant Time: Feb 09 04:45:25 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN020719MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Feb 09 04:41:24 2019
Response via : Initial Calibration

