

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN021119\
 Data File : BN004377.D
 Acq On : 11 Feb 2019 14:37
 Operator : JU/SJ
 Sample : K1410-03
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 X2F73

Quant Time: Feb 11 15:25:26 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN021119.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 11 14:04:55 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.75	152	1336	0.40	ng/ul	0.00
2) Naphthalene-d8	10.54	136	4954	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.39	164	2723	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.14	188	7252	0.40	ng/ul	0.00
16) Chrysene-d12	21.33	240	8011	0.40	ng/ul	0.00
20) Perylene-d12	23.60	264	9395	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.13	152	3050	0.39	ng/ul	0.00
14) Fluoranthene-d10	19.17	212	8594	0.38	ng/ul	0.00
Target Compounds				Ovalue		
3) Naphthalene	10.59	128	9982	0.702	ng/ul	99
5) 2-Methylnaphthalene	12.20	142	4748	0.467	ng/ul	100
9) Fluorene	15.44	166	8432	0.653	ng/ul	99
11) Pentachlorophenol	16.78	266	5311	3.439	ng/ul	97
13) Anthracene	17.27	178	18036	0.908	ng/ul	99
15) Fluoranthene	19.20	202	20588	0.713	ng/ul	99
17) Pyrene	19.56	202	23809	0.923	ng/ul	100
19) Chrysene	21.37	228	18615	0.621	ng/ul	100
21) Benzo(b)fluoranthene	22.92	252	19537	0.520	ng/ul	100
23) Benzo(a)pyrene	23.50	252	24717	0.736	ng/ul	97
25) Dibenzo(a,h)anthracene	25.91	278	55312	1.546	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN021119\
Data File : BN004377.D
Acq On : 11 Feb 2019 14:37
Operator : JU/SJ
Sample : K1410-03
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
X2F73

Quant Time: Feb 11 15:25:26 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN021119.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Feb 11 14:04:55 2019
Response via : Initial Calibration

