

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN021119\
 Data File : BN004379.D
 Acq On : 11 Feb 2019 15:47
 Operator : JU/SJ
 Sample : K1410-08
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 X2F78

Quant Time: Feb 11 17:02:23 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN021119.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 11 14:04:55 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.75	152	1380	0.40	ng/ul	0.00
2) Naphthalene-d8	10.54	136	5518	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.39	164	3142	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.14	188	8540	0.40	ng/ul	0.00
16) Chrysene-d12	21.33	240	10803	0.40	ng/ul	0.00
20) Perylene-d12	23.60	264	11360	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.13	152	2533	0.29	ng/ul	0.00
14) Fluoranthene-d10	19.17	212	9936	0.37	ng/ul	0.00
Target Compounds				Ovalue		
3) Naphthalene	10.59	128	8839	0.558	ng/ul	99
5) 2-Methylnaphthalene	12.20	142	5958	0.526	ng/ul	100
9) Fluorene	15.44	166	9523	0.639	ng/ul	99
11) Pentachlorophenol	16.78	266	3780	2.079	ng/ul	97
12) Phenanthrene	17.17	178	19799	0.713	ng/ul	100
15) Fluoranthene	19.20	202	22680	0.667	ng/ul	99
18) Benzo(a)anthracene	21.32	228	29127	0.836	ng/ul	99
22) Benzo(k)fluoranthene	22.96	252	30912	0.693	ng/ul	98
24) Indeno(1,2,3-cd)pyrene	25.89	276	30572	0.573	ng/ul#	90
26) Benzo(g,h,i)perylene	26.59	276	55266	1.141	ng/ul	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN021119\
Data File : BN004379.D
Acq On : 11 Feb 2019 15:47
Operator : JU/SJ
Sample : K1410-08
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
X2F78

Quant Time: Feb 11 17:02:23 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN021119.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Feb 11 14:04:55 2019
Response via : Initial Calibration

