

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN021319\
 Data File : BN004393.D
 Acq On : 13 Feb 2019 15:26
 Operator : JU/SJ
 Sample : MDL-S-BL
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 MDL-S-BL

Quant Time: Feb 15 17:31:10 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN020719MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Feb 14 06:33:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.75	152	26514	20.00	ng/ul	0.00
18) Naphthalene-d8	10.53	136	123218	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.39	164	79213	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.13	188	197933	20.00	ng/ul	0.00
78) Chrysene-d12	21.33	240	267278	20.00	ng/ul	0.00
86) Perylene-d12	23.61	264	323092	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.25	96	3333	7.56	ng/uL	0.00
5) Phenol-d5	6.90	99	69593	29.45	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.09	67	39194	33.06	ng/ul	0.00
9) 2-Chlorophenol-d4	7.27	132	62462	32.59	ng/ul	0.00
13) 4-Methylphenol-d8	8.44	113	60889	29.98	ng/ul	0.00
19) Nitrobenzene-d5	8.90	128	29403	36.05	ng/ul	0.00
22) 2-Nitrophenol-d4	9.62	143	32059	37.65	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.15	165	63773	33.04	ng/ul	0.00
29) 4-Chloroaniline-d4	10.67	131	100702	47.17	ng/ul	0.00
44) Dimethylphthalate-d6	13.80	166	269550	40.12	ng/ul	0.00
47) Acenaphthylene-d8	14.08	160	303861	38.18	ng/ul	0.00
52) 4-Nitrophenol-d4	14.58	143	42123	32.75	ng/ul	0.00
58) Fluorene-d10	15.38	176	237494	40.16	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.50	200	35363	30.64	ng/ul	0.00
71) Anthracene-d10	17.23	188	392953	41.42	ng/ul	0.00
79) Pyrene-d10	19.54	212	493060	41.16	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.47	264	721764	42.57	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN021319\
Data File : BN004393.D
Acq On : 13 Feb 2019 15:26
Operator : JU/SJ
Sample : MDL-S-BL
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MDL-S-BL

Quant Time: Feb 15 17:31:10 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN020719MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Feb 14 06:33:13 2019
Response via : Initial Calibration

