

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN021419\
 Data File : BN004427.D
 Acq On : 14 Feb 2019 19:12
 Operator : JU/SJ
 Sample : MDL-S-02
 Misc : 0.07 PPM
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 MDL-S-02

Quant Time: Feb 15 01:23:37 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN021119.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 11 14:04:55 2019
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.75	152	834	0.40	ng/ul	0.00
2) Naphthalene-d8	10.53	136	3341	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.38	164	2073	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.13	188	5924	0.40	ng/ul	0.00
16) Chrysene-d12	21.33	240	7809	0.40	ng/ul	0.00
20) Perylene-d12	23.60	264	8211	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.12	152	1835	0.35	ng/ul	0.00
14) Fluoranthene-d10	19.16	212	6402	0.34	ng/ul	0.00
Target Compounds				Qvalue		
3) Naphthalene	10.58	128	655	0.068	ng/ul#	79
5) 2-Methylnaphthalene	12.20	142	463	0.068	ng/ul	98
7) Acenaphthylene	14.11	152	579	0.061	ng/ul#	80
8) Acenaphthene	14.45	153	550	0.067	ng/ul	98
9) Fluorene	15.43	166	665	0.068	ng/ul#	93
11) Pentachlorophenol	16.78	266	39	0.031	ng/ul#	88
12) Phenanthrene	17.17	178	1294	0.067	ng/ul#	89
13) Anthracene	17.27	178	1021	0.063	ng/ul#	85
15) Fluoranthene	19.20	202	1615	0.068	ng/ul	93
17) Pyrene	19.56	202	1615	0.064	ng/ul#	89
18) Benzo(a)anthracene	21.32	228	1647	0.065	ng/ul	95
19) Chrysene	21.37	228	2017	0.069	ng/ul	95
21) Benzo(b)fluoranthene	22.92	252	2305	0.070	ng/ul	73
22) Benzo(k)fluoranthene	22.96	252	2081	0.065	ng/ul#	71
23) Benzo(a)pyrene	23.50	252	1997	0.068	ng/ul#	63
24) Indeno(1,2,3-cd)pyrene	25.89	276	2625	0.068	ng/ul#	98
25) Dibenzo(a,h)anthracene	25.90	278	2129	0.068	ng/ul#	81
26) Benzo(g,h,i)perylene	26.60	276	2444	0.070	ng/ul#	90


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN021419\
Data File : BN004427.D
Acq On : 14 Feb 2019 19:12
Operator : JU/SJ
Sample : MDL-S-02
Misc : 0.07 PPM
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MDL-S-02

Quant Time: Feb 15 01:23:37 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN021119.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Feb 11 14:04:55 2019
Response via : Initial Calibration





#3

Naphthalene

Concen: 0.068 ng/ul

RT: 10.58 min Scan# 2356

Delta R.T. -0.01 min

Lab File: BN004427.D

Acq: 14 Feb 2019 19:12

Instrument :

BNA_N

ClientSampleId :

MDL-S-02


Tgt Ion:128 Resp: 655

Ion Ratio Lower Upper




128 100

129 20.9 9.8 14.6#

127 22.3 11.4 17.0#



Abundance Ion 128.00 (127.70 to 128.70): E



#5

2-Methylnaphthalene

Concen: 0.068 ng/ul

RT: 12.20 min Scan# 2650

Delta R.T. -0.01 min

Lab File: BN004427.D


Acq: 14 Feb 2019 19:12

Tgt Ion:142 Resp: 463



Ion Ratio Lower Upper


142 100

141 91.6 71.8 107.6



Abundance Ion 142.00 (141.70 to 142.70): E





#7

Acenaphthylene

Concen: 0.061 ng/ul

RT: 14.11 min Scan# 3021

Delta R.T. -0.00 min

Lab File: BN004427.D

Acq: 14 Feb 2019 19:12

Instrument :

BNA_N

ClientSampleId :

MDL-S-02


Tgt Ion:152 Resp: 579

Ion Ratio Lower Upper

152 100

151 29.5 16.6 24.8#

153 22.4 11.1 16.7#



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E

400

300

200

100



0

14.11

14.10

14.20

Time-->



#8

Acenaphthene

Concen: 0.067 ng/ul

RT: 14.45 min Scan# 3095

Delta R.T. -0.00 min

Lab File: BN004427.D

Acq: 14 Feb 2019 19:12


Tgt Ion:153 Resp: 550

Ion Ratio Lower Upper

153 100

152 52.1 38.4 57.6

154 89.2 71.4 107.0



Abundance Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E

Ion 154.00 (153.70 to 154.70): E

400

300

200

100


0


14.45

14.40

14.50

Time-->





#9

Fluorene

Concen: 0.068 ng/ul

RT: 15.43 min Scan# 3308

Delta R.T. -0.00 min

Lab File: BN004427.D

Acq: 14 Feb 2019 19:12

Instrument :

BNA_N

ClientSampleId :

MDL-S-02





Tgt Ion:166 Resp: 665

Ion Ratio Lower Upper

166 100

165 103.9 78.9 118.3

167 22.9 11.6 17.4#



#11

Pentachlorophenol

Concen: 0.031 ng/ul

RT: 16.78 min Scan# 3604

Delta R.T. -0.00 min

Lab File: BN004427.D

Acq: 14 Feb 2019 19:12




Tgt Ion:266 Resp: 39


Ion Ratio Lower Upper

266 100

264 48.7 52.3 78.5#

268 59.0 49.0 73.6





#12

Phenanthrene

Concen: 0.067 ng/ul

RT: 17.17 min Scan# 3698

Delta R.T. -0.00 min

Lab File: BN004427.D

Acq: 14 Feb 2019 19:12

Instrument :

BNA_N

ClientSampleId :

MDL-S-02


Tgt Ion:178 Resp: 1294

Ion Ratio Lower Upper

178 100

179 20.8 12.6 19.0#




176 24.4 15.8 23.6#



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 176.00 (175.70 to 176.70): E



#13

Anthracene

Concen: 0.063 ng/ul

RT: 17.27 min Scan# 3720

Delta R.T. -0.00 min

Lab File: BN004427.D

Acq: 14 Feb 2019 19:12


Tgt Ion:178 Resp: 1021

Ion Ratio Lower Upper

178 100

179 23.5 13.0 19.4#



176 25.7 15.5 23.3#




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 176.00 (175.70 to 176.70): E





#15

Fluoranthene

Concen: 0.068 ng/ul

RT: 19.20 min Scan# 4162

Delta R.T. -0.00 min

Lab File: BN004427.D

Acq: 14 Feb 2019 19:12

Instrument :

BNA_N

ClientSampleId :

MDL-S-02


Tgt Ion:202 Resp: 1615

Ion Ratio Lower Upper

202 100

101 12.4 0.0 30.4




100 10.6 0.0 27.6



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 100.00 (99.70 to 100.70): B



#17

Pyrene

Concen: 0.064 ng/ul

RT: 19.56 min Scan# 4240

Delta R.T. -0.00 min

Lab File: BN004427.D

Acq: 14 Feb 2019 19:12


Tgt Ion:202 Resp: 1615

Ion Ratio Lower Upper

202 100

101 15.8 9.0 13.6#



100 12.7 7.4 11.0#




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 100.00 (99.70 to 100.70): B





#18

Benzo(a)anthracene

Concen: 0.065 ng/ul

RT: 21.32 min Scan# 4721

Delta R.T. 0.00 min

Lab File: BN004427.D

Acq: 14 Feb 2019 19:12

Instrument :

BNA_N

ClientSampleId :

MDL-S-02


Tgt Ion:228 Resp: 1647

Ion Ratio Lower Upper




228 100

229 22.4 15.7 23.5

226 29.3 22.1 33.1



Abundance Ion 228.00 (227.70 to 228.70): E
Ion 229.00 (228.70 to 229.70): E
Ion 226.00 (225.70 to 226.70): E



#19

Chrysene

Concen: 0.069 ng/ul

RT: 21.37 min Scan# 4738

Delta R.T. -0.00 min

Lab File: BN004427.D

Acq: 14 Feb 2019 19:12


Tgt Ion:228 Resp: 2017

Ion Ratio Lower Upper



228 100


226 32.7 23.8 35.8

229 21.8 15.8 23.6



Abundance Ion 228.00 (227.70 to 228.70): E
Ion 226.00 (225.70 to 226.70): E
Ion 229.00 (228.70 to 229.70): E





#21

Benzo(b)fluoranthene

Concen: 0.070 ng/ul

RT: 22.92 min Scan# 5268

Delta R.T. 0.00 min

Lab File: BN004427.D

Acq: 14 Feb 2019 19:12





Instrument :

BNA_N

ClientSampleId :

MDL-S-02

Tgt Ion:	252	Resp:	2305
Ion	Ratio	Lower	Upper
252	100		
253	38.3	0.0	48.0
125	20.3	0.0	22.8



#22

Benzo(k)fluoranthene

Concen: 0.065 ng/ul




RT: 22.96 min Scan# 5283


Delta R.T. -0.00 min

Lab File: BN004427.D

Acq: 14 Feb 2019 19:12

Tgt Ion:	252	Resp:	2081
Ion	Ratio	Lower	Upper
252	100		
253	39.7	19.5	29.3#
125	21.4	9.1	13.7#





#23

Benzo(a)pyrene

Concen: 0.068 ng/u1

RT: 23.50 min Scan# 5467

Delta R.T. -0.00 min

Lab File: BN004427.D

Acq: 14 Feb 2019 19:12

Instrument :

BNA_N

ClientSampleId :

MDL-S-02


Tgt Ion:252 Resp: 1997

Ion Ratio Lower Upper

252 100

253 46.2 20.4 30.6#




125 25.3 10.6 15.8#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#24

Indeno(1,2,3-cd)pyrene

Concen: 0.068 ng/u1

RT: 25.89 min Scan# 6247

Delta R.T. 0.00 min

Lab File: BN004427.D

Acq: 14 Feb 2019 19:12


Tgt Ion:276 Resp: 2625

Ion Ratio Lower Upper

276 100

138 23.0 19.1 28.7



227 0.0 0.2 0.4#




Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 227.00 (226.70 to 227.70): E





#25

Dibenzo(a,h)anthracene

Concen: 0.068 ng/ul

RT: 25.90 min Scan# 6251

Delta R.T. -0.00 min

Lab File: BN004427.D

Acq: 14 Feb 2019 19:12

Instrument :

BNA_N

ClientSampleId :

MDL-S-02


Tgt Ion: 278 Resp: 2129

Ion Ratio Lower Upper

278 100

139 22.1 13.8 20.6#




279 38.0 20.4 30.6#



Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E



#26

Benzo(g,h,i)perylene

Concen: 0.070 ng/ul

RT: 26.60 min Scan# 6457

Delta R.T. 0.00 min

Lab File: BN004427.D

Acq: 14 Feb 2019 19:12


Tgt Ion: 276 Resp: 2444

Ion Ratio Lower Upper

276 100

138 25.9 16.8 25.2#

277 29.4 19.4 29.0#



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E

