


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021720\
Data File : BN009729.D
Acq On : 17 Feb 2020 09:29
Operator : CG/JU
Sample : DFTPP61
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP61

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN021220MA.M
Title : SVOA CALIBRATION
Last Update : Mon Feb 17 00:47:25 2020



AutoFind: Scans 2430, 2431, 2432; Background Corrected with Scan 2424

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	65.3	143232	PASS
68	69	0.00	2	2.0	2089	PASS
69	198	0.00	100	48.2	105802	PASS
70	69	0.00	2	0.5	515	PASS
127	198	10	80	50.3	110258	PASS
197	198	0.00	2	0.5	1103	PASS
198	198	100	100	100.0	219413	PASS
199	198	5	9	6.8	14852	PASS
275	198	10	60	25.2	55200	PASS
365	198	1	100	3.0	6519	PASS
441	198	0.01	100	8.5	18682	PASS
442	198	50	100	57.5	126141	PASS
443	442	15	24	20.5	25885	PASS

m/z	Abundance
37.00	541.0
38.10	1658.0
39.10	12284.0
40.00	1443.0
40.90	318.0
44.00	4679.0
44.90	461.0
49.10	1437.0
50.00	36632.0
51.00	155392.0
52.00	8301.0
53.10	311.0
55.00	886.0
56.00	4884.0
57.00	9722.0
58.00	475.0
61.00	1904.0
62.00	2174.0
63.00	5269.0
64.00	764.0
65.10	2055.0
68.00	2437.0
69.00	108032.0
70.00	733.0
73.00	1288.0
74.00	12682.0
75.00	18584.0
76.00	5575.0
77.00	117880.0
78.00	8779.0
79.00	9160.0
80.00	6004.0
81.00	8703.0
82.00	1937.0
83.00	1858.0
85.00	1071.0
86.00	3304.0
87.10	1189.0
87.90	306.0
91.00	1668.0
92.00	2188.0
93.00	15401.0
94.10	703.0
95.90	444.0
97.20	540.0
98.00	11691.0
99.00	9743.0
100.00	956.0
101.00	4872.0
102.90	1274.0
103.90	2933.0
105.00	2656.0
106.00	956.0
107.00	34456.0
108.00	4971.0
109.00	913.0
110.00	57736.0
110.90	9011.0
111.90	1593.0
116.00	1733.0
117.00	31840.0
117.90	2306.0
118.90	466.0
119.90	416.0
121.90	2299.0
122.90	3107.0
123.90	1710.0
124.90	1538.0
127.00	105600.0
128.00	8150.0
129.00	41448.0
130.00	3627.0
131.00	851.0
132.80	314.0
133.90	1290.0
134.90	3662.0
135.90	1421.0
137.00	1995.0
138.00	342.0
140.00	400.0
140.90	6129.0
141.90	1668.0
143.00	1057.0
144.00	404.0
146.00	1220.0
147.00	2941.0
148.00	6631.0
148.90	2034.0
149.90	458.0
151.20	838.0
151.90	371.0
153.00	1439.0

Instrument :
BNA_N
ClientSampleId :
DFTPP61

154.00	1509.0
155.00	3579.0
156.00	4445.0
156.90	858.0
157.90	1254.0
159.00	867.0
160.00	1513.0
161.00	1933.0
161.90	743.0
164.90	2223.0
166.00	1299.0
167.00	11815.0
168.00	8028.0
168.90	1238.0
169.90	472.0
170.80	454.0
172.00	965.0
173.00	1146.0
174.00	2121.0
175.00	3887.0
176.00	1414.0
177.00	2121.0
177.90	1022.0
178.90	8331.0
180.00	6322.0
180.90	2678.0
181.80	450.0
183.90	957.0
184.90	3966.0
186.00	28904.0
187.00	8439.0
187.90	1417.0
188.90	2348.0
191.00	1114.0
191.90	2834.0
192.90	2804.0
193.90	639.0
194.90	318.0
195.90	5182.0
196.80	1306.0
197.90	197440.0
199.00	13621.0
200.00	1051.0
201.30	747.0
201.60	700.0
202.90	2132.0
204.00	7811.0
205.00	13851.0
206.00	56336.0
207.00	8335.0
207.90	1734.0
209.00	1092.0
210.20	984.0
211.20	1657.0
211.90	533.0
214.90	574.0
216.00	1454.0
216.90	15308.0
217.90	1647.0
221.00	5864.0
222.90	2873.0
224.00	29120.0
225.10	8132.0
226.00	977.0
227.00	13974.0
227.90	1946.0
228.80	2112.0
231.00	931.0
234.00	1030.0
235.00	697.0
236.00	410.0
237.00	2304.0
238.80	424.0
240.00	438.0
241.00	533.0
242.00	1381.0
242.90	1336.0
244.00	21696.0
245.00	2770.0
246.00	3716.0
246.90	963.0
249.00	760.0
253.10	526.0
255.00	111896.0
256.00	16792.0
257.00	1512.0
258.00	6851.0
258.90	1106.0
265.00	3023.0
265.80	371.0
271.90	402.0
273.00	3561.0
274.00	7892.0
275.00	46664.0
276.00	6039.0
277.00	3908.0

Instrument :
BNA_N
ClientSampleId :
DFTPP61

278.00	1455.0
283.00	468.0
285.10	553.0
293.00	1271.0
296.00	14295.0
296.90	2116.0
303.00	1177.0
303.90	387.0
313.80	729.0
314.90	1412.0
316.10	1043.0
320.80	596.0
322.00	318.0
323.10	3856.0
324.00	650.0
327.00	858.0
328.00	540.0
331.80	319.0
334.00	2596.0
335.00	621.0
340.90	448.0
345.80	494.0
346.90	362.0
352.00	1281.0
352.90	566.0
354.00	1288.0
364.90	4639.0
366.00	749.0
372.00	1909.0
372.90	483.0
382.90	560.0
383.80	305.0
401.90	818.0
403.00	1224.0
420.80	885.0
421.90	627.0
422.90	5229.0
424.00	2828.0
441.00	13151.0
442.00	85176.0
443.00	18176.0
444.00	1589.0

Instrument :
BNA_N
ClientSampleId :
DFTPP61

m/z	Abundance
37.10	808.0
38.10	1934.0
39.10	12448.0
39.90	1322.0
41.00	403.0
44.00	4308.0
44.90	691.0
49.00	1233.0
50.00	43424.0
51.00	169536.0
52.00	8302.0
53.00	652.0
55.00	875.0
56.00	5370.0
57.00	10793.0
57.90	470.0
61.00	1962.0
62.00	2038.0
63.00	6293.0
64.00	751.0
65.00	3327.0
68.00	2253.0
69.00	128480.0
70.00	813.0
73.00	1730.0
74.00	13974.0
75.00	21592.0
76.00	7240.0
77.00	139968.0
78.10	10103.0
79.00	11383.0
80.00	7770.0
81.00	10960.0
82.00	2646.0
83.00	2503.0
83.90	364.0
84.90	1842.0
86.00	3178.0
87.00	1095.0
88.00	555.0
91.00	2227.0
92.00	2892.0
93.00	19560.0
93.90	1012.0
96.10	903.0
96.90	311.0
98.00	13710.0
99.00	9649.0
100.00	1116.0
101.00	5411.0
101.90	437.0
103.00	1644.0
104.00	3693.0
104.90	3513.0
106.00	1317.0
107.00	42592.0
108.00	6127.0
109.10	1564.0
110.00	70640.0
111.00	11334.0
111.90	1859.0
112.90	462.0
116.00	2345.0
117.00	40064.0
118.00	2855.0
120.00	366.0
121.90	2833.0
123.00	4176.0
124.00	2138.0
124.90	1480.0
127.00	134720.0
128.00	9500.0
129.00	56224.0
130.00	4116.0
130.90	960.0
132.00	506.0
133.90	1847.0
135.00	3985.0
136.00	1688.0
137.10	2337.0
138.00	343.0
140.00	981.0
140.90	7862.0
141.90	2554.0
143.00	1475.0
143.80	388.0
144.90	434.0
146.10	1470.0
147.00	4056.0
147.90	9717.0
148.90	2261.0
149.90	578.0

Instrument :
BNA_N
ClientSampleId :
DFTPP61

151.10	933.0
151.90	471.0
152.90	2455.0
153.90	1766.0
155.00	4487.0
156.00	5976.0
157.00	1268.0
157.90	1901.0
158.90	955.0
159.90	2589.0
161.00	3412.0
162.00	1110.0
162.90	341.0
163.80	346.0
165.00	2773.0
166.00	2414.0
167.00	15301.0
168.00	8095.0
169.00	1328.0
170.00	468.0
170.90	702.0
171.90	1333.0
173.00	1470.0
173.90	2993.0
175.00	5927.0
175.90	1621.0
177.00	2805.0
177.90	1123.0
179.00	10860.0
180.00	8484.0
181.00	3284.0
181.90	882.0
182.90	336.0
184.00	1002.0
185.00	5916.0
186.00	39256.0
187.00	11319.0
188.00	1622.0
188.90	2186.0
189.90	367.0
190.80	981.0
192.00	2763.0
193.00	4461.0
194.00	856.0
194.90	312.0
196.00	8785.0
197.10	1796.0
197.90	262144.0
198.90	18376.0
200.00	1335.0
200.80	313.0
201.60	1119.0
203.00	2362.0
204.00	11050.0
205.00	20256.0
206.00	72592.0
207.00	11757.0
207.90	3382.0
208.90	1169.0
210.30	1043.0
210.90	2638.0
211.80	742.0
214.90	1106.0
216.00	1571.0
216.90	18568.0
217.90	2284.0
219.00	453.0
219.70	525.0
221.00	8814.0
221.90	2447.0
223.00	5280.0
224.00	43712.0
225.00	9786.0
226.10	1320.0
227.00	17496.0
228.00	2516.0
229.00	4048.0
230.00	535.0
230.90	1366.0
232.90	315.0
233.90	1243.0
235.00	1608.0
235.90	632.0
237.00	1863.0
238.00	343.0
239.00	669.0
240.00	658.0
241.00	1049.0
242.00	2427.0
243.00	2080.0
244.00	29344.0
245.00	4047.0
246.00	6658.0
246.90	1404.0
247.90	395.0
248.90	1469.0

Instrument :
BNA_N
ClientSampleId :
DFTPP61

249.90	373.0
250.90	394.0
252.90	653.0
255.00	159168.0
256.00	22456.0
256.90	2506.0
257.90	10521.0
258.90	1626.0
260.00	334.0
265.00	3178.0
266.10	374.0
270.90	696.0
271.90	403.0
273.00	4281.0
274.00	11656.0
275.00	67736.0
276.00	9309.0
277.00	5745.0
277.90	1399.0
280.90	351.0
283.00	702.0
284.00	601.0
285.00	990.0
292.90	1325.0
293.80	423.0
296.00	21104.0
296.90	3096.0
302.00	350.0
302.90	2315.0
303.90	432.0
307.90	483.0
313.90	877.0
315.00	1932.0
316.10	1157.0
320.90	801.0
321.90	305.0
323.00	6076.0
323.90	1083.0
326.90	1017.0
327.90	697.0
332.00	677.0
332.90	397.0
334.00	4147.0
335.00	1116.0
341.10	764.0
342.00	324.0
345.90	1032.0
346.90	408.0
352.00	2052.0
352.90	1654.0
354.00	1992.0
364.90	7714.0
366.00	1371.0
371.10	488.0
372.00	3086.0
373.00	733.0
383.00	755.0
384.00	328.0
389.90	552.0
402.00	1448.0
402.90	2495.0
403.90	672.0
421.10	1407.0
422.00	1326.0
423.00	9549.0
424.00	3388.0
425.10	532.0
441.00	20224.0
442.00	152832.0
443.00	29672.0
444.00	3857.0

Instrument :
BNA_N
ClientSampleId :
DFTPP61

m/z	Abundance
37.10	495.0
38.00	1303.0
39.00	8376.0
40.00	1192.0
44.00	4756.0
48.90	1000.0
50.00	27544.0
51.10	104768.0
52.10	5500.0
55.10	504.0
56.00	2961.0
57.00	7500.0
58.10	346.0
61.00	1151.0
62.10	1316.0
63.10	4253.0
63.90	527.0
65.00	1558.0
68.10	1577.0
69.00	80896.0
73.00	1253.0
74.00	8544.0
75.00	13081.0
76.10	4768.0
77.00	88296.0
78.00	5967.0
79.00	6346.0
80.00	4636.0
81.00	6987.0
81.90	1709.0
82.90	1248.0
84.90	1204.0
86.00	2066.0
87.00	1082.0
88.00	379.0
91.00	1538.0
92.00	1840.0
93.00	11346.0
94.00	627.0
95.90	532.0
98.00	9029.0
99.00	6889.0
99.90	485.0
101.00	2853.0
102.90	1304.0
104.00	2515.0
104.90	2103.0
105.90	747.0
107.00	28600.0
108.00	3962.0
109.10	607.0
110.00	46520.0
111.00	6380.0
111.80	676.0
113.00	349.0
115.90	1665.0
117.00	25912.0
117.90	2391.0
119.90	357.0
121.90	1910.0
123.00	3093.0
124.00	1565.0
124.90	1235.0
127.00	90456.0
128.00	6741.0
129.00	36088.0
130.00	2982.0
130.90	446.0
132.00	404.0
133.90	1069.0
135.00	3131.0
135.90	1152.0
136.90	1640.0
137.90	403.0
140.00	629.0
141.00	4929.0
142.00	1667.0
142.90	1116.0
143.90	441.0
145.10	368.0
146.00	900.0
147.00	2859.0
148.00	6723.0
149.00	1492.0
150.00	340.0
150.90	612.0
153.00	1726.0
153.90	1087.0
155.00	2953.0
156.00	3313.0
157.10	805.0
158.00	1048.0

Instrument :
BNA_N
ClientSampleId :
DFTPP61

158.90	564.0
159.90	1640.0
161.00	2755.0
161.90	851.0
164.90	1883.0
166.00	1673.0
167.00	9959.0
168.00	5753.0
169.00	1048.0
170.10	491.0
171.00	429.0
171.90	1019.0
172.90	1397.0
174.00	2456.0
175.00	4092.0
176.00	979.0
176.90	2135.0
178.00	771.0
178.90	8519.0
180.00	5196.0
181.00	1871.0
184.00	621.0
184.90	3767.0
186.00	28368.0
187.00	8388.0
188.00	798.0
188.90	1963.0
190.00	601.0
190.90	817.0
191.90	2448.0
193.00	2745.0
193.90	890.0
196.00	5229.0
197.00	1514.0
197.90	198656.0
198.90	12561.0
200.00	999.0
200.60	353.0
201.70	613.0
203.10	1632.0
204.00	7030.0
205.00	13887.0
206.00	54400.0
207.00	8359.0
208.00	2475.0
209.00	781.0
209.80	928.0
210.90	2335.0
215.00	618.0
215.90	1079.0
216.90	14640.0
218.00	1945.0
221.00	6869.0
221.90	2014.0
223.00	3345.0
224.00	28688.0
225.00	7095.0
226.00	950.0
227.00	13386.0
228.00	2334.0
229.00	3296.0
231.00	1278.0
234.00	818.0
235.00	995.0
235.90	617.0
236.90	1244.0
238.00	349.0
238.80	552.0
239.90	443.0
241.00	928.0
242.00	1727.0
243.00	1656.0
244.00	22152.0
245.00	2905.0
245.90	5515.0
246.90	1102.0
248.90	1038.0
250.90	365.0
252.90	639.0
254.00	1060.0
255.00	124448.0
256.00	17648.0
257.00	1407.0
258.00	7341.0
259.00	1016.0
259.80	324.0
265.00	3227.0
266.00	476.0
273.00	3592.0
274.00	8767.0
275.00	51200.0
276.00	6460.0
276.90	4678.0
277.90	849.0
283.10	683.0
283.90	535.0

Instrument :
BNA_N
ClientSampleId :
DFTPP61

284.90	753.0
293.00	1115.0
295.00	308.0
296.00	16696.0
297.00	2379.0
301.90	307.0
303.00	2020.0
304.00	490.0
314.00	821.0
314.90	1739.0
315.90	788.0
316.90	413.0
321.00	505.0
321.90	369.0
323.00	4443.0
324.00	914.0
326.90	864.0
328.00	426.0
332.10	355.0
333.10	606.0
334.00	3483.0
335.00	699.0
341.00	746.0
345.90	1003.0
352.00	1541.0
353.00	972.0
353.90	1315.0
355.10	365.0
365.00	7206.0
365.90	1273.0
370.90	441.0
372.00	2625.0
373.00	717.0
382.90	747.0
390.90	317.0
401.90	1217.0
402.90	1576.0
403.90	515.0
421.00	1174.0
421.80	978.0
422.90	8954.0
423.90	2707.0
424.90	456.0
441.00	22672.0
442.00	140416.0
443.00	29808.0
444.00	3084.0

Instrument :
BNA_N
ClientSampleId :
DFTPP61

Scan 2424 (17.245 min): BN009729.D
DFTPP61

m/z	Abundance
39.90	936.0
44.00	4422.0
73.00	558.0
207.00	867.0

Instrument :
BNA_N
ClientSampleId :
DFTPP61