

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN022020\
 Data File : BN009788.D
 Acq On : 20 Feb 2020 20:13
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 21 02:50:28 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN021920.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 19 15:57:43 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	156#	0.00
2	1,4-Dioxane	0.410	0.421	-2.7	161#	-0.02
3	n-Nitrosodimethylamine	0.673	0.770	-14.4	202#	-0.04
4 S	2-Fluorophenol	0.889	0.981	-10.3	178#	-0.02
5 S	Phenol-d6	1.082	1.173	-8.4	177#	-0.02
6	bis(2-Chloroethyl)ether	1.052	1.046	0.6	161#	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	159#	-0.01
8 S	Nitrobenzene-d5	0.333	0.353	-6.0	179#	-0.03
9	Naphthalene	1.091	1.060	2.8	159#	-0.03
10	Hexachlorobutadiene	0.240	0.232	3.3	153#	0.00
11 SURR	2-Methylnaphthalene-d10	0.774	0.751	3.0	160#	-0.02
12	2-Methylnaphthalene	0.748	0.727	2.8	161#	-0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	151#	-0.01
14 S	2,4,6-Tribromophenol	0.154	0.155	-0.6	171#	-0.01
15 S	2-Fluorobiphenyl	1.510	1.500	0.7	156#	-0.02
16	Acenaphthylene	1.705	1.594	6.5	157#	-0.02
17	Acenaphthene	1.199	1.137	5.2	150	-0.01
18	Fluorene	1.582	1.417	10.4	140	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	136	-0.01
20	4-Bromophenyl-phenylether	0.279	0.200	28.3#	103	-0.01
21	Hexachlorobenzene	0.249	0.249	0.0	142	-0.01
22	Pentachlorophenol	0.064	0.076	-18.7	216#	-0.03
23	Phenanthrene	1.190	1.076	9.6	133	-0.01
24	Anthracene	1.006	0.925	8.1	139	-0.01
25 SURR	Fluoranthene-d10	1.384	1.218	12.0	128	0.00
26	Fluoranthene	1.425	1.287	9.7	132	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	138	-0.01
28	Pvrene	1.519	1.423	6.3	134	0.00
29 S	Terphenyl-d14	0.956	1.781	-86.3#	264#	-0.01
30	Benzo(a)anthracene	1.327	1.228	7.5	144	-0.02
31	Chrysene	1.536	1.379	10.2	127	-0.01
32	Bis(2-ethylhexyl)phthalate	0.648	0.682	-5.2	162#	-0.01
33	Indeno(1,2,3-cd)pyrene	1.599	1.444	9.7	125	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	131	-0.01
35	Benzo(b)fluoranthene	1.462	1.358	7.1	133	-0.01
36	Benzo(k)fluoranthene	1.592	1.419	10.9	123	-0.01
37 C	Benzo(a)pyrene	1.334	1.318	1.2	133	-0.02
38	Dibenzo(a,h)anthracene	1.314	1.186	9.7	126	-0.02
39	Benzo(a,h,i)perylene	1.400	1.256	10.3	125	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN022020\
Data File : BN009788.D
Acq On : 20 Feb 2020 20:13
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Feb 21 02:50:28 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN021920.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 19 15:57:43 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			